

246219864

Boldly "X" this box if there is qualified data on this page.

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID:

0920-8

Project:

2018-04-30-23 (PL R180) 4

- Notes:** (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

3/29/18

DO DEP SOP FT 1500	Name	Date	Time (ET)	Temp. (°C)	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	MATTHEW BEARDEN	5/8/18	0700	23.6	8.48	8.49	100.1		1.1	(P) F	(L) F
ICV			0702	23.6	8.48	8.49	100.0			(P) F	(L) F
CCV	Ben Bearden	5/8	1645	26.5	8.04	7.88	97.8			(P) F	(L) F
CCV										P F	L F

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time (ET)	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	M.B	5/8	0706	2711453	10/19	1000	1000	(P) F	(L) F
ICV		5/8	0726	1711050	10/19	100	104	(P) F	(L) F
CCV	BP	5/8	1647	2707875	7/19	500	497	(P) F	(L) F
CCV								P F	L F

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time (ET)	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	MB	5/8	0710	177435	10/19	7.0	7.0	-29.2	(P) F	(L) F
Calibr.			0718	177759	10/20	4.0	4.0	143.7	(P) F	(L) F
Calibr.			0722	173794	5/19	10.0	10.0	-203.2	(P) F	(L) F
ICV			0724	173794	5/19	10.0	10.0		(P) F	(L) F
CCV	BP	5/8	1650	177435	10/19	7.0	7.14	-37	(P) F	(L) F
CCV									P F	L F

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$;
 mV pH 10 Range -180 ± 50 ; Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air				0.000			

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater

Notes: