From: Heidi Parker
To: Wood-McGrath, Avril
Subject: RE: T1811638-AMENDED
Date: Wednesday, August 8, 2018 1:08:23 PM

HI Avril
This is what the analyst said

" When I checked it I did not look at the queue report, just the posted data. When they changed the intranet the benchsheet was dumping from the wrong column putting in calculated results and not raw counts so lims was adding the dilution factor twice. That was wrong and fixed. I thought we had caught all the errors. Please unprotect the sample and it's sample dilutions and I can fix that."

Heidi Parker
Project Manager
Advanced Environmental Laboratories, Inc.
Florida's Largest Laboratory Network!
9610 Princess Palm Avenue
Tampa, FL 33619
813-630-9616 Voice
813-630-4327 Fax
Website: www.aellab.com

-----Original Message-----
From: Wood-McGrath, Avril [mailto:Avril.Wood-McGrath@dep.state.fl.us]
Sent: Wednesday, August 08, 2018 1:03 PM
To: Heidi Parker
Subject: RE: T1811638-AMENDED

Hi Heidi, Can you tell why the result was changed in case I need documentation?

-----Original Message-----
From: Heidi Parker [mailto:hparker@aellab.com]
Sent: Wednesday, August 8, 2018 12:58 PM
To: Wood-McGrath, Avril <Avril.Wood-McGrath@dep.state.fl.us>
Cc: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Subject: RE: T1811638-AMENDED

Heidi Parker
Project Manager
Advanced Environmental Laboratories, Inc.
Florida's Largest Laboratory Network!
9610 Princess Palm Avenue
Tampa, FL 33619
813-630-9616 Voice
813-630-4327 Fax
Website: www.aellab.com

-----Original Message-----
From: Wood-McGrath, Avril [mailto:Avril.Wood-McGrath@dep.state.fl.us]
Sent: Monday, August 06, 2018 4:36 PM
To: Heidi Parker (hparker@aellab.com)
Cc: DEP-Overflowmicro
Subject: T1811638

Hi Heidi,

Can you double check this result?

G2SW0029

Wares Creek @9th Ave

T1811638003

7/12/2018

12:30

SM 9230 C (MF)

Enterococcus

#/100 mL

7000

B

10

[cid:image003.png@01D1BBD9.1BBA8140]<http://www.dep.state.fl.us/>

Avril Wood-McGrath
OPS Environmental Specialist II
FDEP Laboratory - Biology Program
Florida Department of Environmental Protection
Avril.Wood-McGrath@floridadep.gov<mailto:Avril.Wood-McGrath@floridadep.gov>
Office: 850.245.8193

[Dep Customer
Survey]<http://survey.dep.state.fl.us/?refemail=Avril.Wood-McGrath@dep.state
.fl.us>
[Dep Customer
Survey]<http://survey.dep.state.fl.us/?refemail=Avril.Wood-McGrath@dep.state
.fl.us>