

**Boldly
Box If
Qualified
This n**

४०

2018-02-12-5

Project

martin RUN
Smp

- Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature/Quarterny/At 1400

Date of Last Temperature Verification 21/04/2018

DO Acceptance criteria from Table 1: 0.3 mg/L.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-7

Steady-state & Galvanic Sensors: DO Gain & Charge N/A

Conductivity Acceptance criteria $\pm 5\%$

pH Acceptance criteria ± 0.2 SU;

mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$:

mV pH 10 Range -180 $\pm$ 50;

If mV are recorded: slope from 7 to 10 172.2, slope from 4 to 7 182.6 (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? Yes / No If yes, list date of last quarterly depth verification 04/04/2018

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or $\pm 0.05m$, whichever is greater.

COMMENTS: