

Chemical Analysis Report

SE-DIST-2018-07-25-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **SMP- Taco Run**
Request ID: **RQ-2018-05-14-59**
Customer: **SE-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
2199 South Rock Road
Fort Pierce, FL 34945
Attn: Matthew Sewell

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator

Date Certified: 13-AUG-2018 17:35



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: C-54 W OF I-95

Collection Date/Time: 07/24/2018 13:50

Field ID: G5SE0057

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2010531	EPA 180.1 Rev. 2.0	Turbidity	2.0		NTU	P349130	
	EPA 300.0 Rev. 2.1	Chloride	110		mg Cl/L	P349410	
		Sulfate	26		mg SO4/L	P349412	
		Color (true)	150		PCU	P349165	
	SM 2320 B-97	Total Alkalinity	132		mg CaCO3/L	P349352	
	SM 2540 C-97	TDS	394		mg/L	P349673	
	SM 2540 D-97	TSS	4	I	mg/L	P349366	
	SM 4500 F-C-97	Fluoride	0.21		mg F/L	P349356	
2010532	EPA 350.1 Rev. 2.0	Ammonia-N	0.007		mg N/L	P349322	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.0		mg N/L	P349377	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P349331	
	EPA 365.1 Rev. 2.0	Total-P	0.073		mg P/L	P349382	
	SM 5310 B-00	Organic Carbon	19		mg C/L	P349296	
2010533	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.032		mg P/L	P349139	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: MICCO DITCHES WEIR OFF DOTTIE

Collection Date/Time: 07/24/2018 14:20

Field ID: G5SE0063

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2010528	EPA 180.1 Rev. 2.0	Turbidity	3.5		NTU	P349130	
	EPA 300.0 Rev. 2.1	Chloride	48		mg Cl/L	P349410	
		Sulfate	11		mg SO4/L	P349412	
	SM 2120 B	Color (true)	89		PCU	P349165	
	SM 2320 B-97	Total Alkalinity	56		mg CaCO3/L	P349352	
	SM 2540 C-97	TDS	169		mg/L	P349673	
	SM 2540 D-97	TSS	5	I	mg/L	P349366	
	SM 4500 F-C-97	Fluoride	0.11		mg F/L	P349356	
2010529	EPA 350.1 Rev. 2.0	Ammonia-N	0.008		mg N/L	P349322	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.90		mg N/L	P349377	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.006	I	mg N/L	P349331	
	EPA 365.1 Rev. 2.0	Total-P	0.14		mg P/L	P349382	
	SM 5310 B-00	Organic Carbon	16		mg C/L	P349296	
2010530	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.078		mg P/L	P349139	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 350.1 Rev. 2.0: The result was confirmed on 07/27/2018.

Sample Location: MICCO DITCHES WEIR OFF DOTTIE

Collection Date/Time: 07/24/2018 14:25

Field ID: G5SE0063_DUP

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2010534	EPA 180.1 Rev. 2.0	Turbidity	3.3		NTU	P349130	
	EPA 300.0 Rev. 2.1	Chloride	46		mg Cl/L	P349410	
		Sulfate	11		mg SO4/L	P349412	
	SM 2120 B	Color (true)	89		PCU	P349165	
	SM 2320 B-97	Total Alkalinity	56		mg CaCO3/L	P349352	
	SM 2540 C-97	TDS	176		mg/L	P349673	
	SM 2540 D-97	TSS	4	I	mg/L	P349366	
	SM 4500 F-C-97	Fluoride	0.11		mg F/L	P349356	
2010535	EPA 350.1 Rev. 2.0	Ammonia-N	0.011		mg N/L	P349322	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.90		mg N/L	P349377	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.008	I	mg N/L	P349331	
	EPA 365.1 Rev. 2.0	Total-P	0.14		mg P/L	P349382	
	SM 5310 B-00	Organic Carbon	16		mg C/L	P349296	
2010536	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.078		mg P/L	P349139	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 350.1 Rev. 2.0: The result was confirmed on 07/27/2018.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P349130

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P349410

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P349412

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P349322

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P349377

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P349331

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P349382

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P349139

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P349165

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report

Method Blank Results

Reference Method: SM 2320 B-97
Batch ID: P349352

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-97
Batch ID: P349673

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P349366

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P349356

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P349296

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P349410

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	100		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P349412

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	99.8		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P349322

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	102		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P349377

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	109		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P349331

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P349382

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	97.4		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P349139

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	99.1		P	90 - 110

Reference Method: SM 2120 B
Batch ID: P349165

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	101		P	85 - 115

Reference Method: SM 2320 B-97
Batch ID: P349352

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	98 - 107

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 4500 F-C-97
Batch ID: P349356

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	96.5		P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P349296

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	91.6		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P349410

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2010201	Chloride	103		P	90 - 110
2010534	Chloride	101		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P349412

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2010201	Sulfate	103		P	90 - 110
2010534	Sulfate	99.5		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P349322

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2010204	Ammonia-N	100	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P349377

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2010493	Kjeldahl Nitrogen	99.7	92.6	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P349331

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2010480	NO2NO3-N	102	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P349382

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2010617	Total-P	93.1	92.3	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P349139

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2010213	O-Phosphate-P	99.0	98.5	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P349356

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2010817	Fluoride	96.7	96.7	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P349296

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2010223	Organic Carbon	99.2		P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P349130

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2010503	Turbidity	7.10	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P349410

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2010201	Chloride	3.66	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P349412

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2010201	Sulfate	5.18	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P349322

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2010204	Ammonia-N	0.875	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P349377

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2010493	Kjeldahl Nitrogen	4.62	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P349331

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2010480	NO2NO3-N	0.487	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P349382

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2010617	Total-P	0.770	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P349139

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2010213	O-Phosphate-P	0.506	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2120 B
Batch ID: P349165

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2010264	Color (true)	3.12	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P349352

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2010817	Total Alkalinity	0.574	Sample	P	0 - 1.6

Reference Method: SM 2540 C-97
Batch ID: P349673

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2010614	TDS	3.56	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P349356

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2010817	Fluoride	0.0	Spike	P	0 - 3.5

Reference Method: SM 5310 B-00
Batch ID: P349296

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2010223	Organic Carbon	0.0842	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A85472

Included Lab Sample IDs: 2010528, 2010531, 2010534

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	102	102	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A85473

Included Lab Sample IDs: 2010530, 2010533, 2010536

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	97.6	98.1	P/P	90 - 110
O-Phosphate-P	98.6	98.8	P/P	90 - 110

Reference Method: SM 2120 B

Run ID: A85480

Included Lab Sample IDs: 2010528, 2010531, 2010534

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	103	104	P/P	90 - 115
Color (true)	104	105	P/P	90 - 115

Reference Method: SM 5310 B-00

Run ID: A85527

Included Lab Sample IDs: 2010529, 2010532, 2010535

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	95.8	96.2	P/P	90 - 110
Organic Carbon	96.2	97.0	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A85536

Included Lab Sample IDs: 2010529, 2010532, 2010535

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	102	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A85538

Included Lab Sample IDs: 2010529, 2010532, 2010535

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A85546

Included Lab Sample IDs: 2010528, 2010531, 2010534

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	102	99.5	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 4500 F-C-97

Run ID: A85546

Included Lab Sample IDs: 2010528, 2010531, 2010534

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	98.6	101	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A85560

Included Lab Sample IDs: 2010528, 2010531, 2010534

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	104	104	P/P	90 - 110
Chloride	98.8	104	P/P	90 - 110
Sulfate	104	105	P/P	90 - 110
Sulfate	99.5	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A85574

Included Lab Sample IDs: 2010529, 2010532, 2010535

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	93.4	97.9	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A85579

Included Lab Sample IDs: 2010529, 2010532, 2010535

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	93.5	99.4	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					7.10	
EPA 300.0 Rev. 2.1	Chloride	100	103	101		3.66	
	Sulfate	99.8	103	99.5		5.18	
EPA 350.1 Rev. 2.0	Ammonia-N	102	100	101			0.875
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	109	99.7	92.6			4.62
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	102	102	103			0.487
EPA 365.1 Rev. 2.0	Total-P	97.4	93.1	92.3			0.770
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	99.1	99.0	98.5			0.506
SM 2120 B	Color (true)	101				3.12	
SM 2320 B-97	Total Alkalinity	103				0.574	
SM 2540 C-97	TDS					3.56	
SM 4500 F-C-97	Fluoride	96.5	96.7	96.7			0.0
SM 5310 B-00	Organic Carbon	91.6	99.2				0.0842

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	2010528, 2010531, 2010534
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	2010528, 2010531, 2010534
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	2010528, 2010531, 2010534
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2010529, 2010532, 2010535
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2010529, 2010532, 2010535
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2010529, 2010532, 2010535
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2010529, 2010532, 2010535
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2010530, 2010533, 2010536
SM 2120 B / E31780	True Color measured at 450 nm	2010528, 2010531, 2010534
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2010528, 2010531, 2010534
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	2010528, 2010531, 2010534
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2010528, 2010531, 2010534
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2010528, 2010531, 2010534
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2010529, 2010532, 2010535

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	07/25/2018			07/25/2018 16:44	William L. Brown IV	2010528, 2010531, 2010534
EPA 300.0 Rev. 2.1	07/25/2018			07/30/2018 13:35	Lipi Saha	2010534
	07/25/2018			07/30/2018 15:46	Lipi Saha	2010528
	07/25/2018			07/30/2018 15:58	Lipi Saha	2010531
EPA 350.1 Rev. 2.0	07/25/2018			07/27/2018 13:49	Ping Hua	2010529
	07/25/2018			07/27/2018 13:50	Ping Hua	2010532
	07/25/2018			07/27/2018 13:52	Ping Hua	2010535
EPA 351.2 Rev. 2.0	07/25/2018	07/30/2018 10:15	Adrian Shelby	07/31/2018 10:55	Joseph Suarez	2010529
	07/25/2018	07/30/2018 10:15	Adrian Shelby	07/31/2018 10:57	Joseph Suarez	2010532
	07/25/2018	07/30/2018 10:15	Adrian Shelby	07/31/2018 10:58	Joseph Suarez	2010535
EPA 353.2 Rev. 2.0	07/25/2018			07/27/2018 12:51	Lauren Bottorf	2010529
	07/25/2018			07/27/2018 12:52	Lauren Bottorf	2010535
	07/25/2018			07/27/2018 12:55	Lauren Bottorf	2010532
EPA 365.1 Rev. 2.0	07/25/2018	07/30/2018 11:20	Sima Feizi	07/31/2018 09:14	Beverly Y. Sanders	2010529
	07/25/2018	07/30/2018 11:20	Sima Feizi	07/31/2018 09:15	Beverly Y. Sanders	2010532
	07/25/2018	07/30/2018 11:20	Sima Feizi	07/31/2018 09:16	Beverly Y. Sanders	2010535
EPA 365.1 Rev. 2.0 dissolved	07/25/2018			07/25/2018 17:08	Julia Storbeck	2010530, 2010536
	07/25/2018			07/25/2018 17:51	Julia Storbeck	2010533
SM 2120 B	07/25/2018			07/25/2018 16:59	Tanya Welborn	2010528
	07/25/2018			07/25/2018 17:00	Tanya Welborn	2010534
	07/25/2018			07/25/2018 17:20	Tanya Welborn	2010531
SM 2320 B-97	07/25/2018			07/26/2018 23:22	Dale L. Simmons	2010528
	07/25/2018			07/26/2018 23:32	Dale L. Simmons	2010531
	07/25/2018			07/26/2018 23:42	Dale L. Simmons	2010534
SM 2540 C-97	07/25/2018			07/28/2018 20:08	Yijie Li	2010528, 2010531, 2010534
SM 2540 D-97	07/25/2018			07/28/2018 23:02	Yijie Li	2010528, 2010531, 2010534
SM 4500 F-C-97	07/25/2018			07/26/2018 23:22	Dale L. Simmons	2010528
	07/25/2018			07/26/2018 23:32	Dale L. Simmons	2010531
	07/25/2018			07/26/2018 23:42	Dale L. Simmons	2010534
SM 5310 B-00	07/25/2018			07/26/2018 19:34	Megan Treadwell	2010529
	07/25/2018			07/26/2018 19:47	Megan Treadwell	2010532
	07/25/2018			07/26/2018 20:31	Megan Treadwell	2010535