



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2018

Data Qualified?

Yes / No

WIN Station Name: Upper Dianne Lake Middle

Date: 3/7/2018

WIN/ Field ID: G1WA0041

WBID: 546B

Latitude: 30.59986

(mm/dd/yyyy)

County: Leon

ORG ID: 21FLWQA

Longitude: -84.23778

(Decimal Degrees)

Receiving Body of Water: Lake Iamonia

RQ-2018- 03-05-26

Sampling Team Members						WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check	Sampling Equipment				
Ben Ralys							X				☒ Sample Container	☐ Van Dorn	☐ Pole		
Jason Storrs						X		X	X	X	☐ Carboy	☒ Syringe			
											☐ Dipper	☐ Other			
											Depth Measured with <u>YSI</u>				
											Equipment ID <u>Biology #2</u>				
											Collection Method				
											☐ Wading	☒ Canoe/Kayak			
											☐ From Shore	☐ Electric Motor			
											☐ Other	☐ Gasoline Motor			
Water Level: Dry/ Pooled/ Low / <u>Normal</u> / High / Flooded						Meter ID# <u>Biology #2</u>			Matrix: <u>Fresh</u> / Marine / DI						
Photos: Yes / No		Flow: No Flow / Intestinal / Flowing / <u>NA</u>		Tide: Rising/ Falling/ Slack/ <u>NA</u>		Tide Times: High <u>NA</u> Low <u>NA</u>									
Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Sp COND (µmhos/cm)	Temp. (C°)					
<u>EST</u>	<u>1040</u>	<u>1.3</u>	<u>0.3</u>	<u>1.3" L"</u>	<u>10.55</u>	<u>111.8</u>	<u>8.80</u>	<u>0.04</u>	<u>92</u>	<u>18.14</u>					
CEN	<u>1038</u>		<u>—</u>	<u>—</u>	<u>—</u>	<u>—</u>	<u>—</u>	<u>—</u>	<u>—</u>	<u>—</u>	<u>—</u>				
Biological Samples Collected: <u>None</u> HA SCI RPS Algae ID LVS LVI Other:															
SBIO ID Numbers:		SBIO-ID:		HA-ID:		RPS-ID:		Macrophyte-ID:		LIMS#: <u>1972790</u>					
Parameter Suite		Parameter Methods				Preservation		Lot#		Expiration		pH Check			
Preserved Nutrients		☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☒ Ice ☒ H2SO4		<u>SA7341020</u>		<u>12/7/18</u>		☒ <2			
Metals		☐ W-ICPMS ☐ W-ICP				☐ Ice ☐ HNO3		<u>10827054</u>				☐ <2			
Filtered Nutrients		☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter				☒ Ice									
Chlorophyll		☒ CHLSUITE-W				☒ Ice									
Physical/Aggregate (Fresh)		☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-TDS				☒ Ice									
Physical/Aggregate (Marine)		☐ Alkalinity / W-F / W-TSS / TURBIDITY				☐ Ice									
Macroinvertebrates		☐ MI-FW-QLDC				☐ Formalin									
Periphyton		☐ ALGAE-ID				☐ Ice									
Microbiology		☒ E Coli ☐ F Coli ☐ Enterococci				☒ Ice									
Molecular		☐ PCR-FILTER ☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-GULL2 ☐ PCR-GFD				☐ Ice									
Tracers		☐ W-SUC-AA-R ☐ W-UHERB-AA ☐ W-AHERB-AA				☐ Ice									
Pesticides		☐ W-PSNP-TQ ☐ W-UHERB-AA ☐ W-AHERB-AA ☐ W-CARB-AA				☐ Ice									
		☐ W-PCL-SQ-R ☐ W-PYR-AA-R ☐ W-GLYPH ☐ W-SUC-AA-R				☐ Other									

Contains: 1.1 H2SO4
Lot No.: SA7341020
Expires: 12/07/18
See Safety Data Sheet (SDS)

SULFURIC ACID <51%
775-185
1450-600-2711

May be fatal. Causes severe skin burns and eye damage. Do not breathe. Do not get in eyes, on clothing, or on skin. Use only outdoors or in a well-ventilated area. Wash contaminated clothing before reuse. Avoid contact with skin and eyes. If contact occurs, flush with plenty of water for at least 15 minutes. If inhaled, remove to fresh air and breathe normally. If swallowed, drink water. Do not induce vomiting. If the label is not legible, the container should be treated as containing a hazardous material. Store locked up. Keep away from children. Dispose of contents and container in accordance with applicable regulations.

Sticker(s) Here (ie)

Notes:

A-Kit
FW Bacteria

Upper Dianne
Lake Middle

WIN ID/Field ID

G1WA0041

Signature:

Date:

Enter Field Parameters, Verify Station ID in LIMS DMT:

QC Station ID, Field Parameters in LIMS DMT:

Scan/Save Field Sheets

Migrate Data to WIN:

Verify Data in WIN:

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID: Biology #2

RQ: 2018-03-05-26

Project: SMP

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site _____

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification _____

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Ben Ralys	3/7/18	7:10	20.96	8.93	8.92	100.0	62.5	-	(P) / F	(L) / F
ICV	Ben Ralys	3/7/18	7:12	20.97	8.93	8.92	100.0	61.4	-	(P) / F	(L) / F
CCV	Ben Ralys	3/7/18	1:38	21.55	8.82	8.94	101.3	60.4	-	(P) / F	(L) / F
CCV										P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Ben Ralys	3/7/18	7:14	171005D	10/2018	1,000	999	(P) / F	(L) / F
ICV	Ben Ralys	3/7/18	7:14	171005D	10/2018	1,000	1,001	(P) / F	(L) / F
CCV	Ben Ralys	3/7/18	1:40	171005E	10/2018	100	99	(P) / F	(L) / F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Ben Ralys	3/7/18	7:18	170614B	12/2018	7.0	6.98	2.3	(P) / F	(L) / F
Calibr.	Ben Ralys	3/7/18	7:20	171005A	10/2018	4.0	3.98	178.9	(P) / F	(L) / F
Calibr.	Ben Ralys	3/7/18	7:22	171005C	4/2019	10.0	9.81	-151.0	(P) / F	(L) / F
ICV	Ben Ralys	3/7/18	7:25	171005C	4/2019	10.0	9.81	-148.8	(P) / F	(L) / F
CCV	Ben Ralys	3/7/18	1:43	171005A	10/2018	4.0	3.94	178.9	(P) / F	(L) / F
CCV									P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? Yes / No If yes, list date of last quarterly depth verification _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

Customer: WAS-SECT

Project ID: SMP

Requester: Curtis Musson

Collected By: Ben Rabks, Jason Storns

Field Report Prepared By: Ben Rabks, Jason Storns

Sampling Agency: FDEP

Location		WIN ID/Field ID		Pettys Gulf		WIN ID/Field ID		G1WA0022		Bottle Group(s)	
Field ID										A	
WIN/STORET #				lake Middle							
Latitude		30.58915		Longitude		-84.22637		Temp (C)		18.40	
		☒ dd ☐ dms				☒ dd ☐ dms		pH		6.92	
		Salinity (ppt)		0.03		Comments		Diss. Oxygen		84.4	
								Sample Depth		0.3	
								Sp. Conductance (umho/cm)		58	
Location		WIN ID/Field ID		Pine Hill		WIN ID/Field ID		G1WA0054		Bottle Group(s)	
Field ID										A	
WIN/STORET #				Lake Middle							
Latitude		30.58492		Longitude		-84.2195		Temp (C)		17.88	
		☒ dd ☐ dms				☒ dd ☐ dms		pH		5.94	
		Salinity (ppt)		0.02		Comments		Diss. Oxygen		87.7	
								Sample Depth		0.3	
								Sp. Conductance (umho/cm)		44	
Location		WIN ID/Field ID		Lower Dianne		WIN ID/Field ID		G1WA0040		Bottle Group(s)	
Field ID										A	
WIN/STORET #				Lake Middle							
Latitude		30.59355		Longitude		-84.23534		Temp (C)		18.55	
		☒ dd ☐ dms				☒ dd ☐ dms		pH		7.34	
		Salinity (ppt)		0.04		Comments		Diss. Oxygen		90.1	
								Sample Depth		0.3	
								Sp. Conductance (umho/cm)		85	
Location		WIN ID/Field ID		Upper Dianne		WIN ID/Field ID		G1WA0041		Bottle Group(s)	
Field ID										A	
WIN/STORET #				Lake Middle							
Latitude		30.59986		Longitude		-84.23778		Temp (C)		18.14	
		☒ dd ☐ dms				☒ dd ☐ dms		pH		8.80	
		Salinity (ppt)		0.04		Comments		Diss. Oxygen		111.8	
								Sample Depth		0.3	
								Sp. Conductance (umho/cm)		92	

Relinquished By:	Date/Time: 3/2/18 1323	Shipping Method: Drop-off	Number of Coolers Shipped:	Received By:	Date/Time: 03-07-18 1323
------------------	------------------------	---------------------------	----------------------------	--------------	--------------------------

Group: A	Bottle Type: ALKALINITY TURBIDITY W-CL-IC W-COLOR W-F W-SO4-IC W-TDS W-TSS	P-1L	# of Bottles: 5	Preservative: ICE	Enter Number of Bottles Sent to Lab
Group: A	Bottle Type: CHLSUITE-W	BP-1L	# of Bottles: 5	Preservative: ICE	Enter Number of Bottles Sent to Lab
Group: A	Bottle Type: ECOLI-18QT	P-250ML-WO-THI	# of Bottles: 5	Preservative: ICE	Enter Number of Bottles Sent to Lab
Group: A	Bottle Type: W-NH3 W-NO2NO3 W-S-A-TP W-TKN W-TOC	P-500ML	# of Bottles: 5	Preservative: ICE And H2SO4	Enter Number of Bottles Sent to Lab
Group: A	Bottle Type: W-PO4-F	P-125ML	# of Bottles: 5	Preservative: FILTER-ICE	Enter Number of Bottles Sent to Lab

Cooler Packing Worksheet for Request: RQ-2018-03-05-26

NE Tallahassee lakes

Ship Cooler On: 3/2/2018

Customer/Project: WAS-SECT/SMP

Assigned To: SHAIK_A

Requester: Curtis Musson

Priority: 3

850-245-8453
2600 Blair Stone Rd

Tallahassee, FL 32399

Attn: Curtis Musson

Comments:

No preservatives needed.

Group A: Freshwater Kit + Ecoli

Packing Notes
No FedEx labels
no trash bags
2 Submittal forms

Requested Analyses:

Bottle Group: A

of Sites: 5

Container ID: P-1L

Qty: 5 Preservation: ICE

Lot # 00069122

Description: Plastic Bottle 1L

Analysis

ALKALINITY
TURBIDITY
W-CL-IC
W-COLOR
W-F
W-SO4-IC
W-TDS
W-TSS

Description

Alkalinity in aqueous matrices as mg CaCO3/L
Turbidity in aqueous matrices
Chloride in aqueous matrices
True Color measured at 450 nm
Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
Sulfate in aqueous matrices
Total Dissolved Solids in aqueous matrices
Total Suspended Solids in aqueous matrices

Container ID: BP-1L

Qty: 5 Preservation: ICE

Lot # 1214297

Description: Brown Plastic Bottle 1L

Analysis

CHLSUITE-W

Description

Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Container ID: P-250ML-WO-THI

Qty: 5 Preservation: ICE

Lot # 417167

Description:

Analysis

ECOLI-18QT

Description

Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems

Container ID: P-500ML

Qty: 5 Preservation: ICE And H2SO4

Lot # 5507 0302

Description: Plastic Bottle 500 mL

Affix YELLOW sulfuric acid dot to container.

Analysis

W-NH3

W-NO2NO3

W-S-A-TP

W-TKN

W-TOC

Description

Ammonia in aqueous matrices as mg N/L

Nitrite/Nitrate in aqueous matrices as mg N/L

Total Phosphorus in aqueous matrices as mg P/L

Total Kjeldahl Nitrogen in aqueous matrices

Nonpurgeable Organic Carbon in aqueous matrices

Container ID: P-125ML

Qty: 5 Preservation: FILTER-ICE

Lot # 5506 9732

Description: Plastic Bottle 125 mL

Analysis

W-PO4-F

Description

Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

Cooler Packed By: ARJ

Number of Coolers: 2

Date: 2-28-18

Kit must also include:

- ☒ Field Sheets
- ☒ Temperature Control Bottle
- ☒ FedEx Bills, if applicable
- ☒ Plastic Bags
- ☒ Zip Ties

If Preservation Included:

ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____

PLEASE RETURN ALL COOLERS FOR RQ-2018-03-05-26