

Chemical Analysis Report

WAS-SECT-2018-08-08-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Killarney Lakes**
Request ID: **RQ-2018-08-06-20**
Customer: **WAS-SECT**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
Tallahassee, FL 32399
Attn: Curtis Musson

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 27-AUG-2018 16:38

A handwritten signature in black ink, appearing to read "Liang Lin", is shown on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: LOWER DIANNE LAKE MIDDLE

Collection Date/Time: 08/08/2018 11:05

Field ID: G1WA0040

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2015542	EPA 180.1 Rev. 2.0	Turbidity	2.6	A	NTU	P349817	
	EPA 300.0 Rev. 2.1	Chloride	6.4		mg Cl/L	P350007	
		Sulfate	0.65		mg SO4/L	P350008	
	SM 2120 B	Color (true)	14		PCU	P349825	
	SM 2320 B-97	Total Alkalinity	26		mg CaCO3/L	P350106	
	SM 2540 C-97	TDS	45		mg/L	P350318	
	SM 2540 D-97	TSS	3	I	mg/L	P350310	
	SM 4500 F-C-97	Fluoride	0.056	I	mg F/L	P350110	
2015546	EPA 350.1 Rev. 2.0	Ammonia-N	0.003	I	mg N/L	P350236	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.54		mg N/L	P350154	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P350251	
	EPA 365.1 Rev. 2.0	Total-P	0.016		mg P/L	P350140	
	SM 5310 B-00	Organic Carbon	5.2		mg C/L	P350246	
2015550	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P349901	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: UPPER DIANNE LAKE MIDDLE

Collection Date/Time: 08/08/2018 10:00

Field ID: G1WA0041

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2015543	EPA 180.1 Rev. 2.0	Turbidity	2.2		NTU	P349816	
	EPA 300.0 Rev. 2.1	Chloride	5.4		mg Cl/L	P350034	
		Sulfate	1.3		mg SO4/L	P350035	
	SM 2120 B	Color (true)	16		PCU	P349825	
	SM 2320 B-97	Total Alkalinity	37		mg CaCO3/L	P350106	
	SM 2540 C-97	TDS	59		mg/L	P350318	
	SM 2540 D-97	TSS	3	I	mg/L	P350310	
	SM 4500 F-C-97	Fluoride	0.089	I	mg F/L	P350110	
2015547	EPA 350.1 Rev. 2.0	Ammonia-N	0.003	I	mg N/L	P350236	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.52		mg N/L	P350154	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P350251	
	EPA 365.1 Rev. 2.0	Total-P	0.030		mg P/L	P350140	
	SM 5310 B-00	Organic Carbon	4.8		mg C/L	P350246	
2015551	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P349901	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: PINE HILL LAKE MIDDLE

Collection Date/Time: 08/08/2018 12:22

Field ID: G1WA0054

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2015544	EPA 180.1 Rev. 2.0	Turbidity	4.4		NTU	P349816	
	EPA 300.0 Rev. 2.1	Chloride	4.9		mg Cl/L	P350034	
		Sulfate	0.52	I	mg SO4/L	P350035	
	SM 2120 B	Color (true)	17		PCU	P349825	
	SM 2320 B-97	Total Alkalinity	10		mg CaCO3/L	P350106	
	SM 2540 C-97	TDS	32		mg/L	P350318	
	SM 2540 D-97	TSS	6	I	mg/L	P350310	
	SM 4500 F-C-97	Fluoride	0.043	I	mg F/L	P350110	
2015548	EPA 350.1 Rev. 2.0	Ammonia-N	0.003	I	mg N/L	P350236	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.67		mg N/L	P350155	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P350251	
	EPA 365.1 Rev. 2.0	Total-P	0.018		mg P/L	P350140	
	SM 5310 B-00	Organic Carbon	6.7		mg C/L	P350246	
2015552	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P349901	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: PETTY GULF LAKE MIDDLE

Collection Date/Time: 08/08/2018 12:03

Field ID: G1WA0022

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2015545	EPA 180.1 Rev. 2.0	Turbidity	2.0		NTU	P349816	
	EPA 300.0 Rev. 2.1	Chloride	4.3		mg Cl/L	P350034	
		Sulfate	0.53	I	mg SO4/L	P350035	
	SM 2120 B	Color (true)	19		PCU	P349825	
	SM 2320 B-97	Total Alkalinity	16		mg CaCO3/L	P350106	
	SM 2540 C-97	TDS	30		mg/L	P350318	
	SM 2540 D-97	TSS	3	I	mg/L	P350310	
	SM 4500 F-C-97	Fluoride	0.045	I	mg F/L	P350110	
2015549	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	I	mg N/L	P350236	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.45		mg N/L	P350155	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P350251	
	EPA 365.1 Rev. 2.0	Total-P	0.016		mg P/L	P350140	
	SM 5310 B-00	Organic Carbon	5.2		mg C/L	P350246	
2015553	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P349901	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P349816

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P349817

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350007

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350008

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350034

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350035

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P350236

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P350154

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P350155

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P350251

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P350140

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P349901

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P349825

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P350106

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P350318

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P350310

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P350110

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P350246

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350007

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	97.9		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350008

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	95.7		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350034

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	99.8		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350035

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	100		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P350236

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	96.4		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P350154

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	101		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P350155

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	105		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P350251

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO ₂ NO ₃ -N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P350140

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	98.8		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P349901

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	96.9		P	90 - 110

Reference Method: SM 2120 B
Batch ID: P349825

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	97.6		P	85 - 115

Reference Method: SM 2320 B-97
Batch ID: P350106

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	98 - 107

Reference Method: SM 4500 F-C-97
Batch ID: P350110

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	99.3		P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P350246

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	94.8		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350007

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2014634	Chloride	99.1		P	90 - 110
2015234	Chloride	97.0		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350008

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2014634	Sulfate	98.1		P	90 - 110
2015234	Sulfate	95.4		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350034

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2015697	Chloride	99.4		P	90 - 110
2015702	Chloride	97.9		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350035

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2015697	Sulfate	99.8		P	90 - 110
2015702	Sulfate	97.1		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P350236

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2015546	Ammonia-N	97.1	98.2	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P350154

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2015546	Kjeldahl Nitrogen	95.6	98.2	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P350155

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2015549	Kjeldahl Nitrogen	99.3	102	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P350251

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2015537	NO2NO3-N	93.7	95.2	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P350140

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
---------------	-----------	---------	---------	-----------	----------------

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P350140

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2015547	Total-P	100	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P349901

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2015550	O-Phosphate-P	94.0	95.7	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P350110

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2015773	Fluoride	99.2	97.9	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P350246

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2015409	Organic Carbon	104		P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P349816

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015134	Turbidity	4.07	Sample	P	0 - 20

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P349817

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015542	Turbidity	1.16	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350007

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2014634	Chloride	6.40	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350008

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2014634	Sulfate	7.04	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350034

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015702	Chloride	0.557	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350035

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015702	Sulfate	0.365	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P350236

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015546	Ammonia-N	1.03	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P350154

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015546	Kjeldahl Nitrogen	1.80	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P350155

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015549	Kjeldahl Nitrogen	1.76	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P350251

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015537	NO2NO3-N	1.59	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P350140

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015547	Total-P	0.681	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P349901

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015550	O-Phosphate-P	1.79	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P349825

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015308	Color (true)	0.610	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P350106

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015773	Total Alkalinity	0.492	Sample	P	0 - 1.6

Reference Method: SM 2540 C-97
Batch ID: P350318

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015347	TDS	4.30	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P350110

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015773	Fluoride	0.986	Spike	P	0 - 3.5

Quality Assurance Report Precision

Reference Method: SM 5310 B-00
Batch ID: P350246

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015409	Organic Carbon	0.0640	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A85712

Included Lab Sample IDs: 2015542

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	104	99.6	P/P	90 - 110
Sulfate	104	98.6	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A85756

Included Lab Sample IDs: 2015542, 2015543, 2015544, 2015545

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	102	103	P/P	95 - 105
Turbidity	103	102	P/P	95 - 105

Reference Method: SM 2120 B

Run ID: A85759

Included Lab Sample IDs: 2015542, 2015543, 2015544, 2015545

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	97.9	98.5	P/P	90 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A85798

Included Lab Sample IDs: 2015550, 2015551, 2015552, 2015553

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	93.6	93.1	P/P	90 - 110
O-Phosphate-P	94.0	92.9	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A85845

Included Lab Sample IDs: 2015543, 2015544, 2015545

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	100	P/P	90 - 110
Sulfate	101	99.5	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A85868

Included Lab Sample IDs: 2015542, 2015543, 2015544, 2015545

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	103	101	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A85868

Included Lab Sample IDs: 2015542, 2015543, 2015544, 2015545

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	102	100	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A85923

Included Lab Sample IDs: 2015546, 2015547, 2015548, 2015549

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	98.5	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A85924

Included Lab Sample IDs: 2015546, 2015547, 2015548, 2015549

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	101	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A85932

Included Lab Sample IDs: 2015546, 2015547, 2015548, 2015549

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	100	99.8	P/P	90 - 110
Ammonia-N	100	100	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A85934

Included Lab Sample IDs: 2015546, 2015547, 2015548, 2015549

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A85935

Included Lab Sample IDs: 2015546, 2015547, 2015548, 2015549

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	96.0	97.1	P/P	90 - 110
Organic Carbon	97.1	96.0	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					4.07	
	Turbidity					1.16	
EPA 300.0 Rev. 2.1	Chloride	97.9	99.1	97.0		6.40	
	Chloride	99.8	97.9	99.4		0.557	
	Sulfate	95.7	98.1	95.4		7.04	
	Sulfate	100	97.1	99.8		0.365	
EPA 350.1 Rev. 2.0	Ammonia-N	96.4	97.1	98.2			1.03
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	101	95.6	98.2			1.80
	Kjeldahl Nitrogen	105	99.3	102			1.76
EPA 353.2 Rev. 2.0	NO2NO3-N	102	93.7	95.2			1.59
EPA 365.1 Rev. 2.0	Total-P	98.8	100	101			0.681
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	96.9	94.0	95.7			1.79
SM 2120 B	Color (true)	97.6				0.610	
SM 2320 B-97	Total Alkalinity	103				0.492	
SM 2540 C-97	TDS					4.30	
SM 4500 F-C-97	Fluoride	99.3	99.2	97.9			0.986
SM 5310 B-00	Organic Carbon	94.8	104				0.0640

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	2015542, 2015543, 2015544, 2015545
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	2015542, 2015543, 2015544, 2015545
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	2015542, 2015543, 2015544, 2015545
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2015546, 2015547, 2015548, 2015549
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2015546, 2015547, 2015548, 2015549
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2015546, 2015547, 2015548, 2015549
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2015546, 2015547, 2015548, 2015549
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2015550, 2015551, 2015552, 2015553
SM 2120 B / E31780	True Color measured at 450 nm	2015542, 2015543, 2015544, 2015545
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO3/L	2015542, 2015543, 2015544, 2015545
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	2015542, 2015543, 2015544, 2015545
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2015542, 2015543, 2015544, 2015545
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2015542, 2015543, 2015544, 2015545
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2015546, 2015547, 2015548, 2015549

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	08/08/2018			08/08/2018 15:53	William L. Brown IV	2015542, 2015543, 2015544, 2015545
EPA 300.0 Rev. 2.1	08/08/2018			08/09/2018 17:18	Lipi Saha	2015542
	08/08/2018			08/13/2018 19:04	Lipi Saha	2015543
	08/08/2018			08/13/2018 19:16	Lipi Saha	2015544
	08/08/2018			08/13/2018 19:28	Lipi Saha	2015545
EPA 350.1 Rev. 2.0	08/08/2018			08/16/2018 09:51	Ping Hua	2015546
	08/08/2018			08/16/2018 10:04	Ping Hua	2015547
	08/08/2018			08/16/2018 10:06	Ping Hua	2015548
	08/08/2018			08/16/2018 10:07	Ping Hua	2015549
EPA 351.2 Rev. 2.0	08/08/2018	08/15/2018 10:30	Adrian Shelby	08/16/2018 10:19	Joseph Suarez	2015546
	08/08/2018	08/15/2018 10:30	Adrian Shelby	08/16/2018 10:24	Joseph Suarez	2015547
	08/08/2018	08/15/2018 10:30	Adrian Shelby	08/16/2018 10:29	Joseph Suarez	2015548
	08/08/2018	08/15/2018 10:30	Adrian Shelby	08/16/2018 10:30	Joseph Suarez	2015549
EPA 353.2 Rev. 2.0	08/08/2018			08/16/2018 12:50	Lauren Bottorf	2015546
	08/08/2018			08/16/2018 12:51	Lauren Bottorf	2015547
	08/08/2018			08/16/2018 12:53	Lauren Bottorf	2015548
	08/08/2018			08/16/2018 12:54	Lauren Bottorf	2015549
EPA 365.1 Rev. 2.0	08/08/2018	08/15/2018 13:20	Sima Feizi	08/16/2018 11:49	Beverly Y. Sanders	2015546
	08/08/2018	08/15/2018 13:20	Sima Feizi	08/16/2018 11:50	Beverly Y. Sanders	2015547
	08/08/2018	08/15/2018 13:20	Sima Feizi	08/16/2018 11:51	Beverly Y. Sanders	2015548
	08/08/2018	08/15/2018 13:20	Sima Feizi	08/16/2018 11:52	Beverly Y. Sanders	2015549
EPA 365.1 Rev. 2.0 dissolved	08/08/2018			08/08/2018 16:32	Megan Treadwell	2015550
	08/08/2018			08/08/2018 17:27	Megan Treadwell	2015551
	08/08/2018			08/08/2018 17:28	Megan Treadwell	2015552
	08/08/2018			08/08/2018 17:29	Megan Treadwell	2015553
SM 2120 B	08/08/2018			08/08/2018 17:08	Tanya Welborn	2015542
	08/08/2018			08/08/2018 17:09	Tanya Welborn	2015543
	08/08/2018			08/08/2018 17:10	Tanya Welborn	2015544
	08/08/2018			08/08/2018 17:11	Tanya Welborn	2015545
SM 2320 B-97	08/08/2018			08/14/2018 10:07	Dale L. Simmons	2015542
	08/08/2018			08/14/2018 10:17	Dale L. Simmons	2015543
	08/08/2018			08/14/2018 10:28	Dale L. Simmons	2015544
	08/08/2018			08/14/2018 10:39	Dale L. Simmons	2015545
SM 2540 C-97	08/08/2018			08/10/2018 16:03	Yijie Li	2015542, 2015543, 2015544, 2015545
SM 2540 D-97	08/08/2018			08/10/2018 16:03	Yijie Li	2015542, 2015543, 2015544, 2015545
SM 4500 F-C-97	08/08/2018			08/14/2018 10:07	Dale L. Simmons	2015542
	08/08/2018			08/14/2018 10:17	Dale L. Simmons	2015543
	08/08/2018			08/14/2018 10:28	Dale L. Simmons	2015544
	08/08/2018			08/14/2018 10:39	Dale L. Simmons	2015545
SM 5310 B-00	08/08/2018			08/15/2018 17:48	Megan Treadwell	2015546

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
SM 5310 B-00	08/08/2018			08/15/2018 18:31	Megan Treadwell	2015547
	08/08/2018			08/15/2018 18:47	Megan Treadwell	2015548
	08/08/2018			08/15/2018 19:00	Megan Treadwell	2015549