

10/2/2019		1430	ET	AC15T 1910
B/K	Std	Rdg	err%	P/F
0.27	5.47	5.58	2.01	P
	56.5	57.8	2.30	P
	578	579	0.17	P

11/5/19		8:33	NAV	ACGTX1911
B/K	Std	Rdg	err%	P/F
0.19	5.47	5.27	3.7	P
	56.5	57.8	2.3	P
	578	586	1.2	P

11/5/19		1345	NAV	ACGTX1911
B/K	Std	Rdg	Err%	P/F
0.22	5.47	5.54	1.3	P
	56.5	55.1	2.5	P
	578	583	0.9	P

		Hatch #5 turb			
11/12/19		8:40	NAV	OCB1911	
B/K	Std	Rdg	Err%	P/F	
0.26	5.47	5.61	2.6	P	
	56.5	55.3	2.12	P	
	578	587	1.56	P	

11/12/19		1550	NAV	OCB1911	
B/K	Std	Rdg	Err%	P/F	
0.30	5.47	5.64	3.1	P	
	56.5	56.4	0.2	P	
	578	582	0.69	P	

11/12/19		578	0930	ET	OCB1911
B/K	Std	Rdg	Err%	P/F	
0.27	5.47	5.43	0.73	P	
	56.5	56.7	0.35	P	
	578	586	1.38	P	

Blank	std	Reading	Err%	P/F
0.29	5.47	5.62	2.74	P
	56.5	56.3	0.35	P
	578	582	0.69	P

Blank	std	Reading	Err%	P/F
0.29	5.47	5.52	0.91	P
	56.5	56.0	0.88	P
	578	587	1.56	P

Blank	std	Reading	Err%	P/F
0.33	5.47	5.58	2.01	P
	56.5	57.8	2.30	P
	578	586	1.38	P

Blank	std	Reading	Err%	P/F
0.32	5.47	5.54	1.3	P
	56.5	56.3	0.4	P
	578	577	0.2	P

Blank	std	Reading	Err%	P/F
0.30	5.47	5.56	1.65	P
	56.5	55.0	2.65	P
	578	583	0.87	P