

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: P20 DSS

RQ: 2019-03-11-17

Project: Nassau West

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification _____

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	<u>AKalmbacher</u>	<u>3-13-19</u>	<u>0755</u>	<u>21.5</u>	<u>762.0</u>	<u>8.82</u>	<u>8.85</u>	<u>100.3</u>	<u>—</u>	<u>—</u>	<u>P/F</u>	<u>L/F</u>
ICV			<u>0758</u>	<u>21.5</u>	<u>762.0</u>	<u>8.82</u>	<u>8.86</u>	<u>100.4</u>	<u>—</u>	<u>—</u>	<u>P/F</u>	<u>L/F</u>
CCV	<u>AKalmbacher</u>	<u>3-13-19</u>	<u>1255</u>	<u>20.3</u>	<u>762</u>	<u>9.03</u>	<u>9.16</u>	<u>101.5</u>	<u>—</u>	<u>—</u>	<u>P/F</u>	<u>L/F</u>
CCV											<u>P/F</u>	<u>L/F</u>

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	<u>AKalmbacher</u>	<u>3-13-19</u>	<u>0800</u>	<u>190125B</u>	<u>2/2020</u>	<u>1.000</u>	<u>1.000</u>	<u>P/F</u>	<u>L/F</u>
ICV			<u>0805</u>	<u>180711B</u>	<u>7/2019</u>	<u>100</u>	<u>102.1</u>	<u>P/F</u>	<u>L/F</u>
CCV	<u>AKalmbacher</u>	<u>3-13-19</u>	<u>1300</u>	<u>190125B</u>	<u>2/2020</u>	<u>1.000</u>	<u>1000</u>	<u>P/F</u>	<u>L/F</u>
CCV								<u>P/F</u>	<u>L/F</u>

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	<u>AKalmbacher</u>	<u>3-13-19</u>	<u>0810</u>	<u>2811E61</u>	<u>11/2020</u>	<u>7.</u>	<u>22.3</u>	<u>7.01</u>	<u>-43.8</u>	<u>P/F</u>	<u>L/F</u>
Calibr.			<u>0813</u>	<u>2811D47</u>	<u>11/2020</u>	<u>4.</u>	<u>22.9</u>	<u>4.00</u>	<u>136.9</u>	<u>P/F</u>	<u>L/F</u>
Calibr.			<u>0815</u>	<u>2810F55</u>	<u>4/2020</u>	<u>10.</u>	<u>22.8</u>	<u>10.03</u>	<u>-215.6</u>	<u>P/F</u>	<u>L/F</u>
ICV			<u>0820</u>	<u>2811E61</u>	<u>11/2020</u>	<u>7.0</u>	<u>23.1</u>	<u>7.03</u>	<u>-41.5</u>	<u>P/F</u>	<u>L/F</u>
CCV	<u>AKalmbacher</u>	<u>3-13-19</u>	<u>1305</u>	<u>2811D47</u>	<u>11/2020</u>	<u>4.0</u>	<u>23.0</u>	<u>4.01</u>	<u>135.2</u>	<u>P/F</u>	<u>L/F</u>
CCV										<u>P/F</u>	<u>L/F</u>

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 171.8, slope from 4 to 7 180.7 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	<u>Amy Kalmbacher</u>	<u>3-13-19</u>	<u>0825</u>	<u>0.0</u>	<u>0.0</u>	<u>P/F</u>	<u>L/F</u>

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

gr 3/14/19

BD 3/25/19

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Y51 #3

RQ: 2019-03-11-18

Project: Nassau North Boat

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 2/2019

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	C. Ruben	3-13-19	0745	21.5	760	8.83	8.85	100	—	—	P/F	L/F
ICV			0748	21.6	760	8.81	8.80	100	—	—	P/F	L/F
CCV	J. Parnish	3/13/19	1305	19.9	760	9.1	9.2	100	—	—	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	C. Ruben	3-13-19	0751	190125B	2/20	1000	998	P/F	L/F
ICV			0753	180711B	7/2019	100	104	P/F	L/F
CCV	J. Parnish	3/13/19	1307	0424820	6/2020	24820	24080	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	C. Ruben	3-13-19	0755	2807A76	6/20	7.	23.1	7.02	-22	P/F	L/F
Calibr.			0758	2811D47	11/20	4.	22.8	3.94	155	P/F	L/F
Calibr.			0801	2810F55	4/20	10.	22.7	9.89	-1922	P/F	L/F
ICV			0805	2807A76	6/20	7.0	23.1	7.02	-210	P/F	L/F
CCV	J. Parnish	3/13/19	1310	2811D47	11/20	4.0	23	4.07	150.4	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

for 3/13/19

TS 3/25/19