

CONTINUING CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID: **YSI Pro Series**

This Sheet is for **CALIBRATION CHECKS** March 2018 version

CCV= Continuing Calibration Verification (done in Read mode)

SS= Same Standard Used for Initial Calibration (Exp Date and Lot # not required if Y is circled)

Date 8/22/19 Time 0831 Circle CCV that applies: End of day Mid-Day Morning

Project Name: OCB 1908 Analyst Name: E. TAVAREZ

Dissolved Oxygen

CCV: 25.0 °C 99.4 % 8.22 mg/L **Pass or Fail** P F

IF oxygen fails re-calibrate and fill out the following:

Calibrated Reading: _____ °C _____ % _____ mg/L **P** **F**

ICV: _____ °C _____ % _____ mg/L **P** **F**

Specific Conductance

SS	Exp. Date	Lot #	Standard	Reading	
CCV: <u>Y</u> <u>N</u>			<u>100</u>	<u>100</u>	<u>P</u> <u>F</u>
CCV: <u>Y</u> <u>N</u>			<u>1000</u>	<u>971</u>	<u>P</u> <u>F</u>
CCV: <u>Y</u> <u>N</u>					P F
			μS/cm	μS/cm	

pH

SS	Exp. Date	Lot #	Standard	Temperature	pH	
CCV: <u>Y</u> <u>N</u>			7.00 SU	<u>25.2</u> °C	<u>7.08</u>	SU <u>P</u> <u>F</u>
CCV: <u>Y</u> <u>N</u>			10.00 SU	<u>25.0</u> °C	<u>10.04</u>	SU <u>P</u> <u>F</u>
CCV: <u>Y</u> <u>N</u>			4.00 SU	<u>25.1</u> °C	<u>4.07</u>	SU <u>P</u> <u>F</u>

Notes:

Were the batteries replaced? Y N

Date 8/22/19 Time 1605 Circle CCV that applies: End of day Mid-Day Morning

Project Name: OCB 1908 Analyst Name: E. TAVAREZ

Dissolved Oxygen

CCV: 29.7 °C 100.4 % 7.63 mg/L **Pass or Fail** P F

IF oxygen fails re-calibrate and fill out the following:

Calibrated Reading: _____ °C _____ % _____ mg/L **P** **F**

ICV: _____ °C _____ % _____ mg/L **P** **F**

Specific Conductance

SS	Exp. Date	Lot #	Standard	Reading	
CCV: <u>Y</u> <u>N</u>			<u>100</u>	<u>100</u>	<u>P</u> <u>F</u>
CCV: <u>Y</u> <u>N</u>			<u>1000</u>	<u>971</u>	<u>P</u> <u>F</u>
CCV: <u>Y</u> <u>N</u>					P F
			μS/cm	μS/cm	

pH

SS	Exp. Date	Lot #	Standard	Temperature	pH	
CCV: <u>Y</u> <u>N</u>			7.00 SU	<u>25.1</u> °C	<u>7.09</u>	SU <u>P</u> <u>F</u>
CCV: <u>Y</u> <u>N</u>			10.00 SU	<u>25.1</u> °C	<u>10.00</u>	SU <u>P</u> <u>F</u>
CCV: <u>Y</u> <u>N</u>			4.00 SU	<u>25.0</u> °C	<u>4.04</u>	SU <u>P</u> <u>F</u>

Notes:

Were the batteries replaced? Y N

Acceptance Criteria: (DO ±0.3mg/l) (Specific Conductance ± 5%) (pH ± 0.2SU) (Temperature ±0.5°C)
Record Temperature and pH to 1 decimal place; D.O. to 2 decimal places