

Chemical Analysis Report

NE-DIST-2019-06-04-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Continuous Monitoring Event**
Request ID: **RQ-2019-03-18-16**
Customer: **NE-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
8800 Baymeadows Way West
Suite 100
Jacksonville, FL 32256
Attn: Brian G Davis

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator

Date Certified: 20-JUN-2019 11:21



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

A - Value reported is the mean of two or more determinations.
B - Results based on colony counts outside the acceptable range.
I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
J - Estimated value and/or the analysis did not meet established quality control criteria.
K - Actual value is known to be less than value given.
L - Actual value is known to be greater than value given.
N - Presumptive evidence of presence of material.
O - Sampled, but analysis lost or not performed.
Q - Sample held beyond normal holding time.
T - Value reported is less than the criterion of detection.
U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
V - Analyte was detected in both sample and method blank.
X - Too few individuals to calculate SCI value.
Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: SHERMAN CR @ CONF SAN PABLO R

Collection Date/Time: 06/03/2019 12:50

Field ID: 20030802

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2091025	EPA 180.1 Rev. 2.0	Turbidity	12		NTU	P365009	
	EPA 300.0	Bromide	65		mg Br/L	P365717	
	SM 2320 B-97	Total Alkalinity	120		mg CaCO3/L	P365562	
	SM 2540 D-97	TSS	29	A	mg/L	P365630	
	SM 4500 F-C-97	Fluoride	1.1		mg F/L	P365568	
2091026	EPA 350.1 Rev. 2.0	Ammonia-N	0.010		mg N/L	P365211	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.38	I	mg N/L	P365328	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.010		mg N/L	P365286	
	EPA 365.1 Rev. 2.0	Total-P	0.10		mg P/L	P365283	
	SM 5310 B-00	Organic Carbon	5.1		mg C/L	P365297	
2091027	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.040		mg P/L	P365000	

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P365009

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0
Batch ID: P365717

Component	Result	Code	Units
Bromide	0.013	U	mg Br/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P365211

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P365328

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P365286

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P365283

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P365000

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2320 B-97
Batch ID: P365562

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 D-97
Batch ID: P365630

Component	Result	Code	Units
TSS	3	U	mg/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 4500 F-C-97
Batch ID: P365568

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P365297

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0
Batch ID: P365717

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Bromide	100		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P365211

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	99.8		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P365328

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	99.7		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P365286

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P365283

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	96.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P365000

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	98.1		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P365562

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	105		P	98 - 106

Reference Method: SM 4500 F-C-97
Batch ID: P365568

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.4		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P365297

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	96.7		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0
Batch ID: P365717

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2094494	Bromide	94.5	95.0	P/P	90 - 110
2094770	Bromide	96.3		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P365211

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2091026	Ammonia-N	105	106	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P365328

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2090888	Kjeldahl Nitrogen	108	106	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P365286

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2090890	NO2NO3-N	104	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P365283

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2090892	Total-P	95.0	93.6	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P365000

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2090841	O-Phosphate-P	98.0	98.9	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P365568

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2090850	Fluoride	98.1	98.1	P/P	89 - 106

Reference Method: SM 5310 B-00
Batch ID: P365297

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2091006	Organic Carbon	95.0	95.9	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P365009

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2090998	Turbidity	5.71	Sample	P	0 - 20

Reference Method: EPA 300.0
Batch ID: P365717

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2094494	Bromide	0.424	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P365211

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2091026	Ammonia-N	0.288	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P365328

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2090888	Kjeldahl Nitrogen	1.48	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P365286

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2090890	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P365283

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2090892	Total-P	1.46	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P365000

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2090841	O-Phosphate-P	0.861	Spike	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P365562

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2090850	Total Alkalinity	0.741	Sample	P	0 - 2.8

Quality Assurance Report Precision

Reference Method: SM 2540 D-97
Batch ID: P365630

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2091025	TSS	4.14	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P365568

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2090850	Fluoride	0.0	Spike	P	0 - 2.5

Reference Method: SM 5310 B-00
Batch ID: P365297

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2091006	Organic Carbon	0.750	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A91848

Included Lab Sample IDs: 2091027

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.1	98.9	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A91849

Included Lab Sample IDs: 2091025

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	98.0	98.5	P/P	95 - 105

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A91914

Included Lab Sample IDs: 2091026

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	102	101	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A91947

Included Lab Sample IDs: 2091026

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	99.5	100	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A91953

Included Lab Sample IDs: 2091026

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	95.8	96.2	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A91989

Included Lab Sample IDs: 2091026

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	99.5	98.8	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A92025

Included Lab Sample IDs: 2091026

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	94.2	94.8	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A92072

Included Lab Sample IDs: 2091025

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	105	102	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 4500 F-C-97

Run ID: A92072

Included Lab Sample IDs: 2091025

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.5	100	P/P	90 - 110

Reference Method: EPA 300.0

Run ID: A92088

Included Lab Sample IDs: 2091025

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Bromide	103	102	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery			Precision SMP	MS
			LCS				
EPA 180.1 Rev. 2.0	Turbidity					5.71	
EPA 300.0	Bromide	100	95.0	96.3	94.5		0.424
EPA 350.1 Rev. 2.0	Ammonia-N	99.8	105	106			0.288
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	99.7	108	106			1.48
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	102	104	104			0.0
EPA 365.1 Rev. 2.0	Total-P	96.5	95.0	93.6			1.46
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	98.1	98.0	98.9			0.861
SM 2320 B-97	Total Alkalinity	105				0.741	
SM 2540 D-97	TSS					4.14	
SM 4500 F-C-97	Fluoride	97.4	98.1	98.1			0.0
SM 5310 B-00	Organic Carbon	96.7	95.0	95.9			0.750

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	2091025
EPA 300.0 / E31780	Bromide in aqueous matrices	2091025
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2091026
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2091026
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2091026
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2091026
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2091027
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2091025
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2091025
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2091025
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2091026

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	06/04/2019			06/04/2019 17:24	Julia Storbeck	2091025
EPA 300.0	06/04/2019			06/14/2019 15:40	Lipi Saha	2091025
EPA 350.1 Rev. 2.0	06/04/2019			06/05/2019 14:01	Ping Hua	2091026
EPA 351.2 Rev. 2.0	06/04/2019	06/10/2019 10:00	Adrian Shelby	06/11/2019 13:15	Joseph Suarez	2091026
EPA 353.2 Rev. 2.0	06/04/2019			06/07/2019 10:08	Beverly Y. Sanders	2091026
EPA 365.1 Rev. 2.0	06/04/2019	06/07/2019 11:40	Sima Feizi	06/10/2019 15:27	Julia Storbeck	2091026
EPA 365.1 Rev. 2.0 dissolved	06/04/2019			06/04/2019 15:44	Jhamieka S. Greenwood	2091027
SM 2320 B-97	06/04/2019			06/13/2019 10:11	Dale L. Simmons	2091025
SM 2540 D-97	06/04/2019			06/06/2019 15:49	Yijie Li	2091025
SM 4500 F-C-97	06/04/2019			06/13/2019 08:52	Dale L. Simmons	2091025
SM 5310 B-00	06/04/2019			06/07/2019 02:40	Julia Storbeck	2091026