

Chemical Analysis Report

NE-DIST-2019-12-12-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **DEP Alachua County Makeup**
Request ID: **RQ-2019-12-09-21**
Customer: **NE-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
8800 Baymeadows Way West
Suite 100
Jacksonville, FL 32256
Attn: Jeremy Parrish

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Dr. rer. nat. Bettina Steinbock, Environmental Administrator

Date Certified: 30-DEC-2019 15:14



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

A - Value reported is the mean of two or more determinations.
B - Results based on colony counts outside the acceptable range.
I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
J - Estimated value and/or the analysis did not meet established quality control criteria.
K - Actual value is known to be less than value given.
L - Actual value is known to be greater than value given.
N - Presumptive evidence of presence of material.
O - Sampled, but analysis lost or not performed.
Q - Sample held beyond normal holding time.
T - Value reported is less than the criterion of detection.
U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
V - Analyte was detected in both sample and method blank.
X - Too few individuals to calculate SCI value.
Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Tumblin Creek at SW 9th

Collection Date/Time: 12/11/2019 12:00

Field ID: G1NE0894

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2143344	EPA 180.1 Rev. 2.0	Turbidity	0.65		NTU	P375815	
	EPA 300.0 Rev. 2.1	Chloride	23		mg Cl/L	P376310	
		Sulfate	43		mg SO4/L	P376312	
	SM 2120 B	Color (true)	13		PCU	P375752	
	SM 2320 B-97	Total Alkalinity	123		mg CaCO3/L	P376251	
	SM 2540 C-97	TDS	211		mg/L	P376416	
	SM 2540 D-97	TSS	2	I	mg/L	P376356	
	SM 4500 F-C-97	Fluoride	0.20		mg F/L	P376254	
2143350	EPA 350.1 Rev. 2.0	Ammonia-N	0.007		mg N/L	P376140	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.32		mg N/L	P375909	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.25		mg N/L	P375893	
	EPA 365.1 Rev. 2.0	Total-P	0.17		mg P/L	P375923	
	SM 5310 B-00	Organic Carbon	4.8		mg C/L	P375848	
2143356	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.13		mg P/L	P375765	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Lake Kanapaha ~300M N of 56Th

Collection Date/Time: 12/11/2019 12:55

Field ID: 20030925

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2143345	EPA 180.1 Rev. 2.0	Turbidity	2.0		NTU	P375815	
	EPA 300.0 Rev. 2.1	Chloride	14		mg Cl/L	P376310	
		Sulfate	8.1		mg SO4/L	P376312	
	SM 2120 B	Color (true)	20	A	PCU	P375752	
	SM 2320 B-97	Total Alkalinity	253		mg CaCO3/L	P376251	
	SM 2540 C-97	TDS	227		mg/L	P376417	
	SM 2540 D-97	TSS	2	I	mg/L	P376357	
	SM 4500 F-C-97	Fluoride	0.12		mg F/L	P376254	
2143351	EPA 350.1 Rev. 2.0	Ammonia-N	0.29		mg N/L	P376275	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.60		mg N/L	P375909	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P375893	
	EPA 365.1 Rev. 2.0	Total-P	0.27		mg P/L	P375923	
	SM 5310 B-00	Organic Carbon	3.6		mg C/L	P375848	
2143357	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.22		mg P/L	P375765	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Turkey Creek at NW 59th

Collection Date/Time: 12/11/2019 11:40

Field ID: G1NE0018

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2143346	EPA 180.1 Rev. 2.0	Turbidity	0.75		NTU	P375815	
	EPA 300.0 Rev. 2.1	Chloride	17		mg Cl/L	P376310	
		Sulfate	16		mg SO4/L	P376312	
	SM 2120 B	Color (true)	87		PCU	P375752	
	SM 2320 B-97	Total Alkalinity	85		mg CaCO3/L	P376251	
	SM 2540 C-97	TDS	158		mg/L	P376417	
	SM 2540 D-97	TSS	2	U	mg/L	P376357	
	SM 4500 F-C-97	Fluoride	0.33		mg F/L	P376254	
2143352	EPA 350.1 Rev. 2.0	Ammonia-N	0.003	I	mg N/L	P376275	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.44		mg N/L	P375909	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P375893	
	EPA 365.1 Rev. 2.0	Total-P	0.80		mg P/L	P375923	
	SM 5310 B-00	Organic Carbon	13		mg C/L	P375848	
2143358	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.69		mg P/L	P375765	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Biven Arm outlet at331

Collection Date/Time: 12/11/2019 13:00

Field ID: G3NE0012

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2143347	EPA 180.1 Rev. 2.0	Turbidity	1.3		NTU	P375815	
	EPA 300.0 Rev. 2.1	Chloride	9.9		mg Cl/L	P376310	
		Sulfate	6.8		mg SO4/L	P376312	
	SM 2120 B	Color (true)	40		PCU	P375753	
	SM 2320 B-97	Total Alkalinity	96		mg CaCO3/L	P376251	
	SM 2540 C-97	TDS	137		mg/L	P376417	
	SM 2540 D-97	TSS	2	I	mg/L	P376357	
	SM 4500 F-C-97	Fluoride	0.14		mg F/L	P376254	
2143353	EPA 350.1 Rev. 2.0	Ammonia-N	0.63		mg N/L	P376275	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.4		mg N/L	P375909	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.26		mg N/L	P375893	
	EPA 365.1 Rev. 2.0	Total-P	0.25		mg P/L	P375923	
	SM 5310 B-00	Organic Carbon	11		mg C/L	P375848	
2143359	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.20		mg P/L	P375765	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Sweetwater CR. Above Wetland

Collection Date/Time: 12/11/2019 12:40

Field ID: G1NE0001

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2143348	EPA 180.1 Rev. 2.0	Turbidity	0.70		NTU	P375815	
	EPA 300.0 Rev. 2.1	Chloride	140		mg Cl/L	P376310	
		Sulfate	130		mg SO4/L	P376312	
	SM 2120 B	Color (true)	15		PCU	P375753	
	SM 2320 B-97	Total Alkalinity	122		mg CaCO3/L	P376251	
	SM 2540 C-97	TDS	575		mg/L	P376417	
	SM 2540 D-97	TSS	2	U	mg/L	P376357	
	SM 4500 F-C-97	Fluoride	0.33		mg F/L	P376254	
2143354	EPA 350.1 Rev. 2.0	Ammonia-N	0.051		mg N/L	P376275	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.98		mg N/L	P375909	
	EPA 353.2 Rev. 2.0	NO2NO3-N	4.7		mg N/L	P375893	
	EPA 365.1 Rev. 2.0	Total-P	0.22		mg P/L	P375923	
	SM 5310 B-00	Organic Carbon	7.4		mg C/L	P375848	
2143360	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.14		mg P/L	P375765	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Possum CR. 400m Downstream of NW 16th ave.

Collection Date/Time: 12/11/2019 10:25

Field ID: G1NE0930

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2143349	EPA 180.1 Rev. 2.0	Turbidity	0.75		NTU	P375816	
	EPA 300.0 Rev. 2.1	Chloride	26	A	mg Cl/L	P376406	
		Sulfate	16	A	mg SO4/L	P376408	
	SM 2120 B	Color (true)	29		PCU	P375753	
	SM 2320 B-97	Total Alkalinity	83		mg CaCO3/L	P376251	
	SM 2540 C-97	TDS	166	A	mg/L	P376417	
	SM 2540 D-97	TSS	2	IA	mg/L	P376357	
	SM 4500 F-C-97	Fluoride	0.38		mg F/L	P376254	
2143355	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	I	mg N/L	P376275	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.26		mg N/L	P375909	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.49		mg N/L	P375894	
	EPA 365.1 Rev. 2.0	Total-P	1.1		mg P/L	P375923	
	SM 5310 B-00	Organic Carbon	3.7		mg C/L	P375848	
2143361	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.95		mg P/L	P375765	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P375815

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P375816

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P376310

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P376312

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P376406

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P376408

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P376140

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P376275

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P375909

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P375893

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P375894

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P375923

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P375765

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P375752

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2120 B
Batch ID: P375753

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P376251

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P376416

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 C-97
Batch ID: P376417

Component	Result	Code	Units
TDS	15	U	mg/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 2540 D-97
Batch ID: P376356

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P376357

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P376254

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P375848

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P376310

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	105		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P376312

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	104		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P376406

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	98.6		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P376408

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	99.3		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P376140

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	98.5		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P376275

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	97.7		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P375909

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	98.8		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P375893

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P375894

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P375923

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P375765

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	98.6		P	90 - 110

Reference Method: SM 2120 B
Batch ID: P375752

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	99.2		P	85 - 115

Reference Method: SM 2120 B
Batch ID: P375753

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	97.9		P	85 - 115

Reference Method: SM 2320 B-97
Batch ID: P376251

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	98 - 106

Reference Method: SM 4500 F-C-97
Batch ID: P376254

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.5		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P375848

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	101		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P376310

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2142804	Chloride	104		P	90 - 110
2143344	Chloride	105		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P376312

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2142804	Sulfate	104		P	90 - 110
2143344	Sulfate	107		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P376406

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2143349	Chloride	99.1		P	90 - 110
2144045	Chloride	98.9		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P376408

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2143349	Sulfate	99.4		P	90 - 110
2144045	Sulfate	99.9		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P376140

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2143253	Ammonia-N	100	97.7	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P376275

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2143352	Ammonia-N	98.6	99.4	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P375909

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2143350	Kjeldahl Nitrogen	94.1	93.5	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P375893

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2143253	NO2NO3-N	98.5	99.5	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P375894

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
---------------	-----------	---------	---------	-----------	----------------

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P375894

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2143371	NO2NO3-N	96.0	95.6	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P375923

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2143316	Total-P	108	105	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P375765

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2143319	O-Phosphate-P	94.2	96.3	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P376254

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2143445	Fluoride	101	100	P/P	89 - 106

Reference Method: SM 5310 B-00
Batch ID: P375848

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2143351	Organic Carbon	99.2	99.2	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P375815

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2143239	Turbidity	3.85	Sample	P	0 - 20

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P375816

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2143402	Turbidity	1.37	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P376310

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2142804	Chloride	0.216	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P376312

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2142804	Sulfate	0.417	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P376406

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2143349	Chloride	0.133	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P376408

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2143349	Sulfate	0.0434	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P376140

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2143253	Ammonia-N	2.03	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P376275

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2143352	Ammonia-N	0.765	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P375909

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2143350	Kjeldahl Nitrogen	0.471	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P375893

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2143253	NO2NO3-N	1.01	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P375894

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2143371	NO2NO3-N	0.434	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P375923

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2143316	Total-P	3.36	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P375765

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2143319	O-Phosphate-P	2.20	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P375752

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2143345	Color (true)	2.49	Sample	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P375753

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2143402	Color (true)	1.05	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P376251

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2143445	Total Alkalinity	0.0509	Sample	P	0 - 2.8

Quality Assurance Report Precision

Reference Method: SM 2540 C-97
Batch ID: P376416

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2143239	TDS	5.88	Sample	P	0 - 10

Reference Method: SM 2540 C-97
Batch ID: P376417

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2143349	TDS	4.20	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P376254

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2143445	Fluoride	0.979	Spike	P	0 - 2.5

Reference Method: SM 5310 B-00
Batch ID: P375848

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2143351	Organic Carbon	0.0853	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B

Run ID: A96307

Included Lab Sample IDs: 2143344, 2143345, 2143346, 2143347, 2143348, 2143349

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	100	99.0	P/P	90 - 115
Color (true)	99.0	100	P/P	90 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A96310

Included Lab Sample IDs: 2143356, 2143357, 2143358, 2143359, 2143360, 2143361

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	94.9	95.2	P/P	90 - 110
O-Phosphate-P	95.2	94.6	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A96325

Included Lab Sample IDs: 2143344, 2143345, 2143346, 2143347, 2143348, 2143349

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	99.5	P/P	95 - 105
Turbidity	99.5	100	P/P	95 - 105

Reference Method: SM 5310 B-00

Run ID: A96344

Included Lab Sample IDs: 2143350, 2143351, 2143352, 2143353, 2143354, 2143355

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	101	94.6	P/P	90 - 110
Organic Carbon	98.4	96.4	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A96357

Included Lab Sample IDs: 2143350, 2143351, 2143352, 2143353, 2143354, 2143355

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	101	101	P/P	90 - 110
NO2NO3-N	101	102	P/P	90 - 110
NO2NO3-N	102	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A96395

Included Lab Sample IDs: 2143350, 2143351, 2143352, 2143353, 2143354, 2143355

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	104	99.8	P/P	90 - 110
Kjeldahl Nitrogen	98.9	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A96431

Included Lab Sample IDs: 2143350, 2143351, 2143352, 2143353, 2143354, 2143355

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	100	101	P/P	90 - 110
Total-P	101	101	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A96431

Included Lab Sample IDs: 2143350, 2143351, 2143352, 2143353, 2143354, 2143355

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	103	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A96459

Included Lab Sample IDs: 2143350

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	100	98.9	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A96509

Included Lab Sample IDs: 2143344, 2143345, 2143346, 2143347, 2143348, 2143349

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	102	P/P	90 - 110
Total Alkalinity	102	101	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A96509

Included Lab Sample IDs: 2143344, 2143345, 2143346, 2143347, 2143348, 2143349

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	101	P/P	90 - 110
Fluoride	101	103	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A96515

Included Lab Sample IDs: 2143351, 2143352, 2143353, 2143354, 2143355

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	98.0	98.9	P/P	90 - 110
Ammonia-N	99.3	100	P/P	90 - 110
Ammonia-N	99.3	98.0	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A96533

Included Lab Sample IDs: 2143344, 2143345, 2143346, 2143347, 2143348

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	102	102	P/P	90 - 110
Chloride	102	103	P/P	90 - 110
Chloride	103	102	P/P	90 - 110
Sulfate	101	102	P/P	90 - 110
Sulfate	102	103	P/P	90 - 110
Sulfate	103	101	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A96570

Included Lab Sample IDs: 2143349

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
-----------	---------	---------	------------	----------------

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A96570

Included Lab Sample IDs: 2143349

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	97.5	96.9	P/P	90 - 110
Sulfate	98.5	97.8	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision	
					LCS	MS
EPA 180.1 Rev. 2.0	Turbidity				3.85	
	Turbidity				1.37	
EPA 300.0 Rev. 2.1	Chloride	105	104	105	0.216	
	Chloride	98.6	99.1	98.9	0.133	
	Sulfate	104	104	107	0.417	
	Sulfate	99.3	99.4	99.9	0.0434	
EPA 350.1 Rev. 2.0	Ammonia-N	98.5	100	97.7		2.03
	Ammonia-N	97.7	98.6	99.4		0.765
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	98.8	94.1	93.5		0.471
EPA 353.2 Rev. 2.0	NO2NO3-N	100	98.5	99.5		1.01
	NO2NO3-N	100	96.0	95.6		0.434
EPA 365.1 Rev. 2.0	Total-P	104	108	105		3.36
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	98.6	94.2	96.3		2.20
SM 2120 B	Color (true)	99.2			2.49	
	Color (true)	97.9			1.05	
SM 2320 B-97	Total Alkalinity	103			0.0509	
SM 2540 C-97	TDS				5.88	
	TDS				4.20	
SM 4500 F-C-97	Fluoride	97.5	101	100		0.979
SM 5310 B-00	Organic Carbon	101	99.2	99.2		0.0853

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	2143344, 2143345, 2143346, 2143347, 2143348, 2143349
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	2143344, 2143345, 2143346, 2143347, 2143348, 2143349
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	2143344, 2143345, 2143346, 2143347, 2143348, 2143349
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2143350, 2143351, 2143352, 2143353, 2143354, 2143355
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2143350, 2143351, 2143352, 2143353, 2143354, 2143355
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2143350, 2143351, 2143352, 2143353, 2143354, 2143355
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2143350, 2143351, 2143352, 2143353, 2143354, 2143355
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2143356, 2143357, 2143358, 2143359, 2143360, 2143361
SM 2120 B / E31780	True Color measured at 450 nm	2143344, 2143345, 2143346, 2143347, 2143348, 2143349
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO3/L	2143344, 2143345, 2143346, 2143347, 2143348, 2143349
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	2143344, 2143345, 2143346, 2143347, 2143348, 2143349
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2143344, 2143345, 2143346, 2143347, 2143348, 2143349
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2143344, 2143345, 2143346, 2143347, 2143348, 2143349
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2143350, 2143351, 2143352, 2143353, 2143354, 2143355

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	12/12/2019			12/12/2019 17:19	Yen Ta	2143344, 2143345, 2143346, 2143347, 2143348, 2143349
EPA 300.0 Rev. 2.1	12/12/2019			12/21/2019 00:41	Lipi Saha	2143344
	12/12/2019			12/21/2019 04:17	Lipi Saha	2143345
	12/12/2019			12/21/2019 04:53	Lipi Saha	2143346
	12/12/2019			12/21/2019 05:05	Lipi Saha	2143347
	12/12/2019			12/21/2019 05:17	Lipi Saha	2143348
	12/12/2019			12/24/2019 22:33	Jhamieka S. Greenwood	2143349
EPA 350.1 Rev. 2.0	12/12/2019			12/17/2019 14:10	Ping Hua	2143350
	12/12/2019			12/18/2019 11:06	Ping Hua	2143354
	12/12/2019			12/18/2019 13:32	Ping Hua	2143351
	12/12/2019			12/18/2019 13:33	Ping Hua	2143352
	12/12/2019			12/18/2019 13:44	Ping Hua	2143353
	12/12/2019			12/18/2019 13:45	Ping Hua	2143355
EPA 351.2 Rev. 2.0	12/12/2019	12/16/2019 12:35	Sima Feizi	12/17/2019 12:34	Jhamieka S. Greenwood	2143350
	12/12/2019	12/16/2019 12:35	Sima Feizi	12/17/2019 12:46	Jhamieka S. Greenwood	2143351
	12/12/2019	12/16/2019 12:35	Sima Feizi	12/17/2019 12:48	Jhamieka S. Greenwood	2143352
	12/12/2019	12/16/2019 12:35	Sima Feizi	12/17/2019 12:49	Jhamieka S. Greenwood	2143353
	12/12/2019	12/16/2019 12:35	Sima Feizi	12/17/2019 12:51	Jhamieka S. Greenwood	2143354
	12/12/2019	12/16/2019 12:35	Sima Feizi	12/17/2019 12:53	Jhamieka S. Greenwood	2143355
EPA 353.2 Rev. 2.0	12/12/2019			12/13/2019 11:08	Beverly Y. Sanders	2143350
	12/12/2019			12/13/2019 11:09	Beverly Y. Sanders	2143351
	12/12/2019			12/13/2019 11:10	Beverly Y. Sanders	2143352
	12/12/2019			12/13/2019 11:11	Beverly Y. Sanders	2143353
	12/12/2019			12/13/2019 11:19	Beverly Y. Sanders	2143355
	12/12/2019			12/13/2019 11:43	Beverly Y. Sanders	2143354
EPA 365.1 Rev. 2.0	12/12/2019	12/16/2019 14:17	Yen Ta	12/17/2019 13:58	Benjamin T. Nordmann	2143350
	12/12/2019	12/16/2019 14:17	Yen Ta	12/17/2019 14:02	Benjamin T. Nordmann	2143351
	12/12/2019	12/16/2019 14:17	Yen Ta	12/17/2019 14:03	Benjamin T. Nordmann	2143352
	12/12/2019	12/16/2019 14:17	Yen Ta	12/17/2019 14:04	Benjamin T. Nordmann	2143353
	12/12/2019	12/16/2019 14:17	Yen Ta	12/17/2019 14:05	Benjamin T. Nordmann	2143354
	12/12/2019	12/16/2019 14:17	Yen Ta	12/17/2019 15:18	Benjamin T. Nordmann	2143355
EPA 365.1 Rev. 2.0 dissolved	12/12/2019			12/12/2019 17:14	Samantha Davila	2143356
	12/12/2019			12/12/2019 17:15	Samantha Davila	2143357
	12/12/2019			12/12/2019 17:16	Samantha Davila	2143358
	12/12/2019			12/12/2019 17:17	Samantha Davila	2143359
	12/12/2019			12/12/2019 17:22	Samantha Davila	2143360
	12/12/2019			12/12/2019 17:23	Samantha Davila	2143361
SM 2120 B	12/12/2019			12/12/2019 16:46	Madeline Funaro	2143344
	12/12/2019			12/12/2019 16:47	Madeline Funaro	2143345

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
SM 2120 B	12/12/2019			12/12/2019 16:48	Madeline Funaro	2143346
	12/12/2019			12/12/2019 16:52	Madeline Funaro	2143347, 2143348
	12/12/2019			12/12/2019 16:53	Madeline Funaro	2143349
SM 2320 B-97	12/12/2019			12/17/2019 18:08	Dale L. Simmons	2143344
	12/12/2019			12/17/2019 18:22	Dale L. Simmons	2143345
	12/12/2019			12/17/2019 18:34	Dale L. Simmons	2143346
	12/12/2019			12/17/2019 19:15	Dale L. Simmons	2143347
	12/12/2019			12/17/2019 19:28	Dale L. Simmons	2143348
	12/12/2019			12/17/2019 19:41	Dale L. Simmons	2143349
	12/12/2019			12/17/2019 16:36	Yijie Li	2143344, 2143345, 2143346, 2143347, 2143348, 2143349
SM 2540 C-97	12/12/2019					
SM 2540 D-97	12/12/2019			12/17/2019 16:36	Yijie Li	2143344, 2143345, 2143346, 2143347, 2143348, 2143349
SM 4500 F-C-97	12/12/2019			12/17/2019 18:08	Dale L. Simmons	2143344
	12/12/2019			12/17/2019 18:22	Dale L. Simmons	2143345
	12/12/2019			12/17/2019 18:34	Dale L. Simmons	2143346
	12/12/2019			12/17/2019 19:15	Dale L. Simmons	2143347
	12/12/2019			12/17/2019 19:28	Dale L. Simmons	2143348
	12/12/2019			12/17/2019 19:41	Dale L. Simmons	2143349
SM 5310 B-00	12/12/2019			12/13/2019 18:58	Madeline Funaro	2143351
	12/12/2019			12/13/2019 19:52	Madeline Funaro	2143353
	12/12/2019			12/14/2019 00:16	Madeline Funaro	2143350
	12/12/2019			12/14/2019 00:29	Madeline Funaro	2143352
	12/12/2019			12/14/2019 00:42	Madeline Funaro	2143354
	12/12/2019			12/14/2019 00:55	Madeline Funaro	2143355