

Chemical Analysis Report

CEN-DIST-2019-05-09-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Lk Van / Bass / Panther X3**
Request ID: **RQ-2019-04-29-45**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Mike Linger

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator

Date Certified: 29-MAY-2019 14:03



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Lake Griffin @ Center

Collection Date/Time: 05/08/2019 10:42

Field ID: G2CE0207

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2085033	EPA 180.1 Rev. 2.0	Turbidity	1.8		NTU	P363735	
	EPA 300.0 Rev. 2.1	Chloride	15		mg Cl/L	P363882	
		Sulfate	7.5		mg SO4/L	P363885	
	SM 2120 B	Color (true)	21		PCU	P363727	
	SM 2320 B-97	Total Alkalinity	55		mg CaCO3/L	P364259	
	SM 2540 C-97	TDS	106		mg/L	P364145	
	SM 2540 D-97	TSS	2	I	mg/L	P364240	
	SM 4500 F-C-97	Fluoride	0.066	I	mg F/L	P364263	
2085035	EPA 350.1 Rev. 2.0	Ammonia-N	0.004	I	mg N/L	P364044	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.64		mg N/L	P364228	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P364119	
	EPA 365.1 Rev. 2.0	Total-P	0.017		mg P/L	P363966	
	SM 5310 B-00	Organic Carbon	8.5		mg C/L	P364050	
2085037	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P363701	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Lake Bass @ Center of Lake

Collection Date/Time: 05/08/2019 11:46

Field ID: G4CE0147

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2085034	EPA 180.1 Rev. 2.0	Turbidity	11		NTU	P363735	
	EPA 300.0 Rev. 2.1	Chloride	22		mg Cl/L	P363882	
		Sulfate	22		mg SO4/L	P363885	
	SM 2120 B	Color (true)	14		PCU	P363727	
	SM 2320 B-97	Total Alkalinity	44		mg CaCO3/L	P364259	
	SM 2540 C-97	TDS	113		mg/L	P364145	
	SM 2540 D-97	TSS	6	I	mg/L	P364240	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P364263	
2085036	EPA 350.1 Rev. 2.0	Ammonia-N	0.004	I	mg N/L	P364044	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.2		mg N/L	P364076	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P364120	
	EPA 365.1 Rev. 2.0	Total-P	0.041		mg P/L	P363966	
	SM 5310 B-00	Organic Carbon	6.4		mg C/L	P364050	
2085038	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P363702	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P363735

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P363882

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P363885

Component	Result	Code	Units
Sulfate	0.20	U	mg SO ₄ /L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P364044

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P364076

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P364228

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P364119

Component	Result	Code	Units
NO ₂ NO ₃ -N	0.004	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P364120

Component	Result	Code	Units
NO ₂ NO ₃ -N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P363966

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P363701

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P363702

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P363727

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P364259

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-97
Batch ID: P364145

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P364240

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P364263

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P364050

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P363882

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	101		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P363885

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	101		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P364044

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	102		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P364076

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	101		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P364228

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	104		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P364119

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	99.0		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P364120

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	99.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P363966

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P363701

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	100		P	90 - 110

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P363702

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.9		P	90 - 110

Reference Method: SM 2120 B
Batch ID: P363727

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	99.6		P	85 - 115

Reference Method: SM 2320 B-97
Batch ID: P364259

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	98 - 106

Reference Method: SM 4500 F-C-97
Batch ID: P364263

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.5		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P364050

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	98.1		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P363882

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2084669	Chloride	100		P	90 - 110
2085021	Chloride	101		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P363885

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2084669	Sulfate	100		P	90 - 110
2085021	Sulfate	99.8		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P364044

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2085035	Ammonia-N	101	102	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P364076

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2085064	Kjeldahl Nitrogen	105	113	P/F	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P364228

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2085280	Kjeldahl Nitrogen	110	109	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P364119

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2084905	NO2NO3-N	99.4	98.4	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P364120

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2085036	NO2NO3-N	102	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P363966

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2085036	Total-P	99.6	99.9	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P363701

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2085037	O-Phosphate-P	90.3	91.7	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P363702

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2085189	O-Phosphate-P	94.2	95.9	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P364263

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2085112	Fluoride	100	98.1	P/P	89 - 106

Reference Method: SM 5310 B-00
Batch ID: P364050

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2085035	Organic Carbon	98.2	97.3	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P363735

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2084902	Turbidity	0.242	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P363882

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2084669	Chloride	0.262	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P363885

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2084669	Sulfate	0.129	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P364044

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2085035	Ammonia-N	1.17	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P364076

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2085064	Kjeldahl Nitrogen	5.20	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P364228

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2085280	Kjeldahl Nitrogen	0.294	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P364119

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2084905	NO2NO3-N	0.952	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P364120

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2085036	NO2NO3-N	0.0	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P363966

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2085036	Total-P	0.227	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P363701

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2085037	O-Phosphate-P	1.54	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P363702

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2085189	O-Phosphate-P	1.79	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P363727

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2085062	Color (true)	4.53	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P364259

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2085112	Total Alkalinity	0.366	Sample	P	0 - 2.8

Reference Method: SM 2540 C-97
Batch ID: P364145

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2085027	TDS	6.42	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P364263

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2085112	Fluoride	1.45	Spike	P	0 - 2.5

Reference Method: SM 5310 B-00
Batch ID: P364050

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2085035	Organic Carbon	0.641	Spike	P	0 - 20

Quality Assurance Report

Precision

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A91331

Included Lab Sample IDs: 2085037, 2085038

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	98.0	97.7	P/P	90 - 110
O-Phosphate-P	98.3	99.9	P/P	90 - 110

Reference Method: SM 2120 B

Run ID: A91338

Included Lab Sample IDs: 2085033, 2085034

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	102	102	P/P	90 - 115

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A91343

Included Lab Sample IDs: 2085033, 2085034

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	99.5	P/P	95 - 105

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A91404

Included Lab Sample IDs: 2085033, 2085034

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	102	103	P/P	90 - 110
Sulfate	103	103	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A91468

Included Lab Sample IDs: 2085035, 2085036

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	103	104	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A91470

Included Lab Sample IDs: 2085035, 2085036

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	99.1	99.4	P/P	90 - 110
Organic Carbon	99.4	99.2	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A91494

Included Lab Sample IDs: 2085035, 2085036

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	99.5	99.5	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A91516

Included Lab Sample IDs: 2085036

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	102	96.8	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A91522

Included Lab Sample IDs: 2085035, 2085036

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	99.8	99.2	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A91544

Included Lab Sample IDs: 2085033, 2085034

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	102	99.8	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A91544

Included Lab Sample IDs: 2085033, 2085034

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.0	99.5	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A91560

Included Lab Sample IDs: 2085035

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	101	100	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision	
					LCS	MS
EPA 180.1 Rev. 2.0	Turbidity				0.242	
EPA 300.0 Rev. 2.1	Chloride	101	100	101	0.262	
	Sulfate	101	100	99.8	0.129	
EPA 350.1 Rev. 2.0	Ammonia-N	102	101	102		1.17
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	101	105	113		5.20
	Kjeldahl Nitrogen	104	110	109		0.294
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	99.0	99.4	98.4		0.952
	NO ₂ NO ₃ -N	99.5	102	102		0.0
EPA 365.1 Rev. 2.0	Total-P	100	99.6	99.9		0.227
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	100	90.3	91.7		1.54
	O-Phosphate-P	97.9	94.2	95.9		1.79
SM 2120 B	Color (true)	99.6			4.53	
SM 2320 B-97	Total Alkalinity	102			0.366	
SM 2540 C-97	TDS				6.42	
SM 4500 F-C-97	Fluoride	98.5	100	98.1		1.45
SM 5310 B-00	Organic Carbon	98.1	98.2	97.3		0.641

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	2085033, 2085034
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	2085033, 2085034
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	2085033, 2085034
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2085035, 2085036
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2085035, 2085036
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2085035, 2085036
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2085035, 2085036
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2085037, 2085038
SM 2120 B / E31780	True Color measured at 450 nm	2085033, 2085034
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2085033, 2085034
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	2085033, 2085034
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2085033, 2085034
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2085033, 2085034
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2085035, 2085036

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	05/09/2019			05/09/2019 16:45	Sima Feizi	2085033, 2085034
EPA 300.0 Rev. 2.1	05/09/2019			05/14/2019 06:43	Lipi Saha	2085033
	05/09/2019			05/14/2019 06:56	Lipi Saha	2085034
EPA 350.1 Rev. 2.0	05/09/2019			05/15/2019 11:54	Ping Hua	2085035
	05/09/2019			05/15/2019 12:03	Ping Hua	2085036
EPA 351.2 Rev. 2.0	05/09/2019	05/16/2019 12:45	Adrian Shelby	05/17/2019 16:08	Joseph Suarez	2085036
	05/09/2019	05/20/2019 10:15	Jhamieka S. Greenwood	05/21/2019 09:45	Joseph Suarez	2085035
EPA 353.2 Rev. 2.0	05/09/2019			05/17/2019 09:31	Beverly Y. Sanders	2085035
	05/09/2019			05/17/2019 09:38	Beverly Y. Sanders	2085036
EPA 365.1 Rev. 2.0	05/09/2019	05/15/2019 11:26	Joseph Suarez	05/17/2019 14:10	Rosalyn Ballman	2085035
	05/09/2019	05/15/2019 11:26	Joseph Suarez	05/17/2019 14:11	Rosalyn Ballman	2085036
EPA 365.1 Rev. 2.0 dissolved	05/09/2019			05/09/2019 14:33	Jhamieka S. Greenwood	2085037
	05/09/2019			05/09/2019 15:21	Jhamieka S. Greenwood	2085038
SM 2120 B	05/09/2019			05/09/2019 14:51	Blanca Fach	2085033, 2085034
SM 2320 B-97	05/09/2019			05/18/2019 06:13	Dale L. Simmons	2085033
	05/09/2019			05/18/2019 06:27	Dale L. Simmons	2085034
SM 2540 C-97	05/09/2019			05/10/2019 16:05	Yijie Li	2085033, 2085034
SM 2540 D-97	05/09/2019			05/10/2019 16:05	Yijie Li	2085033, 2085034
SM 4500 F-C-97	05/09/2019			05/18/2019 06:13	Dale L. Simmons	2085033
	05/09/2019			05/18/2019 06:27	Dale L. Simmons	2085034
SM 5310 B-00	05/09/2019			05/15/2019 22:57	Julia Storbeck	2085035
	05/09/2019			05/16/2019 00:12	Julia Storbeck	2085036