

Chemical Analysis Report

CEN-DIST-2019-11-21-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **SJR ab Lk Winder + HA**
Request ID: **RQ-2019-11-18-58**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Mike Linger

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Kerry Tate, Ph.D., Environmental Administrator

Date Certified: 11-DEC-2019 11:16

A handwritten signature in black ink that reads "Kerry Tate". The signature is written in a cursive, flowing style with a horizontal line above the first few letters of the last name.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Wekiva River Upstream of Little Wekiva Confluence

Collection Date/Time: 11/20/2019 11:39

Field ID: G2CE0205

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2138154	EPA 180.1 Rev. 2.0	Turbidity	1.3		NTU	P374651	
	EPA 300.0 Rev. 2.1	Chloride	17		mg Cl/L	P375347	
		Sulfate	22		mg SO4/L	P375351	
	SM 2120 B	Color (true)	37	A	PCU	P374635	
	SM 2320 B-97	Total Alkalinity	124	A	mg CaCO3/L	P374989	
	SM 2540 C-97	TDS	199		mg/L	P375144	
	SM 2540 D-97	TSS	3	I	mg/L	P375128	
	SM 4500 F-C-97	Fluoride	0.15		mg F/L	P374994	
2138155	EPA 350.1 Rev. 2.0	Ammonia-N	0.005		mg N/L	P374934	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.37		mg N/L	P374850	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.88		mg N/L	P374756	
	EPA 365.1 Rev. 2.0	Total-P	0.10		mg P/L	P375121	
	SM 5310 B-00	Organic Carbon	4.9		mg C/L	P374888	
2138156	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.089		mg P/L	P374598	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P374651

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P375347

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P375351

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P374934

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P374850

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P374756

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P375121

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P374598

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P374635

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report

Method Blank Results

Reference Method: SM 2320 B-97
Batch ID: P374989

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-97
Batch ID: P375144

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P375128

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P374994

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P374888

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P375347

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	101		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P375351

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	100		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P374934

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	103		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P374850

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	101		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P374756

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P375121

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	107		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P374598

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	96.1		P	90 - 110

Reference Method: SM 2120 B
Batch ID: P374635

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	100		P	85 - 115

Reference Method: SM 2320 B-97
Batch ID: P374989

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	98 - 106

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: SM 4500 F-C-97
Batch ID: P374994

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.2		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P374888

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	106		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P375347

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2138054	Chloride	99.2		P	90 - 110
2138446	Chloride	99.4		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P375351

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2138054	Sulfate	97.6		P	90 - 110
2138446	Sulfate	96.2		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P374934

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2137261	Ammonia-N	102	104	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P374850

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2138058	Kjeldahl Nitrogen	94.5	102	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P374756

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2138062	NO2NO3-N	100	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P375121

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2138055	Total-P	101	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P374598

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2138031	O-Phosphate-P	101	96.7	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P374994

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2138154	Fluoride	97.9	96.6	P/P	89 - 106

Reference Method: SM 5310 B-00
Batch ID: P374888

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2137969	Organic Carbon	106	105	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P374651

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2138050	Turbidity	1.76	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P375347

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2138446	Chloride	0.439	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P375351

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2138446	Sulfate	0.674	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P374934

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2137261	Ammonia-N	1.35	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P374850

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2138058	Kjeldahl Nitrogen	7.00	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P374756

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2138062	NO2NO3-N	0.837	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P375121

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2138055	Total-P	1.49	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P374598

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2138031	O-Phosphate-P	4.45	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2120 B
Batch ID: P374635

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2138154	Color (true)	1.21	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P374989

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2138154	Total Alkalinity	0.125	Sample	P	0 - 2.8

Reference Method: SM 2540 C-97
Batch ID: P375144

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2138053	TDS	3.88	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P374994

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2138154	Fluoride	0.960	Spike	P	0 - 2.5

Reference Method: SM 5310 B-00
Batch ID: P374888

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2137969	Organic Carbon	0.632	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A95880

Included Lab Sample IDs: 2138156

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	101	P/P	90 - 110

Reference Method: SM 2120 B

Run ID: A95896

Included Lab Sample IDs: 2138154

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	101	101	P/P	90 - 115

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A95900

Included Lab Sample IDs: 2138154

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	101	101	P/P	95 - 105

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A95937

Included Lab Sample IDs: 2138155

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	101	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A95973

Included Lab Sample IDs: 2138155

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	102	102	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A95997

Included Lab Sample IDs: 2138155

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A96013

Included Lab Sample IDs: 2138155

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	97.9	98.6	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A96025

Included Lab Sample IDs: 2138154

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	103	101	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 4500 F-C-97

Run ID: A96025

Included Lab Sample IDs: 2138154

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.5	100	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A96087

Included Lab Sample IDs: 2138154

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	102	P/P	90 - 110
Sulfate	101	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A96108

Included Lab Sample IDs: 2138155

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	97.0	99.0	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision	
					LCS	MS
EPA 180.1 Rev. 2.0	Turbidity				1.76	
EPA 300.0 Rev. 2.1	Chloride	101	99.4	99.2	0.439	
	Sulfate	100	96.2	97.6	0.674	
EPA 350.1 Rev. 2.0	Ammonia-N	103	102	104		1.35
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	101	94.5	102		7.00
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	100	100	102		0.837
EPA 365.1 Rev. 2.0	Total-P	107	101	103		1.49
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	96.1	101	96.7		4.45
SM 2120 B	Color (true)	100			1.21	
SM 2320 B-97	Total Alkalinity	103			0.125	
SM 2540 C-97	TDS				3.88	
SM 4500 F-C-97	Fluoride	97.2	97.9	96.6		0.960
SM 5310 B-00	Organic Carbon	106	106	105		0.632

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	2138154
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	2138154
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	2138154
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2138155
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2138155
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2138155
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2138155
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2138156
SM 2120 B / E31780	True Color measured at 450 nm	2138154
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2138154
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	2138154
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2138154
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2138154
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2138155

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	11/21/2019			11/21/2019 17:41	Samantha Davila	2138154
EPA 300.0 Rev. 2.1	11/21/2019			12/04/2019 21:26	Lipi Saha	2138154
EPA 350.1 Rev. 2.0	11/21/2019			11/26/2019 11:28	Ping Hua	2138155
EPA 351.2 Rev. 2.0	11/21/2019	11/26/2019 12:44	Sima Feizi	11/27/2019 15:11	Jhamieka S. Greenwood	2138155
EPA 353.2 Rev. 2.0	11/21/2019			11/25/2019 12:14	Beverly Y. Sanders	2138155
EPA 365.1 Rev. 2.0	11/21/2019	12/03/2019 13:30	Lipi Saha	12/04/2019 12:24	Benjamin T. Nordmann	2138155
EPA 365.1 Rev. 2.0 dissolved	11/21/2019			11/21/2019 13:27	Lipi Saha	2138156
SM 2120 B	11/21/2019			11/21/2019 16:34	Blanca Fach	2138154
SM 2320 B-97	11/21/2019			11/27/2019 11:09	Dale L. Simmons	2138154
SM 2540 C-97	11/21/2019			11/25/2019 15:56	Yijie Li	2138154
SM 2540 D-97	11/21/2019			11/25/2019 15:56	Yijie Li	2138154
SM 4500 F-C-97	11/21/2019			11/27/2019 10:59	Dale L. Simmons	2138154
SM 5310 B-00	11/21/2019			11/26/2019 07:52	Lauren Lees	2138155