

December 11, 2019

Michael Linger
Department of Environmental Protection
3319 Maguire Blvd
Suite 232
Orlando, FL 32803


RE: Workorder: A1910068 SMP

Dear Michael Linger:

Enclosed are the analytical results for sample(s) received by the laboratory on Tuesday, November 26, 2019. Results reported herein conform to the most current NELAC standards, where applicable, unless otherwise narrated in the body of the report. The analytical results for the samples contained in this report were submitted for analysis as outlined by the Chain of Custody and results pertain only to these samples.

If you have any questions concerning this report, please feel free to contact me.

Sincerely,



Brandon O'Hara - Lab Manager
BOhara@AELLab.com

Enclosures

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Advanced Environmental Laboratories, Inc.



SAMPLE SUMMARY

Workorder: A1910068 SMP

Lab ID	Sample ID	Matrix	Date Collected	Date Received
A1910068001	G2CE0202	Water	11/26/2019 10:12	11/26/2019 11:44

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Advanced Environmental Laboratories, Inc.



ANALYTICAL RESULTS

Workorder: A1910068 SMP

Lab ID: **A1910068001**

Date Received: 11/26/19 11:44 Matrix: Water

Sample ID: **G2CE0202**

Date Collected: 11/26/19 10:12

Sample Description:

Location:

Parameters	Results	Qual	Units	DF	Adjusted PQL	Adjusted MDL	Analyzed	Lab
Microbiology								
Analysis Desc: Escherichia Coli,EPA 1603,Water			Analytical Method: EPA 1603					
E. Coli	20	1,B	#/100 mL	10	10	10	11/26/2019 14:24	A

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Advanced Environmental Laboratories, Inc.



ANALYTICAL RESULTS QUALIFIERS

Workorder: A1910068 SMP

PARAMETER QUALIFIERS

- U The compound was analyzed for but not detected.
- I The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- [1] Time In Incubator #1: 11/26/2019 14:24
Time Out Incubator #1: 11/26/2019 16:31
Time In Incubator #2: 11/26/2019 16:35
Time Out Incubator #2: 11/27/2019 14:45
- B Results based upon colony counts outside the acceptable range.

LAB QUALIFIERS

- A DOH Certification #E53076(AEL-A)(FL NELAC Certification)

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Advanced Environmental Laboratories, Inc.



QUALITY CONTROL DATA

Workorder: A1910068 SMP

QC Batch:	MICa/1639	Analysis Method:	EPA 1603
QC Batch Method:	EPA 1603	Prepared:	
Associated Lab Samples:	A1910068001		

METHOD BLANK: 3317190

Parameter	Units	Blank Result	Reporting Limit Qualifiers
Microbiology			
E. Coli	#/100 mL	1	1 U

QUALITY CONTROL DATA QUALIFIERS

Workorder: A1910068 SMP

QUALITY CONTROL PARAMETER QUALIFIERS

- U The compound was analyzed for but not detected.
- I The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- B Results based upon colony counts outside the acceptable range.

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Advanced Environmental Laboratories, Inc.

QUALITY CONTROL DATA CROSS REFERENCE TABLE

Workorder: A1910068 SMP

Lab ID	Sample ID	Prep Method	Prep Batch	Analysis Method	Analysis Batch
A1910068001	G2CE0202			EPA 1603	MICa/1639


CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Advanced Environmental Laboratories, Inc.





- | | |
|---|--|
| ✓ | Altamonte Springs: 380 Northlake Blvd., Suite 1046 • Altamonte Springs, FL 32714 • 407.351.3333 |
| | Gainesville: 4965 SW 41st Blvd. • Gainesville, FL 32608 • 352.377.1111 |
| | Jacksonville: 6681 Southpoint Pkwy. • Jacksonville, FL 32216 • 904.426.1111 |
| | Miramar: 10200 USA Today Way • Miramar, FL 33025 • 954.889.2222 |
| | Tallahassee: 2639 North Monroe Street, Suite D • Tallahassee, FL 32301 • 904.497.1111 |
| | Tampa: 9610 Princess Palm Ave. • Tampa, FL 33619 • 813.630.9610 |



Wednesday, December 11, 2019 8:10:14 AM
Page 7 of 7