From: katie sanderslabs.net
To: Sprague, Kara
Subject: Re: 1903350
Date: Monday, April 1, 2019 8:54:19 AM

For Ecoli it was the value 49 large 47 small wells so 2419 . For TC they were > but you do not get that reported. Thanks sorry for the delay I have been out so much since two my field people quit trying help keep that department afloat. But we finally made some hires so with training we will be back on track. I am sure you know the struggle of finding good people. Thanks

Katie Strothman
Laboratory Director/ QCM
Sanders Laboratories, Inc.
10090 Bavaria Rd
Ft Myers, FL 33913
941-915-7447 C
239-590-0337 O
Monday - Thursday : 8:30 am - 4:00 pm
Friday: 8:30 am - 1:00 pm
Saturday / Sunday: By appointment only
Closed Holidays
I will be on vacation form March 14-22nd. Please email Jeff at jeff@sanderslabs.net if you have any pressing issues. Join me on LinkedIn. We offer Legionella testing. Sanders will always match or beat any quoted price.
The information contained in this email is privileged and confidential information, intended only for the use of the individual (s) distributed to. If the reader of this message is not the intended recipient, you are hereby notified that any distribution or copying of this is strictly prohibited. If you have received this in error, please notify promptly.

From: Sprague, Kara <Kara.Sprague@FloridaDEP.gov>
Sent: Friday, March 29, 2019 11:20:06 AM
To: katie sanderslabs.net
Subject: 1903350

Hello,

Can you take a look at this report? For field ID G1SD0050, was the tray maxed out? Or, was that the actual result from the mpn generator? If it was maxed out, can you add the J with the comment of "result may be higher than reported"?

Thanks Katie!

Kara

Kara Sprague
Environmental Specialist I
FDEP Laboratory - Biology Program
Florida Department of Environmental Protection
Kara.Sprague@Floridadep.gov
Office: 850.245.8826