


BENCHMARK EA SOUTH



ANALYTICAL TEST REPORT

THESE RESULTS MEET NELAC STANDARDS

NELAC Certification # E85086

Submission Number : S 19040044

Florida Dept Of Env Protection
2600 Blair Stone Rd
Tallahassee, FL 32399

Project Name : RQ-2019-04-01-64
Date Received : 04/02/2019
Time Received : 1408

Submission: S 19040044

Sample Number: 001 Sample Description: G2SD0033 - Alligator Creek South
Sample Date: 04/02/2019 Sample Method: Grab
Sample Time: 0925

Parameter	Result	Units	MDL	PQL	Procedure	Preparation Date Time	Analysis Date Time	Analyst
E- COLI BY IDEXX QUANTITRAY	109	#/100ML	10	10	SM9223B	04/02/2019 15:20	04/03/2019 15:50	MM

Melinda Merchant

04/08/2019

Dale D. Dixon / Laboratory Director
Melinda Merchant / Laboratory Manager

Date

DATA QUALIFIERS THAT MAY APPLY:

A = Value reported is an average of two or more determinations.
B = Results based upon colony counts outside the ideal range.
H = Value based on field kit determination. Results may not be accurate.
I = Reported value is between the laboratory MDL and the PQL.
J0 = All Quanti-Tray wells were positive. Result may be higher than reported.
J1 = Est. value surrogate recovery limits exceeded.
J2 = Est. value. No quality control criteria exists for component.
J3 = Est. value quality control criteria for precision or accuracy not met.
J4 = Est. value. Sample matrix interference suspected.
J5 = Est. value. Data questionable due to improper lab or field protocols
K = Off-scale low. Value is known to be < the value reported.
L = Off-scale high. Value is known to be > the value reported

N = Presumptive evidence of presence of material.
Q = Sample held beyond accepted hold time(15 minutes for pH).
T = Value reported is < MDL. Reported for informational purposes only and shall not be used in statistical analysis.
U = Analyte analyzed but not detected at the value indicated.
V = Analyte detected in sample and method blank.
Y = Analysis performed on an improperly preserved sample. Data may be inaccurate.
Z = Too many colonies were present (TNTC); the numeric value represents the filtration volume.
? = Data rejected and should not be used. Some or all of QC data were outside criteria, and the Presence or absence of the analyte cannot be determined from the data.
! = Data deviate from historically established concentration ranges.
* = Not reported due to interference.

NOTES:

PQL = 4xMDL. Except for Micro PQL=MDL

X = Value exceed MCL.

1 dilution was performed on each sample.

For questions and comments regarding these results, please contact Melinda Merchant at (941) 625-3137

These results relate only to the samples included in this report.

1001 Corporate Ave, Suite 102 North Port, FL 34289 Phone: (941) 625-3137 Fax: (941) 240-3071

Benchmark EnviroAnalytical, Inc.
1001 Corporate Ave, Suite 102
North Port, FL 34289
(941) 625-3137 / (941) 240-3071 fax

Client Name: FDEP
Address: 2600 Blair Stone Rd.
Tallahassee, FL, 32399
Phone: (850) 245-8085
Email: dep-overflowmicro@dep.state.fl.us

Bill to: Frances Moore
P.O. #: Lab B16970

Project Name: RQ-2019-04-01-64
Project Run/Location: HA's - Alligator & Hudson

Laboratory Submission #: S19040044

Station ID	Surface Water Sample Matrix ² (Fresh / Marine) Sample Type	Enterococci E.coli (SM9223B Quantitray)	Laboratory Sample #
		E. coli = 250 mL ; Entero = 120 mL ; Fecal Coliform = 120 mL	
G2SD0033-Alligator Creek South	F / M / Grab	E.coli / Entero / F.coli	1
	F / M / Grab	E.coli / Entero / F.coli Date & Time: 04/02/2019 0925	
	F / M / Grab	E.coli / Entero / F.coli Date & Time:	
	F / M / Grab	E.coli / Entero / F.coli Date & Time:	
	F / M / Grab	E.coli / Entero / F.coli Date & Time:	
	F / M / Grab	E.coli / Entero / F.coli Date & Time:	
	F / M / Grab	E.coli / Entero / F.coli Date & Time:	
	F / M / Grab	E.coli / Entero / F.coli Date & Time:	
	F / M / Grab	E.coli / Entero / F.coli Date & Time:	
	F / M / Grab	E.coli / Entero / F.coli Date & Time:	
	F / M / Grab	E.coli / Entero / F.coli Date & Time:	
	F / M / Grab	E.coli / Entero / F.coli Date & Time:	

¹ Sample Type is used to indicate whether the sample was a grab (G) or whether it was a composite (C).
² Sample Matrix is used to indicate whether the sample is being discharged to drinking water (DW), groundwater (GW), surface water (SW), fresh surface water (FSW), saline surface water (SSW), soil sediment (SDMNT), or sludge (SLDG).
³ Container Type is used to indicate whether the container is plastic (P) or glass (G).
⁴ Sample must be refrigerated or stored in ice after collection. The maximum temperature during storage should be 6°C (41°F). Under "Preservative," list any preservatives that were added to the sample container.

Instructions:
1. Each bottle has a label identifying sample ID, premeasured preservative contained in the bottle, sample type, client ID, and parameters for analysis.
2. The following information should be added to each bottle label after collection with permanent black ink: date and time of collection, sampler's name or initials, and any field number or ID.
3. All bottles not containing preservative may be rinsed with appropriate sample prior to collection.
4. The client is responsible for documentation of the sampling event. Please note special sampling events on the sample custody form.

Laboratory Sample Acceptability:
pH < 2.0 BEAS Temperature: 1.9°C

1	Collection: Hannah Jagan	Date: 04/02/19	Time: 1408	Received By:	Date:	Time:
2	Relinquished by: W. J. Jagan	Date: 04/02/19	Time: 1408	Received By: W. J. Jagan	Date: 4/2/19	Time: 1408
3	Relinquished by:	Date:	Time:	Received By:	Date:	Time:
4	Relinquished by:	Date:	Time:	Received By:	Date:	Time:
5	Relinquished by:	Date:	Time:	Received By:	Date:	Time: