

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID: 123470

RQ: 2019-06-17-33

Project: Cont. Monitoring

- Notes:** (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site _____

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 5/6/19

DO FDEP SOP FT 1500	Name	Date	Time	Temp	DO Chart mg/L	Meter DO mg/L	WDO	DO Charge	Probe Gain	Pass	Lab
Calibr.										P / F	L / F
ICV										P / F	L / F
CCV	J. Potts	6/13/19	1000	25.4	8.203	8.25	100.4	1.02	1.02	P / F	L / F
CCV	J. Potts	6/20/19	1100	26.0	8.114	8.10	99.6	—	—	P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec Cond FT 1200	Name	Date	Time	DO	Expir Date	Standard Conductivity	Temp	Pass	Lab
Calibr.	J. Potts	6/13/19	1002	190405D	4/2020	1000	1000	P / F	L / F
ICV			1004	190405C	4/2020	100	104	P / F	L / F
CCV	J. Potts	6/20/19	1102	190405C	4/2020	100	99	P / F	L / F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH FDEP SOP FT 1100	Name	Date	Time	pH	Expir Date	pH Buffer	Temp	Pass	Lab
Calibr.						7.0		P / F	L / F
Calibr.						4.0		P / F	L / F
Calibr.						10.0		P / F	L / F
ICV								P / F	L / F
CCV								P / F	L / F
CCV								P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? Yes / No If yes, list date of last quarterly depth verification 5/6/19

Depth Sensor Daily Calibration (ICV)	Name	Date	Time	Calibrated Value (m)	ICV Value (m)	Pass	Lab
Pressure mode in air	J. Potts	6/13/19	1006	0.00	0.00	P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: