

Chemical Analysis Report

NW-DIST-2019-08-08-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Bioassessment**
Request ID: **RQ-2019-08-05-89**
Customer: **NW-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
160 Government Center
Suite 208
Pensacola, FL 32502-5794
Attn: Cheryl Bunch

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 04-SEP-2019 09:39

A handwritten signature in black ink, reading "Colin Wright", is displayed on a light blue rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals, Nutrients and Pesticides.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: CANEY CREEK ABOVE HWY 2A

Collection Date/Time: 08/07/2019 11:45

Field ID: G4NW0397

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2109676	EPA 180.1 Rev. 2.0	Turbidity	13	A	NTU	P368737	
	EPA 300.0 Rev. 2.1	Chloride	5.4		mg Cl/L	P369215	
		Sulfate	0.64		mg SO4/L	P369218	
	SM 2120 B	Color (true)	42	A	PCU	P368727	
	SM 2320 B-97	Total Alkalinity	5.4		mg CaCO3/L	P369142	
	SM 2540 C-97	TDS	36		mg/L	P368993	
	SM 2540 D-97	TSS	26		mg/L	P368996	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P369146	
2109678	EPA 350.1 Rev. 2.0	Ammonia-N	0.016		mg N/L	P369166	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.46		mg N/L	P368929	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.26		mg N/L	P368955	
	EPA 365.1 Rev. 2.0	Total-P	0.067		mg P/L	P368872	
	SM 5310 B-00	Organic Carbon	3.3		mg C/L	P368814	
2109680	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.032		mg P/L	P368770	
2109689	EPA 8321B	2,4-D	0.0020	U	ug/L	P368497	
		Acesulfame K**	0.10	U	ug/L	P368684	
		AMPA**	0.10	U	ug/L	P368684	
		Sucralose**	0.071		ug/L	P368684	
		Acetaminophen**	0.0080	U	ug/L	P368497	
		Endothall**	0.25	U	ug/L	P368684	
		Acetamiprid**	0.0020	U	ug/L	P368497	
		Glufosinate**	0.10	U	ug/L	P368684	
		Glyphosate**	0.10	U	ug/L	P368684	
		Afidopyropen**	0.0020	UJ	ug/L	P368497	
		Bentazon	8.1E-04	I	ug/L	P368497	
		Benzovindiflupyr**	0.0020	UJ	ug/L	P368497	
		Carbamazepine**	8.0E-04	U	ug/L	P368497	
		Clothianidin**	0.0020	U	ug/L	P368497	
		Dinotefuran**	0.0020	U	ug/L	P368497	
		Diuron	0.0020	U	ug/L	P368497	
		Fenuron	0.016	U	ug/L	P368497	
		Fluridone**	8.0E-04	U	ug/L	P368497	
		Imazapyr**	0.0040	U	ug/L	P368497	MS
		Hydrocodone**	0.0020	U	ug/L	P368497	
		Ibuprofen**	0.020	U	ug/L	P368497	
		Imidacloprid	0.0020	U	ug/L	P368497	MS
		Linuron	0.0040	U	ug/L	P368497	
		Mandestrobin**	0.0020	UJ	ug/L	P368497	
		MCP	0.0020	U	ug/L	P368497	MS
		Naproxen**	0.0080	U	ug/L	P368497	
		Primidone**	0.0040	U	ug/L	P368497	
		Pyraclostrobin**	0.0020	U	ug/L	P368497	
		Thiamethoxam**	0.0020	U	ug/L	P368497	MS

Field ID: G4NW0397

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2109689	EPA 8321B	Tolfenpyrad**	0.0020	U	ug/L	P368497	
		Triclopyr**	0.0040	U	ug/L	P368497	
2109693	EPA 200.7 Rev. 4.4	Barium	27.3		ug/L	P368852	
		Calcium	2.05		mg/L	P368852	
		Iron	820		ug/L	P368852	
		Magnesium	1.50		mg/L	P368852	
		Potassium	0.95	I	mg/L	P368852	
		Sodium	2.3		mg/L	P368852	
		Zinc	5.0	U	ug/L	P368852	
	EPA 200.8 Rev. 5.4	Aluminum	522		ug/L	P368852	
		Antimony	0.050	U	ug/L	P368852	
		Arsenic	0.37		ug/L	P368852	
		Boron**	7.8	I	ug/L	P368852	
		Cadmium	0.020	U	ug/L	P368852	
		Chromium	0.44	I	ug/L	P368852	
		Copper	0.40	U	ug/L	P368852	
		Lead	0.60		ug/L	P368852	
		Nickel	0.50	U	ug/L	P368852	
		Selenium	0.20	U	ug/L	P368852	
		Silver	0.010	U	ug/L	P368852	
		Thallium	0.10	U	ug/L	P368852	
2109703	EPA 8270E	Acetochlor	0.25	U	ng/L	P368795	
		Alachlor	0.25	U	ng/L	P368795	
		Ametryn	0.61	U	ng/L	P368795	
		Atrazine	0.25	U	ng/L	P368795	
		Atrazine Desethyl	1.5	U	ng/L	P368795	
		Azinphos Methyl	5.1	U	ng/L	P368795	
		Bromacil	0.51	U	ng/L	P368795	
		Butylate	0.40	U	ng/L	P368795	
		Chlorpyrifos Ethyl	0.25	U	ng/L	P368795	
		Chlorpyrifos Methyl	0.10	U	ng/L	P368795	
		Cyanazine	3.3	UJ	ng/L	P368795	CCV
		Demeton	7.1	U	ng/L	P368795	
		Diazinon	0.13	U	ng/L	P368795	
		Dimethenamid**	0.10	U	ng/L	P368795	MS, RPD
		Disulfoton	0.51	UJ	ng/L	P368795	RPD
		Dithiopyr**	0.10	U	ng/L	P368795	
		EPTC	1.3	U	ng/L	P368795	
		Ethion	0.15	U	ng/L	P368795	
		Ethoprop	0.10	U	ng/L	P368795	
		Fenamiphos	0.51	U	ng/L	P368795	
		Fipronil	0.25	U	ng/L	P368795	
		Fipronil Desulfinyl**	0.13	U	ng/L	P368795	
		Fipronil Sulfide	0.15	U	ng/L	P368795	
		Fipronil Sulfone	0.15	U	ng/L	P368795	

Field ID: G4NW0397

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2109703	EPA 8270E	Fonofos	0.20	U	ng/L	P368795	
		Hexazinone	0.51	U	ng/L	P368795	
		Malathion	0.35	U	ng/L	P368795	MS
		Metaxyl	0.30	U	ng/L	P368795	
		Metolachlor	0.51	U	ng/L	P368795	
		Metribuzin	0.76	U	ng/L	P368795	MS
		Mevinphos	0.51	U	ng/L	P368795	
		Molinate	0.30	U	ng/L	P368795	
		Norflurazon	0.51	U	ng/L	P368795	
		Oxadiazon**	0.10	U	ng/L	P368795	
		Parathion Ethyl	0.25	U	ng/L	P368795	
		Parathion Methyl	0.56	U	ng/L	P368795	
		Pendimethalin	1.0	U	ng/L	P368795	
		Phorate	0.25	U	ng/L	P368795	
		Prodiamine**	0.18	U	ng/L	P368795	
		Prometon	0.91	U	ng/L	P368795	MS, RPD
		Prometryn	0.20	U	ng/L	P368795	
		Simazine	0.51	U	ng/L	P368795	
		Tebuconazole**	0.51	UJ	ng/L	P368795	MS, RPD
		Terbufos	0.10	U	ng/L	P368795	
		Terbuthylazine	0.43	U	ng/L	P368795	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 8321B: J - second source check standards for some analytes were not available for analysis.

EPA 8270E: Refer to the narrative for an explanation of QC Codes.

Sample Location: GUM CREEK

Collection Date/Time: 08/07/2019 14:45

Field ID: G4NW0382

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2109677	EPA 180.1 Rev. 2.0	Turbidity	2.7		NTU	P368737	
	EPA 300.0 Rev. 2.1	Chloride	3.6		mg Cl/L	P369215	
		Sulfate	0.85		mg SO4/L	P369218	
	SM 2120 B	Color (true)	31		PCU	P368727	
	SM 2320 B-97	Total Alkalinity	1.7	I	mg CaCO3/L	P369142	
	SM 2540 C-97	TDS	29		mg/L	P368993	
	SM 2540 D-97	TSS	2	U	mg/L	P368996	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P369146	
2109679	EPA 350.1 Rev. 2.0	Ammonia-N	0.005		mg N/L	P369166	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.22		mg N/L	P368929	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.64		mg N/L	P368956	
	EPA 365.1 Rev. 2.0	Total-P	0.009		mg P/L	P368872	
	SM 5310 B-00	Organic Carbon	2.8		mg C/L	P368815	
2109681	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P368770	
2109690	EPA 8321B	2,4-D	0.0020	U	ug/L	P368497	
		Acesulfame K**	0.10	U	ug/L	P368684	
		AMPA**	0.10	U	ug/L	P368684	
		Sucralose**	0.010	U	ug/L	P368684	
		Acetaminophen**	0.0080	U	ug/L	P368497	
		Endothall**	0.25	U	ug/L	P368684	
		Acetamiprid**	0.0020	U	ug/L	P368497	
		Glufosinate**	0.10	U	ug/L	P368684	
		Glyphosate**	0.10	U	ug/L	P368684	
		Afidopyropen**	0.0020	UJ	ug/L	P368497	
		Bentazon	8.9E-04	I	ug/L	P368497	
		Benzovindiflupyr**	0.0020	UJ	ug/L	P368497	
		Carbamazepine**	8.0E-04	U	ug/L	P368497	
		Clothianidin**	0.0020	U	ug/L	P368497	
		Dinotefuran**	0.0020	U	ug/L	P368497	
		Diuron	0.0020	U	ug/L	P368497	
		Fenuron	0.016	U	ug/L	P368497	
		Fluridone**	8.0E-04	U	ug/L	P368497	
		Imazapyr**	0.0040	U	ug/L	P368497	MS
		Hydrocodone**	0.0020	U	ug/L	P368497	
		Ibuprofen**	0.020	U	ug/L	P368497	
		Imidacloprid	0.012		ug/L	P368497	MS
		Linuron	0.0040	U	ug/L	P368497	
		Mandestrobin**	0.0020	UJ	ug/L	P368497	
		MCP	0.0020	U	ug/L	P368497	MS
		Naproxen**	0.0080	U	ug/L	P368497	
		Primidone**	0.0040	U	ug/L	P368497	
		Pyraclostrobin**	0.0020	U	ug/L	P368497	
		Thiamethoxam**	0.0020	U	ug/L	P368497	MS

Field ID: G4NW0382

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2109690	EPA 8321B	Tolfenpyrad**	0.0020	U	ug/L	P368497	
		Triclopyr**	0.0040	U	ug/L	P368497	
2109694	EPA 200.7 Rev. 4.4	Barium	23.8		ug/L	P368852	
		Calcium	1.17		mg/L	P368852	
		Iron	300		ug/L	P368852	
		Magnesium	1.0		mg/L	P368852	
		Potassium	0.54	I	mg/L	P368852	
		Sodium	1.9	I	mg/L	P368852	
		Zinc	5.0	U	ug/L	P368852	
	EPA 200.8 Rev. 5.4	Aluminum	125		ug/L	P368852	
		Antimony	0.050	U	ug/L	P368852	
		Arsenic	0.19	I	ug/L	P368852	
		Boron**	7.8	I	ug/L	P368852	
		Cadmium	0.020	U	ug/L	P368852	
		Chromium	0.40	U	ug/L	P368852	
		Copper	0.40	U	ug/L	P368852	
		Lead	0.20	U	ug/L	P368852	
		Nickel	0.50	U	ug/L	P368852	
		Selenium	0.20	U	ug/L	P368852	
		Silver	0.010	U	ug/L	P368852	
		Thallium	0.10	U	ug/L	P368852	
2109704	EPA 8270E	Acetochlor	0.26	U	ng/L	P368795	
		Alachlor	0.26	U	ng/L	P368795	
		Ametryn	0.62	U	ng/L	P368795	
		Atrazine	0.49	I	ng/L	P368795	
		Atrazine Desethyl	1.6	U	ng/L	P368795	
		Azinphos Methyl	5.2	U	ng/L	P368795	
		Bromacil	0.52	U	ng/L	P368795	
		Butylate	0.42	U	ng/L	P368795	
		Chlorpyrifos Ethyl	0.26	U	ng/L	P368795	
		Chlorpyrifos Methyl	0.10	U	ng/L	P368795	
		Cyanazine	3.4	UJ	ng/L	P368795	CCV
		Demeton	7.3	U	ng/L	P368795	
		Diazinon	0.13	U	ng/L	P368795	
		Dimethenamid**	0.10	U	ng/L	P368795	MS, RPD
		Disulfoton	0.52	UJ	ng/L	P368795	RPD
		Dithiopyr**	0.10	U	ng/L	P368795	
		EPTC	1.3	U	ng/L	P368795	
		Ethion	0.16	U	ng/L	P368795	
		Ethoprop	0.10	U	ng/L	P368795	
		Fenamiphos	0.52	U	ng/L	P368795	
		Fipronil	0.26	U	ng/L	P368795	
		Fipronil DesulfinyI**	0.13	U	ng/L	P368795	
		Fipronil Sulfide	0.16	U	ng/L	P368795	
		Fipronil Sulfone	0.16	U	ng/L	P368795	

Field ID: G4NW0382

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2109704	EPA 8270E	Fonofos	0.21	U	ng/L	P368795	
		Hexazinone	0.52	U	ng/L	P368795	
		Malathion	0.36	U	ng/L	P368795	MS
		Metaxyl	0.31	U	ng/L	P368795	
		Metolachlor	0.52	U	ng/L	P368795	
		Metribuzin	0.78	U	ng/L	P368795	MS
		Mevinphos	0.52	U	ng/L	P368795	
		Molinate	0.31	U	ng/L	P368795	
		Norflurazon	0.52	U	ng/L	P368795	
		Oxadiazon**	0.10	U	ng/L	P368795	
		Parathion Ethyl	0.26	U	ng/L	P368795	
		Parathion Methyl	0.57	U	ng/L	P368795	
		Pendimethalin	1.0	U	ng/L	P368795	
		Phorate	0.26	U	ng/L	P368795	
		Prodiamine**	0.18	U	ng/L	P368795	
		Prometon	0.93	U	ng/L	P368795	MS, RPD
		Prometryn	0.21	U	ng/L	P368795	
		Simazine	0.52	U	ng/L	P368795	
		Tebuconazole**	0.52	UJ	ng/L	P368795	MS, RPD
		Terbufos	0.10	U	ng/L	P368795	
		Terbuthylazine	0.44	U	ng/L	P368795	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 8321B: J - second source check standards for some analytes were not available for analysis.

EPA 8270E: Refer to the narrative for an explanation of QC Codes.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P368737

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P368852

Component	Result	Code	Units
Barium	1.0	U	ug/L
Calcium	0.075	U	mg/L
Iron	30	U	ug/L
Magnesium	0.040	U	mg/L
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P368852

Component	Result	Code	Units
Aluminum	5.0	U	ug/L
Antimony	0.050	U	ug/L
Arsenic	0.050	U	ug/L
Boron	2.0	U	ug/L
Cadmium	0.020	U	ug/L
Chromium	0.40	U	ug/L
Copper	0.40	U	ug/L
Lead	0.20	U	ug/L
Nickel	0.50	U	ug/L
Selenium	0.20	U	ug/L
Silver	0.010	U	ug/L
Thallium	0.10	U	ug/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P369215

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P369218

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P369166

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P368929

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P368955

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P368956

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P368872

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P368770

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 8270E
 Batch ID: P368795

Component	Result	Code	Units
Acetochlor	0.25	U	ng/L
Alachlor	0.25	U	ng/L
Ametryn	0.60	U	ng/L
Atrazine	0.25	U	ng/L
Atrazine Desethyl	1.5	U	ng/L
Azinphos Methyl	5.0	U	ng/L
Bromacil	0.50	U	ng/L
Butylate	0.40	U	ng/L
Chlorpyrifos Ethyl	0.25	U	ng/L
Chlorpyrifos Methyl	0.10	U	ng/L
Cyanazine	3.2	U	ng/L
Demeton	7.0	U	ng/L
Diazinon	0.12	U	ng/L
Dimethenamid	0.10	U	ng/L
Disulfoton	0.50	U	ng/L
Dithiopyr	0.10	U	ng/L
EPTC	1.2	U	ng/L
Ethion	0.15	U	ng/L
Ethoprop	0.10	U	ng/L
Fenamiphos	0.50	U	ng/L
Fipronil	0.25	U	ng/L
Fipronil Desulfinyl	0.12	U	ng/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8270E
Batch ID: P368795

Component	Result	Code	Units
Fipronil Sulfide	0.15	U	ng/L
Fipronil Sulfone	0.15	U	ng/L
Fonofos	0.20	U	ng/L
Hexazinone	0.50	U	ng/L
Malathion	0.35	U	ng/L
Metalaxyl	0.30	U	ng/L
Metolachlor	0.50	U	ng/L
Metribuzin	0.75	U	ng/L
Mevinphos	0.50	U	ng/L
Molinate	0.30	U	ng/L
Norflurazon	0.50	U	ng/L
Oxadiazon	0.10	U	ng/L
Parathion Ethyl	0.25	U	ng/L
Parathion Methyl	0.55	U	ng/L
Pendimethalin	1.0	U	ng/L
Phorate	0.25	U	ng/L
Prodiamine	0.18	U	ng/L
Prometon	0.90	U	ng/L
Prometryn	0.20	U	ng/L
Simazine	0.50	U	ng/L
Tebuconazole	0.50	U	ng/L
Terbufos	0.10	U	ng/L
Terbutylazine	0.42	U	ng/L

Reference Method: EPA 8321B
Batch ID: P368497

Component	Result	Code	Units
2,4-D	0.0020	U	ug/L
Acetaminophen	0.0080	U	ug/L
Acetamiprid	0.0020	U	ug/L
Afidopyropen	0.0020	U	ug/L
Bentazon	8.0E-04	U	ug/L
Benzovindiflupyr	0.0020	U	ug/L
Carbamazepine	8.0E-04	U	ug/L
Clothianidin	0.0020	U	ug/L
Dinotefuran	0.0020	U	ug/L
Diuron	0.0020	U	ug/L
Fenuron	0.016	U	ug/L
Fluridone	8.0E-04	U	ug/L
Hydrocodone	0.0020	U	ug/L
Ibuprofen	0.020	U	ug/L
Imazapyr	0.0040	U	ug/L
Imidacloprid	0.0020	U	ug/L
Linuron	0.0040	U	ug/L
Mandestrobin	0.0020	U	ug/L
MCPP	0.0020	U	ug/L
Naproxen	0.0080	U	ug/L
Primidone	0.0040	U	ug/L
Pyraclostrobin	0.0020	U	ug/L
Thiamethoxam	0.0020	U	ug/L
Tolfenpyrad	0.0020	U	ug/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 8321B
Batch ID: P368497

Component	Result	Code	Units
Triclopyr	0.0040	U	ug/L

Reference Method: EPA 8321B
Batch ID: P368684

Component	Result	Code	Units
Acesulfame K	0.10	U	ug/L
AMPA	0.10	U	ug/L
Endothall	0.25	U	ug/L
Glufosinate	0.10	U	ug/L
Glyphosate	0.10	U	ug/L
Sucralose	0.010	U	ug/L

Reference Method: SM 2120 B
Batch ID: P368727

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P369142

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P368993

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P368996

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P369146

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P368814

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 5310 B-00
Batch ID: P368815

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P368852

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Barium	98.8		P	85 - 115
Calcium	105		P	85 - 115
Iron	99.9		P	85 - 115
Magnesium	106		P	85 - 115
Potassium	103		P	85 - 115
Sodium	112		P	85 - 115
Zinc	95.3		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P368852

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aluminum	100		P	85 - 115
Antimony	102		P	85 - 115
Arsenic	104		P	85 - 115
Boron	101		P	85 - 115
Cadmium	106		P	85 - 115
Chromium	105		P	85 - 115
Copper	101		P	85 - 115
Lead	105		P	85 - 115
Nickel	101		P	85 - 115
Selenium	105		P	85 - 115
Silver	106		P	85 - 115
Thallium	105		P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P369215

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	97.2		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P369218

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	102		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P369166

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	105		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P368929

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	102		P	90 - 110

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P368955

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P368956

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P368872

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P368770

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.7		P	90 - 110

Reference Method: EPA 8270E
 Batch ID: P368795

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Acetochlor	96.8	87.1	P/P	69 - 122
Alachlor	90.6	80.4	P/P	71 - 120
Ametryn	92.1	71.3	P/P	47 - 138
Atrazine	100	99.5	P/P	74 - 124
Atrazine Desethyl	89.9	95.0	P/P	38 - 138
Azinphos Methyl	128	122	P/P	69 - 154
Bromacil	44.9	43.0	P/P	39 - 121
Butylate	57.7	47.0	P/P	27 - 87.0
Chlorpyrifos Ethyl	94.0	90.4	P/P	69 - 118
Chlorpyrifos Methyl	101	96.9	P/P	70 - 120
Cyanazine	236	248	P/P	20 - 250
Demeton	66.1	50.0	P/P	39 - 118
Diazinon	107	96.9	P/P	68 - 127
Dimethenamid	90.3	75.9	P/P	66 - 125
Disulfoton	78.5	57.2	P/P	52 - 126
Dithiopyr	92.8	80.7	P/P	67 - 127
EPTC	59.5	45.5	P/P	24 - 88.0
Ethion	95.7	87.2	P/P	51 - 132
Ethoprop	99.7	88.8	P/P	58 - 117
Fenamiphos	49.9	55.6	P/P	38 - 139
Fipronil	94.2	71.6	P/P	61 - 124
Fipronil Desulfinyl	91.4	80.5	P/P	47 - 120
Fipronil Sulfide	89.8	80.3	P/P	56 - 119
Fipronil Sulfone	89.6	78.1	P/P	50 - 117
Fonofos	95.5	90.0	P/P	60 - 121
Hexazinone	98.5	99.1	P/P	55 - 137
Malathion	102	89.6	P/P	66 - 126
Metalaxyl	98.2	87.3	P/P	33 - 140

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 8270E
Batch ID: P368795

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Metolachlor	91.4	82.7	P/P	69 - 119
Metribuzin	189	197	P/P	31 - 238
Mevinphos	91.9	81.0	P/P	42 - 113
Molinate	75.4	70.0	P/P	32 - 96.0
Norflurazon	96.9	84.6	P/P	54 - 123
Oxadiazon	91.0	82.8	P/P	63 - 128
Parathion Ethyl	99.0	86.6	P/P	72 - 111
Parathion Methyl	107	96.9	P/P	74 - 126
Pendimethalin	88.5	82.4	P/P	61 - 131
Phorate	95.3	86.6	P/P	51 - 115
Prodiamine	84.1	77.7	P/P	38 - 144
Prometon	85.5	79.4	P/P	54 - 119
Prometryn	93.3	83.0	P/P	63 - 120
Simazine	108	106	P/P	75 - 126
Tebuconazole	70.5	50.0	P/P	20 - 119
Terbufos	93.6	86.5	P/P	64 - 116
Terbutylazine	100	91.0	P/P	73 - 117

Reference Method: EPA 8321B
Batch ID: P368497

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2,4-D	101		P	40 - 160
Acetaminophen	76.9		P	30 - 160
Acetamiprid	86.8		P	40 - 160
Afidopyropen	72.0		P	40 - 160
Bentazon	144		P	30 - 160
Benzovindiflupyr	75.7		P	40 - 160
Carbamazepine	94.5		P	40 - 160
Clothianidin	88.9		P	40 - 160
Dinotefuran	88.8		P	40 - 160
Diuron	77.9		P	40 - 160
Fenuron	102		P	40 - 160
Fluridone	86.7		P	40 - 160
Hydrocodone	90.3		P	40 - 160
Ibuprofen	82.9		P	30 - 160
Imazapyr	100		P	40 - 160
Imidacloprid	88.8		P	40 - 160
Linuron	71.3		P	40 - 160
Mandestrobin	85.3		P	40 - 160
MCPP	115		P	40 - 160
Naproxen	99.4		P	40 - 160
Primidone	67.6		P	40 - 160
Pyraclostrobin	72.9		P	30 - 160
Thiamethoxam	86.4		P	40 - 160
Tolfenpyrad	47.3		P	40 - 160
Triclopyr	92.8		P	30 - 160

Reference Method: EPA 8321B
Batch ID: P368684

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 8321B
Batch ID: P368684

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Acesulfame K	96.2		P	40 - 150
AMPA	95.1		P	40 - 150
Endothall	131		P	40 - 150
Glufosinate	114		P	40 - 150
Glyphosate	101		P	40 - 140
Sucralose	105		P	40 - 150

Reference Method: SM 2120 B
Batch ID: P368727

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	104		P	85 - 115

Reference Method: SM 2320 B-97
Batch ID: P369142

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	104		P	98 - 106

Reference Method: SM 4500 F-C-97
Batch ID: P369146

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.9		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P368814

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	100		P	85 - 115

Reference Method: SM 5310 B-00
Batch ID: P368815

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	101		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P368852

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109729	Barium	97.3	96.7	P/P	80 - 120
2109729	Calcium	102	104	P/P	80 - 120
2109729	Iron	97.4	98.0	P/P	80 - 120
2109729	Magnesium	102	103	P/P	80 - 120
2109729	Potassium	105	106	P/P	80 - 120
2109729	Sodium	104	103	P/P	80 - 120
2109729	Zinc	94.5	93.0	P/P	80 - 120
2110452	Barium	96.6		P	80 - 120
2110452	Calcium	97.8		P	80 - 120
2110452	Iron	98.1		P	80 - 120
2110452	Magnesium	98.6		P	80 - 120
2110452	Potassium	99.2		P	80 - 120
2110452	Sodium	105		P	80 - 120
2110452	Zinc	94.4		P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P368852

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109729	Aluminum	99.3	101	P/P	80 - 120
2109729	Antimony	103	103	P/P	80 - 120
2109729	Arsenic	105	104	P/P	80 - 120
2109729	Boron	99.4	101	P/P	80 - 120
2109729	Cadmium	105	107	P/P	80 - 120
2109729	Chromium	105	105	P/P	80 - 120
2109729	Copper	103	103	P/P	80 - 120
2109729	Lead	104	105	P/P	80 - 120
2109729	Nickel	101	100	P/P	80 - 120
2109729	Selenium	108	105	P/P	80 - 120
2109729	Silver	106	106	P/P	80 - 120
2109729	Thallium	105	105	P/P	80 - 120
2110452	Aluminum	100		P	80 - 120
2110452	Antimony	103		P	80 - 120
2110452	Arsenic	105		P	80 - 120
2110452	Boron	99.1		P	80 - 120
2110452	Cadmium	104		P	80 - 120
2110452	Chromium	104		P	80 - 120
2110452	Copper	104		P	80 - 120
2110452	Lead	105		P	80 - 120
2110452	Nickel	102		P	80 - 120
2110452	Selenium	105		P	80 - 120
2110452	Silver	106		P	80 - 120
2110452	Thallium	107		P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P369215

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109676	Chloride	97.4	97.6	P/P	90 - 110
2110277	Chloride	97.0		P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P369218

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109676	Sulfate	101	102	P/P	90 - 110
2110277	Sulfate	101		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P369166

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109724	Ammonia-N	105	104	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P368929

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109772	Kjeldahl Nitrogen	103	111	P/F	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P368955

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109614	NO2NO3-N	101	101	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P368956

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109772	NO2NO3-N	96.2		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P368872

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109724	Total-P	101	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P368770

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109725	O-Phosphate-P	97.6	97.9	P/P	90 - 110

Reference Method: EPA 8270E
Batch ID: P368795

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109503	Acetochlor	90.7	74.7	P/P	63 - 129
2109503	Alachlor	84.3	78.7	P/P	65 - 127
2109503	Ametryn	93.0	77.9	P/P	40 - 164
2109503	Atrazine	105	101	P/P	66 - 135
2109503	Atrazine Desethyl	106	112	P/P	14 - 157
2109503	Azinphos Methyl	119	91.0	P/P	49 - 180
2109503	Bromacil	77.7	70.7	P/P	48 - 130
2109503	Butylate	50.1	55.6	P/P	10 - 94.0
2109503	Chlorpyrifos Ethyl	91.3	82.1	P/P	58 - 123

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8270E
Batch ID: P368795

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109503	Chlorpyrifos Methyl	96.6	90.1	P/P	58 - 125
2109503	Cyanazine	245	242	P/P	24 - 253
2109503	Demeton	81.1	79.6	P/P	25 - 149
2109503	Diazinon	103	97.7	P/P	59 - 144
2109503	Dimethenamid	79.3	39.9	P/F	46 - 139
2109503	Disulfoton	95.9	79.8	P/P	54 - 132
2109503	Dithiopyr	87.9	76.7	P/P	64 - 130
2109503	EPTC	45.2	58.9	P/P	10 - 92.0
2109503	Ethion	88.0	72.2	P/P	30 - 148
2109503	Ethoprop	98.1	105	P/P	50 - 130
2109503	Fenamiphos	57.5	54.4	P/P	49 - 142
2109503	Fipronil	78.0	74.5	P/P	46 - 140
2109503	Fipronil Desulfinyl	86.7	78.4	P/P	30 - 121
2109503	Fipronil Sulfide	83.6	79.7	P/P	31 - 136
2109503	Fipronil Sulfone	79.0	62.1	P/P	16 - 136
2109503	Fonofos	96.5	101	P/P	55 - 129
2109503	Hexazinone	92.1	81.6	P/P	56 - 141
2109503	Malathion	48.0	64.8	F/P	56 - 135
2109503	Metalaxyl	93.6	83.7	P/P	54 - 135
2109503	Metolachlor	86.0	72.1	P/P	51 - 140
2109503	Metribuzin	239	294	P/F	45 - 254
2109503	Mevinphos	92.8	79.4	P/P	33 - 131
2109503	Molinate	67.7	64.0	P/P	14 - 98.0
2109503	Norflurazon	87.9	76.6	P/P	34 - 136
2109503	Oxadiazon	86.1	74.1	P/P	67 - 116
2109503	Parathion Ethyl	91.8	90.1	P/P	66 - 118
2109503	Parathion Methyl	99.9	99.4	P/P	64 - 136
2109503	Pendimethalin	85.9	83.1	P/P	59 - 144
2109503	Phorate	98.6	99.5	P/P	42 - 130
2109503	Prodiamine	96.4	85.5	P/P	42 - 148
2109503	Prometon	87.1	48.2	P/F	59 - 134
2109503	Prometryn	90.5	81.2	P/P	54 - 135
2109503	Simazine	110	106	P/P	53 - 156
2109503	Tebuconazole	16.3	4.62	P/F	10 - 131
2109503	Terbufos	99.7	102	P/P	57 - 131
2109503	Terbutylazine	99.1	92.2	P/P	68 - 125

Reference Method: EPA 8321B
Batch ID: P368497

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109232	2,4-D	143	150	P/P	40 - 160
2109232	Acetaminophen	92.4	95.9	P/P	30 - 160
2109232	Acetamiprid	113	134	P/P	40 - 160
2109232	Afidopyropen	108	119	P/P	40 - 160
2109232	Bentazon	106	106	P/P	30 - 160
2109232	Benzovindiflupyr	87.0	93.9	P/P	40 - 160
2109232	Carbamazepine	111	132	P/P	40 - 160
2109232	Clothianidin	111	130	P/P	40 - 160
2109232	Dinotefuran	123	138	P/P	40 - 160
2109232	Diuron	90.0	103	P/P	40 - 160
2109232	Fenuron	105	124	P/P	40 - 160

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8321B
 Batch ID: P368497

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109232	Fluridone	90.6	108	P/P	40 - 160
2109232	Hydrocodone	116	134	P/P	40 - 160
2109232	Ibuprofen	96.6	113	P/P	30 - 160
2109232	Imazapyr	161	181	F/F	40 - 160
2109232	Imidacloprid	155	183	P/F	40 - 160
2109232	Linuron	76.8	91.8	P/P	40 - 160
2109232	Mandestrobin	88.6	108	P/P	40 - 160
2109232	MCPP	169	175	F/F	40 - 160
2109232	Naproxen	90.6	109	P/P	40 - 160
2109232	Primidone	96.8	121	P/P	40 - 160
2109232	Pyraclostrobin	85.8	88.2	P/P	30 - 160
2109232	Thiamethoxam	144	168	P/F	40 - 160
2109232	Tolfenpyrad	51.8	48.4	P/P	40 - 160
2109232	Triclopyr	122	104	P/P	30 - 160

Reference Method: EPA 8321B
 Batch ID: P368684

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109647	Acesulfame K	109	101	P/P	40 - 150
2109647	AMPA	116	107	P/P	40 - 150
2109647	Endothall	143	143	P/P	40 - 150
2109647	Glufosinate	123	118	P/P	40 - 150
2109647	Glyphosate	114	126	P/P	40 - 140
2109647	Sucralose	101	105	P/P	40 - 150

Reference Method: SM 4500 F-C-97
 Batch ID: P369146

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2110276	Fluoride	99.3	99.3	P/P	89 - 106

Reference Method: SM 5310 B-00
 Batch ID: P368814

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109514	Organic Carbon	99.8	100	P/P	85 - 115

Reference Method: SM 5310 B-00
 Batch ID: P368815

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109679	Organic Carbon	99.2	99.7	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P368737

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109676	Turbidity	7.58	Sample	P	0 - 20

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P368852

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109729	Barium	0.607	Spike	P	0 - 20
2109729	Calcium	2.32	Spike	P	0 - 20
2109729	Iron	0.661	Spike	P	0 - 20
2109729	Magnesium	0.659	Spike	P	0 - 20
2109729	Potassium	0.435	Spike	P	0 - 20
2109729	Sodium	0.597	Spike	P	0 - 20
2109729	Zinc	1.62	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P368852

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109729	Aluminum	1.27	Spike	P	0 - 20
2109729	Antimony	0.140	Spike	P	0 - 20
2109729	Arsenic	0.730	Spike	P	0 - 20
2109729	Boron	1.40	Spike	P	0 - 20
2109729	Cadmium	1.34	Spike	P	0 - 20
2109729	Chromium	0.179	Spike	P	0 - 20
2109729	Copper	0.0753	Spike	P	0 - 20
2109729	Lead	0.434	Spike	P	0 - 20
2109729	Nickel	0.861	Spike	P	0 - 20
2109729	Selenium	2.68	Spike	P	0 - 20
2109729	Silver	0.228	Spike	P	0 - 20
2109729	Thallium	0.712	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P369215

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109676	Chloride	0.194	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P369218

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109676	Sulfate	0.544	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P369166

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109724	Ammonia-N	1.13	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P368929

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109772	Kjeldahl Nitrogen	4.46	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P368955

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109614	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P368956

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109772	NO2NO3-N	0.168	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P368872

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109724	Total-P	2.65	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P368770

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109725	O-Phosphate-P	0.307	Spike	P	0 - 20

Reference Method: EPA 8270E
Batch ID: P368795

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109503	Acetochlor	19.4	Spike	P	0 - 30
2109503	Alachlor	6.95	Spike	P	0 - 30
2109503	Ametryn	17.7	Spike	P	0 - 30
2109503	Atrazine	3.53	Spike	P	0 - 30
2109503	Atrazine Desethyl	5.99	Spike	P	0 - 30
2109503	Azinphos Methyl	26.4	Spike	P	0 - 30
2109503	Bromacil	9.41	Spike	P	0 - 30
2109503	Butylate	10.4	Spike	P	0 - 30
2109503	Chlorpyrifos Ethyl	10.6	Spike	P	0 - 30

Quality Assurance Report

Precision

Reference Method: EPA 8270E
Batch ID: P368795

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109503	Chlorpyrifos Methyl	6.98	Spike	P	0 - 30
2109503	Cyanazine	1.04	Spike	P	0 - 30
2109503	Demeton	1.91	Spike	P	0 - 30
2109503	Diazinon	5.28	Spike	P	0 - 30
2109503	Dimethenamid	66.0	Spike	F	0 - 30
2109503	Disulfoton	18.4	Spike	P	0 - 30
2109503	Dithiopyr	13.6	Spike	P	0 - 30
2109503	EPTC	26.3	Spike	P	0 - 30
2109503	Ethion	19.7	Spike	P	0 - 30
2109503	Ethoprop	7.25	Spike	P	0 - 30
2109503	Fenamiphos	5.58	Spike	P	0 - 30
2109503	Fipronil	4.59	Spike	P	0 - 30
2109503	Fipronil Desulfinyl	10.1	Spike	P	0 - 30
2109503	Fipronil Sulfide	4.81	Spike	P	0 - 30
2109503	Fipronil Sulfone	23.9	Spike	P	0 - 30
2109503	Fonofos	4.24	Spike	P	0 - 30
2109503	Hexazinone	7.84	Spike	P	0 - 30
2109503	Malathion	29.9	Spike	P	0 - 30
2109503	Metalaxyl	11.1	Spike	P	0 - 30
2109503	Metolachlor	17.5	Spike	P	0 - 30
2109503	Metribuzin	20.5	Spike	P	0 - 30
2109503	Mevinphos	15.5	Spike	P	0 - 30
2109503	Molinate	5.61	Spike	P	0 - 30
2109503	Norflurazon	13.1	Spike	P	0 - 30
2109503	Oxadiazon	15.0	Spike	P	0 - 30
2109503	Parathion Ethyl	1.84	Spike	P	0 - 30
2109503	Parathion Methyl	0.590	Spike	P	0 - 30
2109503	Pendimethalin	3.22	Spike	P	0 - 30
2109503	Phorate	0.922	Spike	P	0 - 30
2109503	Prodiamine	12.0	Spike	P	0 - 30
2109503	Prometon	57.5	Spike	F	0 - 30
2109503	Prometryn	10.9	Spike	P	0 - 30
2109503	Simazine	3.84	Spike	P	0 - 30
2109503	Tebuconazole	112	Spike	F	0 - 30
2109503	Terbufos	2.67	Spike	P	0 - 30
2109503	Terbuthylazine	7.16	Spike	P	0 - 30
LFB	Acetochlor	10.5	LCS	P	0 - 30
LFB	Alachlor	11.9	LCS	P	0 - 30
LFB	Ametryn	25.5	LCS	P	0 - 30
LFB	Atrazine	0.443	LCS	P	0 - 30
LFB	Atrazine Desethyl	5.60	LCS	P	0 - 30
LFB	Azinphos Methyl	5.06	LCS	P	0 - 30
LFB	Bromacil	4.22	LCS	P	0 - 30
LFB	Butylate	20.5	LCS	P	0 - 30
LFB	Chlorpyrifos Ethyl	3.92	LCS	P	0 - 30
LFB	Chlorpyrifos Methyl	4.35	LCS	P	0 - 30
LFB	Cyanazine	4.84	LCS	P	0 - 30
LFB	Demeton	27.7	LCS	P	0 - 30
LFB	Diazinon	9.55	LCS	P	0 - 30
LFB	Dimethenamid	17.3	LCS	P	0 - 30
LFB	Disulfoton	31.3	LCS	F	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8270E
Batch ID: P368795

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
LFB	Dithiopyr	14.0	LCS	P	0 - 30
LFB	EPTC	26.6	LCS	P	0 - 30
LFB	Ethion	9.31	LCS	P	0 - 30
LFB	Ethoprop	11.5	LCS	P	0 - 30
LFB	Fenamiphos	10.9	LCS	P	0 - 30
LFB	Fipronil	27.3	LCS	P	0 - 30
LFB	Fipronil Desulfinyl	12.8	LCS	P	0 - 30
LFB	Fipronil Sulfide	11.2	LCS	P	0 - 30
LFB	Fipronil Sulfone	13.8	LCS	P	0 - 30
LFB	Fonofos	5.92	LCS	P	0 - 30
LFB	Hexazinone	0.574	LCS	P	0 - 30
LFB	Malathion	12.8	LCS	P	0 - 30
LFB	Metaxyl	11.8	LCS	P	0 - 30
LFB	Metolachlor	9.98	LCS	P	0 - 30
LFB	Metribuzin	4.10	LCS	P	0 - 30
LFB	Mevinphos	12.6	LCS	P	0 - 30
LFB	Molinate	7.44	LCS	P	0 - 30
LFB	Norflurazon	13.5	LCS	P	0 - 30
LFB	Oxadiazon	9.39	LCS	P	0 - 30
LFB	Parathion Ethyl	13.3	LCS	P	0 - 30
LFB	Parathion Methyl	10.4	LCS	P	0 - 30
LFB	Pendimethalin	7.14	LCS	P	0 - 30
LFB	Phorate	9.57	LCS	P	0 - 30
LFB	Prodiamine	7.86	LCS	P	0 - 30
LFB	Prometon	7.41	LCS	P	0 - 30
LFB	Prometryn	11.8	LCS	P	0 - 30
LFB	Simazine	2.39	LCS	P	0 - 30
LFB	Tebuconazole	34.1	LCS	F	0 - 30
LFB	Terbufos	7.96	LCS	P	0 - 30
LFB	Terbutylazine	9.52	LCS	P	0 - 30

Reference Method: EPA 8321B
Batch ID: P368497

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109232	2,4-D	5.08	Spike	P	0 - 30
2109232	Acetaminophen	3.74	Spike	P	0 - 30
2109232	Acetamiprid	16.4	Spike	P	0 - 30
2109232	Afidopyropen	9.65	Spike	P	0 - 30
2109232	Bentazon	0.515	Spike	P	0 - 30
2109232	Benzovindiflupyr	7.56	Spike	P	0 - 30
2109232	Carbamazepine	15.8	Spike	P	0 - 30
2109232	Clothianidin	15.6	Spike	P	0 - 30
2109232	Dinotefuran	11.3	Spike	P	0 - 30
2109232	Diuron	12.6	Spike	P	0 - 30
2109232	Fenuron	17.2	Spike	P	0 - 30
2109232	Fluridone	16.4	Spike	P	0 - 30
2109232	Hydrocodone	14.2	Spike	P	0 - 30
2109232	Ibuprofen	15.8	Spike	P	0 - 30
2109232	Imazapyr	11.6	Spike	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8321B
Batch ID: P368497

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109232	Imidacloprid	16.0	Spike	P	0 - 30
2109232	Linuron	17.8	Spike	P	0 - 30
2109232	Mandestrobin	19.5	Spike	P	0 - 30
2109232	MCP	3.28	Spike	P	0 - 30
2109232	Naproxen	18.2	Spike	P	0 - 30
2109232	Primidone	22.4	Spike	P	0 - 30
2109232	Pyraclostrobin	2.80	Spike	P	0 - 30
2109232	Thiamethoxam	14.9	Spike	P	0 - 30
2109232	Tolfenpyrad	6.75	Spike	P	0 - 30
2109232	Triclopyr	15.7	Spike	P	0 - 30

Reference Method: EPA 8321B
Batch ID: P368684

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109647	Acesulfame K	7.61	Spike	P	0 - 30
2109647	AMPA	8.20	Spike	P	0 - 30
2109647	Endothall	0.0838	Spike	P	0 - 30
2109647	Glufosinate	3.98	Spike	P	0 - 30
2109647	Glyphosate	10.1	Spike	P	0 - 30
2109647	Sucralose	4.44	Spike	P	0 - 30

Reference Method: SM 2120 B
Batch ID: P368727

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109676	Color (true)	1.93	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P369142

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2110276	Total Alkalinity	1.15	Sample	P	0 - 2.8

Reference Method: SM 2540 C-97
Batch ID: P368993

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109612	TDS	1.02	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P369146

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2110276	Fluoride	0.0	Spike	P	0 - 2.5

Quality Assurance Report Precision

Reference Method: SM 5310 B-00
Batch ID: P368814

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109514	Organic Carbon	0.536	Spike	P	0 - 20

Reference Method: SM 5310 B-00
Batch ID: P368815

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109679	Organic Carbon	0.396	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Surrogates

Lab Sample ID: 2109689
Field Sample ID: G4NW0397

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	2,4-D-d3	154	P	30 - 160
EPA 8321B	Acetaminophen-d4	45.0	P	30 - 160
EPA 8321B	Carbamazepine-d10	113	P	30 - 160
EPA 8321B	Diuron-d6	76.0	P	30 - 160
EPA 8321B	Endothall-d6	153	P	30 - 160
EPA 8321B	Endothall-d6	153	P	30 - 160
EPA 8321B	Glyphosate 13C 15N	86.6	P	30 - 160
EPA 8321B	Glyphosate 13C 15N	86.6	P	30 - 160
EPA 8321B	Ibuprofen-d3	122	P	30 - 160
EPA 8321B	Primidone-d5	96.1	P	30 - 160
EPA 8321B	Sucralose-d6	82.2	P	30 - 160
EPA 8321B	Sucralose-d6	82.2	P	30 - 160

Lab Sample ID: 2109690
Field Sample ID: G4NW0382

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	2,4-D-d3	130	P	30 - 160
EPA 8321B	Acetaminophen-d4	90.7	P	30 - 160
EPA 8321B	Carbamazepine-d10	112	P	30 - 160
EPA 8321B	Diuron-d6	93.0	P	30 - 160
EPA 8321B	Endothall-d6	152	P	30 - 160
EPA 8321B	Endothall-d6	152	P	30 - 160
EPA 8321B	Glyphosate 13C 15N	106	P	30 - 160
EPA 8321B	Glyphosate 13C 15N	106	P	30 - 160
EPA 8321B	Ibuprofen-d3	112	P	30 - 160
EPA 8321B	Primidone-d5	103	P	30 - 160
EPA 8321B	Sucralose-d6	88.7	P	30 - 160
EPA 8321B	Sucralose-d6	88.7	P	30 - 160

Lab Sample ID: 2109703
Field Sample ID: G4NW0397

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270E	Simazine-d10	116	P	68 - 130
EPA 8270E	Terbufos-d10	84.6	P	61 - 136

Lab Sample ID: 2109704
Field Sample ID: G4NW0382

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270E	Simazine-d10	111	P	68 - 130
EPA 8270E	Terbufos-d10	100	P	61 - 136

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B

Run ID: A93368

Included Lab Sample IDs: 2109676, 2109677

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	102	104	P/P	90 - 115

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A93369

Included Lab Sample IDs: 2109676, 2109677

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	101	102	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A93388

Included Lab Sample IDs: 2109680, 2109681

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	98.9	P/P	90 - 110
O-Phosphate-P	98.9	100	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A93412

Included Lab Sample IDs: 2109678, 2109679

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	101	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A93452

Included Lab Sample IDs: 2109678

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	100	99.7	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A93456

Included Lab Sample IDs: 2109679

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	98.0	97.5	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A93462

Included Lab Sample IDs: 2109678, 2109679

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	101	101	P/P	90 - 110
NO2NO3-N	101	102	P/P	90 - 110

Reference Method: EPA 8321B

Run ID: A93483

Included Lab Sample IDs: 2109689, 2109690

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
-----------	---------	---------	------------	----------------

Quality Assurance Report Calibration Verification

Reference Method: EPA 8321B

Run ID: A93483

Included Lab Sample IDs: 2109689, 2109690

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
2,4-D	109	122	P/P	50 - 150
Acetaminophen	85.7	77.1	P/P	60 - 160
Acetamiprid	89.4	86.4	P/P	50 - 150
Afidopyropen	79.5	83.9	P/P	50 - 150
Bentazon	108	76.5	P/P	50 - 160
Benzovindiflupyr	87.8	84.4	P/P	50 - 150
Carbamazepine	90.2	91.0	P/P	60 - 150
Clothianidin	85.1	81.6	P/P	60 - 150
Dinotefuran	94.0	90.5	P/P	50 - 150
Diuron	85.9	88.2	P/P	60 - 150
Fenuron	89.9	89.9	P/P	60 - 150
Fluridone	90.0	88.7	P/P	60 - 160
Hydrocodone	87.3	86.9	P/P	60 - 150
Ibuprofen	76.8	91.4	P/P	50 - 160
Imazapyr	92.9	94.9	P/P	50 - 150
Imidacloprid	87.7	87.9	P/P	60 - 150
Linuron	82.6	82.5	P/P	60 - 150
Mandestrobin	87.6	87.0	P/P	50 - 150
MCPP	94.1	101	P/P	50 - 150
Naproxen	128	109	P/P	50 - 160
Primidone	83.2	87.3	P/P	60 - 150
Pyraclostrobin	81.4	81.0	P/P	60 - 150
Thiamethoxam	94.6	92.7	P/P	50 - 150
Tolfenpyrad	109	112	P/P	60 - 150
Triclopyr	101	101	P/P	50 - 150

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A93484

Included Lab Sample IDs: 2109693, 2109694

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Barium	100	97.7	P/P	95 - 105
Calcium	96.5	98.7	P/P	95 - 105
Iron	97.4	97.5	P/P	95 - 105
Magnesium	97.1	96.7	P/P	95 - 105
Potassium	97.0	97.5	P/P	95 - 105
Sodium	99.6	104	P/P	95 - 105
Zinc	99.4	94.6	P/P	95 - 105

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A93486

Included Lab Sample IDs: 2109676, 2109677

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	96.8	97.0	P/P	90 - 110
Sulfate	102	102	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A93507

Included Lab Sample IDs: 2109678, 2109679

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	101	106	P/P	90 - 110

Reference Method: EPA 8321B

Run ID: A93508

Included Lab Sample IDs: 2109689, 2109690

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Acesulfame K	68.0	94.7	P/P	60 - 160
Endothall	136	144	P/P	60 - 160

Reference Method: EPA 8321B

Run ID: A93512

Included Lab Sample IDs: 2109689, 2109690

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
AMPA	110	110	P/P	60 - 160
Glufosinate	120	121	P/P	60 - 160
Glyphosate	110	127	P/P	60 - 160

Reference Method: SM 2320 B-97

Run ID: A93545

Included Lab Sample IDs: 2109676, 2109677

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	103	104	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A93545

Included Lab Sample IDs: 2109676, 2109677

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.5	100	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A93563

Included Lab Sample IDs: 2109678, 2109679

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	107	109	P/P	90 - 110
Ammonia-N	108	108	P/P	90 - 110

Reference Method: EPA 8270E

Run ID: A93584

Included Lab Sample IDs: 2109703, 2109704

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Acetochlor	102		P	80 - 120
Alachlor	103		P	80 - 120
Ametryn	97.4		P	80 - 120
Atrazine	101		P	80 - 120
Atrazine Desethyl	87.3		P	80 - 120

Quality Assurance Report Calibration Verification

Reference Method: EPA 8270E

Run ID: A93584

Included Lab Sample IDs: 2109703, 2109704

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Azinphos Methyl	104		P	80 - 120
Bromacil	102		P	80 - 120
Butylate	94.9		P	80 - 120
Chlorpyrifos Ethyl	99.2		P	80 - 120
Chlorpyrifos Methyl	98.6		P	80 - 120
Cyanazine	76.7*		F	80 - 120
Demeton	99.4		P	80 - 120
Diazinon	99.9		P	80 - 120
Dimethenamid	103		P	80 - 120
Disulfoton	98.0		P	80 - 120
Dithiopyr	101		P	80 - 120
EPTC	99.2		P	80 - 120
Ethion	98.3		P	80 - 120
Ethoprop	97.3		P	80 - 120
Fenamiphos	97.2		P	80 - 120
Fipronil	99.2		P	80 - 120
Fipronil Desulfinyl	103		P	80 - 120
Fipronil Sulfide	98.9		P	80 - 120
Fipronil Sulfone	99.9		P	80 - 120
Fonofos	95.1		P	80 - 120
Hexazinone	98.2		P	80 - 120
Malathion	103		P	80 - 120
Metaxyl	104		P	80 - 120
Metolachlor	109		P	80 - 120
Metribuzin	80.6		P	80 - 120
Mevinphos	101		P	80 - 120
Molinate	99.6		P	80 - 120
Norflurazon	100		P	80 - 120
Oxadiazon	106		P	80 - 120
Parathion Ethyl	95.6		P	80 - 120
Parathion Methyl	96.2		P	80 - 120
Pendimethalin	96.7		P	80 - 120
Phorate	95.5		P	80 - 120
Prodiamine	102		P	80 - 120
Prometon	100		P	80 - 120
Prometryn	102		P	80 - 120
Simazine	97.2		P	80 - 120
Tebuconazole	101		P	80 - 120
Terbufos	96.5		P	80 - 120
Terbuthylazine	99.3		P	80 - 120

Reference Method: EPA 8321B

Run ID: A93637

Included Lab Sample IDs: 2109689, 2109690

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Sucralose	104	113	P/P	60 - 160
Sucralose	96.4	107	P/P	60 - 160

Quality Assurance Report Calibration Verification

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A93758

Included Lab Sample IDs: 2109693, 2109694

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aluminum	100	101	P/P	90 - 110
Antimony	98.9	99.0	P/P	90 - 110
Arsenic	102	101	P/P	90 - 110
Cadmium	102	102	P/P	90 - 110
Chromium	105	105	P/P	90 - 110
Copper	101	100	P/P	90 - 110
Lead	104	104	P/P	90 - 110
Nickel	100	99.8	P/P	90 - 110
Selenium	103	102	P/P	90 - 110
Silver	106	106	P/P	90 - 110
Thallium	99.1	98.6	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A93935

Included Lab Sample IDs: 2109693, 2109694

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Boron	97.3	99.0	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery		MS % Recovery			LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity							7.58	
EPA 200.7 Rev. 4.4	Barium	98.8		97.3	96.7	96.6			0.607
	Calcium	105		102	104	97.8			2.32
	Iron	99.9		97.4	98.0	98.1			0.661
	Magnesium	106		102	103	98.6			0.659
	Potassium	103		105	106	99.2			0.435
	Sodium	112		104	103	105			0.597
	Zinc	95.3		94.5	93.0	94.4			1.62
EPA 200.8 Rev. 5.4	Aluminum	100		99.3	101	100			1.27
	Antimony	102		103	103	103			0.140
	Arsenic	104		105	104	105			0.730
	Boron	101		99.4	101	99.1			1.40
	Cadmium	106		105	107	104			1.34
	Chromium	105		105	105	104			0.179
	Copper	101		103	103	104			0.0753
	Lead	105		104	105	105			0.434
	Nickel	101		101	100	102			0.861
	Selenium	105		108	105	105			2.68
	Silver	106		106	106	106			0.228
	Thallium	105		105	105	107			0.712
EPA 300.0 Rev. 2.1	Chloride	97.2		97.4	97.6	97.0			0.194
	Sulfate	102		101	102	101			0.544
EPA 350.1 Rev. 2.0	Ammonia-N	105		105	104				1.13
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	102		103	111				4.46
EPA 353.2 Rev. 2.0	NO2NO3-N	102		101	101				0.0
	NO2NO3-N	102		96.2					0.168
EPA 365.1 Rev. 2.0	Total-P	104		101	104				2.65
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	97.7		97.6	97.9				0.307
EPA 8270E	Acetochlor	87.1	96.8	90.7	74.7		10.5		19.4
	Alachlor	80.4	90.6	84.3	78.7		11.9		6.95
	Ametryn	92.1	71.3	93.0	77.9		25.5		17.7
	Atrazine	100	99.5	105	101		0.443		3.53
	Atrazine Desethyl	89.9	95.0	106	112		5.60		5.99
	Azinphos Methyl	128	122	119	91.0		5.06		26.4
	Bromacil	44.9	43.0	77.7	70.7		4.22		9.41
	Butylate	57.7	47.0	50.1	55.6		20.5		10.4
	Chlorpyrifos Ethyl	94.0	90.4	91.3	82.1		3.92		10.6
	Chlorpyrifos Methyl	96.9	101	96.6	90.1		4.35		6.98
	Cyanazine	248	236	245	242		4.84		1.04
	Demeton	66.1	50.0	81.1	79.6		27.7		1.91
	Diazinon	96.9	107	103	97.7		9.55		5.28
	Dimethenamid	75.9	90.3	79.3	39.9		17.3		66.0
	Disulfoton	57.2	78.5	95.9	79.8		31.3		18.4
	Dithiopyr	80.7	92.8	87.9	76.7		14.0		13.6
	EPTC	45.5	59.5	45.2	58.9		26.6		26.3
	Ethion	87.2	95.7	88.0	72.2		9.31		19.7
	Ethoprop	88.8	99.7	98.1	105		11.5		7.25
	Fenamiphos	55.6	49.9	57.5	54.4		10.9		5.58
	Fipronil	71.6	94.2	78.0	74.5		27.3		4.59
	Fipronil Desulfinyl	80.5	91.4	86.7	78.4		12.8		10.1
	Fipronil Sulfide	80.3	89.8	83.6	79.7		11.2		4.81
	Fipronil Sulfone	78.1	89.6	79.0	62.1		13.8		23.9
	Fonofos	95.5	90.0	96.5	101		5.92		4.24
	Hexazinone	98.5	99.1	92.1	81.6		0.574		7.84

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery		MS % Recovery		LCS	Precision SMP	MS
EPA 8270E	Malathion	102	89.6	48.0	64.8	12.8		29.9
	Metalaxyl	98.2	87.3	93.6	83.7	11.8		11.1
	Metolachlor	91.4	82.7	86.0	72.1	9.98		17.5
	Metribuzin	189	197	239	294	4.10		20.5
	Mevinphos	91.9	81.0	92.8	79.4	12.6		15.5
	Molinate	75.4	70.0	67.7	64.0	7.44		5.61
	Norflurazon	96.9	84.6	87.9	76.6	13.5		13.1
	Oxadiazon	91.0	82.8	86.1	74.1	9.39		15.0
	Parathion Ethyl	99.0	86.6	91.8	90.1	13.3		1.84
	Parathion Methyl	107	96.9	99.9	99.4	10.4		0.590
	Pendimethalin	88.5	82.4	85.9	83.1	7.14		3.22
	Phorate	95.3	86.6	98.6	99.5	9.57		0.922
	Prodiamine	84.1	77.7	96.4	85.5	7.86		12.0
	Prometon	85.5	79.4	87.1	48.2	7.41		57.5
	Prometryn	93.3	83.0	90.5	81.2	11.8		10.9
	Simazine	108	106	110	106	2.39		3.84
	Tebuconazole	70.5	50.0	16.3	4.62	34.1		112
	Terbufos	93.6	86.5	99.7	102	7.96		2.67
	Terbuthylazine	100	91.0	99.1	92.2	9.52		7.16
EPA 8321B	2,4-D	101		143	150			5.08
	Acesulfame K	96.2		109	101			7.61
	Acetaminophen	76.9		92.4	95.9			3.74
	Acetamiprid	86.8		113	134			16.4
	Afidopyropen	72.0		108	119			9.65
	AMPA	95.1		116	107			8.20
	Bentazon	144		106	106			0.515
	Benzovindiflupyr	75.7		87.0	93.9			7.56
	Carbamazepine	94.5		111	132			15.8
	Clothianidin	88.9		111	130			15.6
	Dinotefuran	88.8		123	138			11.3
	Diuron	77.9		90.0	103			12.6
	Endothall	131		143	143			0.0838
	Fenuron	102		105	124			17.2
	Fluridone	86.7		90.6	108			16.4
	Glufosinate	114		123	118			3.98
	Glyphosate	101		114	126			10.1
	Hydrocodone	90.3		116	134			14.2
	Ibuprofen	82.9		96.6	113			15.8
	Imazapyr	100		161	181			11.6
	Imidacloprid	88.8		155	183			16.0
	Linuron	71.3		76.8	91.8			17.8
	Mandestrobin	85.3		88.6	108			19.5
	MCPP	115		169	175			3.28
	Naproxen	99.4		90.6	109			18.2
	Primidone	67.6		96.8	121			22.4
	Pyraclostrobin	72.9		85.8	88.2			2.80
	Sucralose	105		101	105			4.44
	Thiamethoxam	86.4		144	168			14.9
	Tolfenpyrad	47.3		51.8	48.4			6.75
	Triclopyr	92.8		122	104			15.7
SM 2120 B	Color (true)	104					1.93	
SM 2320 B-97	Total Alkalinity	104					1.15	
SM 2540 C-97	TDS						1.02	
SM 4500 F-C-97	Fluoride	97.9		99.3	99.3			0.0

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision LCS	SMP	MS
SM 5310 B-00	Organic Carbon	100	99.8	100			0.536
	Organic Carbon	101	99.2	99.7			0.396

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	2109676, 2109677
EPA 200.7 Rev. 4.4 / E31780	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	2109693, 2109694
EPA 200.8 Rev. 5.4 / E31780	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	2109693, 2109694
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	2109676, 2109677
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	2109676, 2109677
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2109678, 2109679
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2109678, 2109679
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2109678, 2109679
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2109678, 2109679
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2109680, 2109681
EPA 8270E / E31780	Organonitrogen and phosphorus pesticides in water matrices by GC/MS/MS	2109703, 2109704
EPA 8321B / E31780	Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS	2109689, 2109690
EPA 8321B / E31780	Water soluble organic compounds by HPLC/MS/MS	2109689, 2109690
SM 2120 B / E31780	True Color measured at 450 nm	2109676, 2109677
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2109676, 2109677
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	2109676, 2109677
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2109676, 2109677
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2109676, 2109677
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2109678, 2109679

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	08/08/2019			08/08/2019 17:07	Rosalyn Ballman	2109676, 2109677
EPA 200.7 Rev. 4.4	08/08/2019	08/12/2019 14:45	Elliott D. Healy	08/13/2019 14:31	Betina Topolski	2109693
	08/08/2019	08/12/2019 14:45	Elliott D. Healy	08/13/2019 14:35	Betina Topolski	2109694
EPA 200.8 Rev. 5.4	08/08/2019	08/12/2019 14:45	Elliott D. Healy	08/23/2019 06:03	Martin Acosta	2109693
	08/08/2019	08/12/2019 14:45	Elliott D. Healy	08/23/2019 06:12	Martin Acosta	2109694
	08/08/2019	08/12/2019 14:45	Elliott D. Healy	08/30/2019 16:29	Robert K Palmer	2109693
	08/08/2019	08/12/2019 14:45	Elliott D. Healy	08/30/2019 16:35	Robert K Palmer	2109694
EPA 300.0 Rev. 2.1	08/08/2019			08/16/2019 00:25	Jhamieka S. Greenwood	2109676
	08/08/2019			08/16/2019 01:50	Jhamieka S. Greenwood	2109677
EPA 350.1 Rev. 2.0	08/08/2019			08/15/2019 11:55	Ping Hua	2109679
	08/08/2019			08/15/2019 13:54	Ping Hua	2109678
EPA 351.2 Rev. 2.0	08/08/2019	08/13/2019 12:36	Sima Feizi	08/14/2019 12:03	Joseph Suarez	2109678
	08/08/2019	08/13/2019 12:36	Sima Feizi	08/14/2019 12:04	Joseph Suarez	2109679
EPA 353.2 Rev. 2.0	08/08/2019			08/13/2019 13:00	Beverly Y. Sanders	2109678
	08/08/2019			08/13/2019 13:16	Beverly Y. Sanders	2109679
EPA 365.1 Rev. 2.0	08/08/2019	08/12/2019 13:21	Vijayalakshmi Reddy	08/13/2019 11:47	Lipi Saha	2109678
	08/08/2019	08/12/2019 13:21	Vijayalakshmi Reddy	08/13/2019 12:35	Lipi Saha	2109679
EPA 365.1 Rev. 2.0 dissolved	08/08/2019			08/08/2019 15:00	Jhamieka S. Greenwood	2109681
	08/08/2019			08/08/2019 15:08	Jhamieka S. Greenwood	2109680
EPA 8270E	08/08/2019	08/09/2019 09:00	Julie Wi	08/13/2019 16:35	Vandana T. Nagar	2109703
	08/08/2019	08/09/2019 09:00	Julie Wi	08/13/2019 17:01	Vandana T. Nagar	2109704
EPA 8321B	08/08/2019	08/08/2019 15:30	Rebecca Ware	08/13/2019 12:41	Rebecca Ware	2109689
	08/08/2019	08/08/2019 15:30	Rebecca Ware	08/13/2019 12:55	Rebecca Ware	2109690
	08/08/2019	08/08/2019 15:30	Rebecca Ware	08/13/2019 19:10	Rebecca Ware	2109689
	08/08/2019	08/08/2019 15:30	Rebecca Ware	08/13/2019 19:20	Rebecca Ware	2109690
	08/08/2019	08/08/2019 15:30	Rebecca Ware	08/19/2019 20:36	Rebecca Ware	2109690
	08/08/2019	08/08/2019 15:30	Rebecca Ware	08/20/2019 12:27	Rebecca Ware	2109689
	08/08/2019	08/09/2019 12:00	Fe-Val Siddell	08/14/2019 06:08	Juyoung Kim	2109689
	08/08/2019	08/09/2019 12:00	Fe-Val Siddell	08/14/2019 06:43	Juyoung Kim	2109690
SM 2120 B	08/08/2019			08/08/2019 16:46	Madeline Funaro	2109676
	08/08/2019			08/08/2019 16:49	Madeline Funaro	2109677
SM 2320 B-97	08/08/2019			08/15/2019 11:23	Dale L. Simmons	2109676
	08/08/2019			08/15/2019 11:36	Dale L. Simmons	2109677
SM 2540 C-97	08/08/2019			08/09/2019 15:30	Sima Feizi	2109676, 2109677
SM 2540 D-97	08/08/2019			08/13/2019 15:51	Rosalyn Ballman	2109676, 2109677
SM 4500 F-C-97	08/08/2019			08/15/2019 11:23	Dale L. Simmons	2109676
	08/08/2019			08/15/2019 11:36	Dale L. Simmons	2109677
SM 5310 B-00	08/08/2019			08/09/2019 19:39	Madeline Funaro	2109678
	08/08/2019			08/09/2019 20:16	Madeline Funaro	2109679