



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2019

Data Qualified?

Yes / No

WIN Station Name: Major Lake Center

Date: 04/11/2019
(mm/dd/yyyy)

WIN/ Field ID: G3TLHR0011

WBID: 526B

Latitude: 30.4982

(Decimal Degrees)

County: Washington

ORG ID: 21FLTLHR

Longitude: 85.7144

(Decimal Degrees)

Receiving Body of Water: NA

RQ: 2019-02-18-30

Sampling Team Members					WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check	Sampling Equipment		
J. Burchett P. Blair					✓	✓	✓	✓	✓	☒ Sample Container	☐ Van Dorn	☐ Pole
										☐ Carboy	☒ Syringe	
										☐ Dipper	☐ Other	
										Depth Measured with Equipment ID		
										Collection Method		
										☐ Wading	☐ Canoe/Kayak	
										☐ From Shore	☐ Electric Motor	
										☐ Other	☒ Gasoline Motor	

Water Level: Dry/ Pooled/ Low / Normal / High / Flooded

Meter ID# 154177

Matrix: Fresh / Marine / DI

Photos: Yes / No

Flow: No Flow / Intestinal / Flowing / NA

Tide: Rising/ Falling/ Slack/ NA

Tide Times: High Low

Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Sp COND (µmhos/cm)	Temp. (C°)
EST	1110	7.5	0.3	1.7	8.63	102.3	4.73	0.01	15.1	24.0
CEN	1113		7.0		5.32	53.5	4.51	0.01	17.5	20.9

Biological Samples Collected: None HA SCI RPS Algae ID LVS LVI Other:

SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#:

Parameter Suite	Parameter Methods	Preservation	Lot#	Expiration	pH Check
Preserved Nutrients	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☒ Ice ☒ H2SO4	8344070	12/19/19	☒ <2
Metals	☐ W-ICPMS ☐ W-ICP	☐ Ice ☐ HNO3			☐ <2
Filtered Nutrients	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter	☒ Ice			
Chlorophyll	☒ CHLSUITE-W	☒ Ice			
Physical/Aggregate (Fresh)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS	☒ Ice			
Physical/Aggregate (Marine)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates	☐ MI-FW-QLDC	☐ Formalin			
Periphyton	☐ ALGAE-ID	☐ Ice			
Microbiology	☐ E Coli ☐ F Coli ☐ Enterococci	☐ Ice			
Molecular	☐ PCR-FILTER ☐ PCR-BacR ☐ PCR-GULL2 ☐ PCR-HF183 ☐ PCR-DG3 ☐ PCR-GFD	☐ Ice			
Tracers	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides	☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-E8321-DI ☐ W-PCL-SQ-R ☐ WE8321-MS ☐ W-CARB-AA	☐ Ice			
Phytoplankton	☐ DTY-QN-C ☐ PKD-QN-C	☐ Ice			

Notes:

Place Label Here

Signature: Grade Date: 04/11/19

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: QC Station ID, Field Parameters In LIMS DMT:

Scan/Save Field Sheets (Initials/Date) Migrate Data to WIN: (Initials/Date) Verify Data In WIN: (Initials/Date)

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: 154178

RQ- 2018-02-18-30

Project: SMP

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 4/19

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Paul Blair	4-11-19	8:25	19.8	754.5	9.1	9.06	99.3		0.98	Ⓟ/F	Ⓟ/F
ICV	"	"	8:26	19.8	754.5	9.1	9.06	99.3		0.98	Ⓟ/F	Ⓟ/F
CCV	J. Burdett	✓	1530	26.1	753.5	8.1	8.17	100.7			Ⓟ/F	Ⓟ/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Paul Blair	4-11-19	8:28	180621C	7-19	1000	1001	Ⓟ/F	Ⓟ/F
ICV	"	"	8:30	1805F81	5-20	100	102	Ⓟ/F	Ⓟ/F
CCV	J. Burdett	✓	1532	1805F81	05/20	100	102	Ⓟ/F	Ⓟ/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Paul Blair	4-11-19	8:35	180403C	10-19	7.01	20.0	7.0	-51.0	Ⓟ/F	Ⓟ/F
Calibr.	"	"	8:36	180621D	7-19	4.01	20.6	4.0	126.2	Ⓟ/F	Ⓟ/F
Calibr.	"	"	8:37	180621F	1-20	10.06	20.6	10.0	-213.9	Ⓟ/F	Ⓟ/F
ICV	"	"	8:39	"	"	10.06	19.9	10.0	-212.3	Ⓟ/F	Ⓟ/F
CCV	J. Burdett	✓	1534	180621D	07/19	4.01	20.5	4.12	117.3	Ⓟ/F	Ⓟ/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU;

mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 162.9, slope from 4 to 7 177.2 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 4/19

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: ; Date of last verification:

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Paul Blair	4-11-19	8:45	0.0	0.0	Ⓟ/F	Ⓟ/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

Customer: TLH-ROC

Requester: Michael Eckles

Field Report Prepared By:

Project ID: SMP

Collected By:

S. Burdett Woods

P. B. I. Sampling Agency:

8031

[illegible]

Relinquished By: JB, PB	Date/Time 4/11/19:1529	Shipping Method HD	Number of Coolers Shipped: 1	Received By: Tuber R.	Date/Time 6/11/19 3:30
----------------------------	---------------------------	-----------------------	---------------------------------	--------------------------	---------------------------

Group: A	Bottle Type: ALKALINITY TURBIDITY W-CL-IC W-COLOR W-F W-SO4-IC W-TDS W-TSS	P-1L	# of Bottles: 5	Preservative:	ICE	Enter Number of Bottles Sent to Lab	2
Group: A	Bottle Type: CHLSUITE-W	BP-1L	# of Bottles: 5	Preservative:	ICE	Enter Number of Bottles Sent to Lab	2
Group: A	Bottle Type: W-NH3 W-NO2NO3 W-S-A-TP W-TKN W-TOC	P-500ML	# of Bottles: 5	Preservative: ICE And H2SO4		Enter Number of Bottles Sent to Lab	2
Group: A	Bottle Type: W-PO4-F	P-125ML	# of Bottles: 5	Preservative: FILTER-ICE		Enter Number of Bottles Sent to Lab	2
Group: B	Bottle Type: ALKALINITY TURBIDITY W-CL-IC W-COLOR W-F W-SO4-IC W-TDS W-TSS	P-1L	# of Bottles: 5	Preservative:	ICE	Enter Number of Bottles Sent to Lab	
Group: B	Bottle Type: CHLSUITE-W	BP-1L	# of Bottles: 5	Preservative:	ICE	Enter Number of Bottles Sent to Lab	
Group: B	Bottle Type: ECOLI-18QT	P-250ML-WO-THI	# of Bottles: 5	Preservative:	ICE	Enter Number of Bottles Sent to Lab	
Group: B	Bottle Type:	P-500ML	# of Bottles: 5	Preservative: ICE And H2SO4		Enter Number of Bottles Sent to Lab	