



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # 15A101028

PRE-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 6/18/19 TIME 0730 hrs

POST-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 6/19/19 TIME 0830 hrs

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	8.69	22.340	99.8	P
CALIBRATION	8.69	22.340	99.8	P
PRE-VERIFICATION	8.70	22.220	99.9	P
POST-VERIFICATION	8.52	22.902	99.1	P

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	1000	180621C	7/19	1060.1	P
CALIBRATION	1000	180621C	7/19	1000.0	P
PRE-VERIFICATION	100	2802833	7/20	103.5	P
POST-VERIFICATION	1000	180621C	7/19	1002.5	P

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7.0	180403C	10/19	7.14	23.069	P
CALIBRATION	7.0	180403C	10/19	7.00	23.079	P
CALIBRATION	4.0					
CALIBRATION	10.0	180222A	9/19	10.00	23.633	P
PRE-VERIFICATION	10.0	180222A	9/19	10.00	23.658	P
POST-VERIFICATION	7.0	180403C	10/19	6.97	23.401	P

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. $\pm 5\%$ pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES NO

Was the Dissolved Oxygen Membrane changed? (circle one) YES NO N/A

RQ & SITE IDs

Ichetucknee Main Blue Hole
Grassy Hole
Mission
Mill Pond

SAMPLER NAMES & SIGNATURES (All team members present)

Mike Dalsis [Signature] Gary Maddy [Signature]