

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: 154177

RQ: 2019-08-19-22

Project: 2019 SMP

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 07/19

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Ben Harden	09/05/19	0751	21.2	754.4	8.8	8.82	99.3		1.016	Ⓟ / F	Ⓟ / F
ICV	"	"	0808	20.6	754.5	8.84	8.92	99.4			Ⓟ / F	Ⓟ / F
CCV	"	"	1424	27.3	753.9	8.8	8.85	99.0			Ⓟ / F	Ⓟ / F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Ben Harden	09/05/19	0754	190618F	06/20	1000	1001	Ⓟ / F	Ⓟ / F
ICV	"	"	0755	2908C81	06/21	100.0	102.4	Ⓟ / F	Ⓟ / F
CCV	"	"	1425	2908C81	08/21	100.0	103.5	Ⓟ / F	Ⓟ / F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Ben Harden	09/05/19	0758	190618B	12/20	7.02	21.6	7.01	-32.1	Ⓟ / F	Ⓟ / F
Calibr.	"	"	0801	190618A	12/20	4.00	20.7	4.00	146.8	Ⓟ / F	Ⓟ / F
Calibr.	"	"	0804	180621F	01/20	10.05	20.8	10.05	-135.0	Ⓟ / F	Ⓟ / F
ICV	"	"	0806	"	"	10.05	20.5	10.09	-136.6	Ⓟ / F	Ⓟ / F
CCV	"	"	1435	190405E	10/20	1.00	20.1	1.15	314.4	Ⓟ / F	Ⓟ / F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 152.9, slope from 4 to 7 178.9 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 07/19

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: ; Date of last verification:

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Ben Harden	09/05/19	0746	0.272	0.272	Ⓟ / F	Ⓟ / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: