

Biological Analysis Report

TLH-ROC-2019-07-24-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Run 3**
Request ID: **RQ-2019-07-08-49**
Customer: **TLH-ROC**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
Tallahassee, FL 32399
Attn: Michael Eckles

For additional information please contact
Cheryl Swanson - Administrator
Jacqueline Savage - Microbiology
Diane Power - Bench Biology
Andrew Tintle - Aquatic Toxicology
Puja Jasrotia, Ph.D. - Molecular Biology
Amelia Rankin - Taxonomy (Invertebrate & Algal)
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8177

Certified by: Puja Jasrotia, Environmental Manager

Date Certified: 09-SEP-2019 12:16

A handwritten signature in black ink, appearing to be 'Puja Jasrotia', with a horizontal line underneath it.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Chlorophyll/Grain Size/BOD, Microbiology and Molecular.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: WOODS CREEK @ OSTEEN RD

Collection Date/Time: 07/24/2019 11:05

Field ID: G1TLHR0067

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2105874	SM 10200 H (mod.)	Chlorophyll-a, Corrected	1.6	U	ug/L	P368321	
		Phaeophytin-a	1.8	U	ug/L	P368321	
		Chlorophyll-a, Uncorrected	1.2	U	ug/L	P368321	
2105877	SM 9223 Quanti-Tray	Escherichia Coli-Quanti-Tray	48.7		MPN/100 mL	P368071	

Sample Location: SPRING CREEK @ US98

Collection Date/Time: 07/24/2019 11:35

Field ID: G1TLHR0084

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2105878	SM 9223 Quanti-Tray	Escherichia Coli-Quanti-Tray	86.2		MPN/100 mL	P368071	

Sample Location: AUCILLA RIVER @ MOUTH

Collection Date/Time: 07/24/2019 12:10

Field ID: G1TLHR0091

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2105873	SM 10200 H (mod.)	Chlorophyll-a, Corrected	2.4	I	ug/L	P368321	
		Phaeophytin-a	1.8	U	ug/L	P368321	
		Chlorophyll-a, Uncorrected	2.3	I	ug/L	P368321	
2105876	Enterolert/QT	Enterococci-Quanti-Tray	223.0		MPN/100 mL	P368071	
2105879	SOP-PCR-1.0	HF183-qPCR**	140	U	GEU/100 mL	P368014	LCS
	SOP-PCR-1.2	GULL2-qPCR**	8.4E+03	U	TSC/100 mL	P368014	
	SOP-PCR-1.3	GFD-purified-qPCR**	1.09E+03	J	TSC/100 mL	P369452	LCS, MS
	SOP-PCR-1.4	BacR-qPCR**	1.5E+03	U	TSC/100 mL	P368014	
	SOP-PCR-4.0	HF183 - Human Source Marker**	Above Threshold			P367815	

Ref. Method and Comment:

Enterolert/QT: Precision data is unavailable due to the small amount of analyte in the QC sample.

SOP-PCR-1.2: Precision data unavailable for Analysis Batch A93601

SOP-PCR-1.3: Quality control failure(s) observed. See QA report for details.

Quality Assurance Report Method Blank Results

Reference Method: Enterolert/QT

Batch ID: P368071

Component	Result	Code	Units
Enterococci-Quanti-Tray	1.0	U	MPN/100 mL

Reference Method: SM 10200 H (mod.)

Batch ID: P368321

Component	Result	Code	Units
Chlorophyll-a, Corrected	0.89	U	ug/L
Chlorophyll-a, Uncorrected	0.65	U	ug/L
Phaeophytin-a	0.98	U	ug/L

Reference Method: SM 9223 Quanti-Tray

Batch ID: P368071

Component	Result	Code	Units
-----------	--------	------	-------

Quality Assurance Report Method Blank Results

Reference Method: SM 9223 Quanti-Tray
Batch ID: P368071

Component	Result	Code	Units
Escherichia Coli-Quanti-Tray	1.0	U	MPN/100 mL

Reference Method: SOP-PCR-1.0
Batch ID: P368014

Component	Result	Code	Units
HF183-qPCR	280	U	GEU/100 mL

Reference Method: SOP-PCR-1.2
Batch ID: P368014

Component	Result	Code	Units
GULL2-qPCR	1.7E+04	U	TSC/100 mL

Reference Method: SOP-PCR-1.3
Batch ID: P369452

Component	Result	Code	Units
GFD-purified-qPCR	340	U	TSC/100 mL

Reference Method: SOP-PCR-1.4
Batch ID: P368014

Component	Result	Code	Units
BacR-qPCR	3.0E+03	U	TSC/100 mL

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 10200 H (mod.)
Batch ID: P368321

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chlorophyll-a, Corrected	88.8		P	84 - 116

Reference Method: SOP-PCR-1.0
Batch ID: P368014

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
HF183-qPCR	172	239	P/F	50 - 175

Reference Method: SOP-PCR-1.2
Batch ID: P368014

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
GULL2-qPCR	85.1	77.3	P/P	50 - 300

Reference Method: SOP-PCR-1.3
Batch ID: P369452

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SOP-PCR-1.3
Batch ID: P369452

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
GFD-purified-qPCR	55.0	45.1	P/F	50 - 175

Reference Method: SOP-PCR-1.4
Batch ID: P368014

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
BacR-qPCR	77.3	76.8	P/P	50 - 175

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SOP-PCR-1.0
Batch ID: P368014

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2104380	HF183-qPCR	99.9	124	P/P	50 - 175
2104381	HF183-qPCR	98.2		P	50 - 175
2105231	HF183-qPCR	102		P	50 - 175
2105232	HF183-qPCR	111		P	50 - 175
2105233	HF183-qPCR	141		P	50 - 175
2105234	HF183-qPCR	83.4		P	50 - 175
2105235	HF183-qPCR	96.6		P	50 - 175
2105370	HF183-qPCR	115		P	50 - 175
2105634	HF183-qPCR	102		P	50 - 175
2105879	HF183-qPCR	160		P	50 - 175
2106297	HF183-qPCR	142		P	50 - 175
2106893	HF183-qPCR	143		P	50 - 175
2106895	HF183-qPCR	95.1		P	50 - 175

Reference Method: SOP-PCR-1.2
Batch ID: P368014

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2105231	GULL2-qPCR	57.7		P	50 - 300
2105233	GULL2-qPCR	68.9		P	50 - 300
2105879	GULL2-qPCR	65.7		P	50 - 300

Reference Method: SOP-PCR-1.3
Batch ID: P369452

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2105159	GFD-purified-qPCR	194		F	50 - 175
2105231	GFD-purified-qPCR	123		P	50 - 175
2105233	GFD-purified-qPCR	136		P	50 - 175
2105879	GFD-purified-qPCR	180		F	50 - 175
2108088	GFD-purified-qPCR	186		F	50 - 175
2109108	GFD-purified-qPCR	160	174	P/P	50 - 175
2109620	GFD-purified-qPCR	154		P	50 - 175
2109621	GFD-purified-qPCR	139		P	50 - 175
2109622	GFD-purified-qPCR	206		F	50 - 175
2109623	GFD-purified-qPCR	139		P	50 - 175

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SOP-PCR-1.3

Batch ID: P369452

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109797	GFD-purified-qPCR	149		P	50 - 175

Reference Method: SOP-PCR-1.4

Batch ID: P368014

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2104380	BacR-qPCR	62.3	70.1	P/P	50 - 175
2105231	BacR-qPCR	60.3		P	50 - 175
2105233	BacR-qPCR	65.9		P	50 - 175
2105370	BacR-qPCR	63.1		P	50 - 175
2105879	BacR-qPCR	75.8		P	50 - 175

Quality Assurance Report Precision

Reference Method: SM 10200 H (mod.)

Batch ID: P368321

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2106051	Chlorophyll-a, Corrected	6.82	Sample	P	0 - 25
2106051	Chlorophyll-a, Uncorrected	5.89	Sample	P	0 - 25
2106051	Phaeophytin-a	0.748	Sample	P	0 - 25

Reference Method: SM 9223 Quanti-Tray

Batch ID: P368071

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2105437	Escherichia Coli-Quanti-Tray	23.8	Sample	P	0 - 80
2105838	Escherichia Coli-Quanti-Tray	14.4	Sample	P	0 - 80

Reference Method: SOP-PCR-1.0

Batch ID: P368014

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2104380	HF183-qPCR	21.3	Spike	P	0 - 25

Reference Method: SOP-PCR-1.3

Batch ID: P369452

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109108	GFD-purified-qPCR	8.14	Spike	P	0 - 25

Reference Method: SOP-PCR-1.4

Batch ID: P368014

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
-----------------------	-----------	-----------	-------------------	-----------	----------------

Quality Assurance Report Precision

Reference Method: SOP-PCR-1.4
Batch ID: P368014

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2104380	BacR-qPCR	11.7	Spike	P	0 - 25

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Surrogates

Lab Sample ID: 2105879
Field Sample ID: G1TLHR0091

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
SOP-PCR-1.0	SKETA-qPCR	130	P	50 - 300
SOP-PCR-1.2	SKETA-qPCR	110	P	50 - 300
SOP-PCR-1.4	SKETA-qPCR	108	P	50 - 300

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery		MS % Recovery			LCS	Precision SMP	MS
SM 10200 H (mod.)	Chlorophyll-a, Corrected	88.8						6.82	
	Chlorophyll-a, Uncorrected							5.89	
	Phaeophytin-a							0.748	
SM 9223 Quanti-Tray	Escherichia Coli-Quanti-Tray							23.8	14.4
SOP-PCR-1.0	HF183-qPCR	239	172	142	160	143			21.3
				141					
SOP-PCR-1.2	GULL2-qPCR	77.3	85.1	65.7	68.9	57.7			
SOP-PCR-1.3	GFD-purified-qPCR	45.1	55.0	123	136	180			8.14
				160					
SOP-PCR-1.4	BacR-qPCR	76.8	77.3	75.8	65.9	60.3			11.7
				62.3					

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
Enterolert/QT / E31780	Enterococci by Enterolert Quanti-Tray method - non-chlorinated water systems	2105876
SM 10200 H (mod.) / E31780	Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry	2105873, 2105874
SM 9223 Quanti-Tray / E31780	Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems	2105877, 2105878
SOP-PCR-1.0 / E31780	Detection and quantification of human-specific HF183 Bacteroides genetic marker with quantitative polymerase chain reaction	2105879
SOP-PCR-1.2 / E31780	Detection and quantification of coastal bird specific Catellicoccus marimammalium Gull2 genetic marker with quantitative polymerase chain reaction	2105879
SOP-PCR-1.3 / E31780	Detection and quantification of bird specific Helicobacter GFD genetic marker with quantitative polymerase chain reaction	2105879
SOP-PCR-1.4 / E31780	Detection and quantification of Ruminant specific Bacteroidetes BacR genetic marker with quantitative polymerase chain reaction	2105879
SOP-PCR-4.0 / E31780	Preparation of Samples by Filtration and held for qPCR Analysis	2105879

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
Enterolert/QT	07/24/2019	07/24/2019 16:21	Carina A. Tull	07/25/2019 16:43	Jessica Fetgatter	2105876
SM 10200 H (mod.)	07/24/2019	07/24/2019 16:35	Edwin Gonzalez-Soto	08/01/2019 09:57	Yuliya Danyuk	2105873, 2105874
SM 9223 Quanti-Tray	07/24/2019	07/24/2019 16:57	Carina A. Tull	07/25/2019 11:35	Carina A. Tull	2105877, 2105878
SOP-PCR-1.0	07/24/2019	07/25/2019 15:30	Angela Smith	08/12/2019 10:13	Mailin Lopez	2105879
SOP-PCR-1.2	07/24/2019	07/25/2019 15:30	Angela Smith	08/19/2019 09:21	Mailin Lopez	2105879
SOP-PCR-1.3	07/24/2019	07/25/2019 15:30	Mailin Lopez	08/27/2019 08:19	Mailin Lopez	2105879
SOP-PCR-1.4	07/24/2019	07/25/2019 15:30	Angela Smith	08/13/2019 08:21	Mailin Lopez	2105879
SOP-PCR-4.0	07/24/2019	07/25/2019 15:30	Angela Smith	08/02/2019 16:16	Sarah Menz	2105879