

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Settler

Boldly "X" this box if there are qualified data on this page.

Meter ID:

Cheney

RQ-

2019-09-23-56

Project:

SNP

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

8/12/19

Temperature (Quarterly) FT 1500												
DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Bentley water	9/25/19	717	23.8	758.9	8.45	8.43	99.9			P / F	L / F
ICV			718	23.9	758.9	8.45	8.42	99.8		1.002	P / F	L / F
CCV			1530	23.4	757.6	8.51	8.30	97.4			P / F	L / F
CCV												P / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Optical: DO gain range 0.85 to 1.15 (10 B55 0.75 to 1.05) DO range 0.0 to 10.0										
Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field	
				1904 F49	4/21					
Calibr.	Bentley water	9/25/19	720	2902B78	2/21	1000	1000	P / F	L / F	
ICV			722	2902E32	2/21	100	103	P / F	L / F	
CCV			1531	2902B78	2/21	500	507	P / F	L / F	
CCV									P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Bentley water	9/25/19	734	191304	3/21	7.	23.2	7.00	-26	P / F	L / F
Calibr.			736	192462	5/21	4.	23.4	4.00	149	P / F	L / F
Calibr.			739	191040	3/21	10.	23.5	10.00	-203	P / F	L / F
ICV			742	192462	5/21	4.00	23.7	4.12	145	P / F	L / F
CCV			1533	191304	3/21	7.00	24.1	7.03	-28	P / F	L / F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification:

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: 1; Date of last verification: 8/12/19

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: