

Log-in Checklist

RQ-2019-02-18-21

Login Station ID: # 3

Shipping Method: FedEx | UPS | HD | Greyhound

Date/Time of Receipt: 02/19/2019 / 09:20

Cooler Temperature*	* All Samples in Cooler Preserved with		Number of Sample Containers in Cooler	Evidence Tape				Tracking # (fill out only if waybill copy is damaged)
	HNO ₃	Formalin		Present?		*Intact?		
				Yes	No	Yes	No	
2.1°C			9		X			

* HNO₃ and Formalin persevered samples do **NOT** have a temperature requirement. For all others, if the temperature of a cooler is above 6.0 °C (or above 10.0 °C for W-1-4-DIOX), or if Evidence Seal is damaged. Identify the containers in the affected cooler(s) on back of this form.

****Note:** If "No" is checked below, fill out the back of this form (Field ID, Test IDs, etc.) and generate an NCR.

Chain of Custody/ Field Sheet(s) Included? Yes **X No

Micro-Biology Overflow Chain of Custody/ Field Sheet(s) Included? Yes No **X** Date

CONTAINER CHECK

Evidence Tape on Bottles? Yes No **X If yes, is it intact? Yes No

Caps on tight? Yes **X No **Containers intact? Yes **X** No

Sufficient Sample Volume: Yes **X No

CHLORINE CHECK REQUIRED? (Blue dot on container) Yes No **X** Init: **ABS**

Check the Box marked "Yes" or "No" for the Presence of Chlorine. If Cl detected, an NCR is required.					
Analysis	Yes	No	Analysis	Yes	No
W-1,4 DIOX			W-PC-AA-SF		
W-E8321-DI, -MS, -21			W-NP-AA-SF		
W-BNA (-625, -R)			W-PDQUAT		
W-CARB-AA (-CAB-AA-R)			W-PCL-SQ-R		
W-U2060-MS			W-TR-TQ		
W-PCB-R			W-CT-CI-R		
W-PAH (-R)			W-PSNP-TQ		
W-PFAS-MS					

PRESERVATION CHECK (EXCLUDING SAMPLES SENT TO OVERFLOW LABS, "OV-")

Acid Preserved Samples: pH < 2.0? Yes **X No NA Init: **ABS**

W-1-4-DIOX (Green dot on container) preserved to pH < 4.0? Yes No NA **X Init: **ABS**

Coolers Unpacked/Checked by: **ABS** Date: 02/19/2019 NCRs (Y/N)?

Event ID: **Flint** NCR #
SW-DIST-2019-02-19-04 (SMP)

Log-in Checklist

Samples with Incorrect pH Preservation or presence of Chlorine

Field ID	Bottle Test ID(s)	pH / Cl	Action Taken

Samples in Coolers with Incorrect Preservation or Damaged Evidence Tape

Cooler ID	Field ID	Bottle Test ID(s)

Bottles Not Intact

Field ID	Bottle Test ID(s)	Loose Cap	Damaged Cap		Damaged Container		Evidence Tape Not Intact
			CK	BR	CK	BR	

Samples with Insufficient Volume for Analyses

Field ID	Bottle Test ID(s)	Action Taken

Additional Comments:

Infrared Thermometer ID:	REC_03_2018
pH Paper 0.0 – 3.0 Lot #	
pH Paper 0 – 13 Lot #	215517 EXP:06/01/2020
Chlorine Test Strips Lot #	7226 EXP:2020/07

ENCORE SAMPLES

S-VOC-MS samples in Encores must be frozen within 48 hours of collection.

Were Encore samples received? Yes _____ No X

Date and time placed in freezer _____

Were all samples placed in freezer within 48 hours of collection? Yes _____ If No _____

no, were samples shipped on dry ice? Yes _____ No _____

FOR CLEAN LAB USE ONLY – Total Mercury Preservation Check (W-HG-AF;W-HG-AF-F):

Samples preserved within 48 hours? Yes _____ No _____

Date and time samples preserved: _____

Flint

Central Laboratory Submittal Form

last revised Jan 25, 2018

Customer: SW-DIST

Requester: Benjamin Paswater

Field Report Prepared By:

Project ID: SMP

Collected By:

Sampling Agency:

Location		☐ Comp ☒ Grab		Collection (comp begin or grab) Date <u>2-18-19</u> Time <u>0945</u>		Eastern Central		Composite end Date _____ Time _____		Bottle Group(s) (See pg 2)	
Field ID/WIN	G2SW0004	Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen <u>3.66</u> ☒ mg/L ☐ % sat		Total Res. Chlorine (mg/L)				A	
WIN/		Temp (C) <u>20.82</u>		pH <u>5.59</u>		Sample Depth <u>0.3</u> ☐ ft ☒ m		Sp. Conductance (umho/cm) <u>189</u>			
Site Name:	Flint Creek @ US 301	Salinity (ppt) <u>0.09</u>		Comments							
Latitude	☐ dd ☐ dms	☐ dd ☐ dms									
Location		☐ Comp ☒ Grab		Collection (comp begin or grab) Date <u>2-18-19</u> Time <u>1020</u>		Eastern Central		Composite end Date _____ Time _____		Bottle Group(s)	
Field ID	G2SW0066	Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen <u>4.68</u> ☒ mg/L ☐ % sat		Total Res. Chlorine (mg/L)				B	
WIN/S		Temp (C) <u>21.12</u>		pH <u>5.66</u>		Sample Depth <u>0.3</u> ☐ ft ☒ m		Sp. Conductance (umho/cm) <u>183</u>			
Site Name:	Flint Creek @ Knights Griffin Rd	Salinity (ppt) <u>0.09</u>		Comments							
Latitude	☐ dd ☐ dms	☐ dd ☐ dms									
Location		☐ Comp ☐ Grab		Collection (comp begin or grab) Date _____ Time _____		Eastern Central		Composite end Date _____ Time _____		Bottle Group(s)	
Field ID		Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen ☐ mg/L ☐ % sat		Total Res. Chlorine (mg/L)					
WIN/STORET #		Temp (C) _____		pH _____		Sample Depth ☐ ft ☐ m		Sp. Conductance (umho/cm) _____			
Latitude	☐ dd ☐ dms	Longitude ☐ dd ☐ dms		Salinity (ppt) ☐ dd ☐ dms		Comments					
Location		☐ Comp ☐ Grab		Collection (comp begin or grab) Date _____ Time _____		Eastern Central		Composite end Date _____ Time _____		Bottle Group(s)	
Field ID		Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen ☐ mg/L ☐ % sat		Total Res. Chlorine (mg/L)					
WIN/STORET #		Temp (C) _____		pH _____		Sample Depth ☐ ft ☐ m		Sp. Conductance (umho/cm) _____			
Latitude	☐ dd ☐ dms	Longitude ☐ dd ☐ dms		Salinity (ppt) ☐ dd ☐ dms		Comments					

Relinquished By:	Date/Time	Shipping Method	Number of Coolers Shipped:	Received By:	Date/Time
<u>J. Ashton</u>	<u>2-18-19 1700</u>	<u>FEDEX</u>	<u>1</u>	<u>[Signature]</u>	<u>02-19-19 10:20</u>

Group: A	Bottle Type:	P-1L	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	ALKALINITY TURBIDITY W-CL-IC W-COLOR W-F W-TDS W-TSS						
Group: A	Bottle Type:	BP-1L	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	CHLSUITE-W						
Group: A	Bottle Type:	P-500ML	# of Bottles: 1	Preservative:	HNO3	Enter Number of Bottles Sent to Lab	1
	W-ICP W-ICPMS						
Group: A	Bottle Type:	P-500ML	# of Bottles: 1	Preservative:	ICE And H2SO4	Enter Number of Bottles Sent to Lab	1
	W-NH3 W-NO2NO3 W-S-A-TP W-TKN W-TOC						
Group: A	Bottle Type:	P-125ML	# of Bottles: 1	Preservative:	FILTER-ICE	Enter Number of Bottles Sent to Lab	1
	W-PO4-F W-SO4-IC-F						
Group: B	Bottle Type:	P-1L	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	ALKALINITY TURBIDITY W-CL-IC W-COLOR W-F W-TDS W-TSS						
Group: B	Bottle Type:	BP-1L	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	CHLSUITE-W						
Group: B	Bottle Type:	P-500ML	# of Bottles: 1	Preservative:	ICE And H2SO4	Enter Number of Bottles Sent to Lab	1

Group: B	Bottle Type:	P-500ML	# of Bottles: 1	Preservative: ICE And H2SO4	Enter Number of Bottles Sent to Lab	<u>1</u>
	W-NH3					
	W-NO2NO3					
	W-S-A-TP					
	W-TKN					
	W-TOC					

Group: B	Bottle Type:	P-125ML	# of Bottles: 1	Preservative: FILTER-ICE	Enter Number of Bottles Sent to Lab	<u>1</u>
	W-PO4-F					
	W-SO4-IC-F					

Date of Request: 24-JAN-2019
 Created By: PASWATER_B On: 24-JAN-2019
 Modified By: JASROTIA_P On: 30-JAN-2019
 Customer: SW-DIST
 Project: SMP
 Division: Division of Environmental Assessment and Restoration
 District: Southwest District
 Sampling Event: Flint

Send Coolers To:

Phone: 813-632-7600
 FL Dept. of Environmental Protection
 13051 N. Telecom Parkway
 Temple Terrace, FL 33637-0926
 Attn: Ben Paswater

Program Module Number:
 Priority: 3
 Event Status: R
 Criminal Investigation: NO
 Chemistry Request Reviewed By: WATTS_J On: 29-JAN-2019
 Biology Request Reviewed By: JASROTIA_P On: 30-JAN-2019
 Sampling Kit Required: YES Ship on: 08-FEB-2019
 Sampling Kit Shipped: YES On: 04-FEB-2019
 Sampling Kit Packed By: KLUG_K
 Preservatives Needed: NO
 Date to Receive Samples: 18-FEB-2019
 Received By: SHAIK_A On: 19-FEB-2019

Send Final Report To:

FL Dept. of Environmental Protection
 13051 N. Telecom Parkway
 Temple Terrace, FL 33637-0926
 Attn: Ben Paswater

Report Type: Final Only
 FTP Data: NO
 QC Report: YES
 Date Log: NO
 Authorization Log: NO

Comments: Flint

Bottle Group A (Water) With 1 Samples:

Bottle Type: P-1L	Number of Bottles: 1	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-97) Alkalinity in aqueous matrices as mg CaCO ₃ /L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-COLOR	Template: DEFAULT	(SM 2120 B) True Color measured at 450 nm
Color (true)		
W-F	Template: DEFAULT	(SM 4500 F-C-97) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-TDS	Template: DEFAULT	(SM 2540 C-97) Total Dissolved Solids in aqueous matrices
W-TSS	Template: DEFAULT	(SM 2540 D-97) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample
Bottle Type: BP-1L	Number of Bottles: 1	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry
Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: HNO₃
W-ICP	Template:	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
Barium		
Calcium		
Iron		
Magnesium		
Sodium		
Zinc		
W-ICPMS	Template:	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects
*Boron		
Aluminum		
Antimony		
Arsenic		
Cadmium		

Bottle Group A (Water) With 1 Samples:

Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: HNO3
W-ICPMS	Template:	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects

Chromium
Copper
Lead
Nickel
Selenium
Silver
Thallium

Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: ICE And H2SO4
W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-00) Nonpurgeable Organic Carbon in aqueous matrices

Bottle Type: P-125ML	Number of Bottles: 1	Preserved With: FILTER-ICE
W-PO4-F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L
W-SO4-IC-F	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate, dissolved, in filtered, aqueous matrices

Bottle Group B (Water) With 1 Samples:

Bottle Type: P-1L	Number of Bottles: 1	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-97) Alkalinity in aqueous matrices as mg CaCO3/L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-COLOR	Template: DEFAULT	(SM 2120 B) True Color measured at 450 nm
Color (true)		
W-F	Template: DEFAULT	(SM 4500 F-C-97) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-TDS	Template: DEFAULT	(SM 2540 C-97) Total Dissolved Solids in aqueous matrices
W-TSS	Template: DEFAULT	(SM 2540 D-97) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample

Bottle Type: BP-1L	Number of Bottles: 1	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: ICE And H2SO4
W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-00) Nonpurgeable Organic Carbon in aqueous matrices

Bottle Type: P-125ML	Number of Bottles: 1	Preserved With: FILTER-ICE
W-PO4-F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L
W-SO4-IC-F	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate, dissolved, in filtered, aqueous matrices

* - The laboratory is not NELAP certified for this analyte/method, or certification is not applicable.

551C2/0330/1040

SHIP DATE: 07FEB19
ACTWT: 10.00 LB HAN
CAD: 0446970/CAFE3211

ORIGIN ID: TLHA (813) 470-5772
BEN RASWATER
FL DEP SOUTHWEST DISTRICT
13051 N TELECOM PKWY

TEMPLE TERRACE, FL 336370926
UNITED STATES US

TO JOHN WATTS

FLORIDA DEP/CHEMISTRY SECTION
2600 BLAIRSTONE RD

TALLAHASSEE FL 32399


(850) 246-8077

REF1

REF1

RMA: 01/01/01

FedEx
Express



FedEx


TRK# 0221

4751 8776 4986

TUE - 19 FEB 10:30A
PRIORITY OVERNIGHT

36 TLHA

32399
FL-US TLH



93597606 02/18 561J2/0E3D/2368

Don't forget to use the 10 digit code on the back of the label