

April 10, 2019

Francis Moore
FDEP-Bureau of Labs-Bio Section -Twin Towers Lab Complex A
Rm 216 MS 6515
2600 Blair Stone Rd.
Tallahassee, FL 32399

RE: Workorder: T1904893 SMP

Dear Francis Moore:

Enclosed are the analytical results for sample(s) received by the laboratory on Wednesday, March 20, 2019. Results reported herein conform to the most current NELAC standards, where applicable, unless otherwise narrated in the body of the report. The analytical results for the samples contained in this report were submitted for analysis as outlined by the Chain of Custody and results pertain only to these samples.

If you have any questions concerning this report, please feel free to contact me.

Sincerely,



Heidi Parker - Project Manager
HParker@AELLab.com

Enclosures

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Advanced Environmental Laboratories, Inc.



SAMPLE SUMMARY

Workorder: T1904893 SMP

Lab ID	Sample ID	Matrix	Date Collected	Date Received
T1904893001	G2SW0102	Water	3/20/2019 10:40	3/20/2019 13:51
T1904893002	G2SW0112	Water	3/20/2019 11:15	3/20/2019 13:51

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Advanced Environmental Laboratories, Inc.



ANALYTICAL RESULTS

Workorder: T1904893 SMP

Lab ID: **T1904893001**

Date Received: 03/20/19 13:51 Matrix: Water

Sample ID: **G2SW0102**

Date Collected: 03/20/19 10:40

Sample Description:

Location:

Parameters	Results	Qual	Units	DF	Adjusted PQL	Adjusted MDL	Analyzed	Lab
Microbiology								
Analysis Desc: Escherichia Coli,EPA 1603,Water			Analytical Method: EPA 1603					
E. Coli	96	1	#/100 mL	4	4	4	3/20/2019 16:13	T

Lab ID: **T1904893002**

Date Received: 03/20/19 13:51 Matrix: Water

Sample ID: **G2SW0112**

Date Collected: 03/20/19 11:15

Sample Description:

Location:

Parameters	Results	Qual	Units	DF	Adjusted PQL	Adjusted MDL	Analyzed	Lab
Microbiology								
Analysis Desc: Escherichia Coli,EPA 1603,Water			Analytical Method: EPA 1603					
E. Coli	470	1	#/100 mL	10	10	10	3/20/2019 16:13	T

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Advanced Environmental Laboratories, Inc.



ANALYTICAL RESULTS QUALIFIERS

Workorder: T1904893 SMP

PARAMETER QUALIFIERS

- U The compound was analyzed for but not detected.
- I The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- [1] Incubator 1: 3/20/2019 16:13 - 17:45, Incubator 2: 3/20/19 17:45 - 3/21/19 13:50

LAB QUALIFIERS

- T DOH Certification #E84589(AEL-T)(FL NELAC Certification)

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Advanced Environmental Laboratories, Inc.



QUALITY CONTROL DATA

Workorder: T1904893 SMP

QC Batch:	MICt/3201	Analysis Method:	EPA 1603
QC Batch Method:	EPA 1603	Prepared:	
Associated Lab Samples:	T1904893001, T1904893002		

METHOD BLANK: 3035499

Parameter	Units	Blank Result	Reporting Limit Qualifiers
Microbiology			
E. Coli	#/100 mL	1	1 U

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Advanced Environmental Laboratories, Inc.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Workorder: T1904893 SMP

Lab ID	Sample ID	Prep Method	Prep Batch	Analysis Method	Analysis Batch
T1904893001	G2SW0102			EPA 1603	MICt/3201
T1904893002	G2SW0112			EPA 1603	MICt/3201

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Advanced Environmental Laboratories, Inc.





Advanced
Environmental Laboratories, Inc.

☐ **Gainesville:** 4965 SW 41st Blvd. • Gainesville, FL 32608 • 352.377.2349 • Fax 352.395.6639
☐ **Jacksonville:** 6681 Southpoint Pkwy. • Jacksonville,
☐ **Tallahassee:** 2639 North Monroe Street, Suite D • Ta
☒ **Tampa:** 9610 Princess Palm Ave. • Tampa, FL 33619 •



Client Name: FDEP SW Dist.		Project Name: SMP		BOTTLE SIZE & TYPE Analysis Required E Coli
Address: 13051 N. Telecom Pkwy. Temple Terrace FL 33637		P.O. Number: B370AE		
Phone: 813-470-5770		Project Address:		
Contact: Judy Ashton 813-470-5772		Special Instructions:		
Sampled By:		RQ-2019-03-18-12		
Turn Around Time: ☒ STANDARD ☐ RUSH		☐ ADaPT ☐ EQuIS ☐ Other		
Page: of				

SAMPLE ID	SAMPLE DESCRIPTION	Grab Comp	SAMPLING		MATRIX	NO. COUNT	PRE SER- VATI ON								LABORATORY I.D. #
			DATE	TIME											
Field ID/WIN G2SW0102 	Site Name: Wildcat Creek S Branch@Stephens Rd		3/20/19	1040	SW	1		✓							001
Field ID/WIN G2SW0112 	Site Name: Wildcat Creek S Branch@Scramble Dr N		3/20/19	1115	SW	1		✓							002

Matrix Code: WW = wastewater SW = surface water GW = ground water DW = drinking water O = oil A = air SO = soil SL = sludge Preservation Code: I = ice H=(HCl) S = (H2SO4) N = (HNO3) T = (Sodium Thiosulfate)

Received on Ice ☒ Yes ☐ No ☒ Temp taken from sample ☐ Temp from blank ☒ Where required, pH checked Temperature when received 2.3 (in degrees celcius)

DCN: AD-051 Form last revised 04/30/2015

Device used for measuring Temp by unique identifier (circle IR temp gun used) J: 9A G: LT-1 LT-2 T: 10A A: 3A M: 3A S: 1V

Relinquished by:	Date	Time	Received by:	Date	Time
1 Sarah Myers	3/20/19	1351	[Signature]	3/20/19	1351

FOR DRINKING WATER USE:

PWS ID: