From: Kelsey Ulrich
To: DEP-Overflowmicro; Ashton, Judy
Subject: Re: BenchmarkEA: 19061433 RQ-2019-06-24-15 06.27.19
Date: Thursday, July 11, 2019 10:51:14 AM

Hi Kara,

You're right, the dilution factor should be 1. I talked with the analyst and got that sorted out, so it should be correct from now on. Our database system labels the dilution factor as just "dilution" so there was a bit of confusion as to whether or not that was the same as the "dilution factor" or the amount of sample that was used. Would you like me to resend the spreadsheet and .txt file or are you able to edit it on your side?

I apologize for data entry error! Please let me know if you need anything else!

Thank you,
Kelsey Ulrich
Data Reporting Specialist
Benchmark EA, Inc.
941-723-9986

From: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Sent: Thursday, July 11, 2019 10:40:05 AM
To: Kelsey Ulrich; DEP-Overflowmicro; Ashton, Judy
Subject: RE: BenchmarkEA: 19061433 RQ-2019-06-24-15 06.27.19

Hi Kelsey,

For the fecal result in the report, what volume was filtered for the reported result? It seems like the mdl/pql and dilution factor aren't jiving.

Thanks! Kara

From: Kelsey Ulrich <kelsey.ulrich@benchmarkea.net>
Sent: Wednesday, July 10, 2019 2:35 PM
To: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>; Ashton, Judy <Judy.Ashton@dep.state.fl.us>
Subject: BenchmarkEA: 19061433 RQ-2019-06-24-15 06.27.19

Hello,

Attached is the following completed report:

19061433 RQ-2019-06-24-15 06.27.19

Please feel free to contact me during our normal business hours Mon - Fri 8:30 a.m. to 5 p.m. if you have any questions.

Thank you,
Kelsey Ulrich
Data Reporting Specialist
Benchmark EA, Inc.
941-723-9986