

Chemical Analysis Report

SW-DIST-2019-09-11-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **1654 - Snug Harbor DO**
Request ID: **RQ-2019-08-26-12**
Customer: **SW-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
13051 N. Telecom Parkway
Temple Terrace, FL 33637-0926
Attn: Matthew Bearden

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Dr. rer. nat. Bettina Steinbock, Environmental Administrator

Date Certified: 24-SEP-2019 15:22



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Snug Harbor near Getaway

Collection Date/Time: 09/10/2019 14:00

Field ID: G1SW0109

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2117596	EPA 180.1 Rev. 2.0	Turbidity	2.0		NTU	P370494	
	EPA 300.0	Bromide	35		mg Br/L	P370689	
	SM 2320 B-97	Total Alkalinity	107	A	mg CaCO3/L	P371103	
	SM 2540 D-97	TSS	8	I	mg/L	P370880	
	SM 4500 F-C-97	Fluoride	0.78		mg F/L	P371107	
2117597	EPA 350.1 Rev. 2.0	Ammonia-N	0.005	I	mg N/L	P370666	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.58		mg N/L	P370535	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.005	I	mg N/L	P370709	
	EPA 365.1 Rev. 2.0	Total-P	0.17		mg P/L	P370565	
	SM 5310 B-00	Organic Carbon	7.2		mg C/L	P370679	
2117598	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.14		mg P/L	P370502	

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P370494

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0
Batch ID: P370689

Component	Result	Code	Units
Bromide	0.013	U	mg Br/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P370666

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P370535

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P370709

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P370565

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P370502

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2320 B-97
Batch ID: P371103

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 D-97
Batch ID: P370880

Component	Result	Code	Units
TSS	3	U	mg/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 4500 F-C-97
Batch ID: P371107

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P370679

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0
Batch ID: P370689

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Bromide	94.8		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P370666

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	99.1		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P370535

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	93.0		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P370709

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P370565

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	97.8		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P370502

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	98.0		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P371103

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	104		P	98 - 106

Reference Method: SM 4500 F-C-97
Batch ID: P371107

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	96.6		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P370679

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	104		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0
Batch ID: P370689

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2117129	Bromide	92.6	93.4	P/P	90 - 110
2117707	Bromide	92.5		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P370666

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2117519	Ammonia-N	101	102	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P370535

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2117597	Kjeldahl Nitrogen	102	103	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P370709

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2117531	NO2NO3-N	102	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P370565

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2117519	Total-P	96.4	93.6	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P370502

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2117968	O-Phosphate-P	96.0	97.7	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P371107

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2117596	Fluoride	93.6	96.2	P/P	89 - 106

Reference Method: SM 5310 B-00
Batch ID: P370679

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2117597	Organic Carbon	104	104	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P370494

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2117704	Turbidity	2.33	Sample	P	0 - 20

Reference Method: EPA 300.0
Batch ID: P370689

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2117129	Bromide	0.738	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P370666

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2117519	Ammonia-N	1.15	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P370535

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2117597	Kjeldahl Nitrogen	0.618	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P370709

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2117531	NO2NO3-N	0.370	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P370565

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2117519	Total-P	2.92	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P370502

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2117968	O-Phosphate-P	1.76	Spike	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P371103

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2117596	Total Alkalinity	0.228	Sample	P	0 - 2.8

Quality Assurance Report Precision

Reference Method: SM 2540 D-97
Batch ID: P370880

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2117535	TSS	9.52	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P371107

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2117596	Fluoride	1.95	Spike	P	0 - 2.5

Reference Method: SM 5310 B-00
Batch ID: P370679

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2117597	Organic Carbon	0.451	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0
Run ID: A94097
Included Lab Sample IDs: 2117596

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Bromide	99.1	101	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A94136
Included Lab Sample IDs: 2117596

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	101	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A94137
Included Lab Sample IDs: 2117598

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	98.9	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A94202
Included Lab Sample IDs: 2117597

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	98.4	101	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Run ID: A94212
Included Lab Sample IDs: 2117597

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	98.4	99.0	P/P	90 - 110

Reference Method: SM 5310 B-00
Run ID: A94219
Included Lab Sample IDs: 2117597

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	106	106	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A94233
Included Lab Sample IDs: 2117597

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO ₂ NO ₃ -N	99.5	99.0	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Run ID: A94240
Included Lab Sample IDs: 2117597

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	93.9	103	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 2320 B-97
Run ID: A94392
Included Lab Sample IDs: 2117596

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	104	101	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Run ID: A94392
Included Lab Sample IDs: 2117596

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.5	100	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery			Precision SMP	MS
			LCS				
EPA 180.1 Rev. 2.0	Turbidity					2.33	
EPA 300.0	Bromide	94.8		92.6	93.4		0.738
EPA 350.1 Rev. 2.0	Ammonia-N	99.1		101	102		1.15
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	93.0		102	103		0.618
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	100		102	101		0.370
EPA 365.1 Rev. 2.0	Total-P	97.8		96.4	93.6		2.92
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	98.0		96.0	97.7		1.76
SM 2320 B-97	Total Alkalinity	104				0.228	
SM 2540 D-97	TSS					9.52	
SM 4500 F-C-97	Fluoride	96.6		93.6	96.2		1.95
SM 5310 B-00	Organic Carbon	104		104	104		0.451

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	2117596
EPA 300.0 / E31780	Bromide in aqueous matrices	2117596
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2117597
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2117597
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2117597
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2117597
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2117598
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2117596
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2117596
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2117596
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2117597

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	09/11/2019			09/11/2019 16:13	Samantha Davila	2117596
EPA 300.0	09/11/2019			09/13/2019 14:03	Jhamieka S. Greenwood	2117596
EPA 350.1 Rev. 2.0	09/11/2019			09/13/2019 12:39	Ping Hua	2117597
EPA 351.2 Rev. 2.0	09/11/2019	09/12/2019 13:00	Sima Feizi	09/13/2019 16:54	Julia Storbeck	2117597
EPA 353.2 Rev. 2.0	09/11/2019			09/16/2019 10:58	Beverly Y. Sanders	2117597
EPA 365.1 Rev. 2.0	09/11/2019	09/12/2019 13:45	Lipi Saha	09/13/2019 14:32	Rosalyn Ballman	2117597
EPA 365.1 Rev. 2.0 dissolved	09/11/2019			09/11/2019 16:32	Jhamieka S. Greenwood	2117598
SM 2320 B-97	09/11/2019			09/19/2019 11:42	Dale L. Simmons	2117596
SM 2540 D-97	09/11/2019			09/13/2019 15:21	Yijie Li	2117596
SM 4500 F-C-97	09/11/2019			09/19/2019 08:32	Dale L. Simmons	2117596
SM 5310 B-00	09/11/2019			09/14/2019 06:06	Madeline Funaro	2117597