



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2019

Data Qualified?

Yes

No

10

WIN Station Name: Minnow Creek at Lovewood Rd.

Date: 2/19/19

(mm/dd/yyyy)

WIN/ Field ID: G3WA0008

WBID: 130

Latitude: 30.83682

(Decimal Degrees)

County: Jackson

ORG ID: 21FLWQA

Longitude: -85.52943

(Decimal Degrees)

Receiving Body of Water: Alligator Creek

RQ- 2018-11-05-17

Sampling Team Members				WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check	Sampling Equipment			
Blake Spencer									☒ Sample Container	☐ Van Dorn	☐ Pole	
Avril Wood-McGrath									☐ Carboy	☒ Syringe		
									☐ Dipper	☐ Other		
									Depth Measured with <u>Meter Stick</u>			
									Equipment ID			
									Collection Method			
									☒ Wading	☐ Canoe/Kayak		
									☐ From Shore	☐ Electric Motor		
									☐ Other	☐ Gasoline Motor		
Water Level: Dry/ Pooled/ Low <u>Normal</u> High / Flooded				Meter ID# Biology #2				Matrix: <u>Fresh</u> Marine / DI				
Photos: <u>Yes</u> No				Flow: No Flow / Intestinal / <u>Flowing</u> / NA				Tide: Rising/ Falling/ Slack/ <u>NA</u>				
Tide Times: High <u>NA</u> Low <u>NA</u>												
Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Sp COND (umhos/cm)	Temp. (C°)		
EST	1030	0.4	0.2	VOB	8.10	90.3	6.01	0.02	43	14.92		
CEN												
Biological Samples Collected: <u>None</u> HA SCI RPS Algae ID LVS LVI Other:												
SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#: <u>2063734</u>												
Parameter Suite	Parameter Methods				Preservation	Lot#	Expiration	pH Check				
Preserved Nutrients	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☒ Ice ☒ H ₂ SO ₄	<u>8105AB10022</u>	<u>4/11/19</u>	☒ <2				
Metals	☐ W-ICPMS ☐ W-ICP				☐ Ice ☐ HNO ₃			☐ <2				
Filtered Nutrients	☒ W-PO4-F ☒ Field Filtered 0.45 um Filter				☒ Ice	10827054						
Chlorophyll	☒ CHLSUITE-W				☒ Ice							
Physical/Aggregate (Fresh)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-TDS				☒ Ice							
Physical/Aggregate (Marine)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice							
Macroinvertebrates	☐ MI-FW-QLDC				☐ Formalin							
Periphyton	☐ ALGAE-ID				☐ Ice							
Microbiology	☒ E Coli ☐ F Coli ☐ Enterococci				☒ Ice							
Molecular	☐ PCR-FILTER ☐ PCR-BacR ☐ PCR-GULL2				☐ Ice							
	☐ PCR-HF183 ☐ PCR-DG3 ☐ PCR-GFD											
Tracers	☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice							
Pesticides	☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-E8321-DI				☐ Ice							
	☐ W-PCL-SQ-R ☐ WE8321-MS ☐ W-CARB-AA											
Phytoplankton	☐ DTY-QN-C ☐ PKD-QN-C				☐ Ice							
Notes:												
Signature: <u>[Signature]</u> Date: <u>2/19/19</u>												
LIMS EVENT ID: <u>NAS-SECT-2019-02-20-01</u> WIN Import ID: _____												
Enter Field Parameters, Verify Station ID In LIMS DMT: <u>802 2/20/19</u> QC Station ID, Field Parameters In LIMS DMT: _____												
Scan/Save Field Sheets _____ Migrate Data to WIN: _____ Verify Data In WIN: _____												

LIMS EVENT ID: NAS-SECT-2019-02-20-01

WIN Import ID: _____

Enter Field Parameters, Verify Station ID In LIMS DMT: 802 2/20/19

QC Station ID, Field Parameters In LIMS DMT: _____

Scan/Save Field Sheets _____ (Initials/Date)

Migrate Data to WIN: _____ (Initials/Date)

Verify Data In WIN: _____ (Initials/Date)

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Do not "X" this box if there is qualified data on this page.

Meter ID: Biology #2

RQ- 2018-11-05-17

Project: SMP

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or

Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 11/24/18

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Blake Spencer	2/19/19	0822	19.75	9.129	9.14	100.0	90.0	—	P/F	L/F
ICV	1	1	0824	17.69	9.147	9.14	99.9	90.0	—	P/F	L/F
CCV										P/F	L/F
CCV										P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Blake Spencer	2/19/19	0814	1805F81	05/2020	100	100	P/F	L/F
ICV	1	1	0815	1805F81	05/2020	100	100	P/F	L/F
CCV			1830	180221A	03/2019	500	480	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Blake Spencer	2/19/19	0759	180403C	10/2019	7.0	7.00	-6.4	P/F	L/F
Calibr.		2/19/19	0805	180621A	07/2019	4.0	4.00	164.7	P/F	L/F
Calibr.		2/19/19	0810	180621E	01/2020	10.0	10.00	-178.1	P/F	L/F
ICV		2/19/19	0812	180621F	01/2020	10.0	10.00	-172.9	P/F	L/F
CCV		2/19/19	1833	180621A	07/19	4.0	4.08	166.8	P/F	L/F
CCV									P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? Yes / No If yes, list date of last quarterly depth verification _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air	Blake Spencer	2/19/19	0825	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

DO Membrane fell off for last two sites

Form effective October 1, 2017

Amil wood-McGath

Relinquished By: <i>[Signature]</i>	Date/Time 2/20/19 0757	Shipping Method Hand Delivered	Number of Coolers Shipped: 5	Received By: <i>[Signature]</i>	Date/Time 2.20.19/7:57
--	---------------------------	-----------------------------------	---------------------------------	------------------------------------	---------------------------

Date/Time

Customer: WAS-SECT

Project ID: NWWQI

Requester: Curtis Musson

Field Report Prepared By:

Collected By:

Sampling Agency:

[illegible]

Relinquished By: <i>[Signature]</i>	Date/Time 2/20/0757	Shipping Method <i>Hand Delivered</i>	Number of Coolers Shipped: <i>5</i>	Received By: <i>SK</i>	Date/Time 2-20-19/7157
--	------------------------	--	--	---------------------------	---------------------------

Group: B	Bottle Type: W-SO4-IC W-TDS W-TSS	P-1L	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
Group: B	Bottle Type: CHLSUITE-W	BP-1L	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
Group: B	Bottle Type: ECOLI-18QT	P-250ML-WO-THI	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
Group: B	Bottle Type: PCR-BACR PCR-HF183	QPCR-P-500ML	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
Group: B	Bottle Type: W-CARB-AA	BG-500ML	# of Bottles: 1	Preservative: CH2CICOOH Pre-Preserved		Enter Number of Bottles Sent to Lab	1
Group: B	Bottle Type: W-CT-CI-R W-PCL-SQ-R	BG-500ML	# of Bottles: 4	Preservative:	ICE	Enter Number of Bottles Sent to Lab	4
Group: B	Bottle Type: W-E8321-DI W-E8321-MS	BG-500ML	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
Group: B	Bottle Type: W-ICP W-ICPMS	P-500ML	# of Bottles: 1	Preservative: HNO3 LOT# N68080070 EXP: 03/21/19	HNO3	Enter Number of Bottles Sent to Lab	1
Group: B	Bottle Type: W-NH3 W-NO2NO3 W-S-A-TP W-TKN W-TOC	P-500ML	# of Bottles: 1	Preservative: H2SO4 LOT# S48101020 EXP: 04/11/2019	ICE And H2SO4	Enter Number of Bottles Sent to Lab	1
Group: B	Bottle Type: W-PO4-F	P-125ML	# of Bottles: 1	Preservative:	FILTER-ICE	Enter Number of Bottles Sent to Lab	1
Group: B	Bottle Type: W-PSNP-TQ	BG-1L	# of Bottles: 4	Preservative:	ICE	Enter Number of Bottles Sent to Lab	4

Group: C	Bottle Type:	P-1L	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab
	ALKALINITY					
	TURBIDITY					
	W-CL-IC					
	W-COLOR					
	W-F					
	W-SO4-IC					
	W-TDS					
	W-TSS					
Group: C	Bottle Type:	BP-1L	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab
	CHLSUITE-W					
Group: C	Bottle Type:	P-250ML-WO-THI	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab
	ECOLI-18QT					
Group: C	Bottle Type:	P-500ML	# of Bottles: 1	Preservative:	ICE And H2SO4	Enter Number of Bottles Sent to Lab
	W-NH3					
	W-NO2NO3					
	W-S-A-TP					
	W-TKN					
	W-TOC					
Group: C	Bottle Type:	P-125ML	# of Bottles: 1	Preservative:	FILTER-ICE	Enter Number of Bottles Sent to Lab
	W-PO4-F					