



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2019

Data Qualified?

Yes / No

WIN Station Name: Coon Creek 100 meters upstream of CR 69

Date: 02/27/2019

WIN/ Field ID: G2WA0005

WBID: 885

Latitude: 30.39283938

(mm/dd/yyyy)

County: Calhoun

ORG ID: 21FLWQA

Longitude: -85.06315982

(Decimal Degrees)

Receiving Body of Water: Apalachicola River

RQ- 2019-02-25-48

(Decimal Degrees)

Sampling Team Members				WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check	Sampling Equipment			
Blake Spencer									☒ Sample Container	☐ Van Dorn	☐ Pole	
Avril Wood-McGrath									☐ Carboy	☐ Syringe		
Steve Cafone									☐ Dipper	☐ Other		
									Depth Measured with Equipment ID			
									Collection Method			
									☒ Wading	☐ Canoe/Kayak		
									☐ From Shore	☐ Electric Motor		
									☐ Other	☐ Gasoline Motor		
Water Level: Dry/ Pooled/ Low / <u>Normal</u> / High / Flooded				Meter ID# <u>Biology #2</u>				Matrix: <u>Fresh</u> / Marine / DI				
Photos: Yes / <u>No</u>				Flow: No Flow / Intestinal / Flowing / NA				Tide: Rising/ Falling/ Slack/ <u>NA</u>				
Tide Times: High <u>NA</u> Low <u>NA</u>												
Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Sp COND (µmhos/cm)	Temp. (C°)		
EST	10:55		0.15									
CEN		0.3										
Biological Samples Collected: <u>None</u> HA SCI RPS Algae ID LVS LVI Other:												
SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#:												
Parameter Suite	Parameter Methods					Preservation	Lot#	Expiration	pH Check			
Preserved Nutrients	☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP					☐ Ice ☐ H2SO4			☐ <2			
Metals	☐ W-ICPMS ☐ W-ICP					☐ Ice ☐ HNO3			☐ <2			
Filtered Nutrients	☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter					☐ Ice	18027054					
Chlorophyll	☐ CHLSUTE-W					☐ Ice						
Physical/Aggregate (Fresh)	☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CI-IC / W-COLOR / W-SO4-IC / W-TDS					☐ Ice						
Physical/Aggregate (Marine)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY					☐ Ice						
Macroinvertebrates	☐ MI-FW-QLDC					☐ Formalin						
Periphyton	☐ ALGAE-ID					☐ Ice						
Microbiology	☒ E Coli ☐ F Coli ☐ Enterococci					☒ Ice						
Molecular	☐ PCR-FILTER ☐ PCR-BacR ☐ PCR-GULL2					☐ Ice						
	☐ PCR-HF183 ☐ PCR-DG3 ☐ PCR-GFD					☐ Ice						
Tracers	☐ W-E8321-DI ☐ W-E8321-MS					☐ Ice						
Pesticides	☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-E8321-DI					☐ Ice						
	☐ W-PCL-SQ-R ☐ WE8321-MS ☐ W-CARB-AA					☐ Ice						
Phytoplankton	☐ DTY-QN-C ☐ PKD-QN-C					☐ Ice						
Notes: E. Coli						WIN ID / FILED ID → G2WA0005 Location: Coon Creek 100 meters upstream of CR 69						
Signature:						Date: <u>2/27/19</u>						

LIMS EVENT ID: WAS-SECT-2019-02-27-01

WIN Import ID:

Enter Field Parameters, Verify Station ID in LIMS DMT: 3/14/19

QC Station ID, Field Parameters in LIMS DMT:

Scan/Save Field Sheets

Migrate Data to WIN:

Verify Data in WIN:

(Initials/Date)

(Initials/Date)

(Initials/Date)

(Initials/Date)

Florida Department of Environmental Protection Central Laboratory Submittal Form

Customer: WAS-SECT

Project ID: SMP

Requester: Curtis Musson

Field Report Prepared By: Avril Wood-McGrathCollected By: Blake Spencer, Avril Wood-McGrath, Steve ColoneSampling Agency: FDLP

Location		WIN ID / FILED ID → 		☐ Comp	Collection (comp begin or grab)	☒ Grab	Date <u>2/27/19</u>	Time <u>11:31</u>	☒ Eastern	Composite end	Bottle Group(s)
Field ID		G2WA0012									
WIN/STORET #	Location	Stone Mill Creek Upstream of Dianna St.		Matrix (type, e.g., Salt, Fresh, etc)	Diss. Oxygen	☐ mg/L	Total Res. Chlorine				
				Fresh	0.3	☒ % sat	(mg/L)				
Latitude	☒ dd	Longitude	☒ dms	Temp (C)	pH	Sample Depth	☐ ft	Sp. Conductance			
30.1578847		-85.2305114		17.99	5.4	0.3	☒ m	43			
				Salinity (ppt)	Comments						
				0.02	Dissolved Oxygen is 0.3%; W-CR-U-R not collected						
Location		WIN ID / FILED ID → 		☐ Comp	Collection (comp begin or grab)	☒ Grab	Date <u>2/27/19</u>	Time <u>12:15</u>	☒ Eastern	Composite end	Bottle Group(s)
Field ID		G2WA0010									
WIN/STORET #	Location	Big Branch Downstream of Road 5		Matrix (type, e.g., Salt, Fresh, etc)	Diss. Oxygen	☐ mg/L	Total Res. Chlorine				
				Fresh	10.6	☒ % sat	(mg/L)				
Latitude	☒ dd	Longitude	☒ dms	Temp (C)	pH	Sample Depth	☐ ft	Sp. Conductance			
30.15825339		-85.25200674		17.35	4.92	0.15	☒ m	28			
				Salinity (ppt)	Comments						
				0.01							
Location		WIN ID / FILED ID → 		☐ Comp	Collection (comp begin or grab)	☒ Grab	Date <u>2/27/19</u>	Time <u>10:55</u>	☒ Eastern	Composite end	Bottle Group(s)
Field ID		G2WA0005									
WIN/STORET #	Location	Coon Creek 100 meters upstream of CR 69		Matrix (type, e.g., Salt, Fresh, etc)	Diss. Oxygen	☐ mg/L	Total Res. Chlorine				
				Fresh	—	☒ % sat	(mg/L)				
Latitude	☒ dd	Longitude	☒ dms	Temp (C)	pH	Sample Depth	☐ ft	Sp. Conductance			
30.39283938		-85.06315982		—	—	0.2	☒ m	—			
				Salinity (ppt)	Comments						
				—							
Location		WIN ID / FILED ID → 		☐ Comp	Collection (comp begin or grab)	☒ Grab	Date <u>2/27/19</u>	Time <u>10:10</u>	☒ Eastern	Composite end	Bottle Group(s)
Field ID		G2WA0020									
WIN/STORET #	Location	Wilson Mill Creek		Matrix (type, e.g., Salt, Fresh, etc)	Diss. Oxygen	☐ mg/L	Total Res. Chlorine				
				Fresh	08.7	☒ % sat	(mg/L)				
Latitude	☒ dd	Longitude	☒ dms	Temp (C)	pH	Sample Depth	☐ ft	Sp. Conductance			
30.5777		-85.591		17.52	7.36	0.05	☒ m	77			
				Salinity (ppt)	Comments						
				0.04	W-CR-U-R not collected						

Relinquished By: <u>CORONC</u>	Date/Time: <u>2/27/19 1510</u>	Shipping Method: <u>HAND</u>	Number of Coolers Shipped: <u>2</u>	Received By: <u>SR</u>	Date/Time: <u>2.27.19/15:18</u>
--------------------------------	--------------------------------	------------------------------	-------------------------------------	------------------------	---------------------------------

Group: A	Bottle Type: P-1L	# of Bottles: 2	Preservative: ICE	Enter Number of Bottles Sent to Lab	2
ALKALINITY TURBIDITY W-CL-IC W-COLOR W-F W-SO4-IC W-TDS W-TSS					
Group: A	Bottle Type: BP-1L	# of Bottles: 2	Preservative: ICE	Enter Number of Bottles Sent to Lab	2
CHL SUITE W					
Group: A	Bottle Type: P-250ML-WO-THI	# of Bottles: 2	Preservative: ICE	Enter Number of Bottles Sent to Lab	2
ECOLI-18QT					
Group: A	Bottle Type: BG-500ML	# of Bottles: 6	Preservative: ICE	Enter Number of Bottles Sent to Lab	6
W-CT-CI-R					
Group: A	Bottle Type: P-500ML	# of Bottles: 2	Preservative: HNO3	Enter Number of Bottles Sent to Lab	2
W-ICP W-ICPMS					
			HNO3 LOT# NA8880070 EXP: 03/21/19		
Group: A	Bottle Type: P-500ML	# of Bottles: 2	Preservative: ICE And H2SO4	Enter Number of Bottles Sent to Lab	2
W-NH3 W-NO2NO3 W-S-A-TP W-TKN W-TOC					
			H2SO4 LOT# SA8274058 EXP: 10/01/2019		
Group: A	Bottle Type: P-125ML	# of Bottles: 2	Preservative: FILTER-ICE	Enter Number of Bottles Sent to Lab	2
W-PO4-F					
Group: B	Bottle Type: P-1L	# of Bottles: 1	Preservative: ICE	Enter Number of Bottles Sent to Lab	1
ALKALINITY TURBIDITY W-CL-IC W-COLOR W-F W-SO4-IC					

AWM 2/27/19

Group: B	Bottle Type: P-1L	# of Bottles: 1	Preservative: ICE	Enter Number of Bottles Sent to Lab
W-TDS				
W-TSS				
Group: B	Bottle Type: BP-1L	# of Bottles: 1	Preservative: ICE	Enter Number of Bottles Sent to Lab
CHLSUITE-W				
Group: B	Bottle Type: P-250ML-WO-THI	# of Bottles: 1	Preservative: ICE	Enter Number of Bottles Sent to Lab
ECOLI-18QT				
Group: B	Bottle Type: P-500ML	# of Bottles: 1	Preservative: ICE And H2SO4	Enter Number of Bottles Sent to Lab
W-NH3				
W-NO2NO3				
W-S-A-TP				
W-TKN				
W-TOC				
H2SO4 LOT# SA8274050 EXP: 10/01/2019				
Group: B	Bottle Type: P-125ML	# of Bottles: 1	Preservative: FILTER-ICE	Enter Number of Bottles Sent to Lab
W-PO4-F				
Group: C	Bottle Type: P-250ML-WO-THI	# of Bottles: 1	Preservative: ICE	Enter Number of Bottles Sent to Lab
ECOLI-18QT				

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page

Meter ID: Bio/ogv #2

RQ: 2019-02-25-48

Project: SMP

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Blake Spencer	2/27/19	0735	20.12	9.074	9.07	132.0	67.6	—	P/F	L/F
ICV	Blake Spencer	2/27/19	0737	20.13	9.074	9.06	99.9	66.5	—	P/F	L/F
CCV	Blake Spencer	2/27/19	1442	19.31	9.220	9.41	102.0	59.4	—	P/F	L/F
CCV										P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Blake Spencer	2/27/19	0742	18022A	03/2019	500	500	P/F	L/F
ICV	Blake Spencer	2/27/19	0746	19022A	03/2019	500	500	P/F	L/F
CCV	Blake Spencer	2/27/19	1448	1805F81	05/2020	100	105	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Blake Spencer	2/27/19	0754	180403C	10/2019	7.0	7.00	-9.3	P/F	L/F
Calibr.		2/27/19	0800	180621D	7/2019	4.0	4.00	164.3	P/F	L/F
Calibr.		2/27/19	0804	180621F	01/2020	10.0	9.99	-13.4	P/F	L/F
ICV		2/27/19	0805	180621F	01/2020	10.0	9.99	-13.5	P/F	L/F
CCV	Blake Spencer	2/27/19	1454	180621D	07/2019	4.0	4.12	156.7	P/F	L/F
CCV									P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? Yes / No If yes, list date of last quarterly depth verification

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air	Blake Spencer	2/27/19	0806	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: