

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: 451 #3

RQ: 2019-09-16-14

Project: BMAP 1

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 6/19

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	C Ruben	9-23-19	0820	23.8	760	8.5	8.42	99.5%	—	—	(P) F	(L) F
ICV	↓	↓	0823	23.8	760	8.5	8.40	99.8%	—	—	(P) F	(L) F
CCV	AKalmbacher	9-23-19	1320	23.5	760	8.5	8.52	100.3	—	—	(P) F	(L) F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	C Ruben	9-23-19	0828	19071B	7/20	1000	988	(P) F	(L) F
ICV	↓	↓	0830	190711A	7/20	100	105	(P) F	(L) F
CCV	AKalmbacher	9-23-19	1325	190711B	7/2020	1,000	1,003	(P) F	(L) F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	C Ruben	9-23-19	0832	2902178	1/21	7.	24.6	6.97	19.3	(P) F	(L) F
Calibr.	↓	↓	0838	2904800	3/21	4.	24.7	3.99	146.4	(P) F	(L) F
Calibr.	↓	↓	0842	2810F55	3/20	10.	24.8	10.04	195.5	(P) F	(L) F
ICV	↓	↓	0845	2902778	1/21	7.0	24.4	7.03	24.6	(P) F	(L) F
CCV	AKalmbacher	9-23-19	0830	2810F55	3/2020	10.0	24.3	10.11	199.2	(P) F	(L) F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 175.7, slope from 4 to 7 176.2 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

Data Qualified?

Yes / No

January 2019

WIN Station Name: GOODBYS CR ABOVE WEST BRANCH AT SAN CLERC Date: 9/23/19WIN/Field ID: 20030891 WBID: 2326 A Latitude: 30.22544County: Duval ORG ID: 21FLA Longitude: -81.603861Receiving Body of Water: _____ RQ: 2019-09-16-14

Sampling Team Members					WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
C Ruben					X	X			
A Kalmbacher						X	X	X	

Sampling Equipment		
☒ Sample Container	☐ Van Dorn	☒ Pole
☐ Carboy	☐ Syringe	
☐ Dipper	☐ Other	
Depth Measured with <u>meter</u>		
Equipment ID <u>Ortho #44</u>		

Collection Method		
☐ Wading	☐ Canoe/Kayak	
☒ From Shore	☐ Electric Motor	
☐ Other	☐ Gasoline Motor	

Water Level: Dry / Pooled / Low / Normal / High / FloodedPhotos: Yes / No Flow: No Flow / Intertidal / Flowing / NA

Tide: Rising / Falling / Slack / NA

Tide Times: High Low

Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Sp COND (µmhos/cm)	Temp. (C°)
EST	1210	0.5	0.3	0.5	5.11	60%	7.35	0.18	377	23.9
CEN				5						

Biological Samples Collected: None HA SCI RPS Algae ID LVS LVI Other:

SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#:

Parameter Suite	Parameter Methods	Preservation	Lot#	Expiration	pH Check
Preserved Nutrients	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☒ Ice ☒ H2SO4	SA9010030	4/20	☒ <2
Metals	☐ W-ICPMS ☐ W-ICP	☐ Ice ☐ HNO3			☐ <2
Filtered Nutrients	☒ W-PO4-F ☐ Field Filtered 0.45 µm Filter	☒ Ice	R9BA86875		
Chlorophyll	☒ CHLSUITE-W	☒ Ice			
Physical/Aggregate (Fresh)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-TDS	☒ Ice			
Physical/Aggregate (Marine)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates	☐ MI-FW-QLDC	☐ Formalin			
Periphyton	☐ ALGAE-ID	☐ Ice			
Microbiology	☒ E Coli ☐ F Coli ☐ Enterococci	☒ Ice			
Molecular	☐ PCR-FILTER ☐ PCR-BacR ☐ PCR-GULL2 ☒ PCR-HF183 ☐ PCR-DG3 ☐ PCR-GFD	☒ Ice			
Tracers	☒ W-E8321-DI ☒ W-E8321-MS	☒ Ice			
Pesticides	☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-E8321-DI ☐ W-PCL-SQ-R ☐ WE8321-MS ☐ W-CARB-AA	☐ Ice			
Phytoplankton	☐ DTY-QN-C ☐ PKD-QN-C	☐ Ice			

Notes: _____

WIN ID/Field ID: 20030891

GOODBYS CR ABOVE WEST BRANCH
AT SAN CLERC

Signature: [Signature] Date: 9/23/19

LIMS EVENT ID: _____ WIN Import ID: _____

Enter Field Parameters, Verify Station ID in LIMS DMT: _____ QC Station ID, Field Parameters in LIMS DMT: _____

Scan/Save Field Sheets _____ Migrate Data to WIN: _____ Verify Data in WIN: _____

(Initials/Date)

(Initials/Date)

(Initials/Date)

Bmap 1

Central Laboratory Submittal Form

last revised Jan 25, 2018

Customer: NE-DIST

Requester: Jeremy M. ParrishField Report Prepared By: C. Ruben

Project ID: SMP

Collected By: C. Ruben / A. KalmbacherSampling Agency: FDEP DEAR NE ROC

Location		☐ Comp ☒ Grab		Collection (comp begin or grab) Date <u>9.23.19</u> Time <u>1210</u>		☒ Eastern ☐ Central		Composite end Date _____ Time _____		Bottle Group(s) (See pg 2)	
Field ID		WIN ID/Field ID:  20030891		Matrix (type, e.g., Salt, Fresh, etc) <u>Fresh</u>		Diss. Oxygen <u>5.11</u>		☒ mg/L ☐ % sat		Total Res. Chlorine (mg/L) _____	
WIN/S		GOODBYS CR ABOVE WEST BRANCH AT SAN CLERC		Temp (C) <u>23.9</u>		pH <u>7.35</u>		Sample Depth <u>0.3</u>		☐ ft ☒ m	
Latitude		☐ dd ☐ dms		Salinity (ppt) <u>0.18</u>		Comments		Sp. Conductance (umho/cm) <u>377</u>		<u>B</u>	
Location		☐ Comp ☐ Grab		Collection (comp begin or grab) Date _____ Time _____		☐ Eastern ☐ Central		Composite end Date _____ Time _____		Bottle Group(s)	
Field ID				Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen		☐ mg/L ☐ % sat		Total Res. Chlorine (mg/L)	
WIN/STORET #				Temp (C)		pH		Sample Depth		☐ ft ☐ m	
Latitude		☐ dd ☐ dms		Longitude		☐ dd ☐ dms		Salinity (ppt)		Comments	
Location		☐ Comp ☐ Grab		Collection (comp begin or grab) Date _____ Time _____		☐ Eastern ☐ Central		Composite end Date _____ Time _____		Bottle Group(s)	
Field ID				Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen		☐ mg/L ☐ % sat		Total Res. Chlorine (mg/L)	
WIN/STORET #				Temp (C)		pH		Sample Depth		☐ ft ☐ m	
Latitude		☐ dd ☐ dms		Longitude		☐ dd ☐ dms		Salinity (ppt)		Comments	
Location		☐ Comp ☐ Grab		Collection (comp begin or grab) Date _____ Time _____		☐ Eastern ☐ Central		Composite end Date _____ Time _____		Bottle Group(s)	
Field ID				Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen		☐ mg/L ☐ % sat		Total Res. Chlorine (mg/L)	
WIN/STORET #				Temp (C)		pH		Sample Depth		☐ ft ☐ m	
Latitude		☐ dd ☐ dms		Longitude		☐ dd ☐ dms		Salinity (ppt)		Comments	

Relinquished By: <u>C. Ruben</u>	Date/Time: <u>9/23/19 1500</u>	Shipping Method: <u>Fedex</u>	Number of Coolers Shipped: <u>1</u>	Received By: _____	Date/Time: _____
----------------------------------	--------------------------------	-------------------------------	-------------------------------------	--------------------	------------------

Group: A	Bottle Type:	P-1L	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	_____
	ALKALINITY TURBIDITY W-BR-IC W-F W-TSS						
Group: A	Bottle Type:	BP-1L	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	_____
	CHLSUITE-W						
Group: A	Bottle Type:	P-250ML-WI-THI	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	_____
	NE-ALS-ENQ						
Group: A	Bottle Type:	QPCR-P-500ML	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	_____
	PCR-HF183						
Group: A	Bottle Type:	BG-500ML	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	_____
	W-E8321-DI W-E8321-MS						
Group: A	Bottle Type:	P-500ML	# of Bottles: 1	Preservative:	ICE And H2SO4	Enter Number of Bottles Sent to Lab	_____
	W-NH3 W-NO2NO3 W-S-A-TP W-TKN W-TOC						
Group: A	Bottle Type:	P-125ML	# of Bottles: 1	Preservative:	FILTER-ICE	Enter Number of Bottles Sent to Lab	_____
	W-PO4-F						
Group: B	Bottle Type:	P-1L	# of Bottles: 3	Preservative:	ICE	Enter Number of Bottles Sent to Lab	_____ 1
	ALKALINITY TURBIDITY W-CL-IC W-COLOR W-F W-SO4-IC W-TDS W-TSS						
Group: B	Bottle Type:	BP-1L	# of Bottles: 3	Preservative:	ICE	Enter Number of Bottles Sent to Lab	_____ 1

Group: B	Bottle Type: BP-1L	# of Bottles: 3	Preservative: ICE	Enter Number of Bottles Sent to Lab	1
CHLSUITE-W					
Group: B	Bottle Type: P-250ML-WI-THI	# of Bottles: 3	Preservative: ICE	Enter Number of Bottles Sent to Lab	1 to ALE
NE-ALS-ECQ					
Group: B	Bottle Type: QPCR-P-500ML	# of Bottles: 3	Preservative: ICE	Enter Number of Bottles Sent to Lab	1
PCR-HF183					
Group: B	Bottle Type: BG-500ML	# of Bottles: 3	Preservative: ICE	Enter Number of Bottles Sent to Lab	1
W-E8321-DI W-E8321-MS					
Group: B	Bottle Type: P-500ML	# of Bottles: 3	Preservative: ICE And H2SO4	Enter Number of Bottles Sent to Lab	1
W-NH3 W-NO2NO3 W-S-A-TP W-TKN W-TOC					
Group: B	Bottle Type: P-125ML	# of Bottles: 3	Preservative: FILTER-ICE	Enter Number of Bottles Sent to Lab	1
W-PO4-F					
Group: C	Bottle Type: P-1L	# of Bottles: 1	Preservative: ICE	Enter Number of Bottles Sent to Lab	1
ALKALINITY TURBIDITY W-CL-IC W-COLOR W-F W-SO4-IC W-TDS W-TSS					
Group: C	Bottle Type: BP-1L	# of Bottles: 1	Preservative: ICE	Enter Number of Bottles Sent to Lab	1
CHLSUITE-W					
Group: C	Bottle Type: P-250ML-WI-THI	# of Bottles: 1	Preservative: ICE	Enter Number of Bottles Sent to Lab	1
NE-ALS-ECQ					
Group: C	Bottle Type: P-500ML	# of Bottles: 1	Preservative: HNO3	Enter Number of Bottles Sent to Lab	1
W-ICP					

