

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID:

YSI 60055

RQ:

2019-06-03-26

Project:

SUN-NORTH

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	J Parnish	7/9/19	715	23.3	755	8.5	8.5	99.3	✓	—	P/F	L/F
ICV	J Parnish	7/9/19	716	23.4	755	8.5	8.47	99.4	—	—	P/F	L/F
CCV	B Davis	7/9/19	1355	23.9	754	8.3	8.34	98.9	✓	✓	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	J Parnish	7/9/19	717	190430A	5/20	1000	1000	P/F	L/F
ICV	J Parnish	7/9/19	718	142430	12/19	147	153	P/F	L/F
CCV	B Davis	7/9/19	1357	190430A	5/20	1000	1000	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J Parnish	7/9/19	721	2904C54	3/21	7.	24.6	7.00	62	P/F	L/F
Calibr.	J Parnish	7/9/19	722	2904B00	3/21	4.	24.7	4.01	116	P/F	L/F
Calibr.	J Parnish	7/9/19	723	2901B51	7/20	10.	24.5	10.01	234	P/F	L/F
ICV	J Parnish	7/9/19	725	2904B07	3/21	4.0	24.6	4.03	114	P/F	L/F
CCV	B Davis	7/9/19	1359	2901B51	7/20	10.0	24.1	10.05	-237	P/F	L/F
CCV			1400	2904B00	3/21	4.0	24.4	4.06	112	P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2019

Data Qualified?

Yes / No

WIN Station Name: CANNON CR @ CANNON CR RD/BUSINESS PARK POINT W CR 47

Date: 7/9/19
(mm/dd/yyyy)

WIN/ Field ID: 21030138

WBID: 3520

Latitude: 30.12705

(Decimal Degrees)

County: Columbia

ORG ID: 21FLA

Longitude: -82.65485

(Decimal Degrees)

Receiving Body of Water: Canon creek sink

RQ-2019-06-03-26

Sampling Team Members					WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
B Davis									
J Parvish									

Sampling Equipment		
☒ Sample Container	☐ Van Dorn	☒ Pole
☐ Carboy	☐ Syringe	
☐ Dipper	☐ Other	
Depth Measured with ProPSS		
Equipment ID 0450 #3		

Collection Method		
☐ Wading	☐ Canoe/Kayak	
☒ From Shore	☐ Electric Motor	
☐ Other	☐ Gasoline Motor	

Water Level: Dry/ Pooled/ Low/ Normal/ High/ Flooded

Meter ID# ProPSS

Matrix: Fresh / Marine / DI

Photos: Yes / No

Flow: No Flow / Intertidal / (Flowing) / NA

Tide: Rising/ Falling/ Slack/NA

Tide Times: High Low

Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Sp COND (µmhos/cm)	Temp. (C°)
EST	1055	0.3	0.2	VOB	6.40	78.7	7.15	0.07	143.3	25.8
CEN				(LS)						

Biological Samples Collected: None HA SCI RPS Algae ID LVS LVI Other:

SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#:

Parameter Suite	Parameter Methods	Preservation	Lot#	Expiration	pH Check
Preserved Nutrients	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-P	☒ Ice ☒ H2SO4	5A9129100	5/9/20	☒ <2
Metals	☒ W-ICPMS ☐ W-ICP	☒ Ice ☒ HNO3	NA90900	4/9/20	☒ <2
Filtered Nutrients	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter	☒ Ice	R7HA29007		
Chlorophyll	☒ CHLSUITE-W	☒ Ice			
Physical/Aggregate (Fresh)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS	☒ Ice			
Physical/Aggregate (Marine)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates	☐ MI-FW-QLDC	☐ Formalin			
Periphyton	☐ ALGAE-ID	☐ Ice			
Microbiology	☒ E Coli ☐ F Coli ☐ Enterococci	☒ Ice			
Molecular	☒ PCR-FILTER ☐ PCR-BacR ☐ PCR-GULL2 ☒ PCR-HF183 ☐ PCR-DG3 ☐ PCR-GFD	☒ Ice			
Tracers	☒ W-E8321-DI ☒ W-E8321-MS	☐ Ice			
Pesticides	☒ W-PSNP-TQ ☒ W-CT-CL-R ☒ W-E8321-DI ☒ W-PCL-SQ-R ☒ W-E8321-MS ☒ W-CAR8-AA	☒ Ice			
Phytoplankton	☐ DTY-QN-C ☐ PKD-QN-C	☐ Ice			

Notes:

WIN / Field ID:



21030138

CANNON CR @ CANNON CR RD/
BUSINESS PARK POINT W CR 47

Signature: B. Davis

Date: 7/9/19

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID in LIMS DMT: QC Station ID, Field Parameters in LIMS DMT:

Scan/Save Field Sheets Migrate Data to WIN: Verify Data in WIN:

(Initials/Date)

(Initials/Date)

(Initials/Date)

(Initials/Date)



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2019

Data Qualified?

Yes / No

WIN Station Name: LITTLE CREEK AT US 441

Date: 7/9/19
(mm/dd/yyyy)

WIN/ Field ID: 21010033

WBID: 3368

Latitude: 30.41576
(Decimal Degrees)

County: Columbia

ORG ID: 21FLA

Longitude: -82.6365
(Decimal Degrees)

Receiving Body of Water: Suwannee River

RQ-2019- 06-03-26

Sampling Team Members					WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
<u>B. Davis</u> <u>J. Peltier</u>									

Sampling Equipment		
☐ Sample Container	☐ Van Dorn	☐ Pole
☐ Carboy	☐ Syringe	
☐ Dipper	☐ Other	
Depth Measured with <u>Pro Bss</u>		
Equipment ID <u>0-4043</u>		

Collection Method		
☐ Wading	☐ Canoe/Kayak	
☒ From Shore	☐ Electric Motor	
☐ Other	☐ Gasoline Motor	

Water Level: Dry / Pooled / Low / Normal / High / Flooded

Meter ID# Pro Bss

Matrix: Fresh / Marine / DI

Photos: Yes / No

Flow: No Flow / Intertidal / Flowing / NA

Tide: Rising / Falling / Slack / NA

Tide Times: High — Low —

Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Sp COND (µmhos/cm)	Temp. (C°)
EST	<u>1150</u>	<u>0.2</u>	<u>0.1</u>	<u>VOB</u>	<u>5.88</u>	<u>71.6</u>	<u>6.04</u>	<u>0.04</u>	<u>81.5</u>	<u>25.4</u>
CEN				<u>(S)</u>						

Biological Samples Collected: None HA SCI RPS Algae ID LVS LVI Other:

SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#:

Parameter Suite	Parameter Methods	Preservation	Lot#	Expiration	pH Check
Preserved Nutrients	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☒ Ice ☐ H2SO4	<u>SA9129100</u>	<u>5/9/20</u>	☒ <2
Metals	☐ W-JCPMS ☐ W-ICP	☐ Ice ☐ HNO3			☒ <2
Filtered Nutrients	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter	☒ Ice	<u>R7HA29007</u>		
Chlorophyll	☒ CHL SUITE-W	☒ Ice			
Physical/Aggregate (Fresh)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS	☐ Ice			
Physical/Aggregate (Marine)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates	☐ MI-FW-QLDC	☐ Formalin			
Periphyton	☐ ALGAE-ID	☐ Ice			
Microbiology	☒ E Coli ☐ F Coli ☐ Enterococci	☒ Ice			
Molecular	☒ PCR-FILTER ☐ PCR-BacR ☐ PCR-GULL2 ☒ PCR-HF183 ☐ PCR-DG3 ☐ PCR-GFD	☒ Ice			
Tracers	☒ W-E8321-DI ☐ W-E8321-MS	☒ Ice			
Pesticides	☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-E8321-DI ☐ W-PCL-SQ-R ☐ WE8321-MS ☐ W-CARB-AA	☐ Ice			
Phytoplankton	☐ DTY-QN-C ☐ PKD-QN-C				

Notes:

WIN / Field ID:



21010033

LITTLE CREEK AT US 441

Signature: B. Davis

Date: 7/9/19

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID in LIMS DMT: QC Station ID, Field Parameters in LIMS DMT:

Scan/Save Field Sheets Migrate Data to WIN: Verify Data in WIN:

(Initials/Date)

(Initials/Date)

(Initials/Date)



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2019

Data Qualified?

Yes / No

WIN Station Name: PEACOCK LAKE CENTER

Date: 7/9/19
(mm/dd/yyyy)

WIN/ Field ID: 21020097

WBID: 3438A

Latitude: 30.2357

(Decimal Degrees)

County: Suwannee

ORG ID: 21FLA

Longitude: -82.8983

(Decimal Degrees)

Receiving Body of Water: Rocky Creek

RQ-2019-06-03-26

Sampling Team Members					WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
B Davis J Parish					☒	☒	☒	☒	☒

Sampling Equipment									
☐ Sample Container	☐ Van Dorn	☐ Pole							
☐ Carboy	☐ Syringe								
☐ Dipper	☐ Other								
Depth Measured with <u>Pico DSS</u>									
Equipment ID <u>Outro #3</u>									

Collection Method									
☐ Wading	☒ Canoe/Kayak								
☐ From Shore	☐ Electric Motor								
☐ Other	☐ Gasoline Motor								

Water Level: Dry/ Pooled/ Low <u>Normal</u> / High / Flooded					Meter ID# <u>Pico DSS</u>					Matrix: <u>Fresh</u> Marine / DI	
Photos: Yes <u>No</u>		Flow: No Flow / Intestinal / Flowing / <u>NA</u>			Tide: Rising/ Falling/ Slack/ <u>NA</u>			Tide Times: High — Low <u>✓</u>			

Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Sp COND (µmhos/cm)	Temp. (C°)
EST	955	1.6	0.3	0.9	7.64	102.2	6.92	0.02	55.7	30.7
CEN	957		1.0		5.64	72.8	6.39			

Biological Samples Collected: <u>None</u> HA SCI RPS Algae ID LVS LVI Other:									
SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#:									

Parameter Suite	Parameter Methods	Preservation	Lot#	Expiration	pH Check
Preserved Nutrients	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☒ Ice ☐ H2SO4	SA9129100	5/9/20	☒ <2
Metals	☐ W-ICPMS ☐ W-ICP	☐ Ice ☐ HNO3			☐ <2
Filtered Nutrients	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter	☒ Ice	R7HA29007		
Chlorophyll	☒ CHLSUITE-W	☒ Ice			
Physical/Aggregate (Fresh)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CI-IC / W-COLOR / W-SO4-IC / W-TDS	☒ Ice			
Physical/Aggregate (Marine)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates	☐ MI-FW-QLDC	☐ Formalin			
Periphyton	☐ ALGAE-ID	☐ Ice			
Microbiology	☐ E Coli ☐ F Coli ☐ Enterococci	☐ Ice			
Molecular	☐ PCR-FILTER ☐ PCR-BacR ☐ PCR-GULL2 ☐ PCR-HF183 ☐ PCR-DG3 ☐ PCR-GFD	☐ Ice			
Tracers	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides	☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-E8321-DI ☐ W-PCL-SQ-R ☐ WE8321-MS ☐ W-CARS-AA	☐ Ice			
Phytoplankton	☐ DTY-QN-C ☐ PKD-QN-C				

Notes:

seasweet@yahoo.com for info

WIN / Field ID: 21020097

PEACOCK LAKE CENTER

Signature: B. Davis

Date: 7/9/19

Contains: 1:1 H2SO4
Lot No.: SA9129100
Expires: 05/09/20
See Safety Data Sheet (SDS)

Place Preserved

SULFURIC ACID <51%
DANGER
Corrosive
1:1 H2SO4
Contains: 1:1 H2SO4
Do not breathe vapors or dusts. Avoid contact with skin and eyes. If contact occurs, flush immediately with water for at least 15 minutes. Do not get inside clothing. If swallowed, do not induce vomiting. Call a physician for treatment. Do not eat, drink, or smoke while using this product. Do not use near open flames or heat. Do not mix with other chemicals. Do not dispose of in drains. Store in a cool, dry place. See Safety Data Sheet (SDS) for more information.

LIMS EVENT ID: _____ WIN Import ID: _____

Enter Field Parameters, Verify Station ID in LIMS DMT: _____ QC Station ID, Field Parameters in LIMS DMT: _____

Scan/Save Field Sheets _____ Migrate Data to WIN: _____ Verify Data in WIN: _____

(Initials/Date) (Initials/Date) (Initials/Date)

Suwannee North

Central Laboratory Submittal Form

last revised Jan 25, 2018

Customer: NE-DIST

Requester: Jeremy M. Parrish

Field Report Prepared By:

Project ID: SMP

Collected By:

Sampling Agency:

Location	WIN / Field ID:  21020097	☐ Comp ☒ Grab	Collection (comp begin or grab) Date 7/9/19 Time 955	Eastern Central	Composite end Date Time	Bottle Group(s) (See pg 2)
Field ID	PEACOCK LAKE CENTER	Matrix (type, e.g., Salt, Fresh, etc) FRESH		Diss. Oxygen 7.6	☒ mg/L ☐ % sat	Total Res. Chlorine (mg/L)
WIN/STORE		Temp (C) 30.7	pH 6.92	Sample Depth 0.3	☐ ft ☒ m	Sp. Conductance (umho/cm) 56
Latitude	☐ dd ☐ dms	Longitude	☐ dd ☐ dms	Salinity (ppt) 0.02	Comments	
Location	WIN / Field ID:  21030138	☐ Comp ☒ Grab	Collection (comp begin or grab) Date 7/9/19 Time 1055	Eastern Central	Composite end Date Time	Bottle Group(s)
Field ID	CANNON CR @ CANNON CR RD/ BUSINESS PARK POINT W CR 47	Matrix (type, e.g., Salt, Fresh, etc) FRESH		Diss. Oxygen 6.40	☒ mg/L ☐ % sat	Total Res. Chlorine (mg/L)
WIN/ST		Temp (C) 25.8	pH 7.15	Sample Depth 0.2	☐ ft ☒ m	Sp. Conductance (umho/cm) 143
Latitude	☐ dd ☐ dms	Longitude	☐ dd ☐ dms	Salinity (ppt) 0.07	Comments	
Location	WIN / Field ID:  21010033	☐ Comp ☒ Grab	Collection (comp begin or grab) Date 7/9/19 Time 1150	Eastern Central	Composite end Date Time	Bottle Group(s)
Field ID	LITTLE CREEK AT US 441	Matrix (type, e.g., Salt, Fresh, etc) FRESH		Diss. Oxygen 5.88	☒ mg/L ☐ % sat	Total Res. Chlorine (mg/L)
WIN/STORET		Temp (C) 25.4	pH 6.04	Sample Depth 0.1	☐ ft ☒ m	Sp. Conductance (umho/cm) 81.5
Latitude	☐ dd ☐ dms	Longitude	☐ dd ☐ dms	Salinity (ppt) 0.04	Comments	
Location		☐ Comp ☐ Grab	Collection (comp begin or grab) Date Time	Eastern Central	Composite end Date Time	Bottle Group(s)
Field ID		Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen	☐ mg/L ☐ % sat	Total Res. Chlorine (mg/L)
WIN/STORET #		Temp (C)	pH	Sample Depth	☐ ft ☐ m	Sp. Conductance (umho/cm)
Latitude	☐ dd ☐ dms	Longitude	☐ dd ☐ dms	Salinity (ppt)	Comments	

Relinquished By Jeremy Parrish	Date/Time 7/9/19 1530	Shipping Method Fed ex	Number of Coolers Shipped: 2	Received By:	Date/Time
-----------------------------------	--------------------------	---------------------------	---------------------------------	--------------	-----------

Group: A	Bottle Type:	P-1L	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	ALKALINITY TURBIDITY W-CL-IC W-COLOR W-F W-SO4-IC W-TDS W-TSS						
Group: A	Bottle Type:	BP-1L	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	CHLSUITE-W						
Group: A	Bottle Type:	P-250ML-WI-THI	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1 to ALS
	NE-ALS-ECQ						
Group: A	Bottle Type:	QPCR-P-500ML	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	PCR-FILTER						
Group: A	Bottle Type:	BG-500ML	# of Bottles: 1	Preservative:	CH2CICOOH Pre-Preserved	Enter Number of Bottles Sent to Lab	1
	W-CARB-AA						
Group: A	Bottle Type:	BG-500ML	# of Bottles: 4	Preservative:	ICE	Enter Number of Bottles Sent to Lab	4
	W-CT-CI-R W-PCL-SQ-R						
Group: A	Bottle Type:	BG-500ML	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	W-E8321-DI W-E8321-MS						
Group: A	Bottle Type:	P-500ML	# of Bottles: 1	Preservative:	HNO3	Enter Number of Bottles Sent to Lab	1
	W-ICP W-ICPMS						
Group: A	Bottle Type:	P-500ML	# of Bottles: 1	Preservative:	ICE And H2SO4	Enter Number of Bottles Sent to Lab	1
	W-NH3 W-NO2NO3 W-S-A-TP W-TKN W-TOC						

Group: A	Bottle Type: W-PO4-F	P-125ML	# of Bottles: 1	Preservative: FILTER-ICE	Enter Number of Bottles Sent to Lab	1
Group: A	Bottle Type: W-PSNP-TQ	BG-1L	# of Bottles: 4	Preservative: ICE	Enter Number of Bottles Sent to Lab	4
Group: B	Bottle Type: ALKALINITY TURBIDITY W-CL-IC W-COLOR W-F W-SO4-IC W-TDS W-TSS	P-1L	# of Bottles: 1	Preservative: ICE	Enter Number of Bottles Sent to Lab	1
Group: B	Bottle Type: CHLSUITE-W	BP-1L	# of Bottles: 1	Preservative: ICE	Enter Number of Bottles Sent to Lab	1
Group: B	Bottle Type: NE-ALS-ECQ	P-250ML-WI-THI	# of Bottles: 1	Preservative: ICE	Enter Number of Bottles Sent to Lab	1 to ALS
Group: B	Bottle Type: PCR-FILTER	QPCR-P-500ML	# of Bottles: 1	Preservative: ICE	Enter Number of Bottles Sent to Lab	1
Group: B	Bottle Type: W-E8321-DI W-E8321-MS	BG-500ML	# of Bottles: 1	Preservative: ICE	Enter Number of Bottles Sent to Lab	1
Group: B	Bottle Type: W-NH3 W-NO2NO3 W-S-A-TP W-TKN W-TOC	P-500ML	# of Bottles: 1	Preservative: ICE And H2SO4	Enter Number of Bottles Sent to Lab	1
Group: B	Bottle Type: W-PO4-F	P-125ML	# of Bottles: 1	Preservative: FILTER-ICE	Enter Number of Bottles Sent to Lab	1
Group: C	Bottle Type: ALKALINITY	P-1L	# of Bottles: 1	Preservative: ICE	Enter Number of Bottles Sent to Lab	1

Group: C	Bottle Type:	P-1L	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	TURBIDITY						
	W-CL-IC						
	W-COLOR						
	W-F						
	W-SO4-IC						
	W-TDS						
	W-TSS						

Group: C	Bottle Type:	BP-1L	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	CHLSUITE-W						

Group: C	Bottle Type:	P-500ML	# of Bottles: 1	Preservative:	ICE And H2SO4	Enter Number of Bottles Sent to Lab	1
	W-NH3						
	W-NO2NO3						
	W-S-A-TP						
	W-TKN						
	W-TOC						

Group: C	Bottle Type:	P-125ML	# of Bottles: 1	Preservative:	FILTER-ICE	Enter Number of Bottles Sent to Lab	1
	W-PO4-F						