

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Stitch

RQ- 2019-10-07-80

Project: Southside mallee

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 6/2019

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	J. Parnish	10/3/19	752	23.3	757.4	8.5	8.5	99.7	—	—	P/F	L/F
ICV	J	J	753	23.3	757.4	8.5	8.49	99.6	—	—	P/F	L/F
CCV	Brian Davis	10/3/19	1327	25.1	757.4	8.2	8.29	100.7	—	—	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	J. Parnish	10/3/19	755	190711B	7/20	1000	1000	P/F	L/F
ICV	J	J	759	147430	12/19	100147	150	P/F	L/F
CCV	Brian Davis	10/3/19	1330	190711B	7/20	1000	1001	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J. Parnish	10/3/19	803	2902778	1/21	7.00	23.4	7.01	-8.6	P/F	L/F
Calibr.	J	J	804	2904800	3/21	4.	23.4	4.0	173.1	P/F	L/F
Calibr.	J	J	805	2901851	7/20	10.	23.5	10.02	-179.4	P/F	L/F
ICV	J	J	808	2904800	3/21	4.	23.4	4.07	165.1	P/F	L/F
CCV	Brian Davis	10/3/19	1332	2902778	1/21	7.0	23.6	6.98	-2.8	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2019

Data Qualified?

Yes / No

WIN Station Name: OLDFIELD CR @ JULINGTON CREEK ROAD

Date: 10/3/19
(mm/dd/yyyy)

WIN/ Field ID: 20030684

WBID: 2370

Latitude: 30.138638

(Decimal Degrees)

County: Duval

ORG ID: 21FLA

Longitude: 81.60608

(Decimal Degrees)

Receiving Body of Water: Junington Creek

RQ-2019- 10-07-80

Sampling Team Members						WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check	Sampling Equipment				
<u>Spanish</u> <u>B Davis</u>						☒	☒	☒	☒	☒	☐ Sample Container	☒ Van Dorn	☐ Pole		
											☐ Carboy	☐ Syringe			
											☐ Dipper	☐ Other			
											Depth Measured with <u>STICS</u>				
						Equipment ID <u>0-14047</u>									
Collection Method															
☐ Wading						☐ Canoe/Kayak									
☐ From Shore						☐ Electric Motor									
☐ Other <u>Bridge</u>						☐ Gasoline Motor									
Water Level: Dry/ Pooled/ Low / <u>Normal</u> / High / Flooded						Meter ID# <u>Shitch</u>			Matrix: <u>Fresh</u> / Marine / DI						
Photos: Yes / <u>No</u>		Flow: No Flow / Intestinal / <u>Flowing</u> / NA				Tide: Rising/ Falling/ Slack/ NA			Tide Times: High Low						
Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Sp COND (umhos/cm)	Temp. (C°)					
EST	<u>915</u>	<u>1.2</u>	<u>0.3</u>	<u>0.9</u>	<u>1.78</u>	<u>21.3</u>	<u>6.88</u>	<u>0.16</u>	<u>346</u>	<u>24.4</u>					
CEN															
Biological Samples Collected: <u>None</u> HA SCI RPS Algae ID LVS LVI Other:															
SBIO ID Numbers:		SBIO-ID:		HA-ID:		RPS-ID:		Macrophyte-ID:		LIMS#:					
Parameter Suite		Parameter Methods				Preservation		Lot#		Expiration pH Check					
Preserved Nutrients		☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☒ Ice ☒ H2SO4		<u>SA9010030</u>		<u>4/10/20</u> ☒ <2					
Metals		☐ W-ICPMS ☐ W-ICP				☐ Ice ☐ HNO3				☐ <2					
Filtered Nutrients		☒ W-PO4-F ☒ Field Filtered 0.45 um Filter				☒ Ice		<u>BA8A 8675</u>							
Chlorophyll		☒ CHLSUITE-W				☒ Ice									
Physical/Aggregate (Fresh)		☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS				☒ Ice									
Physical/Aggregate (Marine)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice									
Macroinvertebrates		☐ MI-FW-QLDC				☐ Formalin									
Periphyton		☐ ALGAE-ID				☐ Ice									
Microbiology		☒ E Coli ☐ F Coli ☐ Enterococci				☒ Ice									
Molecular		☐ PCR-FILTER ☐ PCR-BacR ☐ PCR-GULL2 ☐ PCR-HF183 ☐ PCR-DG3 ☐ PCR-GFD				☐ Ice									
Tracers		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice									
Pesticides		☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-E8321-DI ☐ W-PCL-SQ-R ☐ WE8321-MS ☐ W-CARB-AA				☐ Ice									
Phytoplankton		☐ DTY-QN-C ☐ PKD-QN-C				☐ Ice									

Notes:

WIN / Field ID:



20030684

OLDFIELD CR @ JULINGTON CREEK ROAD

Signature: Spanish

Date: 10/3/19

LIMS EVENT ID:

WIN Import ID:

Enter Field Parameters, Verify Station ID in LIMS DMT:

QC Station ID, Field Parameters in LIMS DMT:

Scan/Save Field Sheets

Migrate Data to WIN:

Verify Data in WIN:

(Initials/Date)

(Initials/Date)

(Initials/Date)

Customer: NE-DIST

Requester: Jeremy M. Parrish

Field Report Prepared By:

Project ID: SMP

Collected By:

Sampling Agency:

Location	WIN / Field ID: 20030684	☐ Comp ☒ Grab	Collection (comp begin or grab) Date 10/3/19 Time 915	Eastern Central	Composite end Date	Time	Bottle Group(s) (See pg 2)
Field ID	OLDFIELD CR @ JULINGTON CREEK ROAD	Matrix (type, e.g., Salt, Fresh, etc) Fresh		Diss. Oxygen 1.78	☒ mg/L ☐ % sat	Total Res. Chlorine (mg/L)	A
WIN/S		Temp (C) 24.4	pH 6.88	Sample Depth 0.3	☐ ft ☒ m	Sp. Conductance (umho/cm) 346	
Latitude	☐ dd ☐ dms	Longitude	☐ dd ☐ dms	Salinity (ppt) 0.16	Comments		
Location		☐ Comp ☐ Grab	Collection (comp begin or grab) Date	Eastern Central	Composite end Date	Time	Bottle Group(s)
Field ID		Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen	☐ mg/L ☐ % sat	Total Res. Chlorine (mg/L)	
WIN/STORET #		Temp (C)	pH	Sample Depth	☐ ft ☐ m	Sp. Conductance (umho/cm)	
Latitude	☐ dd ☐ dms	Longitude	☐ dd ☐ dms	Salinity (ppt)	Comments		
Location		☐ Comp ☐ Grab	Collection (comp begin or grab) Date	Eastern Central	Composite end Date	Time	Bottle Group(s)
Field ID		Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen	☐ mg/L ☐ % sat	Total Res. Chlorine (mg/L)	
WIN/STORET #		Temp (C)	pH	Sample Depth	☐ ft ☐ m	Sp. Conductance (umho/cm)	
Latitude	☐ dd ☐ dms	Longitude	☐ dd ☐ dms	Salinity (ppt)	Comments		
Location		☐ Comp ☐ Grab	Collection (comp begin or grab) Date	Eastern Central	Composite end Date	Time	Bottle Group(s)
Field ID		Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen	☐ mg/L ☐ % sat	Total Res. Chlorine (mg/L)	
WIN/STORET #		Temp (C)	pH	Sample Depth	☐ ft ☐ m	Sp. Conductance (umho/cm)	
Latitude	☐ dd ☐ dms	Longitude	☐ dd ☐ dms	Salinity (ppt)	Comments		
Location		☐ Comp ☐ Grab	Collection (comp begin or grab) Date	Eastern Central	Composite end Date	Time	Bottle Group(s)
Field ID		Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen	☐ mg/L ☐ % sat	Total Res. Chlorine (mg/L)	
WIN/STORET #		Temp (C)	pH	Sample Depth	☐ ft ☐ m	Sp. Conductance (umho/cm)	
Latitude	☐ dd ☐ dms	Longitude	☐ dd ☐ dms	Salinity (ppt)	Comments		
Location		☐ Comp ☐ Grab	Collection (comp begin or grab) Date	Eastern Central	Composite end Date	Time	Bottle Group(s)
Field ID		Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen	☐ mg/L ☐ % sat	Total Res. Chlorine (mg/L)	
WIN/STORET #		Temp (C)	pH	Sample Depth	☐ ft ☐ m	Sp. Conductance (umho/cm)	
Latitude	☐ dd ☐ dms	Longitude	☐ dd ☐ dms	Salinity (ppt)	Comments		

Relinquished By: <i>Shmidt</i>	Date/Time: <i>10/3/19 1:00</i>	Shipping Method: <i>Fedex</i>	Number of Coolers Shipped: <i>1</i>	Received By:	Date/Time
--------------------------------	--------------------------------	-------------------------------	-------------------------------------	--------------	-----------

Group: A	Bottle Type:	P-1L	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	ALKALINITY						
	TURBIDITY						
	W-CL-IC						
	W-COLOR						
	W-F						
	W-SO4-IC						
	W-TDS						
	W-TSS						
Group: A	Bottle Type:	BP-1L	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	CHLSUITE-W						
Group: A	Bottle Type:	P-120ML-WC-THI	# of Bottles: 2	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1 to ALS
	NE-ALS-ECQ						
Group: A	Bottle Type:	P-500ML	# of Bottles: 1	Preservative:	ICE And H2SO4	Enter Number of Bottles Sent to Lab	1
	W-NH3						
	W-NO2NO3						
	W-S-A-TP						
	W-TKN						
	W-TOC						
Group: A	Bottle Type:	P-125ML	# of Bottles: 1	Preservative:	FILTER-ICE	Enter Number of Bottles Sent to Lab	1
	W-PO4-F						

CHAIN OF CUSTODY / LABORATORY ANALYSIS REQUEST FORM