

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Pro DSS Stitz RQ- 2019-10-21-28 Project: Nassau West

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 10-23-2019

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	<u>AKalmbacher</u>	<u>10-24-19</u>	<u>0755</u>	<u>22.6</u>	<u>760.2</u>	<u>8.6</u>	<u>8.66</u>	<u>100.0</u>	<u>---</u>	<u>---</u>	<u>P/F</u>	<u>L/F</u>
ICV			<u>0758</u>	<u>22.5</u>	<u>760.2</u>	<u>8.6</u>	<u>8.66</u>	<u>100.1</u>	<u>---</u>	<u>---</u>	<u>P/F</u>	<u>L/F</u>
CCV	<u>AKalmbacher</u>	<u>10-24-19</u>	<u>1330</u>	<u>21.9</u>	<u>760.3</u>	<u>8.8</u>	<u>8.89</u>	<u>101.8</u>	<u>---</u>	<u>---</u>	<u>P/F</u>	<u>L/F</u>
CCV											<u>P/F</u>	<u>L/F</u>

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	<u>AKalmbacher</u>	<u>10-24-19</u>	<u>0800</u>	<u>190711B</u>	<u>7/2020</u>	<u>11000</u>	<u>1001</u>	<u>P/F</u>	<u>L/F</u>
ICV			<u>0805</u>	<u>190711A</u>	<u>7/2020</u>	<u>100</u>	<u>104.7</u>	<u>P/F</u>	<u>L/F</u>
CCV	<u>AKalmbacher</u>	<u>10-24-19</u>	<u>1333</u>	<u>190711B</u>	<u>7/2020</u>	<u>11000</u>	<u>11004</u>	<u>P/F</u>	<u>L/F</u>
CCV								<u>P/F</u>	<u>L/F</u>

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	<u>AKalmbacher</u>	<u>10-24-19</u>	<u>0807</u>	<u>2902778</u>	<u>1/2021</u>	<u>7.</u>	<u>23.9</u>	<u>7.01</u>	<u>9.0</u>	<u>P/F</u>	<u>L/F</u>
Calibr.			<u>0810</u>	<u>2904B00</u>	<u>3/2021</u>	<u>4.</u>	<u>23.8</u>	<u>4.00</u>	<u>170.8</u>	<u>P/F</u>	<u>L/F</u>
Calibr.			<u>0812</u>	<u>2810F55</u>	<u>4/2020</u>	<u>10.</u>	<u>23.7</u>	<u>10.02</u>	<u>181.4</u>	<u>P/F</u>	<u>L/F</u>
ICV			<u>0815</u>	<u>2902778</u>	<u>1/2021</u>	<u>7.0</u>	<u>23.8</u>	<u>7.03</u>	<u>7.7</u>	<u>P/F</u>	<u>L/F</u>
CCV	<u>AKalmbacher</u>	<u>10-24-19</u>	<u>1335</u>	<u>2904B00</u>	<u>3/2021</u>	<u>4.0</u>	<u>24.5</u>	<u>4.07</u>	<u>166.2</u>	<u>P/F</u>	<u>L/F</u>
CCV										<u>P/F</u>	<u>L/F</u>

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 172.4, slope from 4 to 7 179.8 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 10/24/2019

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	<u>AKalmbacher</u>	<u>10-24-19</u>	<u>0817</u>	<u>0.00</u>	<u>0.00</u>	<u>P/F</u>	<u>L/F</u>

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2019

Data Qualified?

Yes / No

WIN Station Name: DAUGHTERY BR @ CREWS RD

Date: 10/24/2019

(mm/dd/yyyy)

WIN/ Field ID: 19010136

WBID: 2236

Latitude: 30.36183

(Decimal Degrees)

County: Nassau

ORG ID: 21FLA

Longitude: -82.21405

(Decimal Degrees)

Receiving Body of Water: Cedar River

RQ-2019- 10-21-28

Sampling Team Members						WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check	Sampling Equipment					
Brian Davis											☒ Sample Container	☐ Van Dorn	☒ Pole			
Amy Kalmbacher											☐ Carboy	☐ Syringe				
											☐ Dipper	☐ Other				
											Depth Measured with <u>5.7CZ</u>					
											Equipment ID <u>Ortho Kit 44</u>					
Collection Method																
☐ Wading												☐ Canoe/Kayak				
☒ From Shore												☐ Electric Motor				
☐ Other												☐ Gasoline Motor				
Water Level: Dry/ Pooled <u>Low</u> / Normal / High / Flooded						Meter ID# <u>Pro DSS Stitch</u>			Matrix: <u>Fresh</u> / Marine / DI							
Photos: Yes / <u>No</u>		Flow: No Flow / Intestinal / <u>Flowing</u> / NA				Tide: Rising/ Falling/ Slack <u>NA</u>			Tide Times: High / Low							
Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Sp COND (µmhos/cm)	Temp. (C°)						
<u>EST</u>	<u>1130</u>	<u>0.3</u>	<u>0.15</u>	<u>VOB</u>	<u>6.11</u>	<u>676</u>	<u>6.64</u>	<u>0.06</u>	<u>137.2</u>	<u>20.3</u>						
CEN				<u>(S)</u>												
Biological Samples Collected: <u>None</u> HA SCI RPS Algae ID LVS LVI Other:																
SBIO ID Numbers:		SBIO-ID:		HA-ID:		RPS-ID:		Macrophyte-ID:		LIMS#:						
Parameter Suite	Parameter Methods					Preservation	Lot#	Expiration	pH Check							
Preserved Nutrients	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP					☒ Ice ☐ H2SO4	<u>SA9203100</u>	<u>7/2020</u>	☒ <2							
Metals	☐ W-ICPMS ☐ W-ICP					☐ Ice ☐ HNO3			☐ <2							
Filtered Nutrients	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter					☒ Ice	<u>R9BA86874</u>									
Chlorophyll	☒ CHLSUITE-W					☒ Ice										
Physical/Aggregate (Fresh)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CI-IC / W-COLOR / W-SO4-IC / W-TDS					☒ Ice										
Physical/Aggregate (Marine)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY					☐ Ice										
Macroinvertebrates	☐ MI-FW-QLDC					☐ Formalin										
Periphyton	☐ ALGAE-ID					☐ Ice										
Microbiology	☒ E Coli ☐ F Coli ☐ Enterococci					☒ Ice										
Molecular	☐ PCR-FILTER ☐ PCR-BacR ☐ PCR-GULL2 ☐ PCR-HF183 ☐ PCR-DG3 ☐ PCR-GFD					☐ Ice										
Tracers	☐ W-E8321-DI ☐ W-E8321-MS					☐ Ice										
Pesticides	☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-E8321-DI ☐ W-PCL-SQ-R ☐ WE8321-MS ☐ W-CARB-AA					☐ Ice										
Phytoplankton	☐ DTY-QN-C ☐ PKD-QN-C					☐ Ice										
Notes:																
Signature: <u>Amy Kalmbacher</u>						Date: <u>10-24-2019</u>										

WIN ID/Field ID: 19010136

DAUGHTERY BR @ CREWS RD

Contains: 1:1 H2SO4
Lot No.: SA9203100
Expires: 07/22/20
See Safety Data Sheet (SDS)

Place Preservation (If Applicable)

LIMS EVENT ID:

WIN Import ID:

Enter Field Parameters, Verify Station ID in LIMS DMT:

QC Station ID, Field Parameters in LIMS DMT:

Scan/Save Field Sheets

Migrate Data to WIN:

Verify Data in WIN:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2019

Data Qualified?

Yes / No

WIN Station Name: ST MARYS S PRONG @ CR 23C

Date: 10/24/2019

(mm/dd/yyyy)

WIN/ Field ID: 19010023

WBID: 2247A

Latitude: 30.33584

(Decimal Degrees)

County: Nassau

ORG ID: 21FLA

Longitude: -82.14423

(Decimal Degrees)

Receiving Body of Water: St. Mary South Prong

RQ-2019- 10-21-28

Sampling Team Members					WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check	Sampling Equipment				
Brian Davis					☒	☒	☒	☒	☒	☒ Sample Container	☐ Van Dorn	☒ Pole		
Amy Kalmbacher					☐	☐	☐	☐	☐	☐ Carboy	☐ Syringe			
					☐	☐	☐	☐	☐	☐ Dipper	☐ Other			
										Depth Measured with <u>Pole</u>				
										Equipment ID <u>Ortho Kit 44</u>				
										Collection Method				
										☐ Wading	☐ Canoe/Kayak			
										☒ From Shore	☐ Electric Motor			
										☐ Other	☐ Gasoline Motor			
Water Level: Dry/ Pooled/ <u>Low</u> / Normal / High / Flooded					Meter ID# <u>57146</u>					Matrix: <u>Fresh</u> Marine / DI				
Photos: Yes / <u>No</u>					Flow: No Flow / Intertidal / <u>Flowing</u> / NA					Tide: Rising/ Falling/ Slack/ <u>NA</u>				
Tide Times: High — Low —														
Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Sp COND (umhos/cm)	Temp. (C°)				
<u>EST</u>	<u>1030</u>	<u>0.4</u>	<u>0.2</u>	<u>VOB (S)</u>	<u>6.82</u>	<u>74.6</u>	<u>7.00</u>	<u>0.10</u>	<u>207.0</u>	<u>19.7</u>				
CEN														
Biological Samples Collected: <u>None</u> HA SCI RPS Algae ID LVS LVI Other:														
SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#:														
Parameter Suite	Parameter Methods					Preservation		Lot#	Expiration	pH Check				
Preserved Nutrients	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP					☒ Ice	☒ H2SO4	<u>SA9203100</u>	<u>7/2020</u>	☒ <2				
Metals	☐ W-ICPMS ☐ W-ICP					☐ Ice	☐ HNO3			☐ <2				
Filtered Nutrients	☒ W-PO4-F ☒ Field Filtered 0.45 um Filter					☒ Ice		<u>RIBA86875</u>						
Chlorophyll	☒ CHLSUITE-W					☒ Ice								
Physical/Aggregate (Fresh)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CI-IC / W-COLOR / W-SO4-IC / W-TDS					☒ Ice								
Physical/Aggregate (Marine)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY					☐ Ice								
Macroinvertebrates	☐ MI-FW-QLDC					☐ Formalin								
Periphyton	☐ ALGAE-ID					☐ Ice								
Microbiology	☒ E Coli ☐ F Coli ☐ Enterococci					☒ Ice								
Molecular	☐ PCR-FILTER ☐ PCR-BacR ☐ PCR-GULL2					☐ Ice								
	☐ PCR-HF183 ☐ PCR-DG3 ☐ PCR-GFD					☐ Ice								
Tracers	☐ W-E8321-DI ☐ W-E8321-MS					☐ Ice								
Pesticides	☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-E8321-DI					☐ Ice								
	☐ W-PCL-SQ-R ☐ WE8321-MS ☐ W-CARB-AA					☐ Ice								
Phytoplankton	☐ DTY-QN-C ☐ PKD-QN-C					☐ Ice								

Notes:

Signature: Kalmbacher

Date: 10-24-2019

LIMS EVENT ID:

WIN Import ID:

Enter Field Parameters, Verify Station ID in LIMS DMT:

QC Station ID, Field Parameters in LIMS DMT:

Scan/Save Field Sheets

Migrate Data to WIN:

Verify Data in WIN:

(Initials/Date)

(Initials/Date)

(Initials/Date)



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2019

Data Qualified?

Yes / No

WIN Station Name: ST MARYS S PRONG @ US 90

Date: 10/24/2019

(mm/dd/yyyy)

WIN/ Field ID: 19010021

WBID: 2247

Latitude: 30.27927

(Decimal Degrees)

County: Nassau

ORG ID: 21FLA

Longitude: -82.14523

(Decimal Degrees)

Receiving Body of Water: St. Mary South Prong

RQ-2019- 10-21-28

Sampling Team Members					WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check	Sampling Equipment				
Brian Davis						X				☒ Sample Container	☐ Van Dorn	☒ Pole		
Amy Kalmbacher					X		X			☐ Carboy	☐ Syringe			
										☐ Dipper	☐ Other			
										Depth Measured with <u>meter</u>				
										Equipment ID <u>Ortho RT 44</u>				
										Collection Method				
										☐ Wading	☐ Canoe/Kayak			
										☒ From Shore	☐ Electric Motor			
										☐ Other	☐ Gasoline Motor			
Water Level: Dry/ Pooled/ Low / <u>Normal</u> / High / Flooded					Meter ID# <u>Pro DSS Switch</u>					Matrix: <u>Fresh</u> / Marine / DI				
Photos: Yes / <u>No</u>		Flow: No Flow / Intestinal / <u>Flowing</u> / NA		Tide: Rising/ Falling/ Slack <u>NA</u>		Tide Times: High		Low						
Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Sp COND (umhos/cm)	Temp. (C°)				
<u>EST</u>	<u>0925</u>	<u>.44</u>	<u>.20</u>	<u>.44</u>	<u>3.87</u>	<u>43.4</u>	<u>6.72</u>	<u>0.11</u>	<u>232.7</u>	<u>21.0</u>				
CEN				<u>S</u>										
Biological Samples Collected: <u>None</u> HA SCI RPS Algae ID LVS LVI Other:														
SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#:														
Parameter Suite	Parameter Methods					Preservation	Lot#	Expiration	pH Check					
Preserved Nutrients	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP					☒ Ice ☒ H2SO4	<u>SA9203100</u>	<u>7/2020</u>	☐ <2					
Metals	☐ W-ICPMS ☐ W-ICP					☐ Ice ☐ HNO3			☐ <2					
Filtered Nutrients	☒ W-PO4-F ☒ Field Filtered 0.45 um Filter					☒ Ice	<u>R9BA86875</u>							
Chlorophyll	☒ CHLSUITE-W					☒ Ice								
Physical/Aggregate (Fresh)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CI-IC / W-COLOR / W-SO4-IC / W-TDS					☒ Ice								
Physical/Aggregate (Marine)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY					☐ Ice								
Macroinvertebrates	☐ MI-FW-QLDC					☐ Formalin								
Periphyton	☐ ALGAE-ID					☐ Ice								
Microbiology	☒ E Coli ☐ F Coli ☐ Enterococci					☒ Ice								
Molecular	☐ PCR-FILTER ☐ PCR-BacR ☐ PCR-GULL2 ☐ PCR-HF183 ☐ PCR-DG3 ☐ PCR-GFD					☐ Ice								
Tracers	☐ W-E8321-DI ☐ W-E8321-MS					☐ Ice								
Pesticides	☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-E8321-DI ☐ W-PCL-SQ-R ☐ WE8321-MS ☐ W-CARB-AA					☐ Ice								
Phytoplankton	☐ DTY-QN-C ☐ PKD-QN-C					☐ Ice								

Notes:

WIN ID/Field ID:



19010021

ST MARYS S PRONG @ US 90

Signature: A. Kalmbacher

Date: 10-24-2019

LIMS EVENT ID:

WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT:

QC Station ID, Field Parameters In LIMS DMT:

Scan/Save Field Sheets

Migrate Data to WIN:

Verify Data In WIN:

(Initials/Date)

(Initials/Date)

(Initials/Date)

Nassau West

Central Laboratory Submittal Form

last revised Jan 25, 2018

Customer: NE-DIST

Requester: Jeremy M. Parrish

Field Report Prepared By:

A. Kalmbacher

Project ID: SMP

Collected By: B. Davis

Sampling Agency:

FDEP / NE ROC

WIN ID/Field ID:



19010136

DAUGHTERY BR @ CREWS RD

Latitude	☐ dd ☐ dms	Longitude	☐ dd ☐ dms	Salinity (ppt)	0.06	Comments	
----------	---	-----------	---	----------------	------	----------	--

WIN ID/Field ID:



19010023

ST MARYS S PRONG @ CR 23C

Latitude	☐ dd ☐ dms	Longitude	☐ dd ☐ dms	Salinity (ppt)	0.10	Comments	
----------	---	-----------	---	----------------	------	----------	--

WIN ID/Field ID:



19010021

ST MARYS S PRONG @ US 90

Latitude	☐ dd ☐ dms	Longitude	☐ dd ☐ dms	Salinity (ppt)	21.0	Comments	
----------	---	-----------	---	----------------	------	----------	--

Location	☐ Comp ☐ Grab	Collection (comp begin or grab)	Eastern Central	Composite end	Bottle Group(s)
Field ID	Date	Time	Date	Time	
WIN/STORET #	Matrix (type, e.g., Salt, Fresh, etc)	Diss. Oxygen	☐ mg/L ☐ % sat	Total Res. Chlorine (mg/L)	
Latitude	Temp (C)	pH	Sample Depth	☐ ft ☐ m	Sp. Conductance (umho/cm)
Longitude					
Salinity (ppt)	Comments				

Relinquished By:	Date/Time	Shipping Method	Number of Coolers Shipped:	Received By:	Date/Time
A. Kalmbacher	10-24-2019	Fed Ex	1		

Group: A	Bottle Type:	P-1L	# of Bottles: 3	Preservative:	ICE	Enter Number of Bottles Sent to Lab	3
	ALKALINITY						
	TURBIDITY						
	W-CL-IC						
	W-COLOR						
	W-F						
	W-SO4-IC						
	W-TDS						
	W-TSS						
Group: A	Bottle Type:	BP-1L	# of Bottles: 3	Preservative:	ICE	Enter Number of Bottles Sent to Lab	3
	CHLSUITE-W						
Group: A	Bottle Type:	P-120ML-WC-THI	# of Bottles: 6	Preservative:	ICE	Enter Number of Bottles Sent to Lab	6 Sent to ALS
	NE-ALS-ECQ						
Group: A	Bottle Type:	P-500ML	# of Bottles: 3	Preservative:	ICE And H2SO4	Enter Number of Bottles Sent to Lab	3
	W-NH3						
	W-NO2NO3						
	W-S-A-TP						
	W-TKN						
	W-TOC						
Group: A	Bottle Type:	P-125ML	# of Bottles: 3	Preservative:	FILTER-ICE	Enter Number of Bottles Sent to Lab	3
	W-PO4-F						



CHAIN OF CUSTODY / LABORATORY ANALYSIS REQUEST FORM

SR #
ALS Contact

9143 Phillips Highway, Suite 200. Jacksonville, FL 32256
Phone (904) 739-2277 / 800-695-7222 x06 / FAX (904) 739-2011
www.alsglobal.com

Page 1 of 1

Project Name Nassau West		Project Number RQ-2019-10-21-28		ANALYSIS REQUESTED (Include Method Number and Container Preservative)																			
Report To Jeremy Parrish		Report CC Dep-Overflow micro@		Preservative																Preservative Key			
Company/Address FIDEP - 8800 Baymeadows Way W Jacksonville, FL 32256		Phone # 904-256-1700		FAX #		NUMBER OF CONTAINERS E. Coli																0. None 1. HCL 2. HNO3 3. H2SO4 4. NaOH 5. Zn. Acetate 6. MeOH 7. NaHSO4 8. Other REMARKS	
Sampler's Signature [Signature]		Sampler's Printed Name Amy Kalmbacher																					
CLIENT SAMPLE ID	LAB ID	SAMPLING DATE	TIME	Matrix																	Fresh		
19010021		10-24	0925	SW	2	✓																↓	
19010023		↓	1030	↓	2	✓																	
19010138		↓	1130	↓	2	✓																	
Special Instructions/Comments:				TURNAROUND REQUIREMENTS RUSH (SURCHARGES APPLY) STANDARD REQUESTED FAX DATE				REPORT REQUIREMENTS I. Results Only II. Results + QC Summaries (LCS, DUP, MS/MSD as required) III. Results + QC and Calibration Summaries IV. Data Validation Report with Raw Data Edata Yes No				INVOICE INFORMATION P.O. # Bill to:											
SAMPLE RECEIPT: CONDITION/COOLER TEMP: 28°C				CUSTODY SEALS: Y (N)				REQUESTED REPORT DATE															
Relinquished By Signature [Signature]		Received By Signature [Signature]		Relinquished By Signature [Signature]		Received By Signature [Signature]		Relinquished By Signature [Signature]		Received By Signature [Signature]		Relinquished By Signature [Signature]		Received By Signature [Signature]									
Printed Name Amy Kalmbacher		Printed Name [Signature]		Printed Name [Signature]		Printed Name [Signature]		Printed Name [Signature]		Printed Name [Signature]		Printed Name [Signature]		Printed Name [Signature]									
Firm FIDEP		Firm		Firm		Firm		Firm		Firm		Firm		Firm									
Date/Time 10-24-2019		Date/Time		Date/Time		Date/Time		Date/Time		Date/Time		Date/Time		Date/Time									