

RQ-2019-07-22-14

Field Personnel processing steps

☐ New Field IDs created (when using a new station) Date: _____ initials: ~~NA~~

☐ Create entry in SMPStationCreationFile if new station is used

\\Fldep1\wqap\NW ROC\TMDL\WIN_Station Creation

☐ Date sampled Date: 7/24/19 initials: ME

☐ Update MAT Date: — initials: —

☐ Update "SMP Tracker" Date: 7/24/19 initials: ME

☐ Photos put into archive Date: _____ initials: ~~NA~~

☒ RQ Sheet – from LIMS (display/print request)

☒ Field Sheets initialed/dated @ "Scan/Save Field Sheets" then scanned (with calibration logs) and saved into:

Scans_ToBeProcessed Date: _____ initials: _____

☐ Update "Data Tracker"

\\publicfiles\pubfds\DEAR\WQAPmonitoring\StrategicMonitoring\2019\21FLPNS

Ready for folder system (top)

LIMS processing (to be completed by LIMS data personnel) (MC)

☐ Field parameters entered into LIMS Date: _____ initials: _____

☐ Digital RQ packet put into

SMP 2019/Data/Scanned Field Sheets Date: _____ initials: _____

☐ Digital field Sheets put into archive Date: _____ initials: _____

☐ Date submitted for QA/QC Date: _____ initials: _____

☐ Update "Data Tracker"

Database processing (Cheryl)

☐ Field parameters QA/QC in LIMS Date: _____ initials: _____

☐ Lab results put into archive (CB) Date: _____ initials: _____

☐ Data migrated to WINS Date: _____ initials: _____

☐ Update "Data Tracker"

Notes:

Date of Request: 17-JUN-2019
 Created By: KING_M On: 17-JUN-2019
 Modified By: JASROTIA_P On: 03-JUL-2019
 Customer: NW-DIST
 Project: SMP
 Division: Division of Environmental Assessment and Restoration
 District: Northwest District
 Sampling Event: Western Lakes

Send Coolers To:

Phone: 850-595-0605
 160 Government Center
 Suite 208
 Pensacola, FL 32502-5794
 Attn: Michael King

Program Module Number:
 Priority: 3
 Event Status: S
 Criminal Investigation: NO
 Chemistry Request Reviewed By: AYRES_J On: 03-JUL-2019
 Biology Request Reviewed By: JASROTIA_P On: 03-JUL-2019
 Sampling Kit Required: YES Ship on: 12-JUL-2019
 Sampling Kit Shipped: YES On: 09-JUL-2019
 Sampling Kit Packed By: SHAIK_A
 Preservatives Needed: YES
 Date to Receive Samples: 22-JUL-2019
 Received By:

Send Final Report To:

160 Government Center
 Suite 208
 Pensacola, FL 32502-5794
 Attn: Cheryl Bunch

Report Type: Final Only
 FTP Data: NO
 QC Report: YES
 Date Log: NO
 Authorization Log: NO

Comments: Group A: (WBID 124AF,145, FIELD DUP) Nutrients
 Group B: (FIELD BLANK) Nutrients

2 Boxes Sulfuric needed
 4 sets of labels

1 Box Nitric
 2 sets of labels

Bottle Group A (Water) With 3 Samples:

Bottle Type: P-1L	Number of Bottles: 3	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-97) Alkalinity in aqueous matrices as mg CaCO ₃ /L
Total Alkalinity		
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
Turbidity		
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
Chloride		
W-COLOR	Template: DEFAULT	(SM 2120 B) True Color measured at 450 nm
Color (true)		
W-F	Template: DEFAULT	(SM 4500 F-C-97) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
Fluoride		
W-SO4-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
Sulfate		
W-TDS	Template: DEFAULT	(SM 2540 C-97) Total Dissolved Solids in aqueous matrices
TDS		
W-TSS	Template: DEFAULT	(SM 2540 D-97) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample
TSS		
Bottle Type: BP-1L	Number of Bottles: 3	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Bottle Group A (Water) With 3 Samples:

Bottle Type: BP-1L	Number of Bottles: 3	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry
Chlorophyll-a, Corrected		
Chlorophyll-a, Uncorrected		
Phaeophytin-a		
Bottle Type: P-500ML	Number of Bottles: 3	Preserved With: ICE And H2SO4
W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
Ammonia-N		
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
NO2NO3-N		
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
Total-P		
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
Kjeldahl Nitrogen		
W-TOC	Template: DEFAULT	(SM 5310 B-00) Nonpurgeable Organic Carbon in aqueous matrices
Organic Carbon		
Bottle Type: P-125ML	Number of Bottles: 3	Preserved With: FILTER-ICE
W-PO4-F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L
O-Phosphate-P		

Bottle Group B (Water) With 1 Samples:

Bottle Type: P-1L	Number of Bottles: 1	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-97) Alkalinity in aqueous matrices as mg CaCO3/L
Total Alkalinity		
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
Turbidity		
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
Chloride		
W-COLOR	Template: DEFAULT	(SM 2120 B) True Color measured at 450 nm
Color (true)		
W-F	Template: DEFAULT	(SM 4500 F-C-97) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
Fluoride		
W-SO4-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
Sulfate		
W-TDS	Template: DEFAULT	(SM 2540 C-97) Total Dissolved Solids in aqueous matrices
TDS		
W-TSS	Template: DEFAULT	(SM 2540 D-97) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample
TSS		
Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: ICE And H2SO4
W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
Ammonia-N		
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
NO2NO3-N		
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
Total-P		
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
Kjeldahl Nitrogen		
W-TOC	Template: DEFAULT	(SM 5310 B-00) Nonpurgeable Organic Carbon in aqueous matrices
Organic Carbon		
Bottle Type: P-125ML	Number of Bottles: 1	Preserved With: FILTER-ICE
W-PO4-F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L
O-Phosphate-P		

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: 155077

RQ: 2019-07-22-H

Project: SMP

- Notes:** (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification _____

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.											P / F	L / F
ICV											P / F	L / F
CCV	M. Clark	7/24/19	7:21	25.7	760	8.1	8.25	101.0			(P) F	(L) F
CCV	M. Clark	7/24/19	13:01	23.5	759.8	8.5	8.55	100.6			(P) F	(L) F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	M. Clark	7/24/19	7:23	190618C	6/20	19000	9889	(P) F	(L) F
ICV	M. Clark	7/24/19	7:28	190416A	1/20	1000	982	(P) F	(L) F
CCV	M. Clark	7/24/19	13:03	190618A	6/20	100	103.3	(P) F	(L) F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	M. Clark	7/24/19	7:32	281803	10/20	7.	25.2	7.00	-36.4	(P) F	(L) F
Calibr.	M. Clark	7/24/19	7:34	2902818	1/21	4.	25.0	4.01	146.4	(P) F	(L) F
Calibr.	M. Clark	7/24/19	7:36	49041509	9/20	10.	24.0	1901	-36.5	(P) F	(L) F
ICV	M. Clark	7/24/19	7:38	2902818	1/21	4	24.8	3.96	146.4	(P) F	(L) F
CCV	M. Clark	7/24/19	13:04	281803	10/20	7	21.2	7.05	-35.1	(P) F	(L) F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 170.7, slope from 4 to 7 161.6 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? **YES** / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 6/13/19

If NO, what will be used? (circle one) **Secchi Disk Line / Sonar** Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	M. Clark	7/24/19	7:40	0.0	0.0	(P) F	(L) F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2019

Data Qualified?

Yes

No

WIN Station Name: Locklin Lake Center

Date: 7/24/2019
(mm/dd/yyyy)

WIN/ Field ID: G4NW0437

WBID: 24AF

Latitude: 30.62786

County: Santa Rosa

ORG ID: 21FLPNS

Longitude: -87.04862811

Receiving Body of Water: Blackwater River

RQ-2019-07-22-14

Sampling Team Members					WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check	Sampling Equipment				
N. Mulkey M. Clark					X	X	X	X	X	☒ Sample Container	☐ Van Dorn	☐ Pole		
										☐ Carboy	☒ Syringe			
										☐ Dipper	☐ Other			
										Depth Measured with PropSS				
										Equipment ID				
										Collection Method				
										☐ Wading	☒ Canoe/Kayak			
										☐ From Shore	☐ Electric Motor			
										☐ Other	☐ Gasoline Motor			
Water Level: Dry/ Pooled/ Low / <u>Normal</u> / High / Flooded					Meter ID# 155077					Matrix: <u>Fresh</u> Marine / DI				
Photos: Yes / <u>No</u>					Flow: No Flow / Intestinal / Flowing / <u>NA</u>					Tide: Rising/ Falling/ Slack <u>NA</u>				
Tide Times: High Low														
Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPTH)	Sp COND (µmhos/cm)	Temp. (C°)				
EST	9:00	1.9	0.3	1.9	5.86	72.5	5.74	0.02	44.2	26.3				
					2.55	31.1	5.69	0.02	44.2	25.6				
CEN			1.4	"5"	2.55	31.1	5.69	0.02	44.2	25.6				
Biological Samples Collected: <u>None</u> HA SCI RPS Algae ID LVS LVI Other:														
SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#:														
Parameter Suite	Parameter Methods				Preservation		Lot#	Expiration	pH Check					
Preserved Nutrients	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☒ Ice ☒ H2SO4		5A8344070	12/19	☒ <2					
Metals	☐ W-ICPMS ☐ W-ICP				☐ Ice ☐ HNO3				☐ <2					
Filtered Nutrients	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter				☒ Ice		15159049							
Chlorophyll	☒ CHL-SUITE-W				☒ Ice									
Physical/Aggregate (Fresh)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CI-IC / W-COLOR / W-SO4-IC / W-TDS				☒ Ice									
Physical/Aggregate (Marine)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice									
Macroinvertebrates	☐ MI-FW-QLDC				☐ Formalin									
Periphyton	☐ ALGAE-ID				☐ Ice									
Microbiology	☐ E Coli ☐ F Coli ☐ Enterococci				☐ Ice									
Molecular	☐ PCR-FILTER ☐ PCR-BacR ☐ PCR-GULL2				☐ Ice									
	☐ PCR-HF183 ☐ PCR-DG3 ☐ PCR-GFD													
Tracers	☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice									
Pesticides	☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-E8321-DI				☐ Ice									
	☐ W-PCL-SQ-R ☐ WE8321-MS ☐ W-CARB-AA													
Phytoplankton	☐ DTY-QN-C ☐ PKD-QN-C													

Notes: Launch canoe at end of Lakeshore Dr at lift station
30.62799/-87.05545

Field/WIN ID →



Location:

G4NW0437



LOCKLIN LAKE CENTER

Signature: *Myron C. Clark*

Date: 7/24/19

LIMS EVENT ID:

WIN Import ID:

Enter Field Parameters, Verify Station ID in LIMS DMT:

QC Station ID, Field Parameters in LIMS DMT:

Scan/Save Field Sheets MC 7/24/19

Migrate Data to WIN:

Verify Data in WIN:

(Initials/Date)

(Initials/Date)

(Initials/Date)

QA/QC Sample Information

Duplicate

Time Zone: EST <u>CEN</u>	Time: <u>9:05</u>	Field ID: <u>G4NW0437-DUP</u> (WIN ID_DUP)			
WIN DMT Activity ID:					
Parameter Suite	Parameter Methods	Preservation	Lot#	Expiration	pH Check
Preserved Nutrients	☒ W-NH3 / ☒ W-NO2NO3 / ☒ W-TKN / ☒ W-TOC / ☒ W-S-A-TP	☒ Ice ☒ H ₂ SO ₄	<u>SA8344070</u>	<u>12/19</u>	☒ < 2
Metals	☐ W-ICPMS ☐ W-ICP	☐ Ice ☐ HNO ₃			☐ < 2
Filtered Nutrients	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter	☒ Ice	<u>15159049</u>		
Chlorophyll	☒ CHLSUITE-W	☒ Ice			
Physical/Aggregate (Fresh)	☒ Alkalinity / ☒ W-F / ☒ W-TSS / ☒ TURBIDITY / ☒ W-Cl-IC / ☒ W-COLOR / ☒ W-SO4-IC / ☒ W-TDS	☒ Ice			
Physical/Aggregate (Marine)	☐ Alkalinity / ☐ W-BR-IC / ☐ W-F / ☐ W-TSS / ☐ TURBIDITY	☐ Ice			
Microbiology	☐ E Coli ☐ F Coli ☐ Enterococci	☐ Ice			
Molecular	☐ PCR-FILTER ☐ PCR-BacR ☐ PCR-GULL2 ☐ PCR-HF183 ☐ PCR-DG3 ☐ PCR-GFD	☐ Ice			
Tracers	☐ W-E3821-MS ☐ W-E3821-DI	☐ Ice			
Pesticides	☐ W-PSNP-TQ ☐ W-E3821-DI ☐ W-CT-CL-R ☐ W-PCL-SQ-R ☐ W-E-3821-MS ☐ W-CARB-AA	☐ Ice			

Place Preservative Sticker(s) Here
(If Available)

Blank

Time Zone: EST <u>CEN</u>	Time: <u>9:10</u>	Field ID: <u>FB-24072019</u> (Blank Type_DDMMYYYY)			
DI Source: <u>Pcda Ice Room</u>	Location: <u>Locklin Lake Center</u>	Equipment ID: <u>NVWROE IV</u>			
☐ Field Blank ☐ Equipment Blank ☐ Field Cleaned Equipment Blank					
WIN DMT Blank Batch ID:					
Parameter Suite	Parameter Methods	Preservation	Lot#	Expiration	pH Check
Preserved Nutrients	☒ W-NH3 / ☒ W-NO2NO3 / ☒ W-TKN / ☒ W-TOC / ☒ W-S-A-TP	☒ Ice ☒ H ₂ SO ₄	<u>SA8344070</u>	<u>12/19</u>	☒ < 2
Metals	☐ W-ICPMS ☐ W-ICP	☐ Ice ☐ HNO ₃			☐ < 2
Filtered Nutrients	☒ W-PO4-F ☐ Field Filtered 0.45 µm Filter	☒ Ice	<u>15159049</u>		
Physical/Aggregate (Fresh)	☒ Alkalinity / ☒ W-F / ☒ W-TSS / ☒ TURBIDITY / ☒ W-Cl-IC / ☒ W-COLOR / ☒ W-SO4-IC / ☒ W-TDS	☒ Ice			
Physical/Aggregate (Marine)	☐ Alkalinity / ☐ W-F / ☐ W-TSS / ☐ TURBIDITY	☐ Ice			
Tracers	☐ W-E3821-MS ☐ W-E3821-DI	☐ Ice			
Pesticides	☐ W-PSNP-TQ ☐ W-E3821-DI ☐ W-CT-CL-R ☐ W-PCL-SQ-R ☐ W-E3821-MS ☐ W-CARB-AA	☐ Other			

Note Field ID → <u>G4NW0437_DUP</u> Location: <u>Locklin Lake Center</u> WIN ID → <u>G4NW0437</u>	Field ID: <u>FB_24072019</u> Location: <u>Locklin Lake Center</u>
Signature: <u>[Signature]</u>	Date: <u>7/24/19</u>

LIMS EVENT ID: _____	WIN Import ID: _____
Enter Field Parameters, Verify Station ID In LIMS DMT: _____	QC Station ID, Field Parameters In LIMS DMT: _____
Scan/Save Field Sheets <u>MC 7/24/19</u> (Initials/Date)	Migrate Data to WIN: _____ (Initials/Date)
Verify Data In WIN: _____ (Initials/Date)	



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2019

Data Qualified?

Yes ☒ No ☐

WIN Station Name: Lake Karick (Center) (also 33040210)

Date: 07/24/2019
(mm/dd/yyyy)

WIN/ Field ID: G4NW0274

WBID: 145

Latitude: 30.89633

(Decimal Degrees)

County: Okaloosa

ORG ID: 21FLPNS

Longitude: -86.64424

(Decimal Degrees)

Receiving Body of Water: Deadfall Creek

RQ-2019-07-22-14

Sampling Team Members					WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
N. Mulkem						X		X	X
M. Clark					X		X		

Sampling Equipment		
☒ Sample Container	☐ Van Dorn	☐ Pole
☐ Carboy	☒ Syringe	
☐ Dipper	☐ Other	
Depth Measured with ProDSS		
Equipment ID		

Collection Method		
☐ Wading	☒ Canoe/Kayak	
☐ From Shore	☐ Electric Motor	
☐ Other	☐ Gasoline Motor	

Water Level: Dry/ Pooled/ Low / Normal / High / Flooded

Meter ID# 155077

Matrix: Fresh Marine / DI

Photos: Yes ☒ No ☐

Flow: No Flow / Intestinal / Flowing NA

Tide: Rising/ Falling/ Slack NA

Tide Times: High Low

Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Sp COND (umhos/cm)	Temp. (C°)
EST	10:35	3.9	0.3	1.3	5.74	76.7	5.68	0.01	21.7	30.4
CEN			3.4		0.12	1.4	5.26	0.02	44.1	24.8

Biological Samples Collected: None HA SCI RPS Algae ID LVS LVI Other:

SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#:

Parameter Suite	Parameter Methods	Preservation	Lot#	Expiration	pH Check
Preserved Nutrients	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☒ Ice ☒ H2SO4	SA834400	12/19	☒ <2
Metals	☐ W-ICPMS ☐ W-ICP	☐ Ice ☐ HNO3			☐ <2
Filtered Nutrients	☒ W-PO4-F ☒ Field Filtered 0.45 um Filter	☒ Ice	18159049		
Chlorophyll	☒ CHLSUITE-W	☒ Ice			
Physical/Aggregate (Fresh)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS	☒ Ice			
Physical/Aggregate (Marine)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates	☐ MI-FW-QLDC	☐ Formalin			
Periphyton	☐ ALGAE-ID	☐ Ice			
Microbiology	☐ E Coli ☐ F Coli ☐ Enterococci	☐ Ice			
Molecular	☐ PCR-FILTER ☐ PCR-BacR ☐ PCR-GULL2 ☐ PCR-HF183 ☐ PCR-DG3 ☐ PCR-GFD	☐ Ice			
Tracers	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides	☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-E8321-DI ☐ W-PCL-SQ-R ☐ WE8321-MS ☐ W-CARB-AA	☐ Ice			
Phytoplankton	☐ DTY-QN-C ☐ PKD-QN-C				

Notes:

Field/WIN ID → G4NW0274

Location: Lake Karick (Center) (also 33040210)

Signature: [Signature] Date: 7/24/19

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: QC Station ID, Field Parameters In LIMS DMT:

Scan/Save Field Sheets 7/24/19 Migrate Data to WIN: Verify Data In WIN: