

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: 154177

RQ-2017-06-24-33

Project: SMP

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Ben Harden	06/24/19	08:02	19.8	756.9	9.2	9.20	99.8	/	1.054	P / F	L / F
ICV	"	"	08:03	19.1	756.9	9.2	9.25	99.8	/	"	P / F	L / F
CCV	Paul Blair	6-24-19	16:00	26.2	756.2	8.03	8.01	100.7	/	"	P / F	L / F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Ben Harden	06/24/19	08:05	190405A	04/20	1000	1001	P / F	L / F
ICV	"	"	08:06	24007654	07/20	100	99.8	P / F	L / F
CCV	Paul Blair	6-24-19	16:05	180927D	10-19	500	502	P / F	L / F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Ben Harden	06/24/19	08:09	180621E	01/20	7.02	9.5	-11.7	7.02	P / F	L / F
Calibr.	"	"	08:13	190405B	10/20	4.00	19.0	162.1	4.00	P / F	L / F
Calibr.	"	"	08:15	180621F	01/20	10.07	19.3	-178.6	10.07	P / F	L / F
ICV	"	"	08:16	180621F	01/20	10.07	19.7	-178.4	10.06	P / F	L / F
CCV	Paul Blair	"	16:08	190405B	10-20	4.01	21.8	169.2	3.92	P / F	L / F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 166.9 , slope from 4 to 7 173.8 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 4-19

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: ; Date of last verification:

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Ben Harden	06/24/19	07:57	0.272	0.272	P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2019

Data Qualified?

Yes / No

WIN Station Name: Camp Branch @ US90

Date: 06/24/2019
(mm/dd/yyyy)

WIN ID/ Field ID G3TLHR0005

WBID: 251

Latitude: 30.7628

(Decimal Degrees)

County: Holmes

ORG ID: 21 FLTLHR

Longitude: 85.6461

(Decimal Degrees)

Receiving Body of Water: Open Creek

RQ - 2019-06-24-33

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check	Duplicate Collection	Blank Collection
Jon Woods		X		X				X
Paul Blais			X		X	X	X	

Sampling Equipment		
☒ Sample Container	☐ Van Dorn	☐ Pole
☐ Carboy	☒ Syringe	
☐ Dipper	☐ Other	
Depth Measured with: <u>YSI</u>		
Equipment ID <u>154177</u>		

Collection Method	
☒ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: Dry/ Pooled/ <u>Low</u> / Normal/ High/ Flooded	Meter ID# <u>154177</u>	Matrix: <u>Fresh</u> / Marine / DI
Photos: Yes <u>No</u>	Flow: No Flow / Intertidal <u>Flowing</u> / NA	Tide: Rising/ Falling/ Slack/ <u>NA</u>
Tide Times: High		Low

Field Sample

(Bracket Sample Time)

Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (ppt)	Sp COND (µmhos/cm)	Temp. (C°)
<u>EST</u>	<u>1220</u>	<u>0.3</u>	<u>0.15</u>	<u>0.3</u>	<u>6.94</u>	<u>86.4</u>	<u>7.33</u>	<u>0.12</u>	<u>254.0</u>	<u>26.6</u>
CEN				<u>S</u>						

Biological Samples Collected:

None

HA

SCI

RPS

Algae ID

LVS

LVI

Other:

SBIO ID Numbers:

SBIO-ID:

HA-ID:

RPS-ID:

Macrophyte-ID:

LIMS#:

Parameter Suite	Parameter Methods	Preservation	Lot#	Expiration	pH Check
Preserved Nutrients	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☒ Ice ☒ H2SO4	<u>9010030</u>	<u>4/10/20</u>	☒ <2
Metals	☐ W-ICPMS ☐ W-ICP	☐ Ice ☐ HNO3			☐ <2
Filtered Nutrients	☒ W-PO4-F ☐ Field Filtered 0.45 µm Filter	☒ Ice			
Chlorophyll	☒ CHLSUITE-W	☒ Ice			
Physical/Aggregate (Fresh)	☒ Alkalinity W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-TDS	☒ Ice			
Physical/Aggregate (Marine)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates	☐ MI-FW-QLDC	☐ Formalin			
Periphyton	☐ ALGAE-ID	☐ Ice			
Microbiology	☒ E Coli ☐ F Coli ☐ Enterococci	☒ Ice			
Molecular	☐ PCR-FILTER ☐ PCR-BacR ☐ PCR-GULL2 ☐ PCR-HF183 ☐ PCR-DG3 ☐ PCR-GFD	☐ Ice			
Tracers	☐ W-E3821-MS ☐ W-E3821-DI	☐ Ice			
Pesticides	☐ W-PSNP-TQ ☐ W-E3821-MS ☐ W-CT-CL-R ☐ W-PCL-SQ-R ☐ W-E3821-DI ☐ W-CARB-AA	☐ Ice			
Phytoplankton	☐ DTY-QN-C ☐ PKD-QN-C	☐ Ice			

QA/QC Sample Information

Duplicate

Time Zone: <u>EST</u> CEN		Time: <u>1225</u>		Field ID: <u>G3TLAR0005 - DUP</u> (WIN ID_DUP)	
WIN DMT Activity ID:					
Parameter Suite	Parameter Methods	Preservation	Lot#	Expiration	pH Check
Preserved Nutrients	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☒ Ice ☒ H ₂ SO ₄	<u>9010030</u>	<u>4/10/20</u>	☒ < 2
Metals	☐ W-ICPMS ☐ W-ICP	☐ Ice ☐ HNO ₃			☐ < 2
Filtered Nutrients	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter	☒ Ice			
Chlorophyll	☒ CHLSUTE-W	☒ Ice			
Physical/Aggregate (Fresh)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-TDS	☒ Ice			
Physical/Aggregate (Marine)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Microbiology	☒ E Coli ☐ F Coli ☐ Enterococci	☒ Ice			
Molecular	☐ PCR-FILTER ☐ PCR-BacR ☐ PCR-GULL2 ☐ PCR-HF183 ☐ PCR-DG3 ☐ PCR-GFD	☐ Ice			
Tracers	☐ W-E3821-MS ☐ W-E3821-DI	☐ Ice			
Pesticides	☐ W-PSNP-TQ ☐ W-E3821-DI ☐ W-CT-CL-R ☐ W-PCL-SQ-R ☐ W-E-3821-MS ☐ W-CARB-AA	☐ Ice			

Blank

Time Zone: <u>EST</u> CEN		Time: <u>1215</u>		Field ID: <u>FB - 06242019</u> (Blank Type_DDMMYYYY)	
DI Source: <u>Carboy 1A</u>		Location: <u>C Building</u>			
☒ Field Blank ☐ Equipment Blank ☐ Field Cleaned Equipment Blank		Equipment ID: _____			
WIN DMT Blank Batch ID:					
Parameter Suite	Parameter Methods	Preservation	Lot#	Expiration	pH Check
Preserved Nutrients	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☒ Ice ☒ H ₂ SO ₄	<u>9010030</u>	<u>4/10/20</u>	☒ < 2
Metals	☐ W-ICPMS ☐ W-ICP	☐ Ice ☐ HNO ₃			☐ < 2
Filtered Nutrients	☒ W-PO4-F ☐ Field Filtered 0.45 µm Filter	☒ Ice			
Physical/Aggregate (Fresh)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-TDS	☒ Ice			
Physical/Aggregate (Marine)	☐ Alkalinity / W-F / W-TSS / TURBIDITY	☐ Ice			
Tracers	☐ W-E3821-MS ☐ W-E3821-DI	☐ Ice			
Pesticides	☐ W-PSNP-TQ ☐ W-E3821-DI ☐ W-CT-CL-R ☐ W-PCL-SQ-R ☐ W-E3821-MS ☐ W-CARB-AA	☐ Other			
Microbiology	☒ E.coli ☐ F. Coliform ☐ Enterococci				
Notes:			Place Label Here		
Signature: <u>[Signature]</u>			Date: <u>6/24/19</u>		

LIMS EVENT ID: _____ WIN Import ID: _____

Enter Field Parameters, Verify Station ID in LIMS DMT: _____ QC Station ID, Field Parameters in LIMS DMT: _____

Scan/Save Field Sheets _____ Migrate Data to WIN: _____ Verify Data in WIN: _____

(Initials/Date) (Initials/Date) (Initials/Date)

Group: A	Bottle Type:	P-1L	# of Bottles: 4	Preservative:	ICE	Enter Number of Bottles Sent to Lab	2
	ALKALINITY						
	TURBIDITY						
	W-CL-IC						
	W-COLOR						
	W-F						
	W-SO4-IC						
	W-TDS						
	W-TSS						
Group: A	Bottle Type:	BP-1L	# of Bottles: 4	Preservative:	ICE	Enter Number of Bottles Sent to Lab	2
	CHLSUITE-W						
Group: A	Bottle Type:	P-250ML-WO-THI	# of Bottles: 4	Preservative:	ICE	Enter Number of Bottles Sent to Lab	2
	ECOLI-18QT						
Group: A	Bottle Type:	P-500ML	# of Bottles: 4	Preservative:	ICE And H2SO4	Enter Number of Bottles Sent to Lab	2
	W-NH3				9010030		
	W-NO2NO3				4/9/19 W		
	W-S-A-TP				4/10/20		
	W-TKN						
	W-TOC						
Group: A	Bottle Type:	P-125ML	# of Bottles: 4	Preservative:	FILTER-ICE	Enter Number of Bottles Sent to Lab	2
	W-PO4-F						
Group: B	Bottle Type:	P-250ML-WO-THI	# of Bottles: 3	Preservative:	ICE	Enter Number of Bottles Sent to Lab	
	ECOLI-18QT						
Group: C	Bottle Type:	P-1L	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	ALKALINITY						
	TURBIDITY						
	W-CL-IC						
	W-COLOR						
	W-F						
	W-SO4-IC						
	W-TDS						
	W-TSS						
Group: C	Bottle Type:	P-250ML-WO-THI	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1

Group: C	Bottle Type: P-250ML-WO-THI	# of Bottles: 1	Preservative: ICE	Enter Number of Bottles Sent to Lab
	ECOL-18QT			
Group: C	Bottle Type: P-500ML	# of Bottles: 1	Preservative: ICE And H2SO4	Enter Number of Bottles Sent to Lab
	W-NH3		9010030	
	W-NO2NO3		4/10/20	
	W-S-A-TP			
	W-TKN			
	W-TOC			
Group: C	Bottle Type: P-125ML	# of Bottles: 1	Preservative: FILTER-ICE	Enter Number of Bottles Sent to Lab
	W-PO4-F			

