

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: 154177

RQ: 2019-08-19-25

Project: 2019 SMP

- Notes:** (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 10/19

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Ben Harden	12/11/19	0748	17.7	763.4	9.2	9.18	100.4		0.985	Ⓟ / F	Ⓟ / F
ICV	"	"	0805	17.9	763.9	9.5	9.56	101.1			Ⓟ / F	Ⓟ / F
CCV	"	"	1452	17.6	764.7	7.6	7.73	102.0			Ⓟ / F	Ⓟ / F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Ben Harden	12/11/19	0750	190618F	06/20	1000	1002	Ⓟ / F	Ⓟ / F
ICV	"	"	0752	290881	08/21	1000	102.6	Ⓟ / F	Ⓟ / F
CCV	"	"	1433	290881	08/21	100.0	100.6	Ⓟ / F	Ⓟ / F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Ben Harden	12/11/19	0754	190618B	12/20	7.03	18.7	7.03	-27.2	Ⓟ / F	Ⓟ / F
Calibr.	"	"	0759	191120A	06/21	4.00	17.1	4.00	148.2	Ⓟ / F	Ⓟ / F
Calibr.	"	"	0801	190618C	12/20	10.09	17.4	10.09	-193.4	Ⓟ / F	Ⓟ / F
ICV	"	"	0802	190618C	12/20	10.09	17.8	10.10	-193.9	Ⓟ / F	Ⓟ / F
CCV	"	"	1435	191120A	06/21	4.00	18.6	4.10	141.0	Ⓟ / F	Ⓟ / F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 166.2, slope from 4 to 7 175.4 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 10/19

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Ben Harden	12/11/19	0742	0.00	0.00	Ⓟ / F	Ⓟ / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID:

RQ-

Project:

- Notes:** (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.											P / F	L / F
ICV											P / F	L / F
CCV											P / F	L / F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. **Steady-state & Galvanic Sensors:** DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.								P / F	L / F
ICV								P / F	L / F
CCV								P / F	L / F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.						7.				P / F	L / F
Calibr.						4.				P / F	L / F
Calibr.						10.				P / F	L / F
ICV										P / F	L / F
CCV										P / F	L / F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) **Secchi Disk Line** / **Sonar** Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

Data Qualified?
Yes / No

January 2019

WIN Station Name: Sandy Branch @ SR22

Date: 12/11/2019
(mm/dd/yyyy)

WIN/ Field ID: G3TLHR0027

WBID: 1148

Latitude: 30.1431

(Decimal Degrees)

County: Gulf

ORG ID: 21FLTLHR

Longitude: 85.285

(Decimal Degrees)

Receiving Body of Water: Wetappo Creek

RQ-2019 08-19-2025

Sampling Team Members					WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
Ben Hadden					X	X	X	X	X

Sampling Equipment		
☒ Sample Container	☐ Van Dorn	☐ Pole
☐ Carboy	☐ Syringe	
☐ Dipper	☐ Other	
Depth Measured with YSI 1534647W		
Equipment ID 1534647W 154177		

Collection Method		
☒ Wading	☐ Canoe/Kayak	
☐ From Shore	☐ Electric Motor	
☐ Other	☐ Gasoline Motor	

Water Level: Dry/ Pooled/ Low / Normal / High / Flooded
Photos: Yes / No / Flow: No Flow / Intertidal / Flowing / NA
Meter ID# 154177
Matrix: Fresh / Marine / DI
Tide: Rising/ Falling/ Slack/ NA
Tide Times: High Low

Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Sp COND (µmhos/cm)	Temp. (C°)
EST	1015	0.2	0.1	0.2	6.20	60.9	3.83	0.02	46.1	14.1
CEN				S						

Biological Samples Collected: None HA SCI RPS Algae ID LVS LVI Other:

SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#:

Parameter Suite	Parameter Methods	Preservation	Lot#	Expiration	pH Check
Preserved Nutrients	☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☐ Ice ☐ H2SO4			☐ <2
Metals	☐ W-ICPMS ☐ W-ICP	☐ Ice ☐ HNO3			☐ <2
Filtered Nutrients	☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter	☐ Ice			
Chlorophyll	☐ CHLSUITE-W	☐ Ice			
Physical/Aggregate (Fresh)	☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CI-IC / W-COLOR / W-SO4-IC / W-TDS	☐ Ice			
Physical/Aggregate (Marine)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates	☐ MI-FW-QLDC	☐ Formalin			
Periphyton	☐ ALGAE-ID	☐ Ice			
Microbiology	☒ E Coli ☐ F Coli ☐ Enterococci	☒ Ice			
Molecular	☐ PCR-FILTER ☐ PCR-BacR ☐ PCR-GULL2 ☐ PCR-HF183 ☐ PCR-DG3 ☐ PCR-GFD	☐ Ice			
Tracers	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides	☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-E8321-DI ☐ W-PCL-SQ-R ☐ WE8321-MS ☐ W-CARB-AA	☐ Ice			
Phytoplankton	☐ DTY-QN-C ☐ PKD-QN-C	☐ Ice			

Notes:

Field/WIN ID →
Location ↓



G3TLHR0027



SANDY BRANCH @ SR22

Date: 12/11/19

Signature: [Signature]

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: QC Station ID, Field Parameters In LIMS DMT:

Scan/Save Field Sheets 12-13-19 Migrate Data to WIN: Verify Data In WIN:

Request Number: RQ-2019-0867

**Florida Department of Environmental Protection
Central Laboratory Submittal Form**

last revised Jan 25, 2018

Customer: TLH-ROC

Requester: Michael Eckles

Field Report Prepared By:

Project ID: SMP

Collected By: Woods/Harden

Sampling Agency: 8034

[illegible]

Relinquished By: Woods	Date/Time 12/11/19	Shipping Method HD	Number of Coolers Shipped: 1	Received By: 	Date/Time 12/11/19
---------------------------	-----------------------	-----------------------	---------------------------------	---	-----------------------

Group: A	Bottle Type:	P-1L	# of Bottles: 5	Preservative: ICE	Enter Number of Bottles Sent to Lab
	ALKALINITY				
	TURBIDITY				
	W-CL-IC				
	W-COLOR				
	W-F				
	W-SO4-IC				
	W-TDS				
	W-TSS				
Group: A	Bottle Type:	BP-1L	# of Bottles: 5	Preservative: ICE	Enter Number of Bottles Sent to Lab
	GHLSUTTE-W				
Group: A C	Bottle Type:	P-250ML-WO-THI	# of Bottles: 5	Preservative: ICE	Enter Number of Bottles Sent to Lab 2
	ECOL-18QT				
Group: A	Bottle Type:	P-500ML	# of Bottles: 5	Preservative: ICE And H2SO4	Enter Number of Bottles Sent to Lab
	W-NH3				
	W-NO2NO3				
	W-S-A-TP				
	W-TKN				
	W-TOC				
Group: A	Bottle Type:	P-125ML	# of Bottles: 5	Preservative: FILTER-ICE	Enter Number of Bottles Sent to Lab
	W-PO4-F				
Group: B	Bottle Type:	P-1L	# of Bottles: 1	Preservative: ICE	Enter Number of Bottles Sent to Lab
	ALKALINITY				
	TURBIDITY				
	W-CL-IC				
	W-COLOR				
	W-F				
	W-SO4-IC				
	W-TDS				
	W-TSS				
Group: B	Bottle Type:	P-250ML-WO-THI	# of Bottles: 1	Preservative: ICE	Enter Number of Bottles Sent to Lab
	ECOL-18QT				
Group: B	Bottle Type:	P-500ML	# of Bottles: 1	Preservative: ICE And H2SO4	Enter Number of Bottles Sent to Lab