

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: 154177

RQ-

Project:

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 9/30/19

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	J. Burtchett	11/25/19	759	19.8	756.0	9.1	9.09	99.6	-	1.02	P/F	L/F
ICV	"	"	809	17.8	756.0		9.49	99.7	-	-	P/F	L/F
CCV	"	"		17.8							P/F	L/F
CCV	"	"	1537	21.1	754.5	8.8	8.95	100.7			P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	J. Burtchett	11/25/19	801	190618F	06/20	1000	1000	P/F	L/F
ICV	"	"	811	290881	08/21	100	102	P/F	L/F
CCV	"	"	1539	190618f	06/20	1000	1004	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J. Burtchett	11/25/19	803	190618B	12/20	7.02	17.8	7.03	-29.3	P/F	L/F
Calibr.	"	"	805	190618D	12/20	4.00	17.7	4.00	145.5	P/F	L/F
Calibr.	"	"	807	190618C	12/20	10.09	17.8	10.09	194.4	P/F	L/F
ICV	"	"	813	190618C	12/20	10.09	18.1	10.01	192.1	P/F	L/F
CCV			1541	190618D	12/20	4.00	19.5	4.01	144.8	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU;

mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 9/30/19

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	J. Burtchett	11/25/19	757	0.000	0.000	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID:

RQ-

Project:

- Notes:** (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.											P / F	L / F
ICV											P / F	L / F
CCV											P / F	L / F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. **Steady-state & Galvanic Sensors:** DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.								P / F	L / F
ICV								P / F	L / F
CCV								P / F	L / F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.						7.				P / F	L / F
Calibr.						4.				P / F	L / F
Calibr.						10.				P / F	L / F
ICV										P / F	L / F
CCV										P / F	L / F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification:

If NO, what will be used? (circle one) **Secchi Disk Line / Sonar** Unique ID: ; Date of last verification:

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2019

Data Qualified?

Yes / No

WIN Station Name: **Sugar Creek @ Suwannee River**

Date: **11/25/19**

WIN/ Field ID: **G1TLHR0047**

WBID: **3389**

Latitude: **30.39809905**

County: **HAMILTON**

ORG ID: **21FTLHR**

Longitude: **-82.93431**

Receiving Body of Water: **Suwannee River**

RQ- **2019-11-16-54**

Sampling Team Members					WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
Burt Harden									
Jade Butchett									

Sampling Equipment		
☒ Sample Container	☐ Van Dorn	☐ Pole
☐ Carboy	☐ Syringe	
☐ Dipper	☐ Other	

Depth Measured with _____

Equipment ID _____

Collection Method	
☒ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: Dry/ Pooled/ Low / Normal / High / Flooded

Meter ID# **154177**

Matrix: Fresh / Marine / DI

Photos: Yes / No

Flow: No Flow / Intestinal / Flowing / NA

Tide: Rising/ Falling/ Slack/ NA

Tide Times: High Low

Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Sp COND (µmhos/cm)	Temp. (C°)
EST	1010	0.1	0.05	0.1	11.12	97.5	7.60	0.09	184.9	9.6
CEN				(S)						

Biological Samples Collected: None HA SCI RPS Algae ID LVS LVI Other:

SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#:

Parameter Suite	Parameter Methods	Preservation	Lot#	Expiration	pH Check
Preserved Nutrients	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☒ Ice ☒ H2SO4	SA928100	07/22/20	☒ <2
Metals	☐ W-ICPMS ☐ W-ICP	☐ Ice ☐ HNO3			☐ <2
Filtered Nutrients	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter	☐ Ice			
Chlorophyll	☒ CHLSUITE-W	☒ Ice			
Physical/Aggregate (Fresh)	☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS	☒ Ice			
Physical/Aggregate (Marine)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates	☐ MI-FW-QLDC	☐ Formalin			
Periphyton	☐ ALGAE-ID	☐ Ice			
Microbiology	☒ E Coli ☐ F Coli ☐ Enterococci	☒ Ice			
Molecular	☐ PCR-FILTER ☒ PCR-BacR ☒ PCR-GULL2 ☐ PCR-HF183 ☒ PCR-DG3 ☒ PCR-GFD	☒ Ice			
Tracers	☒ W-E8321-DI ☒ W-E8321-MS	☒ Ice			
Pesticides	☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-E8321-DI ☐ W-PCL-SQ-R ☐ WE8321-MS ☐ W-CARB-AA	☐ Ice			
Phytoplankton	☐ DTY-QN-C ☐ PKD-QN-C				

Notes:

Field/WIN ID →

Location ↓

Sugar Creek @ Suwannee River



G1TLHR0047

Signature: *[Signature]*

Date: **11/25/19**

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: QC Station ID, Field Parameters In LIMS DMT:

Scan/Save Field Sheets **12-9-18** (Initials/Date) Migrate Data to WIN: (Initials/Date) Verify Data In WIN: (Initials/Date)

Group: A	Bottle Type:	P-1L	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	ALKALINITY						
	TURBIDITY						
	W-CL-IC						
	W-COLOR						
	W-F						
	W-SO4-IC						
	W-TDS						
	W-TSS						

Group: A	Bottle Type:	BP-1L	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	CHLSUITE-W						

Group: A	Bottle Type:	P-120ML-WC-THI	# of Bottles: 2	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	ECOLI-18QT						

Group: A	Bottle Type:	QPCR-P-500ML	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	PCR-BACR						
	PCR-DG3						
	PCR-HF183						

Group: A	Bottle Type:	BG-500ML	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	W-E8321-DI						
	W-E8321-MS						

Group: A	Bottle Type:	P-500ML	# of Bottles: 1	Preservative:	ICE And H2SO4	Enter Number of Bottles Sent to Lab	1
	W-NH3						
	W-NO2NO3						
	W-S-A-TP						
	W-TKN						
	W-TOC						

SA 9203100

Group: A	Bottle Type:	P-125ML	# of Bottles: 1	Preservative:	FILTER-ICE	Enter Number of Bottles Sent to Lab	1
	W-PO4-F						

Group: B	Bottle Type:	P-1L	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab	
	ALKALINITY						
	TURBIDITY						
	W-CL-IC						
	W-COLOR						

Group: B	Bottle Type: P-1L	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab
	W-F				
	W-SO4-IC				
	W-TDS				
	W-TSS				
Group: B	Bottle Type: P-120ML-WC-THI	# of Bottles: 2	Preservative:	ICE	Enter Number of Bottles Sent to Lab
	ECOLI-18QT				
Group: B	Bottle Type: QPCR-P-500ML	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab
	PCR-BACR				
	PCR-DG3				
	PCR-HF183				
Group: B	Bottle Type: BG-500ML	# of Bottles: 1	Preservative:	ICE	Enter Number of Bottles Sent to Lab
	W-E8321-DI				
	W-E8321-MS				
Group: B	Bottle Type: P-500ML	# of Bottles: 1	Preservative: ICE And H2SO4		Enter Number of Bottles Sent to Lab
	W-NH3				
	W-NO2NO3				
	W-S-A-TP				
	W-TKN				
	W-TOC				
Group: B	Bottle Type: P-125ML	# of Bottles: 1	Preservative: FILTER-ICE		Enter Number of Bottles Sent to Lab
	W-PO4-F				

