

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID:

Desi

RQ:

2020-01-06-44

Project:

BMAP 2

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 10/2019

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	K. Appleby	1/9/20	749	20.7	772.8	9.1	9.07	101.7	—	1.04	P/F	L/F
ICV	↓	↓	750	21.0	772.7	9.0	9.06	101.8	—	1.04	P/F	L/F
CCV	↓	↓	1522	21.4	771.9	9.0	8.94	101.1	—	1.04	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	K. Appleby	1/9/20	756	191018	10/2020	1000	1000	P/F	L/F
ICV	↓	↓	758	190711A	7/2020	100	103.4	P/F	L/F
CCV	↓	↓	1542	191018	10/2020	1000	997	P/F	L/F
CCV			1547					P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	K. Appleby	1/9/20	803	2902778	1/2021	7.00	22.3	7.00	-32.6	P/F	L/F
Calibr.	↓	↓	805	2907062	7/2021	4.00	22.5	4.00	148.4	P/F	L/F
Calibr.	↓	↓	808	2904E91	10/2020	10.00	22.5	10.00	-203.3	P/F	L/F
ICV	↓	↓	810	2902778	1/2021	7.00	22.6	7.05	-32.4	P/F	L/F
CCV	↓	↓	1550	2907062	7/2021	4.00	23.2	4.08	142.8	P/F	L/F
CCV	↓	↓	1553	2904E91	10/2020	10.00	23.2	10.03	-207.7	P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 $\pm$ 50; mV pH 4 Range +180 $\pm$ 50; mV pH 10 Range -180 $\pm$ 50;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 10/2019

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	K. Appleby	1/9/20	812	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: