

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: ProDSS Desi RQ: 2020-08-10-12 Project: St Augustine

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 6 18 2020

DO DEP SOP FT 1500	Name	Date	Time CT/ET	Temp °C	Barometer mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	K McDonald	8-13-20	0853	23.8	762.1	8.5	8.5	100.3	-	1.057	P/F	L/F
ICV			0856	23.7	762.1	8.5	8.5	100.3	-	1.057	P/F	L/F
CCV			1428	23.4	760.6	8.5	8.45	99.2	-	1.057	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec Cond. FT 1200	Name	Date	Time CT/ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	K McDonald	8-13-20	0859	200407C	04/2021	1000	1000	P/F	L/F
ICV			0902	200407D	04/2021	100	100.8	P/F	L/F
CCV			1434	191025C	11/2020	50,000	49,384	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT/ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	K McDonald	8-13-20	0905	2001C31	01/2022	7.00	24.8	7.00	26.3	P/F	L/F
Calibr.			0906	2910G29	10/2021	4.00	24.9	4.01	157.7	P/F	L/F
Calibr.			0908	2002911	08/2021	10.00	24.8	10.00	-191.3	P/F	L/F
ICV			0910	2910G29	10/2021	4.00	24.8	63.93	159.5	P/F	L/F
CCV			1436	191025C	11/2020	km				P/F	L/F
CCV				2002911	08/2021	10.00	24.3	9.87	-187.6	P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 .

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 6 18 20

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT/ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	K McDonald	8-13-20	0911	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: