From:	Hanna Taylor
To:	Mirabito, Anthony
Subject:	RE: G2006009 GAINESVILLE
Date:	Thursday, July 2, 2020 8:15:46 AM

Good morning,

Sorry about that. The first incubator was from 1455 to 1655 and the second from 1655 to 1305

Best,
Hanna

From: Mirabito, Anthony [mailto:Anthony.Mirabito@FloridaDEP.gov]
Sent: Wednesday, July 1, 2020 11:32 PM
To: Hanna Taylor
Subject: RE: G2006009 GAINESVILLE

Hello Hanna,

It looks like the incubation out time is missing from the report. Would you be able to send that information when you get the chance? Let me know if you have any questions.

Thanks!

Anthony Mirabito
Florida Department of Environmental Protection
FDEP Laboratory - Biology Program
Environmental Specialist I
Anthony.Mirabito@FloridaDEP.gov
Office: 850-245-8193

From: Hanna Taylor <htaylor@aellab.com>
Sent: Tuesday, June 30, 2020 8:45 AM
To: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Subject: G2006009 GAINESVILLE

Good day,

Attached is a waste water report and EDD

Best,

Hanna Taylor,
Project Manager
Advanced Environmental Laboratories, Inc.
Florida's Largest Laboratory Network!

Voice: 352-377-2349, FAX: 352-395-6639
Website: www.aellab.com

NOTICE: This communication may contain privileged or other confidential information. If you are not the intended recipient, or believe that you have received this communication in error, please do not print, copy, retransmit, disseminate, or otherwise use the information. Also, please indicate to the sender that you have received this communication in error, and delete the copy you received.
Website: www.aellab.com