

Chemical Analysis Report

NE-DIST-2020-06-10-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Suwannee North 2020**
Request ID: **RQ-2020-06-08-34**
Customer: **NE-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
8800 Baymeadows Way West
Suite 100
Jacksonville, FL 32256
Attn: Jeremy Parrish

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator

Date Certified: 24-JUN-2020 09:24



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Clear Lake Center

Collection Date/Time: 06/09/2020 11:20

Field ID: G2NE1134

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2175389	EPA 180.1 Rev. 2.0	Turbidity	0.25		NTU	P383043	
	EPA 300.0 Rev. 2.1	Chloride	17		mg Cl/L	P383190	
		Sulfate	13		mg SO4/L	P383194	
	SM 2120 B-2011	Color (true)	2.5	U	PCU	P383048	
	SM 2320 B-2011	Total Alkalinity	0.65	U	mg CaCO3/L	P383113	
	SM 2540 C-2011	TDS	47		mg/L	P383513	
	SM 2540 D-2011	TSS	2	U	mg/L	P383447	
	SM 4500 F-C-2011	Fluoride	0.045	I	mg F/L	P383116	
2175391	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P383170	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.080	U	mg N/L	P383074	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P383063	
	EPA 365.1 Rev. 2.0	Total-P	0.003	I	mg P/L	P383226	
	SM 5310 B-2011	Organic Carbon	1.3		mg C/L	P383118	
2175393	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P383053	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Clear Lake Center

Collection Date/Time: 06/09/2020 11:50

Field ID: FB 06092020

Matrix: W-FIELD-BK

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2175390	EPA 180.1 Rev. 2.0	Turbidity	0.10	U	NTU	P383043	
	EPA 300.0 Rev. 2.1	Chloride	0.20	U	mg Cl/L	P383190	
		Sulfate	0.20	U	mg SO4/L	P383194	
	SM 2120 B-2011	Color (true)	2.5	U	PCU	P383048	
	SM 2320 B-2011	Total Alkalinity	0.65	U	mg CaCO3/L	P383113	
	SM 2540 C-2011	TDS	15	U	mg/L	P383511	
	SM 2540 D-2011	TSS	2	U	mg/L	P383445	
	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P383116	
2175392	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P383171	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.080	U	mg N/L	P383074	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P383063	
	EPA 365.1 Rev. 2.0	Total-P	0.002	U	mg P/L	P383226	
	SM 5310 B-2011	Organic Carbon	0.50	U	mg C/L	P383118	
2175394	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P383053	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P383043

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P383190

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P383194

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P383170

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P383171

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P383074

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P383063

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P383226

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P383053

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 2120 B-2011
Batch ID: P383048

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-2011
Batch ID: P383113

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-2011
Batch ID: P383511

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 C-2011
Batch ID: P383513

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P383445

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P383447

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P383116

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P383118

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P383190

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	99.1		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P383194

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	102		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P383170

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	99.5		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P383171

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	99.8		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P383074

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	102		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P383063

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO ₂ NO ₃ -N	97.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P383226

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	105		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P383053

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.5		P	90 - 110

Reference Method: SM 2120 B-2011
Batch ID: P383048

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	99.4		P	85 - 115

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 2320 B-2011
Batch ID: P383113

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	99.5		P	93 - 104

Reference Method: SM 4500 F-C-2011
Batch ID: P383116

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	101		P	90 - 106

Reference Method: SM 5310 B-2011
Batch ID: P383118

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	105		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P383190

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2175227	Chloride	101		P	90 - 110
2175387	Chloride	103		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P383194

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2175227	Sulfate	101		P	90 - 110
2175387	Sulfate	103		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P383170

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2175391	Ammonia-N	101	103	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P383171

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2175433	Ammonia-N	98.9	100	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P383074

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2175388	Kjeldahl Nitrogen	98.3	97.0	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P383063

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2175388	NO2NO3-N	98.0	98.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P383226

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2175388	Total-P	108	106	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P383053

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2175393	O-Phosphate-P	95.8	95.8	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P383116

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2175565	Fluoride	102	99.8	P/P	94 - 107

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 5310 B-2011
Batch ID: P383118

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2175555	Organic Carbon	103	102	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P383043

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2175448	Turbidity	1.41	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P383190

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2175387	Chloride	0.127	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P383194

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2175387	Sulfate	0.168	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P383170

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2175391	Ammonia-N	1.71	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P383171

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2175433	Ammonia-N	1.31	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P383074

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2175388	Kjeldahl Nitrogen	0.944	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P383063

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2175388	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P383226

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2175388	Total-P	1.44	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P383053

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2175393	O-Phosphate-P	0.0	Spike	P	0 - 20

Reference Method: SM 2120 B-2011
 Batch ID: P383048

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2175432	Color (true)	0.732	Sample	P	0 - 20

Reference Method: SM 2320 B-2011
 Batch ID: P383113

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2175565	Total Alkalinity	0.256	Sample	P	0 - 8.1

Reference Method: SM 2540 C-2011
 Batch ID: P383511

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2175076	TDS	2.88	Sample	P	0 - 10

Reference Method: SM 2540 C-2011
 Batch ID: P383513

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2175418	TDS	1.02	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
 Batch ID: P383116

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2175565	Fluoride	1.38	Spike	P	0 - 2.8

Reference Method: SM 5310 B-2011
 Batch ID: P383118

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2175555	Organic Carbon	0.203	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A99250

Included Lab Sample IDs: 2175389, 2175390

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	101	P/P	95 - 105

Reference Method: SM 2120 B-2011

Run ID: A99251

Included Lab Sample IDs: 2175389, 2175390

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	98.5	100	P/P	90 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A99252

Included Lab Sample IDs: 2175393, 2175394

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	98.1	98.0	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A99259

Included Lab Sample IDs: 2175391, 2175392

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	97.5	98.0	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A99280

Included Lab Sample IDs: 2175389, 2175390

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	94.8	98.3	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A99280

Included Lab Sample IDs: 2175389, 2175390

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	103	103	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A99281

Included Lab Sample IDs: 2175391, 2175392

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	103	104	P/P	90 - 110
Organic Carbon	104	103	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A99295

Included Lab Sample IDs: 2175389, 2175390

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	97.8	98.0	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A99295

Included Lab Sample IDs: 2175389, 2175390

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Sulfate	99.4	99.6	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A99309

Included Lab Sample IDs: 2175391, 2175392

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	100	101	P/P	90 - 110
Ammonia-N	101	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A99330

Included Lab Sample IDs: 2175391, 2175392

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	102	97.5	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A99366

Included Lab Sample IDs: 2175391, 2175392

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	97.0	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision SMP	MS
					LCS	
EPA 180.1 Rev. 2.0	Turbidity				1.41	
EPA 300.0 Rev. 2.1	Chloride	99.1	103	101	0.127	
	Sulfate	102	103	101	0.168	
EPA 350.1 Rev. 2.0	Ammonia-N	99.5	101	103		1.71
	Ammonia-N	99.8	98.9	100		1.31
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	102	98.3	97.0		0.944
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	97.5	98.0	98.0		0.0
EPA 365.1 Rev. 2.0	Total-P	105	108	106		1.44
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	97.5	95.8	95.8		0.0
SM 2120 B-2011	Color (true)	99.4			0.732	
SM 2320 B-2011	Total Alkalinity	99.5			0.256	
SM 2540 C-2011	TDS				2.88	
	TDS				1.02	
SM 4500 F-C-2011	Fluoride	101	102	99.8		1.38
SM 5310 B-2011	Organic Carbon	105	103	102		0.203

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2175389, 2175390
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2175389, 2175390
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2175389, 2175390
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2175391, 2175392
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2175391, 2175392
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2175391, 2175392
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2175391, 2175392
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2175393, 2175394
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2175389, 2175390
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2175389, 2175390
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2175389, 2175390
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2175389, 2175390
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2175389, 2175390
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2175391, 2175392

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	06/10/2020			06/10/2020 15:06	Yen Ta	2175389, 2175390
EPA 300.0 Rev. 2.1	06/10/2020			06/12/2020 22:13	Elliott Rodriguez	2175389
	06/10/2020			06/12/2020 22:26	Elliott Rodriguez	2175390
EPA 350.1 Rev. 2.0	06/10/2020			06/13/2020 14:02	Ping Hua	2175391
	06/10/2020			06/13/2020 14:43	Ping Hua	2175392
EPA 351.2 Rev. 2.0	06/10/2020	06/11/2020 10:02	Sima Feizi	06/15/2020 13:52	Jhamieka S. Greenwood	2175391
	06/10/2020	06/11/2020 10:02	Sima Feizi	06/15/2020 13:53	Jhamieka S. Greenwood	2175392
EPA 353.2 Rev. 2.0	06/10/2020			06/11/2020 08:22	Beverly Y. Sanders	2175391
	06/10/2020			06/11/2020 08:23	Beverly Y. Sanders	2175392
EPA 365.1 Rev. 2.0	06/10/2020	06/15/2020 15:28	Yen Ta	06/17/2020 07:55	Lipi Saha	2175391
	06/10/2020	06/15/2020 15:28	Yen Ta	06/17/2020 07:56	Lipi Saha	2175392
EPA 365.1 Rev. 2.0 dissolved	06/10/2020			06/10/2020 15:58	Carlos Miranda	2175393
	06/10/2020			06/10/2020 16:00	Carlos Miranda	2175394
SM 2120 B-2011	06/10/2020			06/10/2020 16:09	Blanca Fach	2175389
	06/10/2020			06/10/2020 16:10	Blanca Fach	2175390
SM 2320 B-2011	06/10/2020			06/11/2020 19:58	Dale L. Simmons	2175389
	06/10/2020			06/11/2020 20:07	Dale L. Simmons	2175390
SM 2540 C-2011	06/10/2020			06/12/2020 15:15	Yijie Li	2175389, 2175390
SM 2540 D-2011	06/10/2020			06/12/2020 15:15	Yijie Li	2175389, 2175390
SM 4500 F-C-2011	06/10/2020			06/11/2020 19:58	Dale L. Simmons	2175389
	06/10/2020			06/11/2020 20:07	Dale L. Simmons	2175390
SM 5310 B-2011	06/10/2020			06/11/2020 15:31	David Boose	2175391
	06/10/2020			06/11/2020 16:07	David Boose	2175392