

Login Checklist

RQ-2020-08-17-24

Login Station ID: #4

Shipping Method: FedEx UPS | HD | Greyhound

Date/Time of Receipt: 08/13/2020 @ 11:15

*Cooler Temperature	Samples Frozen?	* All Samples in Cooler Preserved with		Number of Sample Containers in Cooler	Evidence Tape				Waybill Tracking #
					Present?		*Intact?		
	YES	HNO ₃	Formalin		Yes	No	Yes	No	
1.2				10		x			1278 4014 2791

COOLER CHECK

*All samples received on ice unless otherwise noted. HNO3 and Formalin preserved samples do NOT have a temperature requirement.

- If the temperature of a cooler exceeds 6.0 oC (above 10.0 oC for W-1-4-DIOX and microbiology) or received without ice, identify affected samples in an NCR. NCR# _____
- DOH coolers can exceed 2.0 °C above the standard temperature (6.0 or 10.0 °C for W-1-4-DIOX) without login confirmation from customer for samples requiring ice preservation; identify affected samples in an NCR. NCR# _____
- If cooler or samples are received with a damaged Evidence Seal; identify the affected samples in an NCR. NCR# _____
- If coolers or samples are received late; identify the affected samples in an NCR. NCR# _____

Chain of Custody/ Field Sheet(s) Included?

Yes ☒ No _____

Micro-Biology Overflow Chain of Custody/ Field Sheet(s) Included?

Yes ☒ No _____

CONTAINER CHECK

Evidence Tape on Bottles?

Yes _____ No _____ NA X

Evidence Tape intact? Yes _____ No _____

If Criminal, photographs taken?

Yes _____ No _____

Containers intact? Yes ☒ No _____

Caps on tight?

Yes ☒ No _____

Sufficient Sample Volume:

Yes ☒ No _____

If NO is checked above, generate an NCR listing affected sample IDs in its appropriate category. NCR# _____

CHLORINE CHECK REQUIRED? (Blue dot on container)

Yes _____ No X Init: Jylup

Chlorine detected? (One container checked per Field ID)

Yes _____ No _____ Init: _____

If chlorine is detected, generate an NCR listing affected sample IDs . NCR# _____

PRESERVATION CHECK (EXCLUDING SAMPLES SENT TO OVERFLOW LABS, "OV-")

**Acid Preserved Samples: pH< 2.0?

Yes ☒ No _____ NA _____ Init: Jylup

**W-1-4-DIOX (Green dot on container) preserved to pH<4.0?

Yes _____ No _____ NA X Init: Jylup

If samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

Coolers Unpacked/Checked by: Jylup Reynolds

Date: 08/13/2020

Event contains NCRs (Y/N)? None

Event ID: Extra Priority allocations NEROC
NE-DIST-2020-08-13-01 (SMP)

Page 1 of 2, Date Received: 13-AUG-2020 11:15

Login Checklist

Login Preservation Equipment

Lot # & Exp: Date

Additional Comments:

Infrared Thermometer ID:	#4 Temp Gun EXP:11/05/2020
pH Test Strips 0 – 6	
pH Paper 0 – 3	
pH Paper 0 – 13	231019 EXP:11/01/2022
Chlorine Test Strips	8261 EXP:07/2021

ENCORE SAMPLES

S-VOC-MS samples in Encores must be frozen within 48 hours of collection.

Were Encore samples received? Yes _____ Init: _____

Date and time placed in freezer _____

Were all samples placed in freezer within 48 hours of collection? Yes _____ No _____

If no, were samples shipped on dry ice? Yes _____ No _____

If S-VOC-MS samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

FOR CLEAN LAB USE ONLY – Total Mercury Preservation Check(W-HG-AF;W-HG-AF-F):

Samples preserved within 48 hours? Yes _____ No _____

Date and time samples preserved: _____

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 8/12/2020 10:55 **Collected By:** Jeremy Parrish, Kelly McDonald
Customer: NE-DIST **Sampling Agency:** FDEP
RQ: RQ-2020-08-17-24 **Project ID:** SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 1

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
Cedar Cr Blanding Blvd Br Rt 21	 20030083
Ortega R Br Roosevelt Blvd Us 17	 20030077

RELINQUISHED BY(SIGNATURE):

Kelly A McDonald

DATE/TIME: 8/12/2020 1600

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:

Tyler Reynolds

Rec'd/Inspected Date:

08/13/2020

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Attach "X" this box if there are qualified data on this page.

Meter ID: SHACH

RQ: 2020-08-17-74

Project: ICW

- Notes: (1) Always wait for meter to stabilize before recording any readings.
(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 6/18/2020

DO	Name	Date	Time	Temp	Baro-	D.O.	Meter	% DO	Probe	Probe	Pass	Lab
DEP SOP			CT-ET	°C	meter	Chart	D.O.		Charge	Gain		
FT 1500					mmHg	mg/L	mg/L				Fail	Field
Calibr.	J.P. msh	8/12/20	812	23.7	762.5	8.5	8.5	100.3	-	1.028	P/F	L/F
ICV	L	8/12/20	813	23.5	762.4	8.5	8.55	100.6	-	1.025	P/F	L/F
CCV	J.P. msh	8/12/20	1300	25.2	762	8.2	8.26	100.5	-	1.025	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4, DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec.	Name	Date	Time	Lot #	Expir.	Standard	Meter	Pass	Lab
Cond.			CT-ET		Date	μ mhos/cm	Reading		
FT 1200							μ mhos/cm	Fail	Field
Calibr.	J.P. msh	8/12/20	814	2004070	4/21	1000	1001	P/F	L/F
ICV	L	8/12/20	815	2004070	4/21	CT	102	P/F	L/F
CCV	J.P. msh	8/12/20	1302	191025A	11/20	14.3	1404	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH	Name	Date	Time	Lot #	Expir.	pH	Temp	Meter	mV	Pass	Lab
DEP SOP			CT-ET		Date	Buffer	°C	Reading			
FT 1100						SU		SU		Fail	Field
Calibr.	J.P. msh	8/12/20	816	200165	1/22	7	25.7	7.0	25.1	P/F	L/F
Calibr.			817	200165	10/21	4	25.7	6.01	10.5	P/F	L/F
Calibr.			819	200165	8/20	10	25.7	6.01	20.1	P/F	L/F
ICV	L	8/12/20	820	200165	10/21	5	25.5	5.0	13.3	P/F	L/F
CCV	J.P. msh	8/12/20	1305	200294	8/21	10	24.6	9.70	20.5	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 .

If mV are recorded: slope from 7 to 10 _____; slope from 4 to 7 _____ (both must be between 185 and 190 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not available)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? Secchi Disk Line / Sonar Unique ID: _____ Date of last verification: _____

Depth Sensor	Name	Date	Time	Calibrated	ICV Value	Pass	Lab
(Daily Calibration & ICV)			CT-ET	Value (0.00 or Offset) meters	meters	Fail	Field
Pressure mode in air	J.P. msh	8/11/20	830	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End

of Day CCV

Yes No

WIN Station Name: Cedar Cr Blanding Blvd Br Rt 21

Date: 8/12/2020

WIN/Field ID: 20030083

WBID: 2213P2

ENR:

Latitude: 30.27308

County: DUVAL

Org ID: 21FLA

Longitude: -81.73333

Receiving Body of Water: Lower St. Johns

RQ: 2020-08-17-24

Sampling Team Members	WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒ Jeremy Parrish	☒	☒	☒	☒	☒
☒ Kelly McDonald	☒	☒	☒	☒	☒
☐	☐	☐	☐	☐	☐
☐	☐	☐	☐	☐	☐

Sampling Equipment	
☒ Sample Container	☐ Carboy Carboy EquipID
☐ Dipper	☐ Van Dorn YDEquipID
☐ Syringe	☒ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☒ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded		Meter ID# STICH	
Photos: No	Flow: FLOW	Tide: FALLING	Tide Times: High: 6 AM Low: NOON
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:			
SBIO ID Numbers:			
Notes:			

Duplicate Sample

Time Zone: EST	Time: N/A	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time: N/A	Field ID: N/A	(Blank Type_DDMMYYYY)
DI Source: N/A	Location: N/A		
DI Type: N/A	Equipment ID: N/A		

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: QC Station ID, Field Parameters In LIMS DMT:

Scan/Save Field Sheets Migrate Data to WIN: Verify Data In WIN:



RQ-2020-08-17-24

Customer: WQAP

Project ID: SMP

Collected By: Jeremy Parrish

Field Report Prepared By: Jeremy Parrish

Sampling Agency: FDEP

Place Label Here						Comments: Microbiology sample(s) submitted to contract laboratory				
WIN ID/Field ID:  20030083										
Location: Cedar Cr Blanding Blvd Br Rt 21										
Matrix: Marine						(Check one) ☒ GRAB ☐ COMPOSITE				
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)
8/12/2020	10:10 EST	4.83	60.5	26.8	7.24	0.3	0.3 ☐ VOB (S)	0.644	157.1	0.07
Central Lab please use top row for login purposes										
EST										
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☒ Ice ☒ H2SO4*		☒ <2	1	D
		Lot #		Sa9354050		Exp:		12/20/2020		
Metals (P-500ML)		☒ W-ICPMS ☒ W-ICP ☒ W-HARD				☒ Ice ☒ HNO3*		☐ <2	1	D
		Lot #		Na9354060		Exp:		12/20/2020		
Filtered Nutrients (P125ML)		☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter				☒ Ice			1	D
		Filter Lot #		R9ba86875						
		Syringe Lot #								
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	D
Physical/Aggregate (Fresh) (P-1L)		☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CHC / W-COLOR / W-SO4-IC / W-TDS				☐ Ice				
Physical/Aggregate (Marine) (P-1L)		☒ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☒ Ice			1	D
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC				☐ Formalin				
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice				
Microbiology (Contract Lab Bottle)		☐ E Coli ☐ F Coli ☒ Enterococci ☒ Contract Lab				☒ Ice			1	D
Molecular (QPCR-P-500ML)		☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD ☐				☐ Ice				
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice				
Pesticides (BG-500ML) (BG-1L)		☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-SQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution)				☐ Ice				
						☐ Ice				
Name: Jeremy Parrish Kelly McDonald					Signature: Kelly McDonald					Date: 8/12/20

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution



RQ-2020-08-17-24

Customer: WQAP

Project ID: SMP

Collected By: Jeremy Parrish, Kelly McDonald

Field Report Prepared By: Jeremy Parrish

Sampling Agency: FDEP

Place Label Here						Comments: Microbiology sample(s) submitted to contract laboratory					
WIN ID/Field ID:											
		20030077									
Location:		Ortega R Br Roosevelt Blvd Us 17									
Matrix: Marine						(Check one) ☒ GRAB ☐ COMPOSITE					
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPTth)	
8/12/2020	10:55 EST	4.62	59.4	28.4	6.89	0.3	0.4 ☐ VOB (S)	2.534	535	0.26	
Central Lab please use top row for login purposes											
	11:00 EST	4.68	60.2	28.3	6.91	2			688	0.33	
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group	
Preserved Nutrients (P-500ML)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☒ Ice ☒ H2SO4*		☒ <2	1	D	
		Lot #	Sa9354050		Exp:	12/20/2020					
Metals (P-500ML)		☒ W-ICPMS ☒ W-ICP ☒ W-HARD				☒ Ice ☒ HNO3*		☒ <2	1	D	
		Lot #	Na9301040		Exp:	10/28/2020					
Filtered Nutrients (P125ML)		☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter				☒ Ice			1	D	
		Filter Lot #	R9ba86875								
		Syringe Lot #									
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	D	
Physical/Aggregate (Fresh) (P-1L)		☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CH-IC / W-COLOR / W-SO4-IC / W-TDS				☐ Ice					
Physical/Aggregate (Marine) (P-1L)		☒ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☒ Ice			1	D	
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC				☐ Formalin					
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice					
Microbiology (Contract Lab Bottle)		☐ E Coli ☐ F Coli ☒ Enterococci ☒ Contract Lab				☒ Ice			1	D	
Molecular (QPCR-P-500ML)		☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD ☐				☐ Ice					
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice					
Pesticides (BG-500ML) (BG-1L)		☐ W-E8321-DI ☐ W-E8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-SQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution)				☐ Ice					
						☐ Ice					
Name: Jeremy Parrish Kelly McDonald				Signature: 				Date: 8/12/20			

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End

of Day CCV

Yes No

WIN Station Name: Ortega R Br Roosevelt Blvd Us 17

Date: 8/12/2020

WIN/Field ID: 20030077

WBID: 2213P1

ENR:

Latitude: 30.27342

County: DUVAL

Org ID: 21FLA

Longitude: -81.71708

Receiving Body of Water: Lower St. Johns

RQ- 2020-08-17-24

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Jeremy Parrish		☒	☒	☒	☒
☒	Kelly McDonald	☒		☒	☒	☒
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy CarboyEquipID
☐ Dipper	☐ Van Dorn VDEquipID
☐ Syringe	☒ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☒ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded		Meter ID# STITCH	
Photos: No	Flow: FLOW	Tide: FALLING	Tide Times: High: 6 AM Low: NOON
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:			
SBIO ID Numbers:			
Notes:			

Duplicate Sample

Time Zone: EST	Time: N/A	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time: N/A	Field ID: N/A	(Blank Type_DDMMYYYY)
DI Source: N/A	Location: N/A		
DI Type: N/A	Equipment ID: N/A		

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: QC Station ID, Field Parameters In LIMS DMT:

Scan/Save Field Sheets Migrate Data to WIN: Verify Data In WIN:



ALS Contact: Mandy Sullivan

9143 Phillips Highway, Suite 200 Jacksonville, FL 32256 / PH(904) 739-2277 / FAX (904) 739-2011

Page 1 of 1[illegible]

Date of Request: 22-JUL-2020
 Created By: PARRISH_J On: 22-JUL-2020
 Modified By: JASROTIA_P On: 28-JUL-2020
 Customer: NE-DIST
 Project: SMP
 Division: Division of Environmental Assessment and Restoration
 District: Northeast District
 Event Description: Extra Priority allocations NERO

Send Coolers To:

Phone: 904-256-1700
 8800 Baymeadows Way West
 Suite 100
 Jacksonville, FL 32256
 Attn: Jeremy Parrish
 Cooler Ship EMail: Jeremy.M.Parrish@dep.state.fl.us

Priority: 3
 Event Status: S
 Criminal Investigation: NO
 Chemistry Request Reviewed By: JASROTIA_P On: 28-JUL-2020
 Biology Request Reviewed By: JASROTIA_P On: 28-JUL-2020
 Sampling Kit Required: YES Ship on: 07-AUG-2020
 Sampling Kit Shipped: YES On: 03-AUG-2020
 Sampling Kit Packed By: REYNOLDS_T
 Preservatives Needed: NO
 Date to Receive Samples: 17-AUG-2020
 Logged In By:

Send Final Report To:

8800 Baymeadows Way West
 Suite 100
 Jacksonville, FL 32256
 Attn: Jeremy Parrish
 Report Notification EMail: Jeremy.M.Parrish@dep.state.fl.us

Comments: Group A: (Surf-Fresh) Nutrients, Metals, Tracers Bacteria
 Group B: (Surf-Fresh) Nutrients, Bacteria
 Group C: (Surf-Fresh) Bacteria
 Group D: (Surf.-Salt) Nutrients, Metals, Tracers, Bacteria

NE-ALS-ECQ

Packing Notes:
 None

Bottle Group A (Water) With 2 Samples:

Bottle Type: P-1L	Number of Bottles: 2	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO ₃ /L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-COLOR	Template: DEFAULT	(SM 2120 B-2011) True Color measured at 450 nm (not valid for NPDES monitoring)
Color (true)		
W-F	Template: DEFAULT	(SM 4500 F-C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO ₄ -IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
W-TDS	Template: DEFAULT	(SM 2540 C-2011) Total Dissolved Solids in aqueous matrices
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample
Bottle Type: BP-1L	Number of Bottles: 2	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry
Bottle Type: P-250ML-WI-THI	Number of Bottles: 2	Preserved With: ICE
NE-ALS-ECQ	Template: DEFAULT	(SM 9223 B Quanti-Tray) E. coli by Colilert/QT by ALS in Jacksonville
Bottle Type: BG-500ML	Number of Bottles: 2	Preserved With: ICE
W-E8321-DI	Template: DEFAULT	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
W-E8321-MS	Template: DEFAULT	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
Bottle Type: P-500ML	Number of Bottles: 2	Preserved With: HNO ₃

Bottle Group A (Water) With 2 Samples:

Bottle Type: P-500ML	Number of Bottles: 2	Preserved With: HNO3
W-HARD	Template: DEFAULT	(SM 2340B) Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3
Hardness		
W-ICP	Template:	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
Barium		
Calcium		
Iron		
Magnesium		
Potassium		
Sodium		
Zinc		
W-ICPMS	Template:	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects
*Boron		
Aluminum		
Antimony		
Arsenic		
Cadmium		
Chromium		
Copper		
Lead		
Nickel		
Selenium		
Silver		
Thallium		

Bottle Type: P-500ML	Number of Bottles: 2	Preserved With: ICE And H2SO4
W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices
Bottle Type: P-125ML	Number of Bottles: 2	Preserved With: FILTER-ICE
W-PO4-F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

Bottle Group B (Water) With 2 Samples:

Bottle Type: P-250ML-WI-THI	Number of Bottles: 2	Preserved With: ICE
NE-ALS-ECQ	Template: DEFAULT	(SM 9223 B Quanti-Tray) E. coli by Colilert/QT by ALS in Jacksonville
Bottle Type: P-500ML	Number of Bottles: 2	Preserved With: ICE And H2SO4
W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices
Bottle Type: P-125ML	Number of Bottles: 2	Preserved With: FILTER-ICE
W-PO4-F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L
Bottle Type: P-1L	Number of Bottles: 2	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO3/L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-COLOR	Template: DEFAULT	(SM 2120 B-2011) True Color measured at 450 nm (not valid for NPDES monitoring)
Color (true)		
W-F	Template: DEFAULT	(SM 4500 F-C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO4-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
W-TDS	Template: DEFAULT	(SM 2540 C-2011) Total Dissolved Solids in aqueous matrices
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample
Bottle Type: BP-1L	Number of Bottles: 2	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Bottle Group C (Water) With 2 Samples:

Bottle Type: P-250ML-WI-THI Number of Bottles: 2 Preserved With: ICE
 NE-ALS-ECQ Template: DEFAULT (SM 9223 B Quanti-Tray) E. coli by Colilert/QT by ALS in Jacksonville

Bottle Group D (Water) With 2 Samples:

Bottle Type: P-1L Number of Bottles: 2 Preserved With: ICE
 ALKALINITY Template: DEFAULT (SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO₃/L
 TURBIDITY Template: DEFAULT (EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
 W-BR-IC Template: DEFAULT (EPA 300.0 Rev. 2.1) Bromide in aqueous matrices
 W-F Template: DEFAULT (SM 4500 F-C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
 W-TSS Template: DEFAULT (SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample

Bottle Type: BP-1L Number of Bottles: 2 Preserved With: ICE
 CHLSUITE-W Template: DEFAULT (SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Bottle Type: P-250ML-WI-THI Number of Bottles: 2 Preserved With: ICE
 NE-ALS-ENQ Template: DEFAULT (ASTM D650399) Enterococci by Enterolert/QT by ALS in Jacksonville

Bottle Type: P-250ML-WI-THI Number of Bottles: 2 Preserved With: ICE
 NE-ALS-FC Template: DEFAULT (SM 9222 D) Fecal coliforms by MF by ALS in Jacksonville

Bottle Type: BG-500ML Number of Bottles: 2 Preserved With: ICE
 W-E8321-DI Template: DEFAULT (EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
 W-E8321-MS Template: DEFAULT (EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS

Bottle Type: P-500ML Number of Bottles: 2 Preserved With: HNO₃
 W-ICP Template: S-METALS (EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects

Iron
 Manganese
 Zinc

W-ICPMS Template: (EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects

Aluminum
 Antimony
 Arsenic
 Cadmium
 Chromium
 Copper
 Lead
 Nickel
 Selenium
 Silver
 Thallium

Bottle Type: P-500ML Number of Bottles: 2 Preserved With: ICE And H₂SO₄
 W-NH₃ Template: DEFAULT (EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
 W-NO₂NO₃ Template: DEFAULT (EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
 W-S-A-TP Template: DEFAULT (EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
 W-TKN Template: DEFAULT (EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
 W-TOC Template: DEFAULT (SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices

Bottle Type: P-125ML Number of Bottles: 2 Preserved With: FILTER-ICE
 W-PO₄-F Template: DEFAULT (EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

* - The laboratory is not NELAP certified for this analyte/method, or certification is not applicable.