

Chemical Analysis Report

NE-DIST-2020-10-22-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Continuous Monitoring Event**
Request ID: **RQ-2020-10-19-70**
Customer: **NE-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
8800 Baymeadows Way West
Suite 100
Jacksonville, FL 32256
Attn: Brian G Davis

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 13-NOV-2020 13:11

A handwritten signature in black ink, appearing to read "Liang Lin", is shown on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Plummer Creek @ Second Island

Collection Date/Time: 10/21/2020 11:15

Field ID: G4NE0301

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2204189	EPA 180.1 Rev. 2.0	Turbidity	3.7		NTU	P389132	
	EPA 300.0 Rev. 2.1	Bromide	0.41		mg Br/L	P389809	
	SM 2320 B-2011	Total Alkalinity	24		mg CaCO3/L	P389230	
	SM 2540 D-2011	TSS	5	I	mg/L	P389603	
	SM 4500 F-C-2011	Fluoride	0.071	I	mg F/L	P389236	
2204190	EPA 350.1 Rev. 2.0	Ammonia-N	0.012		mg N/L	P389212	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.2		mg N/L	P389492	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.008	I	mg N/L	P389250	
	EPA 365.1 Rev. 2.0	Total-P	0.099		mg P/L	P389507	
	SM 5310 B-2011	Organic Carbon	38		mg C/L	P389296	
2204191	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.052		mg P/L	P389120	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P389132

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P389809

Component	Result	Code	Units
Bromide	0.013	U	mg Br/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P389212

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P389492

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P389250

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P389507

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P389120

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2320 B-2011
Batch ID: P389230

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 D-2011
Batch ID: P389603

Component	Result	Code	Units
TSS	2	U	mg/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 4500 F-C-2011
Batch ID: P389236

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P389296

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P389809

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Bromide	105		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P389212

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	97.1		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P389492

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	102		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P389250

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P389507

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	107		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P389120

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.4		P	90 - 110

Reference Method: SM 2320 B-2011
Batch ID: P389230

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	93 - 104

Reference Method: SM 4500 F-C-2011
Batch ID: P389236

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.6		P	90 - 106

Reference Method: SM 5310 B-2011
Batch ID: P389296

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	100		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P389809

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2204189	Bromide	105	100	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P389212

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2204244	Ammonia-N	98.7	97.8	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P389492

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2204123	Kjeldahl Nitrogen	113	115	F/F	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P389250

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2204190	NO2NO3-N	91.5	91.5	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P389507

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2204124	Total-P	104	107	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P389120

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2204107	O-Phosphate-P	94.8	94.8	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P389236

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2204041	Fluoride	99.7	98.0	P/P	94 - 107

Reference Method: SM 5310 B-2011
Batch ID: P389296

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2204495	Organic Carbon	106		P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P389132

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2204140	Turbidity	3.22	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P389809

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2204189	Bromide	2.30	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P389212

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2204244	Ammonia-N	0.705	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P389492

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2204123	Kjeldahl Nitrogen	1.57	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P389250

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2204190	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P389507

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2204124	Total-P	2.63	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P389120

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2204107	O-Phosphate-P	0.0	Spike	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P389230

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2204041	Total Alkalinity	5.18	Sample	P	0 - 8.1

Quality Assurance Report Precision

Reference Method: SM 4500 F-C-2011
Batch ID: P389236

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2204041	Fluoride	1.50	Spike	P	0 - 2.8

Reference Method: SM 5310 B-2011
Batch ID: P389296

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2204495	Organic Carbon	1.42	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A101695

Included Lab Sample IDs: 2204191

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.4	100	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A101705

Included Lab Sample IDs: 2204189

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	98.0	99.5	P/P	95 - 105

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A101740

Included Lab Sample IDs: 2204190

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	97.0	98.2	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A101749

Included Lab Sample IDs: 2204189

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	102	98.9	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A101749

Included Lab Sample IDs: 2204189

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	100	101	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A101763

Included Lab Sample IDs: 2204190

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	99.5	100	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A101781

Included Lab Sample IDs: 2204190

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	101	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A101882

Included Lab Sample IDs: 2204190

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	104	96.7	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A101924

Included Lab Sample IDs: 2204190

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	103	105	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A101986

Included Lab Sample IDs: 2204189

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Bromide	101	99.2	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					3.22	
EPA 300.0 Rev. 2.1	Bromide	105	105	100			2.30
EPA 350.1 Rev. 2.0	Ammonia-N	97.1	98.7	97.8			0.705
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	102	113	115			1.57
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	100	91.5	91.5			0.0
EPA 365.1 Rev. 2.0	Total-P	107	104	107			2.63
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	97.4	94.8	94.8			0.0
SM 2320 B-2011	Total Alkalinity	102				5.18	
SM 4500 F-C-2011	Fluoride	97.6	99.7	98.0			1.50
SM 5310 B-2011	Organic Carbon	100	106				1.42

Reference Method Descriptions

Method	Description	<u>Associated Samples</u>
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2204189
EPA 300.0 Rev. 2.1	Bromide in aqueous matrices	2204189
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2204190
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2204190
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2204190
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2204190
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2204191
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2204189
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2204189
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2204189
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2204190

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	10/22/2020			10/22/2020 13:37	Blanca Fach	2204189
EPA 300.0 Rev. 2.1	10/22/2020			11/04/2020 18:39	Lipi Saha	2204189
EPA 350.1 Rev. 2.0	10/22/2020			10/25/2020 12:46	Ping Hua	2204190
EPA 351.2 Rev. 2.0	10/22/2020	10/29/2020 13:45	David Boose	10/30/2020 13:00	Elliott Rodriguez	2204190
EPA 353.2 Rev. 2.0	10/22/2020			10/26/2020 11:24	Beverly Y. Sanders	2204190
EPA 365.1 Rev. 2.0	10/22/2020	10/29/2020 15:54	Yen Ta	11/02/2020 12:17	Benjamin T. Nordmann	2204190
EPA 365.1 Rev. 2.0 dissolved	10/22/2020			10/22/2020 15:52	Virginia P. Brown	2204191
SM 2320 B-2011	10/22/2020			10/24/2020 06:12	Samantha Davila	2204189
SM 2540 D-2011	10/22/2020			10/27/2020 19:45	Madeline Gruver	2204189
SM 4500 F-C-2011	10/22/2020			10/24/2020 06:12	Samantha Davila	2204189
SM 5310 B-2011	10/22/2020			10/26/2020 17:04	Lauren Lees	2204190