

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID:

Shrimp Trap

RQ:

2020-11-02-52

Project:

SMP

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

09/02/2020

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	<u>Leif Boman</u>	<u>11/02/20</u>	<u>8:00</u>	<u>23.2</u>	<u>758.0</u>	<u>8.6</u>	<u>8.50</u>	<u>99.4</u>	<u>/</u>	<u>1.0583</u>	<u>P/F</u>	<u>L/F</u>
ICV	<u>1</u>	<u>1</u>	<u>1</u>	<u>23.3</u>	<u>758.5</u>	<u>8.5</u>	<u>8.50</u>	<u>99.7</u>	<u>/</u>	<u>1</u>	<u>P/F</u>	<u>L/F</u>
CCV	<u>Victoria Schwartz</u>	<u>11/2/20</u>	<u>15:10</u>	<u>22.3</u>	<u>758.0</u>	<u>8.7</u>	<u>8.80</u>	<u>101.3</u>	<u>/</u>	<u>1</u>	<u>P/F</u>	<u>L/F</u>
CCV											<u>P/F</u>	<u>L/F</u>

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	<u>Leif Boman</u>	<u>11/02/20</u>	<u>8:00</u>	<u>200109C</u>	<u>01/2021</u>	<u>50,000</u>	<u>50,000</u>	<u>P/F</u>	<u>L/F</u>
ICV	<u>1</u>	<u>1</u>	<u>1</u>	<u>200109B</u>	<u>01/2021</u>	<u>10,000</u>	<u>9923</u>	<u>P/F</u>	<u>L/F</u>
CCV	<u>Victoria Schwartz</u>	<u>11/2/20</u>	<u>15:10</u>	<u>201026B</u>	<u>11/2021</u>	<u>10,000</u>	<u>9,899</u>	<u>P/F</u>	<u>L/F</u>
CCV								<u>P/F</u>	<u>L/F</u>

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	<u>Leif Boman</u>	<u>11/02/20</u>	<u>8:00</u>	<u>202114</u>	<u>05/22</u>	<u>7.01</u>	<u>23.1</u>	<u>7.01</u>	<u>-31.4</u>	<u>P/F</u>	<u>L/F</u>
Calibr.	<u>1</u>	<u>1</u>	<u>1</u>	<u>200052</u>	<u>02/22</u>	<u>4.00</u>	<u>23.4</u>	<u>4.00</u>	<u>143.9</u>	<u>P/F</u>	<u>L/F</u>
Calibr.	<u>1</u>	<u>1</u>	<u>1</u>	<u>200109A</u>	<u>07/21</u>	<u>10.02</u>	<u>23.2</u>	<u>10.02</u>	<u>-203.0</u>	<u>P/F</u>	<u>L/F</u>
ICV	<u>1</u>	<u>1</u>	<u>1</u>	<u>200052</u>	<u>02/22</u>	<u>4.00</u>	<u>23.4</u>	<u>3.99</u>	<u>144.1</u>	<u>P/F</u>	<u>L/F</u>
CCV	<u>Victoria Schwartz</u>	<u>11/2/20</u>	<u>15:10</u>	<u>200109A</u>	<u>07/21</u>	<u>10.02</u>	<u>23.2</u>	<u>10.15</u>	<u>-211.1</u>	<u>P/F</u>	<u>L/F</u>
CCV										<u>P/F</u>	<u>L/F</u>

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 171.6, slope from 4 to 7 175.3 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 05/15/2020

If NO, what will be used? (circle one) Seechi Disk Line / Sonar Unique ID: KNOTHROAD; Date of last verification: 04/2/2020

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	<u>Leif Boman</u>	<u>11/02/20</u>	<u>8:00</u>	<u>0.272</u>	<u>0.272</u>	<u>P/F</u>	<u>L/F</u>

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID:

Crank Daddy

RQ-

Project:

HAB

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Natalie Ayala	11/2/20	1401	23.3	758.5	8.5	8.50	99.8	✓	1.06	0/F	0/F
ICV	↓	↓	↓	23.3	758.3	8.5	8.51	99.8	✓	✓	0/F	0/F
CCV	Natalie Ayala	11/2/20	1412	21.8	758.5	8.8	8.92	101.7	✓	✓	0/F	0/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Natalie Ayala	11/02/20	1304	191003	10/20	1000	1000	0/F	0/F
ICV	↓	↓	↓	200821	09/21	100	103.3	0/F	0/F
CCV	Natalie Ayala	11/2/20	1643	200821	09/21	100	103.2	0/F	0/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Natalie Ayala	11/02/20	1308	202114	05/22	7.00	23.6	7.00	-27.4	0/F	0/F
Calibr.	↓	↓	↓	200052	02/22	4.00	23.6	4.00	147.6	0/F	0/F
Calibr.	↓	↓	↓	200109A	07/21	10.02	23.3	10.02	-191.3	0/F	0/F
ICV	↓	↓	↓	200109A	07/21	10.02	23.1	10.99	-191.5	0/F	0/F
CCV	Natalie Ayala	11/02/20	16	200052	02/22	4.00	23.5	4.09	140.5	0/F	0/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 5/15/20

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: Kothand; Date of last verification: 4/2/20

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Natalie Ayala	11/02/20	1317	0.272	0.272	0/F	0/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: