

Chemical Analysis Report

CEN-DIST-2020-08-26-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Halifax**
Request ID: **RQ-2020-07-13-37**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Natalie Ayala

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Kerry Tate, Ph.D., Environmental Administrator

Date Certified: 17-SEP-2020 09:05

A handwritten signature in black ink that reads "Kerry Tate". The signature is written in a cursive, flowing style with a horizontal line above the first name.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals and Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Halifax River @ 1 Km Nne of Dunlawton Ave

Collection Date/Time: 08/25/2020 12:37

Field ID: G5CE0107

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2192084	EPA 180.1 Rev. 2.0	Turbidity	7.7	A	NTU	P386579	
	EPA 300.0 Rev. 2.1	Bromide	32		mg Br/L	P387262	MS
	SM 2320 B-2011	Total Alkalinity	123		mg CaCO3/L	P386668	
	SM 2540 D-2011	TSS	15		mg/L	P386990	
	SM 4500 F-C-2011	Fluoride	0.79		mg F/L	P386794	RPD
2192088	EPA 350.1 Rev. 2.0	Ammonia-N	0.008		mg N/L	P386834	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.75		mg N/L	P386611	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P386745	
	EPA 365.1 Rev. 2.0	Total-P	0.12		mg P/L	P386614	
	SM 5310 B-2011	Organic Carbon	8.7		mg C/L	P386727	
2192092	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.068		mg P/L	P386589	
2192104	EPA 200.7 Rev. 4.4	Chromium	3.4	I	ug/L	P386652	
		Iron	420		ug/L	P386652	
		Manganese	17		ug/L	P386652	
		Zinc	15	U	ug/L	P386652	
	EPA 200.8 Rev. 5.4	Aluminum	347		ug/L	P386652	
		Antimony	0.24	I	ug/L	P386652	
		Arsenic	3.57		ug/L	P386652	
		Cadmium	0.080	U	ug/L	P386652	
		Copper	2.0	I	ug/L	P386652	
		Lead	0.80	U	ug/L	P386652	
		Nickel	2.0	U	ug/L	P386652	
		Selenium	0.80	U	ug/L	P386652	
		Silver	0.040	U	ug/L	P386652	
		Thallium	0.40	U	ug/L	P386652	

Sample Location: Halifax River @ 250m N of Bethune Point Park Boat Ramp **Collection Date/Time: 08/25/2020 13:38**

Field ID: G5CE0108

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2192085	EPA 180.1 Rev. 2.0	Turbidity	8.6		NTU	P386579	
	EPA 300.0 Rev. 2.1	Bromide	22		mg Br/L	P387262	MS
	SM 2320 B-2011	Total Alkalinity	121	A	mg CaCO3/L	P386668	
	SM 2540 D-2011	TSS	13		mg/L	P386990	
	SM 4500 F-C-2011	Fluoride	0.61		mg F/L	P386794	RPD
2192089	EPA 350.1 Rev. 2.0	Ammonia-N	0.026		mg N/L	P387017	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.76		mg N/L	P386611	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.15		mg N/L	P386745	
	EPA 365.1 Rev. 2.0	Total-P	0.16		mg P/L	P386614	
	SM 5310 B-2011	Organic Carbon	11		mg C/L	P386727	
2192093	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.12		mg P/L	P386589	
2192105	EPA 200.7 Rev. 4.4	Chromium	4.1	I	ug/L	P386652	
		Iron	560		ug/L	P386652	
		Manganese	14		ug/L	P386652	
		Zinc	15	U	ug/L	P386652	
	EPA 200.8 Rev. 5.4	Aluminum	422		ug/L	P386652	
		Antimony	0.21	I	ug/L	P386652	
		Arsenic	3.54		ug/L	P386652	
		Cadmium	0.080	U	ug/L	P386652	
		Copper	2.7	I	ug/L	P386652	
		Lead	0.85	I	ug/L	P386652	
		Nickel	2.0	U	ug/L	P386652	
		Selenium	0.80	U	ug/L	P386652	
		Silver	0.040	U	ug/L	P386652	
		Thallium	0.40	U	ug/L	P386652	

Sample Location: Halifax River @ 200m Nne of Main St Bridge

Collection Date/Time: 08/25/2020 14:17

Field ID: G5CE0109

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2192086	EPA 180.1 Rev. 2.0	Turbidity	16		NTU	P386579	
	EPA 300.0 Rev. 2.1	Bromide	20		mg Br/L	P387262	MS
	SM 2320 B-2011	Total Alkalinity	121		mg CaCO3/L	P386668	
	SM 2540 D-2011	TSS	24		mg/L	P386990	
	SM 4500 F-C-2011	Fluoride	0.59		mg F/L	P386794	RPD
2192090	EPA 350.1 Rev. 2.0	Ammonia-N	0.028		mg N/L	P387017	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.72		mg N/L	P386611	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.21		mg N/L	P386745	
	EPA 365.1 Rev. 2.0	Total-P	0.18		mg P/L	P386614	
	SM 5310 B-2011	Organic Carbon	11		mg C/L	P386727	
2192094	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.12		mg P/L	P386589	
2192106	EPA 200.7 Rev. 4.4	Chromium	5.1	I	ug/L	P386652	
		Iron	1.17E+03		ug/L	P386652	
		Manganese	19		ug/L	P386652	
		Zinc	16	I	ug/L	P386652	
	EPA 200.8 Rev. 5.4	Aluminum	833		ug/L	P386652	
		Antimony	0.20	U	ug/L	P386652	
		Arsenic	3.70		ug/L	P386652	
		Cadmium	0.080	U	ug/L	P386652	
		Copper	3.0	I	ug/L	P386652	
		Lead	1.7		ug/L	P386652	
		Nickel	2.0	U	ug/L	P386652	
		Selenium	0.80	U	ug/L	P386652	
		Silver	0.040	U	ug/L	P386652	
		Thallium	0.40	U	ug/L	P386652	

**Sample Location: Halifax River @ 1Km N of Bethune Point Park
Boat Ramp**

Collection Date/Time: 08/25/2020 13:45

Field ID: G5CE0110

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2192087	EPA 180.1 Rev. 2.0	Turbidity	12		NTU	P386579	
	EPA 300.0 Rev. 2.1	Bromide	20		mg Br/L	P387262	MS
	SM 2320 B-2011	Total Alkalinity	121		mg CaCO3/L	P386668	
	SM 2540 D-2011	TSS	17	A	mg/L	P386990	
	SM 4500 F-C-2011	Fluoride	0.59		mg F/L	P386794	RPD
2192091	EPA 350.1 Rev. 2.0	Ammonia-N	0.032		mg N/L	P387017	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.76		mg N/L	P386611	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.20		mg N/L	P386745	
	EPA 365.1 Rev. 2.0	Total-P	0.17		mg P/L	P386614	
	SM 5310 B-2011	Organic Carbon	11		mg C/L	P386727	
2192095	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.12		mg P/L	P386589	
2192107	EPA 200.7 Rev. 4.4	Chromium	5.1	I	ug/L	P386652	
		Iron	950		ug/L	P386652	
		Manganese	17		ug/L	P386652	
		Zinc	16	I	ug/L	P386652	
	EPA 200.8 Rev. 5.4	Aluminum	716		ug/L	P386652	
		Antimony	0.20	U	ug/L	P386652	
		Arsenic	3.54		ug/L	P386652	
		Cadmium	0.080	U	ug/L	P386652	
		Copper	2.8	I	ug/L	P386652	
		Lead	1.2	I	ug/L	P386652	
		Nickel	2.0	U	ug/L	P386652	
		Selenium	0.80	U	ug/L	P386652	
		Silver	0.040	U	ug/L	P386652	
		Thallium	0.40	U	ug/L	P386652	

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P386579

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P386652

Component	Result	Code	Units
Chromium	1.0	U	ug/L
Iron	30	U	ug/L
Manganese	1.0	U	ug/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P386652

Component	Result	Code	Units
Aluminum	5.0	U	ug/L
Antimony	0.050	U	ug/L
Arsenic	0.050	U	ug/L
Cadmium	0.020	U	ug/L
Copper	0.40	U	ug/L
Lead	0.20	U	ug/L
Nickel	0.50	U	ug/L
Selenium	0.20	U	ug/L
Silver	0.010	U	ug/L
Thallium	0.10	U	ug/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387262

Component	Result	Code	Units
Bromide	0.013	U	mg Br/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P386834

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387017

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P386611

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P386745

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P386614

Component	Result	Code	Units
Total-P	0.005	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P386589

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2320 B-2011
Batch ID: P386668

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 D-2011
Batch ID: P386990

Component	Result	Code	Units
TSS	3	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P386794

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P386727

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P386652

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chromium	97.9		P	85 - 115
Iron	105		P	85 - 115
Manganese	102		P	85 - 115
Zinc	99.7		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P386652

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aluminum	99.7		P	85 - 115
Antimony	101		P	85 - 115
Arsenic	102		P	85 - 115
Cadmium	100		P	85 - 115
Copper	95.6		P	85 - 115
Lead	104		P	85 - 115
Nickel	94.7		P	85 - 115
Selenium	101		P	85 - 115
Silver	104		P	85 - 115
Thallium	108		P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387262

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Bromide	96.7		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P386834

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	99.6		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387017

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	98.5		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P386611

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	105		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P386745

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	99.5		P	90 - 110

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P386614

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	108		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P386589

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	98.4		P	90 - 110

Reference Method: SM 2320 B-2011
Batch ID: P386668

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	93 - 104

Reference Method: SM 4500 F-C-2011
Batch ID: P386794

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	101		P	90 - 106

Reference Method: SM 5310 B-2011
Batch ID: P386727

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	95.0		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P386652

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192703	Chromium	98.9	99.7	P/P	70 - 130
2192703	Iron	102	97.9	P/P	70 - 130
2192703	Manganese	101	101	P/P	70 - 130
2192703	Zinc	101	101	P/P	70 - 130

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P386652

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192703	Aluminum	104	102	P/P	70 - 130
2192703	Antimony	107	104	P/P	70 - 130
2192703	Arsenic	118	117	P/P	70 - 130
2192703	Cadmium	105	106	P/P	70 - 130
2192703	Copper	108	107	P/P	70 - 130
2192703	Lead	108	115	P/P	70 - 130
2192703	Nickel	106	109	P/P	70 - 130
2192703	Selenium	118	115	P/P	70 - 130
2192703	Silver	99.0	98.0	P/P	70 - 130
2192703	Thallium	114	118	P/P	70 - 130

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387262

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192715	Bromide	91.5	89.4	P/F	90 - 110
2192716	Bromide	87.1		F	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P386834

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192749	Ammonia-N	104	101	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387017

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192089	Ammonia-N	100	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P386611

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192088	Kjeldahl Nitrogen	108	107	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P386745

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192088	NO2NO3-N	96.8	96.8	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P386614

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192088	Total-P	104	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P386589

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192373	O-Phosphate-P	102	103	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P386794

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192824	Fluoride	100	105	P/P	94 - 107

Reference Method: SM 5310 B-2011
Batch ID: P386727

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2191540	Organic Carbon	94.9	95.2	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P386579

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192084	Turbidity	9.90	Sample	P	0 - 20

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P386652

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192703	Chromium	0.761	Spike	P	0 - 20
2192703	Iron	3.75	Spike	P	0 - 20
2192703	Manganese	0.200	Spike	P	0 - 20
2192703	Zinc	0.340	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P386652

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192703	Aluminum	2.04	Spike	P	0 - 20
2192703	Antimony	2.43	Spike	P	0 - 20
2192703	Arsenic	0.516	Spike	P	0 - 20
2192703	Cadmium	1.67	Spike	P	0 - 20
2192703	Copper	1.07	Spike	P	0 - 20
2192703	Lead	5.50	Spike	P	0 - 20
2192703	Nickel	3.29	Spike	P	0 - 20
2192703	Selenium	3.13	Spike	P	0 - 20
2192703	Silver	0.988	Spike	P	0 - 20
2192703	Thallium	4.15	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387262

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192715	Bromide	1.76	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P386834

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192749	Ammonia-N	1.99	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387017

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192089	Ammonia-N	0.734	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P386611

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192088	Kjeldahl Nitrogen	1.02	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P386745

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192088	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P386614

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192088	Total-P	2.01	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P386589

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192373	O-Phosphate-P	0.672	Spike	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P386668

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192085	Total Alkalinity	1.36	Sample	P	0 - 8.1

Reference Method: SM 2540 D-2011
Batch ID: P386990

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192087	TSS	6.90	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
Batch ID: P386794

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192824	Fluoride	2.82	Spike	F	0 - 2.8

Reference Method: SM 5310 B-2011
Batch ID: P386727

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2191540	Organic Carbon	0.278	Spike	P	0 - 20

Quality Assurance Report

Precision

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A100660

Included Lab Sample IDs: 2192084, 2192085, 2192086, 2192087

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A100663

Included Lab Sample IDs: 2192092, 2192093, 2192094, 2192095

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	96.0	95.7	P/P	90 - 110
O-Phosphate-P	96.1	96.0	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A100697

Included Lab Sample IDs: 2192084, 2192085, 2192086, 2192087

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	99.7	102	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A100700

Included Lab Sample IDs: 2192088, 2192089, 2192090, 2192091

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	105	108	P/P	90 - 110
Kjeldahl Nitrogen	108	99.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A100710

Included Lab Sample IDs: 2192088, 2192089, 2192090, 2192091

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	103	P/P	90 - 110
Total-P	102	102	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A100718

Included Lab Sample IDs: 2192088, 2192089, 2192090, 2192091

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	92.2	92.2	P/P	90 - 110
Organic Carbon	95.0	92.2	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A100725

Included Lab Sample IDs: 2192088, 2192089, 2192090, 2192091

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A100733

Included Lab Sample IDs: 2192104, 2192105, 2192106, 2192107

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chromium	97.4	98.5	P/P	90 - 110
Chromium	98.5	96.6	P/P	90 - 110
Iron	100	97.1	P/P	90 - 110
Iron	97.1	97.4	P/P	90 - 110
Manganese	101	100	P/P	90 - 110
Manganese	99.9	101	P/P	90 - 110
Zinc	98.8	98.8	P/P	90 - 110
Zinc	99.7	98.8	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A100747

Included Lab Sample IDs: 2192084, 2192085, 2192086, 2192087

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	102	104	P/P	90 - 110
Fluoride	104	98.2	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A100761

Included Lab Sample IDs: 2192088

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	101	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A100832

Included Lab Sample IDs: 2192089, 2192090, 2192091

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.6	101	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A100864

Included Lab Sample IDs: 2192104, 2192105, 2192106, 2192107

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aluminum	103	104	P/P	90 - 110
Aluminum	104	104	P/P	90 - 110
Arsenic	101	101	P/P	90 - 110
Arsenic	101	102	P/P	90 - 110
Cadmium	98.7	100	P/P	90 - 110
Cadmium	99.0	98.7	P/P	90 - 110
Copper	94.5	95.4	P/P	90 - 110
Copper	95.4	99.2	P/P	90 - 110
Lead	102	98.8	P/P	90 - 110
Lead	105	102	P/P	90 - 110
Nickel	96.0	98.3	P/P	90 - 110
Nickel	96.2	96.0	P/P	90 - 110
Selenium	103	103	P/P	90 - 110
Selenium	103	104	P/P	90 - 110
Silver	98.6	102	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A100864

Included Lab Sample IDs: 2192104, 2192105, 2192106, 2192107

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Silver	98.9	98.6	P/P	90 - 110
Thallium	105	107	P/P	90 - 110
Thallium	107	102	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A100907

Included Lab Sample IDs: 2192104, 2192105, 2192106, 2192107

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Antimony	101	101	P/P	90 - 110
Antimony	101	100	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A100925

Included Lab Sample IDs: 2192084, 2192085, 2192086, 2192087

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Bromide	101	95.4	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					9.90	
EPA 200.7 Rev. 4.4	Chromium	97.9	98.9	99.7			0.761
	Iron	105	102	97.9			3.75
	Manganese	102	101	101			0.200
	Zinc	99.7	101	101			0.340
EPA 200.8 Rev. 5.4	Aluminum	99.7	104	102			2.04
	Antimony	101	107	104			2.43
	Arsenic	102	118	117			0.516
	Cadmium	100	105	106			1.67
	Copper	95.6	108	107			1.07
	Lead	104	108	115			5.50
	Nickel	94.7	106	109			3.29
	Selenium	101	118	115			3.13
	Silver	104	99.0	98.0			0.988
	Thallium	108	114	118			4.15
EPA 300.0 Rev. 2.1	Bromide	96.7	91.5	89.4	87.1		1.76
EPA 350.1 Rev. 2.0	Ammonia-N	99.6	104	101			1.99
	Ammonia-N	98.5	100	101			0.734
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	105	108	107			1.02
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	99.5	96.8	96.8			0.0
EPA 365.1 Rev. 2.0	Total-P	108	104	102			2.01
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	98.4	102	103			0.672
SM 2320 B-2011	Total Alkalinity	102				1.36	
SM 2540 D-2011	TSS					6.90	
SM 4500 F-C-2011	Fluoride	101	100	105			2.82
SM 5310 B-2011	Organic Carbon	95.0	94.9	95.2			0.278

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2192084, 2192085, 2192086, 2192087
EPA 200.7 Rev. 4.4	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	2192104, 2192105, 2192106, 2192107
EPA 200.8 Rev. 5.4	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	2192104, 2192105, 2192106, 2192107
EPA 300.0 Rev. 2.1	Bromide in aqueous matrices	2192084, 2192085, 2192086, 2192087
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2192088, 2192089, 2192090, 2192091
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2192088, 2192089, 2192090, 2192091
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2192088, 2192089, 2192090, 2192091
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2192088, 2192089, 2192090, 2192091
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2192092, 2192093, 2192094, 2192095
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2192084, 2192085, 2192086, 2192087
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2192084, 2192085, 2192086, 2192087
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2192084, 2192085, 2192086, 2192087
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2192088, 2192089, 2192090, 2192091

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	08/26/2020			08/26/2020 15:10	Yen Ta	2192084, 2192085, 2192086, 2192087
EPA 200.7 Rev. 4.4	08/26/2020	08/27/2020 17:15	Elliott D. Healy	08/31/2020 18:19	Betina Topolski	2192104
	08/26/2020	08/27/2020 17:15	Elliott D. Healy	08/31/2020 18:27	Betina Topolski	2192105
	08/26/2020	08/27/2020 17:15	Elliott D. Healy	08/31/2020 19:02	Betina Topolski	2192106
	08/26/2020	08/27/2020 17:15	Elliott D. Healy	08/31/2020 19:10	Betina Topolski	2192107
EPA 200.8 Rev. 5.4	08/26/2020	08/27/2020 17:15	Elliott D. Healy	09/08/2020 20:23	Alexander Thompson	2192104
	08/26/2020	08/27/2020 17:15	Elliott D. Healy	09/08/2020 21:10	Alexander Thompson	2192105
	08/26/2020	08/27/2020 17:15	Elliott D. Healy	09/08/2020 21:16	Alexander Thompson	2192106
	08/26/2020	08/27/2020 17:15	Elliott D. Healy	09/08/2020 21:23	Alexander Thompson	2192107
	08/26/2020	08/27/2020 17:15	Elliott D. Healy	09/10/2020 12:17	Alexander Thompson	2192104
	08/26/2020	08/27/2020 17:15	Elliott D. Healy	09/10/2020 12:22	Alexander Thompson	2192105
	08/26/2020	08/27/2020 17:15	Elliott D. Healy	09/10/2020 12:54	Alexander Thompson	2192106
	08/26/2020	08/27/2020 17:15	Elliott D. Healy	09/10/2020 12:58	Alexander Thompson	2192107
EPA 300.0 Rev. 2.1	08/26/2020			09/09/2020 19:48	Lipi Saha	2192084
	08/26/2020			09/09/2020 20:05	Lipi Saha	2192085
	08/26/2020			09/09/2020 20:22	Lipi Saha	2192086
	08/26/2020			09/09/2020 20:39	Lipi Saha	2192087
EPA 350.1 Rev. 2.0	08/26/2020			09/01/2020 19:30	Ping Hua	2192088
	08/26/2020			09/05/2020 11:50	Ping Hua	2192089
	08/26/2020			09/05/2020 11:54	Ping Hua	2192090
	08/26/2020			09/05/2020 11:56	Ping Hua	2192091
EPA 351.2 Rev. 2.0	08/26/2020	08/27/2020 10:15	Sima Feizi	08/28/2020 11:06	Virginia P. Brown	2192088
	08/26/2020	08/27/2020 10:15	Sima Feizi	08/28/2020 11:14	Virginia P. Brown	2192089
	08/26/2020	08/27/2020 10:15	Sima Feizi	08/28/2020 11:19	Virginia P. Brown	2192090
	08/26/2020	08/27/2020 10:15	Sima Feizi	08/28/2020 11:21	Virginia P. Brown	2192091
EPA 353.2 Rev. 2.0	08/26/2020			08/31/2020 09:30	Beverly Y. Sanders	2192088
	08/26/2020			08/31/2020 09:36	Beverly Y. Sanders	2192089
	08/26/2020			08/31/2020 09:37	Beverly Y. Sanders	2192090
	08/26/2020			08/31/2020 09:38	Beverly Y. Sanders	2192091
EPA 365.1 Rev. 2.0	08/26/2020	08/27/2020 11:54	Yen Ta	08/28/2020 13:28	Benjamin T. Nordmann	2192088
	08/26/2020	08/27/2020 11:54	Yen Ta	08/28/2020 13:31	Benjamin T. Nordmann	2192089
	08/26/2020	08/27/2020 11:54	Yen Ta	08/28/2020 13:32	Benjamin T. Nordmann	2192090
	08/26/2020	08/27/2020 11:54	Yen Ta	08/28/2020 13:36	Benjamin T. Nordmann	2192091
EPA 365.1 Rev. 2.0 dissolved	08/26/2020			08/26/2020 19:09	Blanca Fach	2192092
	08/26/2020			08/26/2020 19:10	Blanca Fach	2192093
	08/26/2020			08/26/2020 19:15	Blanca Fach	2192095
	08/26/2020			08/26/2020 19:17	Blanca Fach	2192094
SM 2320 B-2011	08/26/2020			08/28/2020 04:03	Dale L. Simmons	2192085
	08/26/2020			08/28/2020 04:16	Dale L. Simmons	2192086
	08/26/2020			08/28/2020 04:28	Dale L. Simmons	2192087

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
SM 2320 B-2011	08/26/2020			08/28/2020 04:40	Dale L. Simmons	2192084
SM 2540 D-2011	08/26/2020			08/27/2020 15:08	Yijie Li	2192084, 2192085, 2192086, 2192087
SM 4500 F-C-2011	08/26/2020			08/29/2020 00:41	Dale L. Simmons	2192085
	08/26/2020			08/29/2020 00:46	Dale L. Simmons	2192086
	08/26/2020			08/29/2020 00:52	Dale L. Simmons	2192087
	08/26/2020			08/29/2020 01:58	Dale L. Simmons	2192084
SM 5310 B-2011	08/26/2020			08/30/2020 12:26	Ping Hua	2192088
	08/26/2020			08/30/2020 12:39	Ping Hua	2192089
	08/26/2020			08/30/2020 13:03	Ping Hua	2192090
	08/26/2020			08/30/2020 13:40	Ping Hua	2192091