

Chemical Analysis Report

CEN-DIST-2020-08-11-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Condel/Lotus/Asher**
Request ID: **RQ-2020-08-10-47**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Kalina Warren

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator

Date Certified: 02-SEP-2020 10:32



NON-CONFORMANCE REPORT INCLUDED

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals and Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Lake Lotus @ Center

Collection Date/Time: 08/10/2020 12:11

Field ID: G2CE0015

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2189507	EPA 180.1 Rev. 2.0	Turbidity	6.7		NTU	P385911	
	EPA 300.0 Rev. 2.1	Chloride	9.1		mg Cl/L	P386273	
		Sulfate	4.4		mg SO4/L	P386277	
	SM 2120 B-2011	Color (true)	49		PCU	P385923	
	SM 2320 B-2011	Total Alkalinity	55	A	mg CaCO3/L	P386228	
	SM 2540 C-2011	TDS	117		mg/L	P386182	
	SM 2540 D-2011	TSS	6	I	mg/L	P386223	
	SM 4500 F-C-2011	Fluoride	0.078	I	mg F/L	P385994	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.007	Y	mg N/L	P386085	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.83	Y	mg N/L	P386141	
2189509	EPA 353.2 Rev. 2.0	NO2NO3-N	0.015		mg N/L	P385938	
	EPA 365.1 Rev. 2.0	Total-P	0.095		mg P/L	P385942	
	SM 5310 B-2011	Organic Carbon	10		mg C/L	P385980	
	EPA 365.1 Rev. 2.0	O-Phosphate-P	0.004	U	mg P/L	P385919	
	dissolved						

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 350.1 Rev. 2.0: Sample was left out at room temperature overnight, should be stored below 6 degC.

EPA 351.2 Rev. 2.0: Sample was left out at room temperature overnight, should be stored below 6 degC.

Sample Location: Lake Asher @ Center Of Lake

Collection Date/Time: 08/10/2020 11:39

Field ID: G2CE0130

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2189508	EPA 180.1 Rev. 2.0	Turbidity	5.4		NTU	P385911	
	EPA 300.0 Rev. 2.1	Chloride	19		mg Cl/L	P386273	
		Sulfate	6.3		mg SO4/L	P386277	
	SM 2120 B-2011	Color (true)	36		PCU	P385923	
	SM 2320 B-2011	Total Alkalinity	34		mg CaCO3/L	P386228	
	SM 2540 C-2011	TDS	103		mg/L	P386182	
	SM 2540 D-2011	TSS	6	I	mg/L	P386223	
	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P385994	
2189510	EPA 350.1 Rev. 2.0	Ammonia-N	0.005	IY	mg N/L	P386085	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.84	Y	mg N/L	P386141	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P385938	
	EPA 365.1 Rev. 2.0	Total-P	0.052		mg P/L	P385942	
	SM 5310 B-2011	Organic Carbon	9.2		mg C/L	P385980	
2189512	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P385919	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 350.1 Rev. 2.0: Sample was left out at room temperature overnight, should be stored below 6 degC.

EPA 351.2 Rev. 2.0: Sample was left out at room temperature overnight, should be stored below 6 degC.

Sample Location: Lake Condel @ Center

Collection Date/Time: 08/10/2020 10:47

Field ID: G4CE0177

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes	
2189504	EPA 180.1 Rev. 2.0	Turbidity	7.8		NTU	P385911		
	EPA 300.0 Rev. 2.1	Chloride	12		mg Cl/L	P386273		
		Sulfate	9.4		mg SO4/L	P386277		
		Color (true)	26	A	PCU	P385923		
	SM 2320 B-2011	Total Alkalinity	41		mg CaCO3/L	P386228		
	SM 2540 C-2011	TDS	92		mg/L	P386182		
	SM 2540 D-2011	TSS	12	I	mg/L	P386223		
	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P385994		
2189505	EPA 350.1 Rev. 2.0	Ammonia-N	0.005	Y	mg N/L	P386085		
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.2	Y	mg N/L	P386141		
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P385938		
	EPA 365.1 Rev. 2.0	Total-P	0.11		mg P/L	P385942		
	SM 5310 B-2011	Organic Carbon	6.9		mg C/L	P385980		
2189506	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P385919		
2189541	EPA 200.7 Rev. 4.4	Barium	9.1		ug/L	P385910		
		Calcium	17.5		mg/L	P385910		
		Iron	37	I	ug/L	P385910		
		Magnesium	2.26		mg/L	P385910		
		Potassium	2.2		mg/L	P385910		
		Sodium	10.1		mg/L	P385910		
		Zinc	5.0	U	ug/L	P385910		
	EPA 200.8 Rev. 5.4	Aluminum	56		ug/L	P385910		
		Antimony	0.11	I	ug/L	P385910		
		Arsenic	0.92		ug/L	P385910		
		Boron**	56.1		ug/L	P385910		
		Cadmium	0.020	U	ug/L	P385910		
		Chromium	0.40	U	ug/L	P385910		
		Copper	0.60	I	ug/L	P385910		
		Lead	0.40	I	ug/L	P385910		
		Nickel	0.50	U	ug/L	P385910		
		Selenium	0.20	U	ug/L	P385910		
		Silver	0.010	U	ug/L	P385910		
		Thallium	0.10	U	ug/L	P385910		
		SM 2340B	Hardness	53.0		mg/L	P385910	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample. The PQL was elevated because of sample matrix interference.

EPA 350.1 Rev. 2.0: Sample was left out at room temperature overnight, should be stored below 6 degC.

EPA 351.2 Rev. 2.0: Sample was left out at room temperature overnight, should be stored below 6 degC.

Non-Conformance Report

NCR ID: 8387

<u>Event(s)</u>	<u>Job(s)</u>	<u>Sample(s)</u>	<u>Test(s)</u>
CEN-DIST-2020-08-11-02	TLH-2020-08-11-38	2189505	
CEN-DIST-2020-08-11-02	TLH-2020-08-11-38	2189509	
CEN-DIST-2020-08-11-02	TLH-2020-08-11-38	2189510	

NCR Type: CUSTODY

NCR Category: General

Observation: Samples were not stored in refrigerator overnight on 08/12/2020.

Resolution: Samples were returned to refrigerator at 10:15 am on 08/13/2020. Data will be qualified as appropriate.

Authorized by/Date: Irina Carbone 9/1/2020

The Non-Conformance Report details exceptions or problems encountered with the events/jobs/samples/test.
Please address questions to:

Chemistry	Colin Wright	(850) 245-8085
Biology	Cheryl Swanson	(850) 245-8177

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P385911

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P385910

Component	Result	Code	Units
Barium	1.0	U	ug/L
Calcium	0.075	U	mg/L
Iron	30	U	ug/L
Magnesium	0.040	U	mg/L
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P385910

Component	Result	Code	Units
Aluminum	5.0	U	ug/L
Antimony	0.050	U	ug/L
Arsenic	0.050	U	ug/L
Boron	2.0	U	ug/L
Cadmium	0.020	U	ug/L
Chromium	0.40	U	ug/L
Copper	0.40	U	ug/L
Lead	0.20	U	ug/L
Nickel	0.50	U	ug/L
Selenium	0.20	U	ug/L
Silver	0.010	U	ug/L
Thallium	0.10	U	ug/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P386273

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P386277

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P386085

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P386141

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P385938

Component	Result	Code	Units
NO ₂ NO ₃ -N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P385942

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P385919

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B-2011
Batch ID: P385923

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-2011
Batch ID: P386228

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-2011
Batch ID: P386182

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P386223

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P385994

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 5310 B-2011
Batch ID: P385980

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P385910

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Barium	101		P	85 - 115
Calcium	104		P	85 - 115
Iron	104		P	85 - 115
Magnesium	104		P	85 - 115
Potassium	104		P	85 - 115
Sodium	102		P	85 - 115
Zinc	103		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P385910

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aluminum	100		P	85 - 115
Antimony	105		P	85 - 115
Arsenic	102		P	85 - 115
Boron	100		P	85 - 115
Cadmium	106		P	85 - 115
Chromium	102		P	85 - 115
Copper	106		P	85 - 115
Lead	108		P	85 - 115
Nickel	105		P	85 - 115
Selenium	100		P	85 - 115
Silver	107		P	85 - 115
Thallium	107		P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P386273

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	97.5		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P386277

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	99.2		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P386085

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	99.2		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P386141

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	105		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P385938

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	99.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P385942

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P385919

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	100		P	90 - 110

Reference Method: SM 2120 B-2011
Batch ID: P385923

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	104		P	85 - 115

Reference Method: SM 2320 B-2011
Batch ID: P386228

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	101		P	93 - 104

Reference Method: SM 4500 F-C-2011
Batch ID: P385994

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.5		P	90 - 106

Reference Method: SM 5310 B-2011
Batch ID: P385980

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	97.4		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P385910

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2189410	Barium	103	103	P/P	80 - 120
2189410	Calcium	104		P	80 - 120
2189410	Iron	104	106	P/P	80 - 120
2189410	Magnesium	106		P	80 - 120
2189410	Potassium	103	104	P/P	80 - 120
2189410	Sodium	103		P	80 - 120
2189410	Zinc	105	105	P/P	80 - 120
2189491	Barium	98.7		P	80 - 120
2189491	Calcium	105		P	80 - 120
2189491	Iron	103		P	80 - 120
2189491	Magnesium	102		P	80 - 120
2189491	Potassium	101		P	80 - 120
2189491	Sodium	103		P	80 - 120
2189491	Zinc	102		P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P385910

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2189410	Aluminum	97.0	100	P/P	80 - 120
2189410	Antimony	103	107	P/P	80 - 120
2189410	Arsenic	101	103	P/P	80 - 120
2189410	Boron	100	104	P/P	80 - 120
2189410	Cadmium	103	107	P/P	80 - 120
2189410	Chromium	104	106	P/P	80 - 120
2189410	Copper	103	104	P/P	80 - 120
2189410	Lead	108	111	P/P	80 - 120
2189410	Nickel	100	105	P/P	80 - 120
2189410	Selenium	97.5	101	P/P	80 - 120
2189410	Silver	103	105	P/P	80 - 120
2189410	Thallium	108	112	P/P	80 - 120
2189491	Aluminum	101		P	80 - 120
2189491	Antimony	105		P	80 - 120
2189491	Arsenic	101		P	80 - 120
2189491	Boron	103		P	80 - 120
2189491	Cadmium	106		P	80 - 120
2189491	Chromium	105		P	80 - 120
2189491	Copper	104		P	80 - 120
2189491	Lead	110		P	80 - 120
2189491	Nickel	103		P	80 - 120
2189491	Selenium	96.5		P	80 - 120
2189491	Silver	107		P	80 - 120
2189491	Thallium	108		P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P386273

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2189370	Chloride	98.8		P	90 - 110
2189507	Chloride	99.1		P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P386277

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2189370	Sulfate	100		P	90 - 110
2189507	Sulfate	101		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P386085

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2189467	Ammonia-N	99.0	103	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P386141

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2189505	Kjeldahl Nitrogen	106	92.2	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P385938

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2189464	NO2NO3-N	97.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P385942

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2189464	Total-P	103	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P385919

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2189506	O-Phosphate-P	98.8	99.4	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P385994

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2189516	Fluoride	99.7	97.9	P/P	94 - 107

Reference Method: SM 5310 B-2011
Batch ID: P385980

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2189375	Organic Carbon	96.2	96.2	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P385911

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2189445	Turbidity	15.2	Sample	P	0 - 20

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P385910

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2189410	Barium	0.133	Spike	P	0 - 20
2189410	Calcium	0.259	Spike	P	0 - 20
2189410	Iron	1.75	Spike	P	0 - 20
2189410	Magnesium	0.858	Spike	P	0 - 20
2189410	Potassium	0.473	Spike	P	0 - 20
2189410	Sodium	2.32	Spike	P	0 - 20
2189410	Zinc	0.0947	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P385910

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2189410	Aluminum	2.49	Spike	P	0 - 20
2189410	Antimony	3.59	Spike	P	0 - 20
2189410	Arsenic	2.54	Spike	P	0 - 20
2189410	Boron	2.87	Spike	P	0 - 20
2189410	Cadmium	4.11	Spike	P	0 - 20
2189410	Chromium	2.29	Spike	P	0 - 20
2189410	Copper	1.30	Spike	P	0 - 20
2189410	Lead	3.20	Spike	P	0 - 20
2189410	Nickel	4.26	Spike	P	0 - 20
2189410	Selenium	3.79	Spike	P	0 - 20
2189410	Silver	2.01	Spike	P	0 - 20
2189410	Thallium	3.65	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P386273

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2189370	Chloride	0.0877	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P386277

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2189370	Sulfate	0.332	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P386085

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2189467	Ammonia-N	1.81	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P386141

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2189505	Kjeldahl Nitrogen	7.01	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P385938

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2189464	NO2NO3-N	0.378	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P385942

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2189464	Total-P	1.51	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P385919

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2189506	O-Phosphate-P	0.605	Spike	P	0 - 20

Reference Method: SM 2120 B-2011
Batch ID: P385923

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2189504	Color (true)	1.78	Sample	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P386228

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2189507	Total Alkalinity	0.255	Sample	P	0 - 8.1

Reference Method: SM 2540 C-2011
Batch ID: P386182

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2189458	TDS	0.583	Sample	P	0 - 10

Quality Assurance Report Precision

Reference Method: SM 4500 F-C-2011
Batch ID: P385994

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2189516	Fluoride	1.48	Spike	P	0 - 2.8

Reference Method: SM 5310 B-2011
Batch ID: P385980

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2189375	Organic Carbon	0.0442	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A100393

Included Lab Sample IDs: 2189504, 2189507, 2189508

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	101	101	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A100394

Included Lab Sample IDs: 2189506, 2189511, 2189512

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	99.0	P/P	90 - 110
O-Phosphate-P	99.0	98.6	P/P	90 - 110

Reference Method: SM 2120 B-2011

Run ID: A100395

Included Lab Sample IDs: 2189504, 2189507, 2189508

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	103	106	P/P	90 - 115

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A100406

Included Lab Sample IDs: 2189505, 2189509, 2189510

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	101	100	P/P	90 - 110
NO2NO3-N	101	102	P/P	90 - 110

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A100428

Included Lab Sample IDs: 2189541

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Barium	101	98.2	P/P	90 - 110
Calcium	101	101	P/P	90 - 110
Iron	102	103	P/P	90 - 110
Magnesium	102	102	P/P	90 - 110
Potassium	101	101	P/P	90 - 110
Sodium	100	101	P/P	90 - 110
Zinc	102	100	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A100436

Included Lab Sample IDs: 2189505, 2189509, 2189510

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	98.0	97.8	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A100438

Included Lab Sample IDs: 2189505, 2189509, 2189510

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	101	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 4500 F-C-2011

Run ID: A100439

Included Lab Sample IDs: 2189504, 2189507, 2189508

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	100	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A100478

Included Lab Sample IDs: 2189505, 2189509, 2189510

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	98.6	98.6	P/P	90 - 110
Ammonia-N	99.7	98.6	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A100483

Included Lab Sample IDs: 2189541

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aluminum	102	102	P/P	90 - 110
Antimony	107	107	P/P	90 - 110
Arsenic	100	99.4	P/P	90 - 110
Boron	100	101	P/P	90 - 110
Cadmium	103	103	P/P	90 - 110
Chromium	102	103	P/P	90 - 110
Copper	98.4	97.7	P/P	90 - 110
Lead	106	108	P/P	90 - 110
Nickel	101	100	P/P	90 - 110
Selenium	99.3	101	P/P	90 - 110
Silver	106	106	P/P	90 - 110
Thallium	103	102	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A100515

Included Lab Sample IDs: 2189505, 2189509, 2189510

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	94.6	96.7	P/P	90 - 110
Kjeldahl Nitrogen	96.7	103	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A100529

Included Lab Sample IDs: 2189504, 2189507, 2189508

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	98.7	104	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A100543

Included Lab Sample IDs: 2189504, 2189507, 2189508

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	101	P/P	90 - 110
Chloride	101	102	P/P	90 - 110
Sulfate	101	102	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A100543

Included Lab Sample IDs: 2189504, 2189507, 2189508

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Sulfate	102	101	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery			Precision SMP	MS
			LCS				
EPA 180.1 Rev. 2.0	Turbidity					15.2	
EPA 200.7 Rev. 4.4	Barium	101	103	103	98.7		0.133
	Calcium	104	104	105			0.259
	Iron	104	104	106	103		1.75
	Magnesium	104	106	102			0.858
	Potassium	104	103	104	101		0.473
	Sodium	102	103	103			2.32
	Zinc	103	105	105	102		0.0947
EPA 200.8 Rev. 5.4	Aluminum	100	97.0	100	101		2.49
	Antimony	105	103	107	105		3.59
	Arsenic	102	101	103	101		2.54
	Boron	100	100	104	103		2.87
	Cadmium	106	103	107	106		4.11
	Chromium	102	104	106	105		2.29
	Copper	106	103	104	104		1.30
	Lead	108	108	111	110		3.20
	Nickel	105	100	105	103		4.26
	Selenium	100	97.5	101	96.5		3.79
	Silver	107	103	105	107		2.01
	Thallium	107	108	112	108		3.65
EPA 300.0 Rev. 2.1	Chloride	97.5	98.8	99.1		0.0877	
	Sulfate	99.2	100	101		0.332	
EPA 350.1 Rev. 2.0	Ammonia-N	99.2	99.0	103			1.81
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	105	106	92.2			7.01
EPA 353.2 Rev. 2.0	NO2NO3-N	99.5	97.5				0.378
EPA 365.1 Rev. 2.0	Total-P	102	103	101			1.51
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	100	98.8	99.4			0.605
SM 2120 B-2011	Color (true)	104				1.78	
SM 2320 B-2011	Total Alkalinity	101				0.255	
SM 2540 C-2011	TDS					0.583	
SM 4500 F-C-2011	Fluoride	98.5	99.7	97.9			1.48
SM 5310 B-2011	Organic Carbon	97.4	96.2	96.2			0.0442

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2189504, 2189507, 2189508
EPA 200.7 Rev. 4.4	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	2189541
EPA 200.8 Rev. 5.4	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	2189541
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2189504, 2189507, 2189508
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2189504, 2189507, 2189508
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2189505, 2189509, 2189510
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2189505, 2189509, 2189510
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2189505, 2189509, 2189510
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2189505, 2189509, 2189510
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2189506, 2189511, 2189512
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2189504, 2189507, 2189508
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO3/L	2189504, 2189507, 2189508
SM 2340B	Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3	2189541
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2189504, 2189507, 2189508

Reference Method Descriptions

Method	Description	<u>Associated Samples</u>
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2189504, 2189507, 2189508
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2189504, 2189507, 2189508
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2189505, 2189509, 2189510

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	08/11/2020			08/11/2020 15:24	Yen Ta	2189504, 2189507, 2189508
EPA 200.7 Rev. 4.4	08/11/2020	08/11/2020 15:55	Elliott D. Healy	08/12/2020 17:47	Betina Topolski	2189541
EPA 200.8 Rev. 5.4	08/11/2020	08/11/2020 15:55	Elliott D. Healy	08/16/2020 01:54	Alexander Thompson	2189541
EPA 300.0 Rev. 2.1	08/11/2020			08/18/2020 01:53	Julia Storbeck	2189507
	08/11/2020			08/18/2020 04:19	Julia Storbeck	2189504
	08/11/2020			08/18/2020 04:31	Julia Storbeck	2189508
EPA 350.1 Rev. 2.0	08/11/2020			08/15/2020 14:43	Ping Hua	2189509
	08/11/2020			08/15/2020 14:45	Ping Hua	2189510
	08/11/2020			08/15/2020 15:01	Ping Hua	2189505
EPA 351.2 Rev. 2.0	08/11/2020	08/17/2020 09:42	David Boose	08/18/2020 14:55	Jhamieka S. Greenwood	2189505
	08/11/2020	08/17/2020 09:42	David Boose	08/18/2020 15:12	Jhamieka S. Greenwood	2189509
	08/11/2020	08/17/2020 09:42	David Boose	08/18/2020 15:14	Jhamieka S. Greenwood	2189510
EPA 353.2 Rev. 2.0	08/11/2020			08/12/2020 09:24	Beverly Y. Sanders	2189509
	08/11/2020			08/12/2020 09:25	Beverly Y. Sanders	2189510
	08/11/2020			08/12/2020 10:08	Beverly Y. Sanders	2189505
EPA 365.1 Rev. 2.0	08/11/2020	08/12/2020 11:41	Yen Ta	08/13/2020 09:47	Benjamin T. Nordmann	2189505
	08/11/2020	08/12/2020 11:41	Yen Ta	08/13/2020 09:48	Benjamin T. Nordmann	2189509
	08/11/2020	08/12/2020 11:41	Yen Ta	08/13/2020 09:49	Benjamin T. Nordmann	2189510
EPA 365.1 Rev. 2.0 dissolved	08/11/2020			08/11/2020 15:37	Jhamieka S. Greenwood	2189506
	08/11/2020			08/11/2020 15:48	Jhamieka S. Greenwood	2189511
	08/11/2020			08/11/2020 15:49	Jhamieka S. Greenwood	2189512
SM 2120 B-2011	08/11/2020			08/11/2020 17:04	Benjamin T. Nordmann	2189504
	08/11/2020			08/11/2020 17:06	Benjamin T. Nordmann	2189507
	08/11/2020			08/11/2020 17:07	Benjamin T. Nordmann	2189508
SM 2320 B-2011	08/11/2020			08/18/2020 15:16	Dale L. Simmons	2189507
	08/11/2020			08/18/2020 15:32	Dale L. Simmons	2189504
	08/11/2020			08/18/2020 15:42	Dale L. Simmons	2189508
SM 2340B	08/11/2020			08/12/2020 17:47	Betina Topolski	2189541
SM 2540 C-2011	08/11/2020			08/12/2020 15:12	Yijie Li	2189504, 2189507, 2189508
SM 2540 D-2011	08/11/2020			08/12/2020 15:12	Yijie Li	2189504, 2189507, 2189508
SM 4500 F-C-2011	08/11/2020			08/12/2020 22:18	Dale L. Simmons	2189504
	08/11/2020			08/12/2020 22:31	Dale L. Simmons	2189507
	08/11/2020			08/12/2020 22:43	Dale L. Simmons	2189508
SM 5310 B-2011	08/11/2020			08/13/2020 00:49	David Boose	2189505
	08/11/2020			08/13/2020 01:01	David Boose	2189509
	08/11/2020			08/13/2020 01:13	David Boose	2189510