

RQ-2020-08-10-55

Login Checklist

Login Station ID: #4

Shipping Method: FedEx UPS | HD | Greyhound

Date/Time of Receipt: 08/19/2020 @ 11:20

*Cooler Temperature	Samples Frozen?	* All Samples in Cooler Preserved with		Number of Sample Containers in Cooler	Evidence Tape				Waybill Tracking #
					Present?		*Intact?		
	YES	HNO ₃	Formalin		Yes	No	Yes	No	
2.3				3		x			1278 40121201

COOLER CHECK

*All samples received on ice unless otherwise noted. HNO₃ and Formalin preserved samples do NOT have a temperature requirement.

- If the temperature of a cooler exceeds 6.0 oC (above 10.0 oC for W-1-4-DIOX and microbiology) or received without ice, identify affected samples in an NCR. NCR# _____
- DOH coolers can exceed 2.0 °C above the standard temperature (6.0 or 10.0 °C for W-1-4-DIOX) without login confirmation from customer for samples requiring ice preservation; identify affected samples in an NCR. NCR# _____
- If cooler or samples are received with a damaged Evidence Seal; identify the affected samples in an NCR. NCR# _____
- If coolers or samples are received late; identify the affected samples in an NCR. NCR# _____

Chain of Custody/ Field Sheet(s) Included?

Yes ☒ No ☐

Micro-Biology Overflow Chain of Custody/ Field Sheet(s) Included?

Yes ☐ No ☒

CONTAINER CHECK

Evidence Tape on Bottles?

Yes ☐ No ☐ NA ☒Evidence Tape intact? Yes ☐ No ☐

If Criminal, photographs taken?

Yes ☐ No ☐Containers intact? Yes ☒ No ☐

Caps on tight?

Yes ☒ No ☐

Sufficient Sample Volume:

Yes ☒ No ☐If **NO** is checked above, generate an NCR listing affected sample IDs in its appropriate category. NCR# _____

CHLORINE CHECK REQUIRED? (Blue dot on container)

Yes ☒ No ☐Init: *Jylyup*

Chlorine detected? (One container checked per Field ID)

Yes ☐ No ☒Init: *Jylyup*

If chlorine is detected, generate an NCR listing affected sample IDs . NCR# _____

PRESERVATION CHECK (EXCLUDING SAMPLES SENT TO OVERFLOW LABS, "OV-")

**Acid Preserved Samples: pH < 2.0?

Yes ☐ No ☐ NA ☒Init: *Jylyup*

**W-1-4-DIOX (Green dot on container) preserved to pH < 4.0?

Yes ☐ No ☐ NA ☒Init: *Jylyup*

If samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

Coolers Unpacked/Checked by: *Jylyup*

Date: 08/19/2020

Event contains NCRs (Y/N)? *None*

Event ID:

Boggy Creek

CEN-DIST-2020-08-19-01 (SMP)

Date Received: 19-AUG-2020 11:20

Login Checklist

Login Preservation Equipment

Lot # & Exp: Date

Additional Comments:

Infrared Thermometer ID:	#4 Temp Gun EXP:11/05/2020
pH Test Strips 0 – 6	
pH Paper 0 – 3	
pH Paper 0 – 13	231019 EXP:11/01/2022
Chlorine Test Strips	8261 EXP:07/2021

ENCORE SAMPLES

S-VOC-MS samples in Encores must be frozen within 48 hours of collection.

Were Encore samples received? Yes _____ Init: _____

Date and time placed in freezer _____

Were all samples placed in freezer within 48 hours of collection? Yes _____ No _____

If no, were samples shipped on dry ice? Yes _____ No _____

If S-VOC-MS samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

FOR CLEAN LAB USE ONLY – Total Mercury Preservation Check(W-HG-AF;W-HG-AF-F):

Samples preserved within 48 hours? Yes _____ No _____

Date and time samples preserved: _____

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 8/18/2020
Customer: CEN-DIST
RQ: RQ-2020-08-10-55

Collected By: Katie March, Leif Boman
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 1

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
Boggy Creek @ 80m Se J Lawson Blvd	 G4CE0181

RELINQUISHED BY(SIGNATURE):



DATE/TIME: 08/18/2020 16:00

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:

Tyler R

Rec'd/Inspected Date:

08/19/2020



RQ-2020-08-10-55

Customer: WQAP

Project ID: SMP

Collected By: Katie March, Leif Boman

Field Report Prepared By: Katie March

Sampling Agency: FDEP

Place Label Here						Comments:				
WIN ID/Field ID:  G4CE0181										
Location: Boggy Creek @ 80m Se J Lawson Blvd										
Matrix: Fresh						(Check one) ☒ GRAB ☐ COMPOSITE				
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)
8/18/2020	11:08 EST	4.32	53.9	26.7	6.64	0.36	1.5 ☐ VOB (S)	2	149.2	0.07
Central Lab please use top row for login purposes										
	11:11 EST	4.31	53.9	26.7	6.51	1.504			149	0.07
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)		☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☐ Ice ☐ H2SO4*		☐ <2		
		Lot #		Exp:						
Metals (P-500ML)		☐ W-ICPMS ☐ W-ICP ☐ W-HARD				☐ Ice ☐ HNO3*		☐ <2		
		Lot #		Exp:						
Filtered Nutrients (P125ML)		☐ W-P04-F ☐ Field Filtered 0.45 µm Filter				☐ Ice				
		Filter Lot #								
		Syringe Lot #								
Chlorophyll (BP-1L)		☐ CHLSUITE-W				☐ Ice				
Physical/Aggregate (Fresh) (P-1L)		☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CI-IC / W-COLOR / W-SO4-IC / W-TDS				☐ Ice				
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice				
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC				☐ Formalin				
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice				
Microbiology (Contract Lab Bottle)		☒ E Coli ☐ F Coli ☐ Enterococci ☐ Contract Lab				☒ Ice			1	A
Molecular (QPCR-P-500ML)		☒ PCR-HF183 ☒ PCR-BacR ☒ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD ☐				☒ Ice			1	A
Tracers (BG-500ML)		☒ W-E8321-DI ☒ W-E8321-MS				☒ Ice			1	A
Pesticides (BG-500ML) (BG-1L)		☐ W-E8321-DI ☐ W-E8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-SQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution)				☐ Ice				
						☐ Ice				
Name: Katie March Leif Boman				Signature:				Date:		

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

Date of Request: 07-AUG-2020
 Created By: BOMAN_L On: 07-AUG-2020
 Modified By: BOMAN_L On: 07-AUG-2020
 Customer: CEN-DIST
 Project: SMP
 Division:
 District: Central District
 Event Description: Boggy Creek

Send Coolers To:

Phone: 407-897-4178
 FL Dept. of Environmental Protection
 3319 Maguire Blvd., Suite 232
 Orlando, FL 32803-3767
 Attn: Natalie Ayala
 Cooler Ship EMail: natalie.ayala@dep.state.fl.us

Priority: 3
 Event Status: P
 Criminal Investigation: NO
 Chemistry Request Reviewed By:
 Biology Request Reviewed By:
 Sampling Kit Required: NO Ship on:
 Sampling Kit Shipped:
 Sampling Kit Packed By:
 Preservatives Needed: NO
 Date to Receive Samples: 10-AUG-2020
 Logged In By:

Send Final Report To:

FL Dept. of Environmental Protection
 3319 Maguire Blvd., Suite 232
 Orlando, FL 32803-3767
 Attn: Kalina Warren
 Report Notification EMail: natalie.ayala@dep.state.fl.us

Comments:

Bottle Group A (Water) With 1 Samples:

Bottle Type:	Number of Bottles:	Preserved With:
ECOLI-18QT Template: DEFAULT	(SM 9223 Quanti-Tray)	Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems
Escherichia Coli-Quanti-Tray		
PCR-BACR Template: DEFAULT	(SOP-PCR-1.4)	Detection and quantification of Ruminant specific Bacteroidetes BacR genetic marker with quantitative polymerase chain reaction
*BacR-qPCR		
PCR-DG3 Template: DEFAULT	(SOP-PCR-1.5)	Detection and quantification of canine-specific DG3 Bacteroides genetic marker with quantitative polymerase chain reaction
*DG3-qPCR		
PCR-HF183 Template: DEFAULT	(SOP-PCR-1.0)	Detection and quantification of human-specific HF183 Bacteroides genetic marker with quantitative polymerase chain reaction
*HF183-qPCR		
W-E8321-DI Template: DEFAULT	(EPA 8321B)	Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
*Acesulfame K		
*AMPA		
*Endothall		
*Glufosinate		
*Glyphosate		
*Sucralose		
W-E8321-MS Template: DEFAULT	(EPA 8321B)	Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
*Acetaminophen		
*Acetamiprid		
*Afidopyropen		
*Benzovindiflupyr		
*Carbamazepine		
*Clothianidin		
*Dinotefuran		
*Fluridone		
*Hydrocodone		

Bottle Group A (Water) With 1 Samples:

Bottle Type:	Number of Bottles:	Preserved With:
W-E8321-MS	Template: DEFAULT	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
*Ibuprofen		
*Imazapyr		
*Mandestrobin		
*Naproxen		
*Primidone		
*Pyraclostrobin		
*Thiamethoxam		
*Tolfenpyrad		
*Triclopyr		
2,4-D		
Bentazon		
Diuron		
Fenuron		
Imidacloprid		
Linuron		
MCP		

* - The laboratory is not NELAP certified for this analyte/method, or certification is not applicable.