

Chemical Analysis Report

CEN-DIST-2020-08-27-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **SJR Above Puzzle Lake**
Request ID: **RQ-2020-08-24-30**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Kalina Warren

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator

Date Certified: 16-SEP-2020 10:15



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals and Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Sjr @ Power Line Crossing

Collection Date/Time: 08/26/2020 11:42

Field ID: G3CE0015

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2192657	EPA 180.1 Rev. 2.0	Turbidity	2.2		NTU	P386641	
	EPA 300.0 Rev. 2.1	Chloride	110		mg Cl/L	P387060	
		Sulfate	16		mg SO4/L	P387065	
		Color (true)	310		PCU	P386653	
	SM 2320 B-2011	Total Alkalinity	66		mg CaCO3/L	P386788	
	SM 2540 C-2011	TDS	396		mg/L	P386808	
	SM 2540 D-2011	TSS	4	I	mg/L	P386812	
	SM 4500 F-C-2011	Fluoride	0.12		mg F/L	P386792	
2192659	EPA 350.1 Rev. 2.0	Ammonia-N	0.032		mg N/L	P387019	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.6		mg N/L	P386687	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.006	I	mg N/L	P386673	
	EPA 365.1 Rev. 2.0	Total-P	0.11		mg P/L	P386678	
	SM 5310 B-2011	Organic Carbon	29		mg C/L	P386726	
2192661	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.066		mg P/L	P386648	
2192691	EPA 200.7 Rev. 4.4	Barium	26.4		ug/L	P386749	
		Calcium	38.4		mg/L	P386749	
		Iron	720		ug/L	P386749	
		Magnesium	9.42		mg/L	P386749	
		Potassium	2.5		mg/L	P386749	
		Sodium	60.8		mg/L	P386749	
		Zinc	5.0	U	ug/L	P386749	
	EPA 200.8 Rev. 5.4	Aluminum	69		ug/L	P386749	
		Antimony	0.063	I	ug/L	P386749	
		Arsenic	1.04		ug/L	P386749	
		Boron**	35.7		ug/L	P386749	
		Cadmium	0.020	U	ug/L	P386749	
		Chromium	0.40	U	ug/L	P386749	
		Copper	0.40	U	ug/L	P386749	
		Lead	0.22	I	ug/L	P386749	
		Nickel	0.50	U	ug/L	P386749	
		Selenium	0.20	U	ug/L	P386749	
		Silver	0.010	U	ug/L	P386749	
		Thallium	0.10	U	ug/L	P386749	
	SM 2340B	Hardness	135		mg/L	P386749	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Sjr @ 250m Nnw Of Power Line Crossing

Collection Date/Time: 08/26/2020 11:08

Field ID: G3CE0031

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2192658	EPA 180.1 Rev. 2.0	Turbidity	2.8	A	NTU	P386641	
	EPA 300.0 Rev. 2.1	Chloride	110		mg Cl/L	P387061	
		Sulfate	17		mg SO4/L	P387066	
		Color (true)	320		PCU	P386653	
	SM 2320 B-2011	Total Alkalinity	67		mg CaCO3/L	P386788	
	SM 2540 C-2011	TDS	402		mg/L	P386808	
	SM 2540 D-2011	TSS	4	I	mg/L	P386812	
	SM 4500 F-C-2011	Fluoride	0.12		mg F/L	P386792	
2192660	EPA 350.1 Rev. 2.0	Ammonia-N	0.035		mg N/L	P387019	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.4		mg N/L	P386687	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.006	I	mg N/L	P386673	
	EPA 365.1 Rev. 2.0	Total-P	0.11		mg P/L	P386678	
	SM 5310 B-2011	Organic Carbon	29		mg C/L	P386726	
2192662	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.067		mg P/L	P386648	
2192692	EPA 200.7 Rev. 4.4	Barium	25.4		ug/L	P386749	
		Calcium	37.0		mg/L	P386749	
		Iron	730		ug/L	P386749	
		Magnesium	9.12		mg/L	P386749	
		Potassium	2.4		mg/L	P386749	
		Sodium	58.7		mg/L	P386749	
		Zinc	5.0	U	ug/L	P386749	
	EPA 200.8 Rev. 5.4	Aluminum	85		ug/L	P386749	
		Antimony	0.057	I	ug/L	P386749	
		Arsenic	1.03		ug/L	P386749	
		Boron**	34.2		ug/L	P386749	
		Cadmium	0.020	U	ug/L	P386749	
		Chromium	0.40	U	ug/L	P386749	
		Copper	0.40	U	ug/L	P386749	
		Lead	0.23	I	ug/L	P386749	
		Nickel	0.50	U	ug/L	P386749	
		Selenium	0.20	U	ug/L	P386749	
		Silver	0.010	U	ug/L	P386749	
		Thallium	0.10	U	ug/L	P386749	
	SM 2340B	Hardness	130		mg/L	P386749	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P386641

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P386749

Component	Result	Code	Units
Barium	1.0	U	ug/L
Calcium	0.075	U	mg/L
Iron	30	U	ug/L
Magnesium	0.040	U	mg/L
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P386749

Component	Result	Code	Units
Aluminum	5.0	U	ug/L
Antimony	0.050	U	ug/L
Arsenic	0.050	U	ug/L
Boron	2.0	U	ug/L
Cadmium	0.020	U	ug/L
Chromium	0.40	U	ug/L
Copper	0.40	U	ug/L
Lead	0.20	U	ug/L
Nickel	0.50	U	ug/L
Selenium	0.20	U	ug/L
Silver	0.010	U	ug/L
Thallium	0.10	U	ug/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387060

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387061

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387065

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387066

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387019

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P386687

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P386673

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P386678

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P386648

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B-2011
Batch ID: P386653

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-2011
Batch ID: P386788

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-2011
Batch ID: P386808

Component	Result	Code	Units
TDS	15	U	mg/L

Quality Assurance Report Method Blank Results

Reference Method: SM 2540 D-2011
Batch ID: P386812

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P386792

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P386726

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P386749

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Barium	97.9		P	85 - 115
Calcium	103		P	85 - 115
Iron	104		P	85 - 115
Magnesium	104		P	85 - 115
Potassium	101		P	85 - 115
Sodium	100		P	85 - 115
Zinc	99.8		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P386749

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aluminum	93.4		P	85 - 115
Antimony	100		P	85 - 115
Arsenic	99.5		P	85 - 115
Boron	94.3		P	85 - 115
Cadmium	96.1		P	85 - 115
Chromium	103		P	85 - 115
Copper	96.8		P	85 - 115
Lead	98.1		P	85 - 115
Nickel	95.6		P	85 - 115
Selenium	97.6		P	85 - 115
Silver	107		P	85 - 115
Thallium	99.4		P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387060

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	97.7		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387061

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	99.6		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387065

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	96.4		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387066

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	98.0		P	90 - 110

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387019

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	101		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P386687

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	104		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P386673

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P386678

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	106		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P386648

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	98.5		P	90 - 110

Reference Method: SM 2120 B-2011
Batch ID: P386653

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	97.9		P	85 - 115

Reference Method: SM 2320 B-2011
Batch ID: P386788

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	93 - 104

Reference Method: SM 4500 F-C-2011
Batch ID: P386792

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.7		P	90 - 106

Reference Method: SM 5310 B-2011
Batch ID: P386726

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	94.2		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P386749

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192874	Barium	104	102	P/P	80 - 120
2192874	Calcium	105	106	P/P	80 - 120
2192874	Iron	108	110	P/P	80 - 120
2192874	Magnesium	109	110	P/P	80 - 120
2192874	Potassium	104	105	P/P	80 - 120
2192874	Sodium	99.9	98.4	P/P	80 - 120
2192874	Zinc	108	106	P/P	80 - 120
2192988	Barium	99.0		P	80 - 120
2192988	Calcium	104		P	80 - 120
2192988	Iron	105		P	80 - 120
2192988	Magnesium	98.9		P	80 - 120
2192988	Potassium	103		P	80 - 120
2192988	Sodium	102		P	80 - 120
2192988	Zinc	103		P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P386749

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192874	Aluminum	94.7	93.5	P/P	80 - 120
2192874	Antimony	100	98.9	P/P	80 - 120
2192874	Arsenic	100	97.9	P/P	80 - 120
2192874	Boron	99.0	97.3	P/P	80 - 120
2192874	Cadmium	98.1	97.3	P/P	80 - 120
2192874	Chromium	103	102	P/P	80 - 120
2192874	Copper	97.5	95.7	P/P	80 - 120
2192874	Lead	101	101	P/P	80 - 120
2192874	Nickel	97.2	95.7	P/P	80 - 120
2192874	Selenium	98.0	96.1	P/P	80 - 120
2192874	Silver	106	107	P/P	80 - 120
2192874	Thallium	101	103	P/P	80 - 120
2192988	Aluminum	93.2		P	80 - 120
2192988	Antimony	105		P	80 - 120
2192988	Arsenic	107		P	80 - 120
2192988	Boron	97.8		P	80 - 120
2192988	Cadmium	101		P	80 - 120
2192988	Chromium	105		P	80 - 120
2192988	Copper	99.0		P	80 - 120
2192988	Lead	102		P	80 - 120
2192988	Nickel	99.2		P	80 - 120
2192988	Selenium	104		P	80 - 120
2192988	Silver	106		P	80 - 120
2192988	Thallium	107		P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387060

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192150	Chloride	98.2		P	90 - 110
2192431	Chloride	99.0		P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387061

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192838	Chloride	99.8		P	90 - 110
2193005	Chloride	99.7		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387065

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192150	Sulfate	96.8		P	90 - 110
2192431	Sulfate	96.9		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387066

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192838	Sulfate	98.8		P	90 - 110
2193005	Sulfate	96.4		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387019

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192467	Ammonia-N	99.6	100	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P386687

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192466	Kjeldahl Nitrogen	105	103	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P386673

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192664	NO2NO3-N	101	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P386678

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192466	Total-P	106	106	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P386648

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192447	O-Phosphate-P	97.1	97.2	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P386792

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192663	Fluoride	98.9	97.1	P/P	94 - 107

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 5310 B-2011
Batch ID: P386726

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192444	Organic Carbon	95.2	94.5	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P386641

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192658	Turbidity	9.25	Sample	P	0 - 20

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P386749

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192874	Barium	1.73	Spike	P	0 - 20
2192874	Calcium	0.712	Spike	P	0 - 20
2192874	Iron	1.61	Spike	P	0 - 20
2192874	Magnesium	0.587	Spike	P	0 - 20
2192874	Potassium	1.16	Spike	P	0 - 20
2192874	Sodium	0.934	Spike	P	0 - 20
2192874	Zinc	1.50	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P386749

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192874	Aluminum	1.03	Spike	P	0 - 20
2192874	Antimony	1.49	Spike	P	0 - 20
2192874	Arsenic	2.51	Spike	P	0 - 20
2192874	Boron	1.61	Spike	P	0 - 20
2192874	Cadmium	0.872	Spike	P	0 - 20
2192874	Chromium	0.742	Spike	P	0 - 20
2192874	Copper	1.88	Spike	P	0 - 20
2192874	Lead	0.404	Spike	P	0 - 20
2192874	Nickel	1.48	Spike	P	0 - 20
2192874	Selenium	1.97	Spike	P	0 - 20
2192874	Silver	0.701	Spike	P	0 - 20
2192874	Thallium	2.03	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387060

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192150	Chloride	0.0899	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387061

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2193005	Chloride	0.302	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387065

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192150	Sulfate	0.458	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387066

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2193005	Sulfate	0.678	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387019

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192467	Ammonia-N	0.752	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P386687

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192466	Kjeldahl Nitrogen	0.699	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P386673

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192664	NO2NO3-N	0.482	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P386678

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192466	Total-P	0.212	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P386648

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192447	O-Phosphate-P	0.103	Spike	P	0 - 20

Reference Method: SM 2120 B-2011
Batch ID: P386653

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192442	Color (true)	9.84	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2320 B-2011
Batch ID: P386788

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192663	Total Alkalinity	0.197	Sample	P	0 - 8.1

Reference Method: SM 2540 C-2011
Batch ID: P386808

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192663	TDS	0.673	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
Batch ID: P386792

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192663	Fluoride	1.43	Spike	P	0 - 2.8

Reference Method: SM 5310 B-2011
Batch ID: P386726

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192444	Organic Carbon	0.640	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A100685

Included Lab Sample IDs: 2192657, 2192658

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	101	101	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A100687

Included Lab Sample IDs: 2192661, 2192662

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	98.0	100	P/P	90 - 110

Reference Method: SM 2120 B-2011

Run ID: A100689

Included Lab Sample IDs: 2192657, 2192658

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	101	100	P/P	90 - 115

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A100698

Included Lab Sample IDs: 2192659, 2192660

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	104	104	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A100718

Included Lab Sample IDs: 2192659, 2192660

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	93.0	94.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A100726

Included Lab Sample IDs: 2192659, 2192660

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	103	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A100735

Included Lab Sample IDs: 2192659, 2192660

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	97.9	98.7	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A100747

Included Lab Sample IDs: 2192657, 2192658

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	101	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 4500 F-C-2011

Run ID: A100747

Included Lab Sample IDs: 2192657, 2192658

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	101	P/P	90 - 110

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A100789

Included Lab Sample IDs: 2192691, 2192692

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Barium	101	98.1	P/P	90 - 110
Calcium	101	100	P/P	90 - 110
Iron	103	103	P/P	90 - 110
Magnesium	102	102	P/P	90 - 110
Potassium	100	100	P/P	90 - 110
Sodium	99.6	100	P/P	90 - 110
Zinc	104	100	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A100832

Included Lab Sample IDs: 2192659, 2192660

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	100	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A100843

Included Lab Sample IDs: 2192657, 2192658

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	102	102	P/P	90 - 110
Chloride	102	101	P/P	90 - 110
Sulfate	101	101	P/P	90 - 110
Sulfate	101	100	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A100913

Included Lab Sample IDs: 2192691, 2192692

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aluminum	96.0	95.8	P/P	90 - 110
Antimony	103	103	P/P	90 - 110
Arsenic	98.2	97.1	P/P	90 - 110
Boron	96.0	97.9	P/P	90 - 110
Cadmium	95.4	95.1	P/P	90 - 110
Chromium	99.8	98.8	P/P	90 - 110
Copper	94.5	93.6	P/P	90 - 110
Lead	95.6	95.8	P/P	90 - 110
Nickel	95.6	94.8	P/P	90 - 110
Selenium	99.6	97.6	P/P	90 - 110
Thallium	94.0	95.3	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A100945

Included Lab Sample IDs: 2192691, 2192692

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Silver	105	104	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery			Precision	
						LCS	SMP MS
EPA 180.1 Rev. 2.0	Turbidity						9.25
EPA 200.7 Rev. 4.4	Barium	97.9	104	102	99.0		1.73
	Calcium	103	105	106	104		0.712
	Iron	104	108	110	105		1.61
	Magnesium	104	109	110	98.9		0.587
	Potassium	101	104	105	103		1.16
	Sodium	100	99.9	98.4	102		0.934
	Zinc	99.8	108	106	103		1.50
EPA 200.8 Rev. 5.4	Aluminum	93.4	93.2	94.7	93.5		1.03
	Antimony	100	105	100	98.9		1.49
	Arsenic	99.5	107	100	97.9		2.51
	Boron	94.3	97.8	99.0	97.3		1.61
	Cadmium	96.1	101	98.1	97.3		0.872
	Chromium	103	105	103	102		0.742
	Copper	96.8	99.0	97.5	95.7		1.88
	Lead	98.1	101	102	101		0.404
	Nickel	95.6	99.2	97.2	95.7		1.48
	Selenium	97.6	104	98.0	96.1		1.97
	Silver	107	106	106	107		0.701
	Thallium	99.4	107	101	103		2.03
EPA 300.0 Rev. 2.1	Chloride	97.7	98.2	99.0		0.0899	
	Chloride	99.6	99.7	99.8		0.302	
	Sulfate	96.4	96.8	96.9		0.458	
	Sulfate	98.0	96.4	98.8		0.678	
EPA 350.1 Rev. 2.0	Ammonia-N	101	99.6	100			0.752
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	104	105	103			0.699
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	104	101	100			0.482
EPA 365.1 Rev. 2.0	Total-P	106	106	106			0.212
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	98.5	97.1	97.2			0.103
SM 2120 B-2011	Color (true)	97.9				9.84	
SM 2320 B-2011	Total Alkalinity	102				0.197	
SM 2540 C-2011	TDS					0.673	
SM 4500 F-C-2011	Fluoride	98.7	98.9	97.1			1.43
SM 5310 B-2011	Organic Carbon	94.2	95.2	94.5			0.640

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2192657, 2192658
EPA 200.7 Rev. 4.4	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	2192691, 2192692
EPA 200.8 Rev. 5.4	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	2192691, 2192692
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2192657, 2192658
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2192657, 2192658
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2192659, 2192660
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2192659, 2192660
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2192659, 2192660
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2192659, 2192660
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2192661, 2192662
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2192657, 2192658
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2192657, 2192658

Reference Method Descriptions

Method	Description	<u>Associated Samples</u>
SM 2340B	Hardness by calculation, in aqueous matrices, reported as mg/L CaCO ₃	2192691, 2192692
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2192657, 2192658
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2192657, 2192658
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2192657, 2192658
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2192659, 2192660

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	08/27/2020			08/27/2020 15:06	Yen Ta	2192657, 2192658
EPA 200.7 Rev. 4.4	08/27/2020	08/31/2020 13:30	Elliott D. Healy	09/02/2020 16:54	Betina Topolski	2192691
	08/27/2020	08/31/2020 13:30	Elliott D. Healy	09/02/2020 17:02	Betina Topolski	2192692
EPA 200.8 Rev. 5.4	08/27/2020	08/31/2020 13:30	Elliott D. Healy	09/10/2020 19:27	Alexander Thompson	2192691
	08/27/2020	08/31/2020 13:30	Elliott D. Healy	09/10/2020 19:34	Alexander Thompson	2192692
	08/27/2020	08/31/2020 13:30	Elliott D. Healy	09/11/2020 14:19	Alexander Thompson	2192691
	08/27/2020	08/31/2020 13:30	Elliott D. Healy	09/11/2020 14:27	Alexander Thompson	2192692
EPA 300.0 Rev. 2.1	08/27/2020			09/05/2020 03:51	Julia Storbeck	2192657
	08/27/2020			09/05/2020 06:04	Julia Storbeck	2192658
EPA 350.1 Rev. 2.0	08/27/2020			09/05/2020 13:06	Ping Hua	2192659
	08/27/2020			09/05/2020 13:08	Ping Hua	2192660
EPA 351.2 Rev. 2.0	08/27/2020	08/28/2020 12:10	David Boose	08/31/2020 13:08	Virginia P. Brown	2192659
	08/27/2020	08/28/2020 12:10	David Boose	08/31/2020 13:10	Virginia P. Brown	2192660
EPA 353.2 Rev. 2.0	08/27/2020			08/28/2020 09:57	Beverly Y. Sanders	2192659
	08/27/2020			08/28/2020 09:58	Beverly Y. Sanders	2192660
EPA 365.1 Rev. 2.0	08/27/2020	08/28/2020 12:36	Yen Ta	08/31/2020 09:36	Benjamin T. Nordmann	2192659
	08/27/2020	08/28/2020 12:36	Yen Ta	08/31/2020 09:37	Benjamin T. Nordmann	2192660
EPA 365.1 Rev. 2.0 dissolved	08/27/2020			08/27/2020 15:47	Virginia P. Brown	2192661
	08/27/2020			08/27/2020 15:48	Virginia P. Brown	2192662
SM 2120 B-2011	08/27/2020			08/27/2020 16:31	Madeline Gruver	2192657
	08/27/2020			08/27/2020 16:32	Madeline Gruver	2192658
SM 2320 B-2011	08/27/2020			08/28/2020 17:59	Dale L. Simmons	2192657
	08/27/2020			08/28/2020 18:12	Dale L. Simmons	2192658
SM 2340B	08/27/2020			09/02/2020 16:54	Betina Topolski	2192691
	08/27/2020			09/02/2020 17:02	Betina Topolski	2192692
SM 2540 C-2011	08/27/2020			08/28/2020 17:05	Madeline Gruver	2192657, 2192658
SM 2540 D-2011	08/27/2020			08/28/2020 19:46	Madeline Gruver	2192657, 2192658
SM 4500 F-C-2011	08/27/2020			08/28/2020 17:59	Dale L. Simmons	2192657
	08/27/2020			08/28/2020 18:12	Dale L. Simmons	2192658
SM 5310 B-2011	08/27/2020			08/30/2020 07:41	Ping Hua	2192659
	08/27/2020			08/30/2020 07:53	Ping Hua	2192660