

Chemical Analysis Report

CEN-DIST-2020-08-25-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Lindley/Marshall/Unnamed X2**
Request ID: **RQ-2020-08-24-48**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Kalina Warren

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Kerry Tate, Ph.D., Environmental Administrator

Date Certified: 16-SEP-2020 11:36

A handwritten signature in black ink that reads "Kerry Tate". The signature is written in a cursive, flowing style with a horizontal line above the "T" and a long, sweeping underline.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals and Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Marshall Lake @ Center

Collection Date/Time: 08/24/2020 14:17

Field ID: G1CE0110

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2191943	EPA 180.1 Rev. 2.0	Turbidity	2.6		NTU	P386526	
	EPA 300.0 Rev. 2.1	Chloride	12		mg Cl/L	P387078	
		Sulfate	13		mg SO4/L	P387079	
	SM 2120 B-2011	Color (true)	36		PCU	P386527	
	SM 2320 B-2011	Total Alkalinity	47		mg CaCO3/L	P386621	
	SM 2540 C-2011	TDS	110		mg/L	P386992	
	SM 2540 D-2011	TSS	3	I	mg/L	P386985	
	SM 4500 F-C-2011	Fluoride	0.070	I	mg F/L	P386623	
2191945	EPA 350.1 Rev. 2.0	Ammonia-N	0.008		mg N/L	P386833	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.0		mg N/L	P386602	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P386542	
	EPA 365.1 Rev. 2.0	Total-P	0.030		mg P/L	P386551	
	SM 5310 B-2011	Organic Carbon	13		mg C/L	P386595	
2191947	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P386533	
2191963	EPA 200.7 Rev. 4.4	Barium	11.7		ug/L	P386556	
		Calcium	20.2		mg/L	P386556	
		Iron	30	U	ug/L	P386556	
		Magnesium	3.51		mg/L	P386556	
		Potassium	3.7		mg/L	P386556	
		Sodium	8.3		mg/L	P386556	
		Zinc	5.0	U	ug/L	P386556	
	EPA 200.8 Rev. 5.4	Aluminum	10	I	ug/L	P386556	
		Antimony	0.15	I	ug/L	P386556	
		Arsenic	1.26		ug/L	P386556	
		Beryllium	0.010	U	ug/L	P386556	
		Boron**	34.4		ug/L	P386556	
		Cadmium	0.020	U	ug/L	P386556	
		Chromium	0.40	U	ug/L	P386556	
		Copper	0.94		ug/L	P386556	
		Lead	0.20	U	ug/L	P386556	
		Nickel	0.50	U	ug/L	P386556	
		Selenium	0.20	U	ug/L	P386556	
		Silver	0.010	U	ug/L	P386556	
		Thallium	0.10	U	ug/L	P386556	
	SM 2340B	Hardness	65.0		mg/L	P386556	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Lindley Lake @ center of southern lobe

Collection Date/Time: 08/24/2020 12:18

Field ID: G2CE0234

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2191944	EPA 180.1 Rev. 2.0	Turbidity	1.4		NTU	P386526	
	EPA 300.0 Rev. 2.1	Chloride	17		mg Cl/L	P387059	
		Sulfate	18		mg SO4/L	P387064	
		Color (true)	22		PCU	P386527	
	SM 2320 B-2011	Total Alkalinity	5.9		mg CaCO3/L	P386621	
	SM 2540 C-2011	TDS	84		mg/L	P386992	
	SM 2540 D-2011	TSS	2	I	mg/L	P386985	
	SM 4500 F-C-2011	Fluoride	0.064	I	mg F/L	P386623	
2191946	EPA 350.1 Rev. 2.0	Ammonia-N	0.005	I	mg N/L	P386833	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.62		mg N/L	P386602	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P386542	
	EPA 365.1 Rev. 2.0	Total-P	0.021		mg P/L	P386551	
	SM 5310 B-2011	Organic Carbon	7.5		mg C/L	P386595	
2191948	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P386533	
2191964	EPA 200.7 Rev. 4.4	Barium	11.9		ug/L	P386556	
		Calcium	6.67		mg/L	P386556	
		Iron	37	I	ug/L	P386556	
		Magnesium	3.26		mg/L	P386556	
		Potassium	5.5		mg/L	P386556	
		Sodium	9.9		mg/L	P386556	
		Zinc	5.0	U	ug/L	P386556	
	EPA 200.8 Rev. 5.4	Aluminum	26		ug/L	P386556	
		Antimony	0.21		ug/L	P386556	
		Arsenic	1.61		ug/L	P386556	
		Beryllium	0.010	U	ug/L	P386556	
		Boron**	40.2		ug/L	P386556	
		Cadmium	0.020	U	ug/L	P386556	
		Chromium	0.40	U	ug/L	P386556	
		Copper	0.89		ug/L	P386556	
		Lead	0.21	I	ug/L	P386556	
		Nickel	0.50	U	ug/L	P386556	
		Selenium	0.20	U	ug/L	P386556	
		Silver	0.010	U	ug/L	P386556	
		Thallium	0.10	U	ug/L	P386556	
		Hardness	30.1		mg/L	P386556	

Ref. Method and Comment:

SM 2540 C-2011: The reported result is consistent with the laboratory conductivity, 135 umhos/cm.

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Unnamed Ditch @ 30m from Boy Scout Camp Rd **Collection Date/Time: 08/24/2020 11:15**

Field ID: G2CE0121

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2191949	EPA 180.1 Rev. 2.0	Turbidity	2.3		NTU	P386526	
	EPA 300.0 Rev. 2.1	Chloride	21		mg Cl/L	P387078	
		Sulfate	2.1		mg SO4/L	P387079	
	SM 2120 B-2011	Color (true)	600		PCU	P386527	
	SM 2320 B-2011	Total Alkalinity	0.65	U	mg CaCO3/L	P386621	
	SM 2540 C-2011	TDS	139		mg/L	P386992	
	SM 2540 D-2011	TSS	3	I	mg/L	P386985	
	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P386623	
2191951	EPA 350.1 Rev. 2.0	Ammonia-N	0.020		mg N/L	P387015	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.4		mg N/L	P386770	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.020		mg N/L	P386542	
	EPA 365.1 Rev. 2.0	Total-P	0.10		mg P/L	P386551	
	SM 5310 B-2011	Organic Carbon	59		mg C/L	P386595	
2191953	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.083		mg P/L	P386533	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Unnamed Ditch @ 30m W of Riverside Dr

Collection Date/Time: 08/24/2020 10:27

Field ID: G5CE0120

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2191950	EPA 180.1 Rev. 2.0	Turbidity	1.4		NTU	P386525	
	EPA 300.0 Rev. 2.1	Chloride	74		mg Cl/L	P387059	
		Sulfate	15		mg SO4/L	P387064	
	SM 2120 B-2011	Color (true)	70		PCU	P386527	
	SM 2320 B-2011	Total Alkalinity	230		mg CaCO3/L	P386621	
	SM 2540 C-2011	TDS	421		mg/L	P386992	
	SM 2540 D-2011	TSS	2	U	mg/L	P386985	
	SM 4500 F-C-2011	Fluoride	0.23		mg F/L	P386623	
2191952	EPA 350.1 Rev. 2.0	Ammonia-N	0.040		mg N/L	P387015	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.89		mg N/L	P386602	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.24		mg N/L	P386542	
	EPA 365.1 Rev. 2.0	Total-P	0.22		mg P/L	P386553	
	SM 5310 B-2011	Organic Carbon	17		mg C/L	P386595	
2191954	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.19		mg P/L	P386533	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P386525

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P386526

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P386556

Component	Result	Code	Units
Barium	1.0	U	ug/L
Calcium	0.075	U	mg/L
Iron	30	U	ug/L
Magnesium	0.040	U	mg/L
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P386556

Component	Result	Code	Units
Aluminum	5.0	U	ug/L
Antimony	0.050	U	ug/L
Arsenic	0.050	U	ug/L
Beryllium	0.010	U	ug/L
Boron	2.0	U	ug/L
Cadmium	0.020	U	ug/L
Chromium	0.40	U	ug/L
Copper	0.40	U	ug/L
Lead	0.20	U	ug/L
Nickel	0.50	U	ug/L
Selenium	0.20	U	ug/L
Silver	0.010	U	ug/L
Thallium	0.10	U	ug/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387059

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387064

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387078

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387079

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P386833

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387015

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P386602

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P386770

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P386542

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P386551

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P386553

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P386533

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B-2011
Batch ID: P386527

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-2011
Batch ID: P386621

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-2011
Batch ID: P386992

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P386985

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P386623

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P386595

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P386556

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Barium	100		P	85 - 115
Calcium	105		P	85 - 115
Iron	104		P	85 - 115
Magnesium	102		P	85 - 115
Potassium	101		P	85 - 115
Sodium	104		P	85 - 115
Zinc	97.9		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P386556

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aluminum	97.3		P	85 - 115
Antimony	102		P	85 - 115
Arsenic	104		P	85 - 115
Beryllium	91.4		P	85 - 115
Boron	99.8		P	85 - 115
Cadmium	101		P	85 - 115
Chromium	106		P	85 - 115
Copper	103		P	85 - 115
Lead	109		P	85 - 115
Nickel	103		P	85 - 115
Selenium	99.3		P	85 - 115
Silver	109		P	85 - 115
Thallium	107		P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387059

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	97.3		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387064

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	96.6		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387078

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	95.2		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387079

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	93.9		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P386833

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	101		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387015

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	100		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P386602

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	105		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P386770

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	95.3		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P386542

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P386551

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	103		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P386553

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	101		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P386533

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	99.9		P	90 - 110

Reference Method: SM 2120 B-2011
Batch ID: P386527

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	98.3		P	85 - 115

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: SM 2320 B-2011
Batch ID: P386621

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	104		P	93 - 104

Reference Method: SM 4500 F-C-2011
Batch ID: P386623

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.1		P	90 - 106

Reference Method: SM 5310 B-2011
Batch ID: P386595

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	100		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P386556

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2191963	Barium	104	103	P/P	80 - 120
2191963	Calcium	102		P	80 - 120
2191963	Iron	108	108	P/P	80 - 120
2191963	Magnesium	102	103	P/P	80 - 120
2191963	Potassium	100	101	P/P	80 - 120
2191963	Sodium	102		P	80 - 120
2191963	Zinc	103	101	P/P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P386556

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2191963	Aluminum	97.6	99.2	P/P	80 - 120
2191963	Antimony	104	104	P/P	80 - 120
2191963	Arsenic	107	107	P/P	80 - 120
2191963	Beryllium	92.5	93.8	P/P	80 - 120
2191963	Boron	99.0	102	P/P	80 - 120
2191963	Cadmium	104	102	P/P	80 - 120
2191963	Chromium	109	109	P/P	80 - 120
2191963	Copper	103	101	P/P	80 - 120
2191963	Lead	111	112	P/P	80 - 120
2191963	Nickel	103	101	P/P	80 - 120
2191963	Selenium	99.6	101	P/P	80 - 120
2191963	Silver	108	108	P/P	80 - 120
2191963	Thallium	110	112	P/P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387059

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2191971	Chloride	97.5		P	90 - 110
2192041	Chloride	97.5		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387064

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2191971	Sulfate	95.2		P	90 - 110
2192041	Sulfate	95.4		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387078

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2191897	Chloride	98.8		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387079

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2191897	Sulfate	97.6		P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P386833

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2191906	Ammonia-N	99.4	100	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387015

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192024	Ammonia-N	98.6	98.4	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P386602

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2191972	Kjeldahl Nitrogen	111	113	F/F	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P386770

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192445	Kjeldahl Nitrogen	101	99.4	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P386542

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2191898	NO2NO3-N	97.6	99.6	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P386551

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2191850	Total-P	104	107	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P386553

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2191952	Total-P	101	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P386533

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2191907	O-Phosphate-P	98.1	97.9	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P386623

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2192055	Fluoride	95.0	96.8	P/P	94 - 107

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 5310 B-2011
Batch ID: P386595

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2191946	Organic Carbon	102	101	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P386525

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2191990	Turbidity	4.34	Sample	P	0 - 20

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P386526

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2191917	Turbidity	3.12	Sample	P	0 - 20

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P386556

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2191963	Barium	0.818	Spike	P	0 - 20
2191963	Calcium	0.532	Spike	P	0 - 20
2191963	Iron	0.0506	Spike	P	0 - 20
2191963	Magnesium	0.745	Spike	P	0 - 20
2191963	Potassium	0.438	Spike	P	0 - 20
2191963	Sodium	0.518	Spike	P	0 - 20
2191963	Zinc	1.23	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P386556

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2191963	Aluminum	1.68	Spike	P	0 - 20
2191963	Antimony	0.730	Spike	P	0 - 20
2191963	Arsenic	0.251	Spike	P	0 - 20
2191963	Beryllium	1.33	Spike	P	0 - 20
2191963	Boron	1.89	Spike	P	0 - 20
2191963	Cadmium	2.34	Spike	P	0 - 20
2191963	Chromium	0.639	Spike	P	0 - 20
2191963	Copper	1.18	Spike	P	0 - 20
2191963	Lead	0.0644	Spike	P	0 - 20
2191963	Nickel	1.55	Spike	P	0 - 20
2191963	Selenium	1.80	Spike	P	0 - 20
2191963	Silver	0.480	Spike	P	0 - 20
2191963	Thallium	1.53	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387059

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2191971	Chloride	1.33	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387064

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2191971	Sulfate	1.13	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387078

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2191897	Chloride	5.22	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387079

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2191897	Sulfate	1.11	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P386833

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2191906	Ammonia-N	0.901	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387015

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192024	Ammonia-N	0.101	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P386602

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2191972	Kjeldahl Nitrogen	0.669	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P386770

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192445	Kjeldahl Nitrogen	0.923	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P386542

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2191898	NO2NO3-N	1.90	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P386551

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2191850	Total-P	2.74	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P386553

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2191952	Total-P	0.288	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P386533

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2191907	O-Phosphate-P	0.204	Spike	P	0 - 20

Reference Method: SM 2120 B-2011
Batch ID: P386527

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2191905	Color (true)	2.78	Sample	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P386621

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192055	Total Alkalinity	0.680	Sample	P	0 - 8.1

Reference Method: SM 2540 C-2011
Batch ID: P386992

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2191848	TDS	2.00	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
Batch ID: P386623

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2192055	Fluoride	0.915	Spike	P	0 - 2.8

Reference Method: SM 5310 B-2011
Batch ID: P386595

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2191946	Organic Carbon	0.0720	Spike	P	0 - 20

Quality Assurance Report Precision

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A100641

Included Lab Sample IDs: 2191943, 2191944, 2191949, 2191950

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: SM 2120 B-2011

Run ID: A100642

Included Lab Sample IDs: 2191943, 2191944, 2191949, 2191950

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	97.6	98.5	P/P	90 - 115
Color (true)	98.5	99.4	P/P	90 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A100644

Included Lab Sample IDs: 2191947, 2191948, 2191953

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	99.4	P/P	90 - 110
O-Phosphate-P	99.8	100	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A100651

Included Lab Sample IDs: 2191945, 2191946, 2191951, 2191952

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	102	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A100665

Included Lab Sample IDs: 2191945, 2191946, 2191951, 2191952

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	97.0	99.6	P/P	90 - 110
Organic Carbon	99.6	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A100666

Included Lab Sample IDs: 2191951, 2191952

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	99.8	P/P	90 - 110
Total-P	101	99.8	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A100675

Included Lab Sample IDs: 2191943, 2191944, 2191949, 2191950

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	101	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 4500 F-C-2011

Run ID: A100675

Included Lab Sample IDs: 2191943, 2191944, 2191949, 2191950

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.5	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A100695

Included Lab Sample IDs: 2191945, 2191946

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	96.2	95.5	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A100708

Included Lab Sample IDs: 2191945, 2191946, 2191952

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	99.6	99.6	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A100715

Included Lab Sample IDs: 2191963, 2191964

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aluminum	100	101	P/P	90 - 110
Aluminum	101	100	P/P	90 - 110
Arsenic	105	105	P/P	90 - 110
Arsenic	105	106	P/P	90 - 110
Boron	101	105	P/P	90 - 110
Boron	97.5	101	P/P	90 - 110
Cadmium	100	101	P/P	90 - 110
Cadmium	101	101	P/P	90 - 110
Chromium	103	104	P/P	90 - 110
Chromium	104	103	P/P	90 - 110
Copper	102	98.9	P/P	90 - 110
Copper	98.9	98.8	P/P	90 - 110
Lead	106	107	P/P	90 - 110
Lead	107	106	P/P	90 - 110
Nickel	102	99.6	P/P	90 - 110
Nickel	99.6	99.1	P/P	90 - 110
Selenium	102	103	P/P	90 - 110
Selenium	98.2	102	P/P	90 - 110
Silver	109	109	P/P	90 - 110
Silver	109	109	P/P	90 - 110
Thallium	103	104	P/P	90 - 110
Thallium	104	101	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A100716

Included Lab Sample IDs: 2191963, 2191964

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Antimony	92.1	101	P/P	90 - 110
Beryllium	94.8	94.8	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A100721

Included Lab Sample IDs: 2191963, 2191964

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Barium	103	104	P/P	90 - 110
Barium	104	105	P/P	90 - 110
Calcium	101	103	P/P	90 - 110
Calcium	103	101	P/P	90 - 110
Iron	104	105	P/P	90 - 110
Iron	106	104	P/P	90 - 110
Magnesium	101	103	P/P	90 - 110
Magnesium	103	101	P/P	90 - 110
Potassium	102	99.9	P/P	90 - 110
Potassium	99.9	101	P/P	90 - 110
Sodium	100	101	P/P	90 - 110
Sodium	103	100	P/P	90 - 110
Zinc	101	103	P/P	90 - 110
Zinc	103	104	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A100761

Included Lab Sample IDs: 2191945, 2191946

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A100780

Included Lab Sample IDs: 2191951

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	98.8	100	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A100832

Included Lab Sample IDs: 2191951, 2191952

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	99.1	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A100843

Included Lab Sample IDs: 2191944, 2191950

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	99.5	101	P/P	90 - 110
Sulfate	99.2	100	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A100847

Included Lab Sample IDs: 2191943, 2191949

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	99.0	98.4	P/P	90 - 110
Sulfate	98.3	96.8	P/P	90 - 110

Quality Assurance Report

Calibration Verification

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					4.34	
	Turbidity					3.12	
EPA 200.7 Rev. 4.4	Barium	100	104	103			0.818
	Calcium	105	102				0.532
	Iron	104	108	108			0.0506
	Magnesium	102	102	103			0.745
	Potassium	101	100	101			0.438
	Sodium	104	102				0.518
	Zinc	97.9	103	101			1.23
EPA 200.8 Rev. 5.4	Aluminum	97.3	97.6	99.2			1.68
	Antimony	102	104	104			0.730
	Arsenic	104	107	107			0.251
	Beryllium	91.4	92.5	93.8			1.33
	Boron	99.8	99.0	102			1.89
	Cadmium	101	104	102			2.34
	Chromium	106	109	109			0.639
	Copper	103	103	101			1.18
	Lead	109	111	112			0.0644
	Nickel	103	103	101			1.55
	Selenium	99.3	99.6	101			1.80
	Silver	109	108	108			0.480
	Thallium	107	110	112			1.53
EPA 300.0 Rev. 2.1	Chloride	97.3	97.5	97.5		1.33	
	Chloride	95.2	98.8			5.22	
	Sulfate	96.6	95.2	95.4		1.13	
	Sulfate	93.9	97.6			1.11	
EPA 350.1 Rev. 2.0	Ammonia-N	101	99.4	100			0.901
	Ammonia-N	100	98.6	98.4			0.101
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	105	111	113			0.669
	Kjeldahl Nitrogen	95.3	101	99.4			0.923
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	102	97.6	99.6			1.90
EPA 365.1 Rev. 2.0	Total-P	103	104	107			2.74
	Total-P	101	101	101			0.288
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	99.9	98.1	97.9			0.204
SM 2120 B-2011	Color (true)	98.3				2.78	
SM 2320 B-2011	Total Alkalinity	104				0.680	
SM 2540 C-2011	TDS					2.00	
SM 4500 F-C-2011	Fluoride	97.1	95.0	96.8			0.915
SM 5310 B-2011	Organic Carbon	100	102	101			0.0720

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2191943, 2191944, 2191949, 2191950
EPA 200.7 Rev. 4.4	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	2191963, 2191964
EPA 200.8 Rev. 5.4	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	2191963, 2191964
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2191943, 2191944, 2191949, 2191950
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2191943, 2191944, 2191949, 2191950
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2191945, 2191946, 2191951, 2191952
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2191945, 2191946, 2191951, 2191952
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2191945, 2191946, 2191951, 2191952

Reference Method Descriptions

Method	Description	Associated Samples
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2191945, 2191946, 2191951, 2191952
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2191947, 2191948, 2191953, 2191954
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2191943, 2191944, 2191949, 2191950
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO3/L	2191943, 2191944, 2191949, 2191950
SM 2340B	Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3	2191963, 2191964
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2191943, 2191944, 2191949, 2191950
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2191943, 2191944, 2191949, 2191950
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2191943, 2191944, 2191949, 2191950
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2191945, 2191946, 2191951, 2191952

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	08/25/2020			08/25/2020 16:17	Yen Ta	2191943, 2191944, 2191949, 2191950
EPA 200.7 Rev. 4.4	08/25/2020	08/26/2020 14:10	Elliott D. Healy	08/27/2020 14:23	Betina Topolski	2191963
	08/25/2020	08/26/2020 14:10	Elliott D. Healy	08/27/2020 15:23	Betina Topolski	2191964
EPA 200.8 Rev. 5.4	08/25/2020	08/26/2020 14:10	Elliott D. Healy	08/29/2020 01:50	Alexander Thompson	2191963
	08/25/2020	08/26/2020 14:10	Elliott D. Healy	08/29/2020 02:36	Alexander Thompson	2191964
	08/25/2020	08/26/2020 14:10	Elliott D. Healy	08/30/2020 00:59	Alexander Thompson	2191964
	08/25/2020	08/26/2020 14:10	Elliott D. Healy	08/30/2020 01:16	Alexander Thompson	2191963
EPA 300.0 Rev. 2.1	08/25/2020			09/04/2020 09:00	Elliott Rodriguez	2191943
	08/25/2020			09/04/2020 09:24	Elliott Rodriguez	2191949
	08/25/2020			09/04/2020 18:45	Julia Storbeck	2191944
	08/25/2020			09/04/2020 18:57	Julia Storbeck	2191950
EPA 350.1 Rev. 2.0	08/25/2020			09/01/2020 18:52	Ping Hua	2191945
	08/25/2020			09/01/2020 18:54	Ping Hua	2191946
	08/25/2020			09/05/2020 10:54	Ping Hua	2191951
	08/25/2020			09/05/2020 10:55	Ping Hua	2191952
EPA 351.2 Rev. 2.0	08/25/2020	08/27/2020 10:15	Sima Feizi	08/28/2020 13:09	Virginia P. Brown	2191945
	08/25/2020	08/27/2020 10:15	Sima Feizi	08/28/2020 13:11	Virginia P. Brown	2191946
	08/25/2020	08/27/2020 10:15	Sima Feizi	08/28/2020 13:13	Virginia P. Brown	2191952
	08/25/2020	09/01/2020 11:35	Sima Feizi	09/02/2020 10:28	Virginia P. Brown	2191951
EPA 353.2 Rev. 2.0	08/25/2020			08/26/2020 09:36	Beverly Y. Sanders	2191945
	08/25/2020			08/26/2020 09:37	Beverly Y. Sanders	2191946
	08/25/2020			08/26/2020 09:38	Beverly Y. Sanders	2191951
	08/25/2020			08/26/2020 09:39	Beverly Y. Sanders	2191952
EPA 365.1 Rev. 2.0	08/25/2020	08/26/2020 12:59	Yen Ta	08/27/2020 08:25	Benjamin T. Nordmann	2191951
	08/25/2020	08/26/2020 12:59	Yen Ta	08/27/2020 08:42	Benjamin T. Nordmann	2191952
	08/25/2020	08/26/2020 12:59	Yen Ta	08/28/2020 09:15	Lipi Saha	2191945
	08/25/2020	08/26/2020 12:59	Yen Ta	08/28/2020 09:16	Lipi Saha	2191946
EPA 365.1 Rev. 2.0 dissolved	08/25/2020			08/25/2020 15:46	Blanca Fach	2191947
	08/25/2020			08/25/2020 15:51	Blanca Fach	2191948
	08/25/2020			08/25/2020 15:52	Blanca Fach	2191953
	08/25/2020			08/25/2020 16:17	Blanca Fach	2191954
SM 2120 B-2011	08/25/2020			08/25/2020 16:03	Benjamin T. Nordmann	2191943, 2191944
	08/25/2020			08/25/2020 16:05	Benjamin T. Nordmann	2191950
	08/25/2020			08/25/2020 16:13	Benjamin T. Nordmann	2191949
SM 2320 B-2011	08/25/2020			08/26/2020 18:22	Dale L. Simmons	2191943
	08/25/2020			08/26/2020 18:29	Dale L. Simmons	2191949
	08/25/2020			08/26/2020 18:42	Dale L. Simmons	2191950
	08/25/2020			08/26/2020 20:24	Dale L. Simmons	2191944
SM 2340B	08/25/2020			08/27/2020 14:23	Betina Topolski	2191963
	08/25/2020			08/27/2020 15:23	Betina Topolski	2191964

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
SM 2540 C-2011	08/25/2020			08/27/2020 15:08	Yijie Li	2191943, 2191944, 2191949, 2191950
SM 2540 D-2011	08/25/2020			08/27/2020 15:08	Yijie Li	2191943, 2191944, 2191949, 2191950
SM 4500 F-C-2011	08/25/2020			08/26/2020 18:22	Dale L. Simmons	2191943
	08/25/2020			08/26/2020 18:29	Dale L. Simmons	2191949
	08/25/2020			08/26/2020 18:42	Dale L. Simmons	2191950
	08/25/2020			08/26/2020 20:24	Dale L. Simmons	2191944
SM 5310 B-2011	08/25/2020			08/26/2020 17:20	David Boose	2191946
	08/25/2020			08/26/2020 18:09	David Boose	2191945
	08/25/2020			08/26/2020 20:39	David Boose	2191951
	08/25/2020			08/26/2020 20:51	David Boose	2191952