

Chemical Analysis Report

CEN-DIST-2020-09-11-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Bugg Spring**
Request ID: **RQ-2020-09-07-52**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Kalina Warren

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Kerry Tate, Ph.D., Environmental Administrator

Date Certified: 01-OCT-2020 08:47

A handwritten signature in black ink that reads "Kerry Tate". The signature is written in a cursive, flowing style with a horizontal line above the first name.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Bugg Spring @ vent

Collection Date/Time: 09/10/2020 11:41

Field ID: G1CE0109

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2195725	EPA 180.1 Rev. 2.0	Turbidity	0.75		NTU	P387277	
	EPA 300.0 Rev. 2.1	Chloride	15		mg Cl/L	P387418	
		Sulfate	7.9		mg SO4/L	P387420	
	SM 2120 B-2011	Color (true)	2.5	U	PCU	P387287	
	SM 2320 B-2011	Total Alkalinity	142		mg CaCO3/L	P387364	
	SM 2540 C-2011	TDS	177		mg/L	P387784	
	SM 2540 D-2011	TSS	2	I	mg/L	P387631	
	SM 4500 F-C-2011	Fluoride	0.062	I	mg F/L	P387366	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.016		mg N/L	P387306	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.14	I	mg N/L	P387477	
2195726	EPA 353.2 Rev. 2.0	NO2NO3-N	0.25		mg N/L	P387468	
	EPA 365.1 Rev. 2.0	Total-P	0.066		mg P/L	P387808	
	SM 5310 B-2011	Organic Carbon	1.3		mg C/L	P387426	
	EPA 365.1 Rev. 2.0	O-Phosphate-P	0.051		mg P/L	P387283	
	dissolved						

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: FB_09102020

Collection Date/Time: 09/10/2020 11:54

Field ID: FB_09102020

Matrix: W-FIELD-BK

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2195728	EPA 180.1 Rev. 2.0	Turbidity	0.10	U	NTU	P387277	
	EPA 300.0 Rev. 2.1	Chloride	0.20	U	mg Cl/L	P387418	
		Sulfate	0.20	U	mg SO4/L	P387420	
	SM 2120 B-2011	Color (true)	2.5	U	PCU	P387287	
	SM 2320 B-2011	Total Alkalinity	0.65	U	mg CaCO3/L	P387364	
	SM 2540 C-2011	TDS	15	U	mg/L	P387784	
	SM 2540 D-2011	TSS	2	U	mg/L	P387631	
	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P387366	
2195729	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P387310	MS
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.080	U	mg N/L	P387477	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P387468	
	EPA 365.1 Rev. 2.0	Total-P	0.002	U	mg P/L	P387808	
	SM 5310 B-2011	Organic Carbon	0.50	U	mg C/L	P387426	
2195730	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P387283	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P387277

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387418

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387420

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387306

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387310

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P387477

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P387468

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P387808

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P387283

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 2120 B-2011
Batch ID: P387287

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-2011
Batch ID: P387364

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-2011
Batch ID: P387784

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P387631

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P387366

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P387426

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387418

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	96.3		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387420

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	96.2		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387306

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	92.5		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387310

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	92.0		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P387477

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	101		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P387468

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P387808

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	105		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P387283

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	102		P	90 - 110

Reference Method: SM 2120 B-2011
Batch ID: P387287

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	96.3		P	85 - 115

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 2320 B-2011
Batch ID: P387364

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	98.7		P	93 - 104

Reference Method: SM 4500 F-C-2011
Batch ID: P387366

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	101		P	90 - 106

Reference Method: SM 5310 B-2011
Batch ID: P387426

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	99.8		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387418

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2195501	Chloride	93.1		P	90 - 110
2195725	Chloride	96.2		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387420

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2195501	Sulfate	95.1		P	90 - 110
2195725	Sulfate	94.8		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387306

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2195677	Ammonia-N	95.0	96.3	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387310

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2195742	Ammonia-N	90.7	89.0	P/F	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P387477

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2195742	Kjeldahl Nitrogen	105	110	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P387468

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2195726	NO2NO3-N	98.8		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P387808

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2195702	Total-P	104	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P387283

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2195767	O-Phosphate-P	98.4	98.8	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P387366

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2195758	Fluoride	98.3	99.2	P/P	94 - 107

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 5310 B-2011
Batch ID: P387426

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2195578	Organic Carbon	98.5	100	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P387277

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2195687	Turbidity	5.88	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387418

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2195501	Chloride	0.0871	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P387420

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2195501	Sulfate	0.114	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387306

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2195677	Ammonia-N	1.21	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387310

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2195742	Ammonia-N	1.76	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P387477

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2195742	Kjeldahl Nitrogen	2.76	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P387468

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2195726	NO2NO3-N	0.220	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P387808

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2195702	Total-P	0.430	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P387283

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2195767	O-Phosphate-P	0.303	Spike	P	0 - 20

Reference Method: SM 2120 B-2011
 Batch ID: P387287

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2195741	Color (true)	0.583	Sample	P	0 - 20

Reference Method: SM 2320 B-2011
 Batch ID: P387364

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2195758	Total Alkalinity	2.82	Sample	P	0 - 8.1

Reference Method: SM 2540 C-2011
 Batch ID: P387784

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2195676	TDS	8.61	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
 Batch ID: P387366

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2195758	Fluoride	0.451	Spike	P	0 - 2.8

Reference Method: SM 5310 B-2011
 Batch ID: P387426

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2195578	Organic Carbon	1.51	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A100939

Included Lab Sample IDs: 2195725, 2195728

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.0	101	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A100940

Included Lab Sample IDs: 2195727, 2195730

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	97.5	97.5	P/P	90 - 110

Reference Method: SM 2120 B-2011

Run ID: A100942

Included Lab Sample IDs: 2195725, 2195728

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	99.3	98.0	P/P	90 - 115

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A100947

Included Lab Sample IDs: 2195726

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	97.6	98.1	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A100948

Included Lab Sample IDs: 2195729

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	94.0	94.6	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A100975

Included Lab Sample IDs: 2195725, 2195728

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	99.0	102	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A100975

Included Lab Sample IDs: 2195725, 2195728

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	102	103	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A100980

Included Lab Sample IDs: 2195725, 2195728

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	101	P/P	90 - 110
Chloride	101	101	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A100980

Included Lab Sample IDs: 2195725, 2195728

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Sulfate	101	101	P/P	90 - 110
Sulfate	101	101	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A101004

Included Lab Sample IDs: 2195726, 2195729

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	101	100	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A101029

Included Lab Sample IDs: 2195726, 2195729

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A101057

Included Lab Sample IDs: 2195726, 2195729

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	107	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A101204

Included Lab Sample IDs: 2195726

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	104	P/P	85 - 115

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A101251

Included Lab Sample IDs: 2195729

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	96.0	95.2	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					5.88	
EPA 300.0 Rev. 2.1	Chloride	96.3	93.1	96.2		0.0871	
	Sulfate	96.2	95.1	94.8		0.114	
EPA 350.1 Rev. 2.0	Ammonia-N	92.5	95.0	96.3			1.21
	Ammonia-N	92.0	90.7	89.0			1.76
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	101	105	110			2.76
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	102	98.8				0.220
EPA 365.1 Rev. 2.0	Total-P	105	104	104			0.430
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	102	98.4	98.8			0.303
SM 2120 B-2011	Color (true)	96.3				0.583	
SM 2320 B-2011	Total Alkalinity	98.7				2.82	
SM 2540 C-2011	TDS					8.61	
SM 4500 F-C-2011	Fluoride	101	99.2	98.3			0.451
SM 5310 B-2011	Organic Carbon	99.8	98.5	100			1.51

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2195725, 2195728
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2195725, 2195728
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2195725, 2195728
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2195726, 2195729
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2195726, 2195729
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2195726, 2195729
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2195726, 2195729
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2195727, 2195730
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2195725, 2195728
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2195725, 2195728
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2195725, 2195728
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2195725, 2195728
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2195725, 2195728
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2195726, 2195729

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	09/11/2020			09/11/2020 14:50	Yen Ta	2195725, 2195728
EPA 300.0 Rev. 2.1	09/11/2020			09/15/2020 23:01	Lipi Saha	2195725
	09/11/2020			09/16/2020 02:18	Lipi Saha	2195728
EPA 350.1 Rev. 2.0	09/11/2020			09/13/2020 14:16	Ping Hua	2195726
	09/11/2020			09/13/2020 16:55	Ping Hua	2195729
EPA 351.2 Rev. 2.0	09/11/2020	09/16/2020 12:40	David Boose	09/17/2020 13:05	Virginia P. Brown	2195726
	09/11/2020	09/16/2020 12:40	David Boose	09/17/2020 13:07	Virginia P. Brown	2195729
EPA 353.2 Rev. 2.0	09/11/2020			09/16/2020 11:10	Beverly Y. Sanders	2195726
	09/11/2020			09/16/2020 11:18	Beverly Y. Sanders	2195729
EPA 365.1 Rev. 2.0	09/11/2020	09/24/2020 11:57	Yen Ta	09/25/2020 14:21	Benjamin T. Nordmann	2195726
	09/11/2020	09/24/2020 11:57	Yen Ta	09/29/2020 11:55	Lipi Saha	2195729
EPA 365.1 Rev. 2.0 dissolved	09/11/2020			09/11/2020 15:59	Samantha Davila	2195727
	09/11/2020			09/11/2020 16:01	Samantha Davila	2195730
SM 2120 B-2011	09/11/2020			09/11/2020 16:03	Madeline Gruver	2195725, 2195728
SM 2320 B-2011	09/11/2020			09/15/2020 00:51	Dale L. Simmons	2195725
	09/11/2020			09/15/2020 01:00	Dale L. Simmons	2195728
SM 2540 C-2011	09/11/2020			09/16/2020 15:09	Yijie Li	2195725, 2195728
SM 2540 D-2011	09/11/2020			09/16/2020 15:09	Yijie Li	2195725, 2195728
SM 4500 F-C-2011	09/11/2020			09/15/2020 00:51	Dale L. Simmons	2195725
	09/11/2020			09/15/2020 01:00	Dale L. Simmons	2195728
SM 5310 B-2011	09/11/2020			09/16/2020 00:33	David Boose	2195726
	09/11/2020			09/16/2020 00:45	David Boose	2195729